

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG082518\
 Data File : BG036421.D
 Acq On : 25 Aug 2018 11:05
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 27 01:04:44 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 23 12:07:02 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.588	0.527	10.4	108	0.00
3	Pyridine	1.652	1.556	5.8	107	0.00
4	n-Nitrosodimethylamine	0.736	0.664	9.8	104	0.00
5 S	2-Fluorophenol	1.261	1.184	6.1	108	0.00
6	Aniline	2.291	2.142	6.5	108	0.00
7 S	Phenol-d6	1.805	1.750	3.0	112	0.00
8	2-Chlorophenol	1.367	1.281	6.3	108	0.00
9	Benzaldehyde	1.243	1.098	11.7	109	0.00
10 C	Phenol	1.889	1.802	4.6	110	0.00
11	bis(2-Chloroethyl)ether	1.498	1.400	6.5	109	0.00
12	1,3-Dichlorobenzene	1.561	1.471	5.8	111	0.00
13 C	1,4-Dichlorobenzene	1.576	1.477	6.3	110	0.00
14	1,2-Dichlorobenzene	1.505	1.422	5.5	111	0.00
15	Benzyl Alcohol	1.455	1.395	4.1	109	0.00
16	2,2'-oxybis(1-Chloropropane	3.020	2.683	11.2	104	0.00
17	2-Methylphenol	1.254	1.192	4.9	111	0.00
18	Hexachloroethane	0.565	0.540	4.4	110	0.00
19 P	n-Nitroso-di-n-propylamine	1.349	1.254	7.0	109	0.00
20	3+4-Methylphenols	1.751	1.705	2.6	113	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
22	Acetophenone	0.525	0.487	7.2	111	0.00
23 S	Nitrobenzene-d5	0.369	0.375	-1.6	119	0.00
24	Nitrobenzene	0.397	0.380	4.3	111	0.00
25	Isophorone	0.732	0.675	7.8	110	0.00
26 C	2-Nitrophenol	0.158	0.147	7.0	111	0.00
27	2,4-Dimethylphenol	0.292	0.271	7.2	112	0.00
28	bis(2-Chloroethoxy)methane	0.449	0.417	7.1	111	0.00
29 C	2,4-Dichlorophenol	0.320	0.318	0.6	117	0.00
30	1,2,4-Trichlorobenzene	0.374	0.353	5.6	115	0.00
31	Naphthalene	1.000	0.933	6.7	112	0.00
32	Benzoic acid	0.202	0.188	6.9	109	0.00
33	4-Chloroaniline	0.458	0.443	3.3	114	0.00
34 C	Hexachlorobutadiene	0.251	0.243	3.2	114	-0.01
35	Caprolactam	0.127	0.123	3.1	111	0.00
36 C	4-Chloro-3-methylphenol	0.371	0.367	1.1	115	0.00
37	2-Methylnaphthalene	0.732	0.708	3.3	115	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
39	1,2,4,5-Tetrachlorobenzene	0.708	0.659	6.9	116	0.00
40 P	Hexachlorocyclopentadiene	0.368	0.332	9.8	111	0.00
41 S	2,4,6-Tribromophenol	0.239	0.249	-4.2	123	0.00
42 C	2,4,6-Trichlorophenol	0.431	0.428	0.7	122	0.00
43	2,4,5-Trichlorophenol	0.460	0.463	-0.7	122	0.00
44 S	2-Fluorobiphenyl	1.401	1.357	3.1	123	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG082518\
 Data File : BG036421.D
 Acq On : 25 Aug 2018 11:05
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 27 01:04:44 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 23 12:07:02 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.660	1.514	8.8	117	0.00
46	2-Chloronaphthalene	1.267	1.191	6.0	118	0.00
47	2-Nitroaniline	0.398	0.397	0.3	118	0.00
48	Acenaphthylene	1.981	1.846	6.8	117	0.00
49	Dimethylphthalate	1.633	1.549	5.1	118	0.00
50	2,6-Dinitrotoluene	0.336	0.328	2.4	120	0.00
51 C	Acenaphthene	1.269	1.193	6.0	118	0.00
52	3-Nitroaniline	0.362	0.366	-1.1	117	0.00
53 P	2,4-Dinitrophenol	0.127	0.134	-5.5	129	0.00
54	Dibenzofuran	1.987	1.875	5.6	119	0.00
55 P	4-Nitrophenol	0.296	0.282	4.7	113	0.00
56	2,4-Dinitrotoluene	0.438	0.463	-5.7	128	0.00
57	Fluorene	1.486	1.393	6.3	117	0.00
58	2,3,4,6-Tetrachlorophenol	0.432	0.447	-3.5	126	0.00
59	Diethylphthalate	1.646	1.560	5.2	117	0.00
60	4-Chlorophenyl-phenylether	0.917	0.890	2.9	123	0.00
61	4-Nitroaniline	0.379	0.382	-0.8	118	0.00
62	Azobenzene	1.675	1.530	8.7	114	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	126	0.00
64	4,6-Dinitro-2-methylphenol	0.088	0.084	4.5	121	0.00
65 c	n-Nitrosodiphenylamine	0.594	0.547	7.9	119	0.00
66	4-Bromophenyl-phenylether	0.234	0.224	4.3	122	0.00
67	Hexachlorobenzene	0.239	0.232	2.9	123	0.00
68	Atrazine	0.223	0.197	11.7	115	0.00
69 C	Pentachlorophenol	0.163	0.160	1.8	116	0.00
70	Phenanthrene	1.016	0.938	7.7	118	0.00
71	Anthracene	1.013	0.931	8.1	116	0.00
72	Carbazole	1.012	0.920	9.1	114	0.00
73	Di-n-butylphthalate	1.127	1.020	9.5	112	0.00
74 C	Fluoranthene	1.239	1.132	8.6	114	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
76	Benzidine	0.638	0.556	12.9	116	0.00
77	Pyrene	1.230	1.142	7.2	115	0.00
78 S	Terphenyl-d14	0.856	0.818	4.4	120	-0.01
79	Butylbenzylphthalate	0.489	0.453	7.4	111	0.00
80	Benzo(a)anthracene	1.212	1.115	8.0	116	0.00
81	3,3'-Dichlorobenzidine	0.483	0.457	5.4	115	0.00
82	Chrysene	1.148	1.060	7.7	115	0.00
83	Bis(2-ethylhexyl)phthalate	0.685	0.629	8.2	112	-0.02
84 c	Di-n-octyl phthalate	1.144	1.067	6.7	112	-0.02
85	Indeno(1,2,3-cd)pyrene	1.334	1.292	3.1	120	-0.03
86 I	Perylene-d12	1.000	1.000	0.0	127	-0.02
87	Benzo(b)fluoranthene	1.111	1.046	5.9	118	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG082518\
Data File : BG036421.D
Acq On : 25 Aug 2018 11:05
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Aug 27 01:04:44 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 23 12:07:02 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.085	1.006	7.3	117	-0.01
89 C	Benzo(a)pyrene	1.067	0.995	6.7	117	-0.01
90	Dibenzo(a,h)anthracene	1.030	0.988	4.1	121	-0.04
91	Benzo(a,h,i)perylene	1.035	0.988	4.5	121	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0