

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/25/2023	BNA_G	<u>BG058848.D</u>
Compound Name	Response	Retention Time
DDT	267290	20.685
DDD	4621	20.291
DDE	189	19.751
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4810	272100	1.77