


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG082620\
 Data File : BG046454.D
 Acq On : 26 Aug 2020 14:25
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG082520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Aug 25 18:00:34 2020



AutoFind: Scans 2496, 2497, 2498; Background Corrected with Scan 2489

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.7	15699	PASS
68	69	0.00	2	1.0	154	PASS
69	198	0.00	100	25.5	15601	PASS
70	69	0.00	2	0.5	84	PASS
127	198	10	80	36.4	22253	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	61074	PASS
199	198	5	9	6.8	4168	PASS
275	198	10	60	27.7	16907	PASS
365	198	1	100	3.6	2189	PASS
441	198	0.01	100	9.8	5986	PASS
442	198	50	100	60.4	36885	PASS
443	442	15	24	18.3	6765	PASS