

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082724\
 Data File : BG062587.D
 Acq On : 27 Aug 2024 20:35
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG082724

Quant Time: Aug 28 01:03:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 28 00:59:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	30.667	23.3	72	0.00
3	Pyridine	40.000	30.012	25.0	70	0.00
4	n-Nitrosodimethylamine	40.000	31.325	21.7	75	0.00
5 S	2-Fluorophenol	80.000	68.936	13.8	99	0.00
6	Aniline	40.000	47.134	-17.8	123	0.00
7 S	Phenol-d6	80.000	70.072	12.4	89	0.00
8	2-Chlorophenol	40.000	36.392	9.0	91	0.00
9	Benzaldehyde	40.000	39.061	2.3	101	0.00
10 C	Phenol	40.000	36.713	8.2	94	0.00
11	bis(2-Chloroethyl)ether	40.000	35.347	11.6	89	0.00
12	1,3-Dichlorobenzene	40.000	33.830	15.4	87	0.00
13 C	1,4-Dichlorobenzene	40.000	34.028	14.9	87	0.00
14	1,2-Dichlorobenzene	40.000	33.510	16.2	84	0.00
15	Benzyl Alcohol	40.000	41.589	-4.0	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.960	5.1	94	-0.01
17	2-Methylphenol	40.000	37.352	6.6	92	0.00
18	Hexachloroethane	40.000	36.317	9.2	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.875	-2.2	95	0.00
20	3+4-Methylphenols	40.000	37.986	5.0	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	37.303	6.7	93	0.00
23 S	Nitrobenzene-d5	80.000	78.832	1.5	95	0.00
24	Nitrobenzene	40.000	39.056	2.4	93	0.00
25	Isophorone	40.000	38.772	3.1	91	0.00
26 C	2-Nitrophenol	40.000	42.664	-6.7	107	0.00
27	2,4-Dimethylphenol	40.000	44.837	-12.1	109	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.815	5.5	93	0.00
29 C	2,4-Dichlorophenol	40.000	37.184	7.0	89	0.00
30	1,2,4-Trichlorobenzene	40.000	35.046	12.4	86	0.00
31	Naphthalene	40.000	33.960	15.1	82	0.00
32	Benzoic acid	40.000	38.082	4.8	98	-0.01
33	4-Chloroaniline	40.000	41.322	-3.3	97	0.00
34 C	Hexachlorobutadiene	40.000	33.929	15.2	87	0.00
35	Caprolactam	40.000	45.455	-13.6	111	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.429	1.4	95	0.00
37	2-Methylnaphthalene	40.000	37.063	7.3	94	0.00
38	1-Methylnaphthalene	40.000	34.584	13.5	86	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	117	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	33.622	15.9	94	0.00
41 P	Hexachlorocyclopentadiene	40.000	51.484	-28.7#	140	0.00
42 S	2,4,6-Tribromophenol	80.000	72.722	9.1	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.364	6.6	99	0.00
44	2,4,5-Trichlorophenol	40.000	34.568	13.6	95	0.00
45 S	2-Fluorobiphenyl	80.000	65.297	18.4	92	0.00
46	1,1'-Biphenyl	40.000	33.816	15.5	96	0.00
47	2-Chloronaphthalene	40.000	32.748	18.1	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082724\
 Data File : BG062587.D
 Acq On : 27 Aug 2024 20:35
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG082724

Quant Time: Aug 28 01:03:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 28 00:59:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.261	-0.7	118	0.00
49	Acenaphthylene	40.000	37.039	7.4	103	0.00
50	Dimethylphthalate	40.000	37.137	7.2	103	0.00
51	2,6-Dinitrotoluene	40.000	35.539	11.2	105	0.00
52 C	Acenaphthene	40.000	34.982	12.5	99	0.00
53	3-Nitroaniline	40.000	43.301	-8.3	116	0.00
54 P	2,4-Dinitrophenol	40.000	39.397	1.5	110	0.00
55	Dibenzofuran	40.000	34.386	14.0	97	0.00
56 P	4-Nitrophenol	40.000	39.462	1.3	108	0.00
57	2,4-Dinitrotoluene	40.000	39.933	0.2	106	0.00
58	Fluorene	40.000	33.641	15.9	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.160	4.6	105	0.00
60	Diethylphthalate	40.000	38.606	3.5	106	0.00
61	4-Chlorophenyl-phenylether	40.000	33.509	16.2	96	0.00
62	4-Nitroaniline	40.000	41.114	-2.8	110	0.00
63	Azobenzene	40.000	35.741	10.6	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.792	-2.0	110	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.302	9.2	102	0.00
67	4-Bromophenyl-phenylether	40.000	34.699	13.3	97	0.00
68	Hexachlorobenzene	40.000	34.193	14.5	97	0.00
69	Atrazine	40.000	50.236	-25.6#	150	0.00
70 C	Pentachlorophenol	40.000	36.053	9.9	95	0.00
71	Phenanthrene	40.000	35.284	11.8	100	0.00
72	Anthracene	40.000	36.720	8.2	102	0.00
73	Carbazole	40.000	34.396	14.0	96	0.00
74	Di-n-butylphthalate	40.000	39.252	1.9	108	0.00
75 C	Fluoranthene	40.000	33.078	17.3	92	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	111	0.00
77	Benzydine	40.000	38.712	3.2	103	0.00
78	Pyrene	40.000	35.114	12.2	93	0.00
79 S	Terphenyl-d14	80.000	70.742	11.6	94	0.00
80	Butylbenzylphthalate	40.000	38.405	4.0	111	0.00
81	Benzo(a)anthracene	40.000	35.632	10.9	95	0.00
82	3,3'-Dichlorobenzidine	40.000	43.437	-8.6	112	0.00
83	Chrysene	40.000	35.025	12.4	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.420	-6.1	107	0.00
85 c	Di-n-octyl phthalate	40.000	41.510	-3.8	109	0.00
86 I	Perylene-d12	20.000	20.000	0.0	114	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.334	9.2	99	0.00
88	Benzo(b)fluoranthene	40.000	33.453	16.4	91	0.00
89	Benzo(k)fluoranthene	40.000	35.332	11.7	94	0.00
90 C	Benzo(a)pyrene	40.000	37.040	7.4	99	0.00
91	Dibenzo(a,h)anthracene	40.000	36.045	9.9	97	-0.01
92	Benzo(g,h,i)perylene	40.000	32.542	18.6	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082724\
Data File : BG062587.D
Acq On : 27 Aug 2024 20:35
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG082724

Quant Time: Aug 28 01:03:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 28 00:59:59 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0