

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG083019\
 Data File : BG042598.D
 Acq On : 30 Aug 2019 13:30
 Operator : HP/JU
 Sample : PB122665BSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Quant Time: Aug 30 14:23:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG082219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 22 14:29:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.99	152	32718	20.00	ng	-0.01
21) Naphthalene-d8	10.82	136	129215	20.00	ng	0.00
39) Acenaphthene-d10	14.66	164	92913	20.00	ng	0.00
64) Phenanthrene-d10	17.41	188	216951	20.00	ng	-0.01
76) Chrysene-d12	21.69	240	247140	20.00	ng	0.00
87) Perylene-d12	24.93	264	266228	20.00	ng	-0.01

System Monitoring Compounds

5) 2-Fluorophenol	5.56	112	183243	113.43	ng	0.00
7) Phenol-d6	7.17	99	225437	103.31	ng	0.00
23) Nitrobenzene-d5	9.16	82	119147	63.72	ng	-0.01
42) 2,4,6-Tribromophenol	16.15	330	144671	109.55	ng	-0.01
45) 2-Fluorobiphenyl	13.28	172	383856	69.87	ng	0.00
79) Terphenyl-d14	20.02	244	773447	67.66	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.41	88	23589	34.990	ng	98
3) Pyridine	3.82	79	52335	30.274	ng	99
4) n-Nitrosodimethylamine	3.73	42	24959	40.423	ng	# 96
6) Aniline	7.31	93	79149	27.194	ng	99
8) 2-Chlorophenol	7.57	128	81590	41.674	ng	97
9) Benzaldehyde	7.12	77	47370	43.360	ng	96
10) Phenol	7.19	94	89019	40.123	ng	93
11) bis(2-Chloroethyl)ether	7.41	93	63620	36.512	ng	96
12) 1,3-Dichlorobenzene	7.89	146	95868	38.492	ng	97
13) 1,4-Dichlorobenzene	8.03	146	97366	38.594	ng	96
14) 1,2-Dichlorobenzene	8.35	146	93549	38.721	ng	98
15) Benzyl Alcohol	8.24	79	59742	38.054	ng	98
16) 2,2'-oxybis(1-Chloropropan	8.53	45	80003	33.875	ng	97
17) 2-Methylphenol	8.45	107	63116	38.623	ng	96
18) Hexachloroethane	9.09	117	32307	37.407	ng	94
19) n-Nitroso-di-n-propylamine	8.80	70	48206	34.099	ng	91
20) 3+4-Methylphenols	8.78	107	84562	37.396	ng	96
22) Acetophenone	8.82	105	110460	39.968	ng	# 99
24) Nitrobenzene	9.20	77	73963	39.062	ng	99
25) Isophorone	9.73	82	133108	36.070	ng	98
26) 2-Nitrophenol	9.93	139	49020	41.312	ng	96
27) 2,4-Dimethylphenol	9.99	122	72884	44.590	ng	98
28) bis(2-Chloroethoxy)methane	10.22	93	84999	36.545	ng	99
29) 2,4-Dichlorophenol	10.47	162	91069	42.585	ng	97
30) 1,2,4-Trichlorobenzene	10.68	180	104237	39.710	ng	98
31) Naphthalene	10.87	128	241559	40.266	ng	99
32) Benzoic acid	10.13	122	30353	29.402	ng	92
33) 4-Chloroaniline	10.98	127	64960	23.457	ng	99
34) Hexachlorobutadiene	11.16	225	70985	40.238	ng	98
35) Caprolactam	11.75	113	24530	34.376	ng	91
36) 4-Chloro-3-methylphenol	12.11	107	80955	39.877	ng	95
37) 2-Methylnaphthalene	12.48	142	193483	40.166	ng	99
38) 1-Methylnaphthalene	12.70	142	181104	39.581	ng	97
40) 1,2,4,5-Tetrachlorobenzene	12.85	216	131861	44.193	ng	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG083019\
 Data File : BG042598.D
 Acq On : 30 Aug 2019 13:30
 Operator : HP/JU
 Sample : PB122665BSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Quant Time: Aug 30 14:23:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG082219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 22 14:29:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	12.83	237	112221	71.600	ng	99
43) 2,4,6-Trichlorophenol	13.09	196	82722	41.016	ng	97
44) 2,4,5-Trichlorophenol	13.17	196	87055	41.653	ng	99
46) 1,1'-Biphenyl	13.49	154	249898	41.608	ng	99
47) 2-Chloronaphthalene	13.53	162	209803	40.512	ng	99
48) 2-Nitroaniline	13.73	65	45497	36.432	ng	97
49) Acenaphthylene	14.38	152	322238	41.359	ng	100
50) Dimethylphthalate	14.11	163	258181	38.511	ng	98
51) 2,6-Dinitrotoluene	14.23	165	62408	40.161	ng	97
52) Acenaphthene	14.72	154	185788	39.271	ng	99
53) 3-Nitroaniline	14.56	138	43092	27.728	ng	91
54) 2,4-Dinitrophenol	14.77	184	64454	62.977	ng	100
55) Dibenzofuran	15.06	168	325697	41.590	ng	99
56) 4-Nitrophenol	14.88	139	86658	78.474	ng	95
57) 2,4-Dinitrotoluene	15.02	165	87729	39.425	ng	97
58) Fluorene	15.71	166	260269	40.658	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.29	232	87588	42.377	ng	# 93
60) Diethylphthalate	15.47	149	247887	37.825	ng	98
61) 4-Chlorophenyl-phenylether	15.70	204	154952	40.154	ng	98
62) 4-Nitroaniline	15.73	138	61798	37.155	ng	94
63) Azobenzene	15.99	77	168833	37.597	ng	96
65) 4,6-Dinitro-2-methylphenol	15.79	198	54610	40.149	ng	98
66) n-Nitrosodiphenylamine	15.91	169	239420	40.129	ng	100
67) 4-Bromophenyl-phenylether	16.59	248	108040	39.260	ng	96
68) Hexachlorobenzene	16.72	284	116034	38.788	ng	97
69) Atrazine	16.86	200	90622	42.951	ng	96
70) Pentachlorophenol	17.07	266	132975	85.551	ng	97
71) Phenanthrene	17.45	178	447901	41.563	ng	99
72) Anthracene	17.54	178	455442	42.335	ng	98
73) Carbazole	17.81	167	404617	42.200	ng	99
74) Di-n-butylphthalate	18.37	149	453886	40.047	ng	100
75) Fluoranthene	19.46	202	607253	46.173	ng	98
77) Benzidine	19.64	184	240552	33.949	ng	100
78) Pyrene	19.83	202	615017	40.590	ng	98
80) Butylbenzylphthalate	20.71	149	221796	37.216	ng	96
81) Benzo(a)anthracene	21.67	228	614745	40.890	ng	97
82) 3,3'-Dichlorobenzidine	21.58	252	216088	35.738	ng	98
83) Chrysene	21.74	228	589926	40.687	ng	99
84) Bis(2-ethylhexyl)phthalate	21.57	149	319933	38.665	ng	97
85) Di-n-octyl phthalate	22.78	149	557295	39.159	ng	99
86) Indeno(1,2,3-cd)pyrene	28.63	276	778831	39.527	ng	# 97
88) Benzo(b)fluoranthene	23.90	252	684560	46.772	ng	96
89) Benzo(k)fluoranthene	23.96	252	647378	46.016	ng	# 98
90) Benzo(a)pyrene	24.77	252	659597	47.492	ng	98
91) Dibenzo(a,h)anthracene	28.69	278	643840	45.785	ng	98
92) Benzo(g,h,i)perylene	29.79	276	656170	46.655	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.99 min Scan# 835
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion:152 Resp: 32718
 Ion Ratio Lower Upper
 152 100
 150 144.9 125.4 188.0
 115 46.9 38.5 57.7




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#2

1,4-Dioxane
 Concen: 34.990 ng
 RT: 3.41 min Scan# 55
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion: 88 Resp: 23589
 Ion Ratio Lower Upper
 88 100
 58 70.6 58.3 87.5
 43 24.9 20.2 30.2



Abundance Ion 88.00 (87.70 to 88.70): BGC
 Ion 58.00 (57.70 to 58.70): BGC
 Ion 43.00 (42.70 to 43.70): BGC





#3

Pyridine

Concen: 30.274 ng

RT: 3.82 min Scan# 124

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD





Tgt Ion: 79 Resp: 52335

Ion Ratio Lower Upper

79 100

52 72.1 57.7 86.5

51 36.9 28.8 43.2



#4

n-Nitrosodimethylamine

Concen: 40.423 ng

RT: 3.73 min Scan# 110

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30




Tgt Ion: 42 Resp: 24959

Ion Ratio Lower Upper

42 100

74 144.0 119.6 179.4

44 5.6 6.5 9.7#





#5

2-Fluorophenol

Concen: 113.430 ng

RT: 5.56 min Scan# 420

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion: 112 Resp: 183243

Ion Ratio Lower Upper

112 100

64 53.1 42.8 64.2




63 28.1 22.7 34.1



Abundance Ion 112.00 (111.70 to 112.70): BGC

Ion 64.00 (63.70 to 64.70): BGC

Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 27.194 ng

RT: 7.31 min Scan# 719

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion: 93 Resp: 79149

Ion Ratio Lower Upper

93 100

66 37.2 29.0 43.6



65 19.1 15.2 22.8




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 66.00 (65.70 to 66.70): BGC

Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 103.310 ng

RT: 7.17 min Scan# 694

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion: 99 Resp: 225437

Ion Ratio Lower Upper




99 100

42 14.9 12.0 18.0

71 35.7 26.3 39.5



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#8

2-Chlorophenol

Concen: 41.674 ng

RT: 7.57 min Scan# 762

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion: 128 Resp: 81590

Ion Ratio Lower Upper



128 100

130 31.1 12.8 52.8

64 37.5 19.2 59.2



Abundance Ion 128.00 (127.70 to 128.70): BGC
Ion 130.00 (129.70 to 130.70): BGC
Ion 64.00 (63.70 to 64.70): BGC





#9

Benzaldehyde

Concen: 43.360 ng

RT: 7.12 min Scan# 687

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion: 77 Resp: 47370

Ion Ratio Lower Upper




77 100

105 111.2 93.4 133.4

106 102.7 88.4 128.4



Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 105.00 (104.70 to 105.70): BGC
Ion 106.00 (105.70 to 106.70): BGC



#10

Phenol

Concen: 40.123 ng

RT: 7.19 min Scan# 698

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion: 94 Resp: 89019

Ion Ratio Lower Upper



94 100

65 27.5 5.6 45.6

66 41.3 16.0 56.0



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC









#11
bis(2-Chloroethyl)ether
Concen: 36.512 ng
RT: 7.41 min Scan# 736
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30




Instrument :
BNA_G
ClientSampleId :
PB122665BSD


Tgt Ion: 93 Resp: 63620
Ion Ratio Lower Upper
93 100
63 70.3 53.8 93.8
95 32.0 13.1 53.1



#12
1,3-Dichlorobenzene
Concen: 38.492 ng
RT: 7.89 min Scan# 817
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Tgt Ion: 146 Resp: 95868
Ion Ratio Lower Upper
146 100
148 63.1 52.2 78.2
75 23.7 18.0 27.0









#13
 1,4-Dichlorobenzene
 Concen: 38.594 ng
 RT: 8.03 min Scan# 841
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion:146 Resp: 97366
 Ion Ratio Lower Upper
 146 100
 148 64.9 49.5 74.3
 111 37.6 28.1 42.1




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E






#14
 1,2-Dichlorobenzene
 Concen: 38.721 ng
 RT: 8.35 min Scan# 896
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion:146 Resp: 93549
 Ion Ratio Lower Upper
 146 100
 148 62.9 51.8 77.6
 111 37.1 30.9 46.3



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 38.054 ng

RT: 8.24 min Scan# 877

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion: 79 Resp: 59742

Ion Ratio Lower Upper




79 100

108 81.4 67.4 101.2

77 63.4 50.8 76.2



Abundance Ion 79.00 (78.70 to 79.70): BGC
Ion 108.00 (107.70 to 108.70): BGC
Ion 77.00 (76.70 to 77.70): BGC



#16

2,2'-oxybis(1-Chloropropane)

Concen: 33.875 ng

RT: 8.53 min Scan# 926

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion: 45 Resp: 80003

Ion Ratio Lower Upper



45 100


77 16.2 0.0 35.6

79 13.2 0.0 31.6



Abundance Ion 45.00 (44.70 to 45.70): BGC
Ion 77.00 (76.70 to 77.70): BGC
Ion 79.00 (78.70 to 79.70): BGC









#17
2-Methylphenol
Concen: 38.623 ng
RT: 8.45 min Scan# 912
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Instrument :
BNA_G
ClientSampleId :
PB122665BSD

Tgt Ion:	107	Resp:	63116
Ion	Ratio	Lower	Upper
107	100		
108	108.5	92.9	139.3
77	42.8	34.4	51.6
79	42.8	33.8	50.6




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BG
Ion 79.00 (78.70 to 79.70): BG








#18
Hexachloroethane
Concen: 37.407 ng
RT: 9.09 min Scan# 1021
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Tgt Ion:	117	Resp:	32307
Ion	Ratio	Lower	Upper
117	100		
119	90.3	73.8	110.8
201	145.0	108.3	162.5



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 34.099 ng

RT: 8.80 min Scan# 973

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD

Tgt Ion: 70 Resp: 48206





Ion Ratio Lower Upper

70 100

42 55.4 38.8 58.2

101 13.3 9.0 13.6

130 29.8 20.0 30.0



#20

3+4-Methylphenols

Concen: 37.396 ng

RT: 8.78 min Scan# 969

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Tgt Ion: 107 Resp: 84562


Ion Ratio Lower Upper


107 100

108 83.6 68.7 108.7

77 31.1 10.2 50.2

79 26.7 6.9 46.9





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.82 min Scan# 1316

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD

Tgt Ion:136 Resp: 129215


Ion Ratio Lower Upper

136 100

137 10.6 8.7 13.1

54 6.5 4.9 7.3

68 3.7 3.6 5.4






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BGC

Ion 68.00 (67.70 to 68.70): BGC



#22

Acetophenone

Concen: 39.968 ng

RT: 8.82 min Scan# 976

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Tgt Ion:105 Resp: 110460


Ion Ratio Lower Upper

105 100

71 2.4 2.8 4.2#

51 22.7 18.7 28.1

120 25.2 20.1 30.1






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BGC

Ion 51.00 (50.70 to 51.70): BGC

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 63.720 ng
 RT: 9.16 min Scan# 1034
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30




Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD


Tgt Ion: 82 Resp: 119147
 Ion Ratio Lower Upper
 82 100
 128 51.7 40.7 61.1
 54 47.6 36.3 54.5



#24
 Nitrobenzene
 Concen: 39.062 ng
 RT: 9.20 min Scan# 1041
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion: 77 Resp: 73963
 Ion Ratio Lower Upper
 77 100
 123 53.4 43.4 65.0
 65 13.7 11.0 16.6









#25
Isophorone
Concen: 36.070 ng
RT: 9.73 min Scan# 1131
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30




Instrument :
BNA_G
ClientSampleId :
PB122665BSD

Tgt Ion: 82 Resp: 133108
Ion Ratio Lower Upper
82 100
95 9.4 6.6 9.8
138 22.7 18.6 27.8



#26
2-Nitrophenol
Concen: 41.312 ng
RT: 9.93 min Scan# 1164
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Tgt Ion: 139 Resp: 49020
Ion Ratio Lower Upper
139 100
109 25.0 20.5 30.7
65 34.6 30.9 46.3









#27
 2,4-Dimethylphenol
 Concen: 44.590 ng
 RT: 9.99 min Scan# 1174
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion:122 Resp: 72884
 Ion Ratio Lower Upper
 122 100
 107 112.7 90.8 136.2
 121 59.5 49.6 74.4




Abundance Ion 122.00 (121.70 to 122.70): E
 Ion 107.00 (106.70 to 107.70): E
 Ion 121.00 (120.70 to 121.70): E






#28
 bis(2-Chloroethoxy)methane
 Concen: 36.545 ng
 RT: 10.22 min Scan# 1213
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion: 93 Resp: 84999
 Ion Ratio Lower Upper
 93 100
 95 33.6 26.6 40.0
 123 13.9 10.1 15.1



Abundance Ion 93.00 (92.70 to 93.70): BGC
 Ion 95.00 (94.70 to 95.70): BGC
 Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 42.585 ng
 RT: 10.47 min Scan# 1257
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30




Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD


Tgt Ion	Ratio	Lower	Upper
162	100		
164	65.6	48.1	88.1
98	26.7	7.6	47.6



#30
 1,2,4-Trichlorobenzene
 Concen: 39.710 ng
 RT: 10.68 min Scan# 1292
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion	Ratio	Lower	Upper
180	100		
182	96.9	78.7	118.1
145	25.5	21.6	32.4





#31

Naphthalene

Concen: 40.266 ng

RT: 10.87 min Scan# 1324

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:128 Resp: 241559

Ion Ratio Lower Upper

128 100

129 11.2 8.8 13.2




127 13.7 11.4 17.0



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 29.402 ng

RT: 10.13 min Scan# 1199

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:122 Resp: 30353

Ion Ratio Lower Upper

122 100

105 109.4 99.6 139.6



77 74.9 59.5 99.5



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BG








#33
4-Chloroaniline
Concen: 23.457 ng
RT: 10.98 min Scan# 1343
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Instrument :
BNA_G
ClientSampleId :
PB122665BSD

Tgt Ion:	127	Resp:	64960
Ion	Ratio	Lower	Upper
127	100		
129	32.9	26.8	40.2
65	25.2	20.7	31.1
92	17.3	13.6	20.4




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BG
Ion 92.00 (91.70 to 92.70): BG






#34
Hexachlorobutadiene
Concen: 40.238 ng
RT: 11.16 min Scan# 1373
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Tgt Ion:	225	Resp:	70985
Ion	Ratio	Lower	Upper
225	100		
223	63.4	49.8	74.6
227	61.7	50.9	76.3



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 34.376 ng
 RT: 11.75 min Scan# 1475
 Delta R.T. -0.00 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30




Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion:113 Resp: 24530
 Ion Ratio Lower Upper
 113 100
 55 131.4 124.6 164.6
 56 100.8 88.3 128.3



#36
 4-Chloro-3-methylphenol
 Concen: 39.877 ng
 RT: 12.11 min Scan# 1536
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion:107 Resp: 80955
 Ion Ratio Lower Upper
 107 100
 144 31.8 23.6 35.4
 142 93.9 71.0 106.4









#37
 2-Methylnaphthalene
 Concen: 40.166 ng
 RT: 12.48 min Scan# 1599
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30




Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.8	71.4	107.2
115	30.8	25.2	37.8



#38
 1-Methylnaphthalene
 Concen: 39.581 ng
 RT: 12.70 min Scan# 1636
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion	Ratio	Lower	Upper
142	100		
141	92.4	76.2	114.2
116	3.5	2.8	4.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.66 min Scan# 1969

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:164 Resp: 92913

Ion Ratio Lower Upper

164 100

162 102.9 79.9 119.9

160 43.8 34.3 51.5






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 44.193 ng

RT: 12.85 min Scan# 1662

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Tgt Ion:216 Resp: 131861


Ion Ratio Lower Upper

216 100

214 77.4 61.4 92.2

179 18.4 15.1 22.7

108 12.0 9.4 14.0





Abundance


Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 71.600 ng

RT: 12.83 min Scan# 1658

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion: 237 Resp: 112221

Ion Ratio Lower Upper

237 100

235 63.1 43.7 83.7




272 13.2 0.0 33.3



Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 109.549 ng

RT: 16.15 min Scan# 2223

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion: 330 Resp: 144671

Ion Ratio Lower Upper

330 100

332 97.6 77.4 116.0



141 26.6 21.4 32.0



Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#43
 2,4,6-Trichlorophenol
 Concen: 41.016 ng
 RT: 13.09 min Scan# 1702
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30




Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion:196 Resp: 82722
 Ion Ratio Lower Upper
 196 100
 198 96.7 80.7 121.1
 200 32.7 25.7 38.5



#44
 2,4,5-Trichlorophenol
 Concen: 41.653 ng
 RT: 13.17 min Scan# 1716
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion:196 Resp: 87055
 Ion Ratio Lower Upper
 196 100
 198 101.8 80.2 120.4
 97 32.5 26.3 39.5
 132 21.7 16.6 24.8








#45
 2-Fluorobiphenyl
 Concen: 69.873 ng
 RT: 13.28 min Scan# 1734
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30




Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD


Tgt Ion:172 Resp: 383856
 Ion Ratio Lower Upper
 172 100
 171 36.8 30.5 45.7
 170 23.7 20.2 30.2



#46
 1,1'-Biphenyl
 Concen: 41.608 ng
 RT: 13.49 min Scan# 1770
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion:154 Resp: 249898
 Ion Ratio Lower Upper
 154 100
 153 42.7 23.0 63.0
 76 7.7 0.0 29.6









#47
 2-Chloronaphthalene
 Concen: 40.512 ng
 RT: 13.53 min Scan# 1778
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30




Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion	Ratio	Lower	Upper
162	100		
127	32.2	26.0	39.0
164	33.3	27.4	41.0



#48
 2-Nitroaniline
 Concen: 36.432 ng
 RT: 13.73 min Scan# 1812
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion	Ratio	Lower	Upper
65	100		
92	73.2	61.8	92.6
138	141.3	111.0	166.6





#49

Acenaphthylene

Concen: 41.359 ng

RT: 14.38 min Scan# 1922

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:152 Resp: 322238

Ion Ratio Lower Upper

152 100

151 21.4 17.0 25.4




153 13.3 10.6 16.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 38.511 ng

RT: 14.11 min Scan# 1876

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:163 Resp: 258181

Ion Ratio Lower Upper

163 100

194 5.2 4.2 6.2



164 10.1 8.8 13.2




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 40.161 ng
 RT: 14.23 min Scan# 1896
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30




Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion	Ratio	Lower	Upper
165	100		
63	36.5	30.6	45.8
89	39.3	32.6	49.0



#52
 Acenaphthene
 Concen: 39.271 ng
 RT: 14.72 min Scan# 1980
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion	Ratio	Lower	Upper
154	100		
153	113.5	90.2	135.4
152	52.5	43.2	64.8









#53
3-Nitroaniline
Concen: 27.728 ng
RT: 14.56 min Scan# 1952
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30




Instrument :
BNA_G
ClientSampleId :
PB122665BSD


Tgt Ion:138 Resp: 43092
Ion Ratio Lower Upper
138 100
108 9.2 8.3 12.5
92 88.6 78.2 117.4



#54
2,4-Dinitrophenol
Concen: 62.977 ng
RT: 14.77 min Scan# 1989
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Tgt Ion:184 Resp: 64454
Ion Ratio Lower Upper
184 100
63 39.5 31.7 47.5
154 52.8 42.0 63.0





#55

Dibenzenofuran

Concen: 41.590 ng

RT: 15.06 min Scan# 2037

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:168 Resp: 325697

Ion Ratio Lower Upper

168 100

139 34.4 27.2 40.8




169 14.0 11.5 17.3



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#56

4-Nitrophenol

Concen: 78.474 ng

RT: 14.88 min Scan# 2007

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:139 Resp: 86658

Ion Ratio Lower Upper

139 100

109 70.5 43.9 83.9



65 74.9 56.3 96.3



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BG







#57
2,4-Dinitrotoluene
Concen: 39.425 ng
RT: 15.02 min Scan# 2031
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Instrument :
BNA_G
ClientSampleId :
PB122665BSD

Tgt Ion	Ratio	Lower	Upper
165	100		
63	27.9	22.5	33.7
89	50.6	37.8	56.6
182	3.6	3.0	4.4





Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BG



Ion 89.00 (88.70 to 89.70): BG

Ion 182.00 (181.70 to 182.70): E



#58
Fluorene
Concen: 40.658 ng
RT: 15.71 min Scan# 2148
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30


Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.0	79.7	119.5
167	14.1	10.9	16.3



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E






#59
 2,3,4,6-Tetrachlorophenol
 Concen: 42.377 ng
 RT: 15.29 min Scan# 2076
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion	Ratio	Lower	Upper
232	100		
131	31.7	22.9	34.3
130	0.5	1.4	2.0
166	25.6	17.0	25.4






Abundance

Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E


Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#60
 Diethylphthalate
 Concen: 37.825 ng
 RT: 15.47 min Scan# 2108
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion	Ratio	Lower	Upper
149	100		
177	26.8	20.4	30.6
150	13.0	10.3	15.5





Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E









#61
 4-Chlorophenyl-phenylether
 Concen: 40.154 ng
 RT: 15.70 min Scan# 2146
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion: 204 Resp: 154952
 Ion Ratio Lower Upper
 204 100
 206 34.3 27.2 40.8
 141 46.7 38.7 58.1




Abundance Ion 204.00 (203.70 to 204.70): E
 Ion 206.00 (205.70 to 206.70): E
 Ion 141.00 (140.70 to 141.70): E






#62
 4-Nitroaniline
 Concen: 37.155 ng
 RT: 15.73 min Scan# 2151
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion: 138 Resp: 61798
 Ion Ratio Lower Upper
 138 100
 92 39.8 24.2 64.2
 108 70.8 46.8 86.8



Abundance Ion 138.00 (137.70 to 138.70): E
 Ion 92.00 (91.70 to 92.70): BG
 Ion 108.00 (107.70 to 108.70): E









#63
Azobenzene
Concen: 37.597 ng
RT: 15.99 min Scan# 2196
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Instrument :
BNA_G
ClientSampleId :
PB122665BSD

Tgt Ion: 77 Resp: 168833
Ion Ratio Lower Upper
77 100
182 44.6 19.9 59.9
105 22.7 2.9 42.9
51 26.6 7.6 47.6




Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 182.00 (181.70 to 182.70): BGC
Ion 105.00 (104.70 to 105.70): BGC
Ion 51.00 (50.70 to 51.70): BGC





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.41 min Scan# 2437
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Tgt Ion: 188 Resp: 216951
Ion Ratio Lower Upper
188 100
94 4.9 4.5 6.7
80 5.2 4.9 7.3



Abundance Ion 188.00 (187.70 to 188.70): BGC
Ion 94.00 (93.70 to 94.70): BGC
Ion 80.00 (79.70 to 80.70): BGC





#65

4,6-Dinitro-2-methylphenol

Concen: 40.149 ng

RT: 15.79 min Scan# 2162

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:198 Resp: 54610

Ion Ratio Lower Upper

198 100

51 23.8 4.2 44.2




105 26.8 8.8 48.8



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BG

Ion 105.00 (104.70 to 105.70): E



#66

n-Nitrosodiphenylamine

Concen: 40.129 ng

RT: 15.91 min Scan# 2183

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:169 Resp: 239420

Ion Ratio Lower Upper

169 100

168 67.5 54.3 81.5



167 35.0 28.2 42.2



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#67
 4-Bromophenyl-phenylether
 Concen: 39.260 ng
 RT: 16.59 min Scan# 2298
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Instrument :
 BNA_G
 ClientSampleId :
 PB122665BSD

Tgt Ion: 248 Resp: 108040
 Ion Ratio Lower Upper
 248 100
 250 95.4 78.9 118.3
 141 51.0 43.3 64.9




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E






#68
 Hexachlorobenzene
 Concen: 38.788 ng
 RT: 16.72 min Scan# 2320
 Delta R.T. -0.01 min
 Lab File: BG042598.D
 Acq: 30 Aug 2019 13:30

Tgt Ion: 284 Resp: 116034
 Ion Ratio Lower Upper
 284 100
 142 25.6 18.2 27.4
 249 32.1 25.0 37.4



Abundance Ion 283.80 (283.50 to 284.50): E
 Ion 142.00 (141.70 to 142.70): E
 Ion 249.00 (248.70 to 249.70): E





#69

Atrazine

Concen: 42.951 ng

RT: 16.86 min Scan# 2344

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:200 Resp: 90622

Ion Ratio Lower Upper

200 100

173 25.1 4.9 44.9




215 54.5 30.7 70.7



Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#70

Pentachlorophenol

Concen: 85.551 ng

RT: 17.07 min Scan# 2379

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:266 Resp: 132975

Ion Ratio Lower Upper

266 100

268 65.8 52.1 78.1



264 65.3 49.4 74.2



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E









#71
Phenanthrene
Concen: 41.563 ng
RT: 17.45 min Scan# 2445
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30




Instrument :
BNA_G
ClientSampleId :
PB122665BSD


Tgt Ion:178 Resp: 447901
Ion Ratio Lower Upper
178 100
176 18.9 15.9 23.9
179 16.0 12.7 19.1



#72
Anthracene
Concen: 42.335 ng
RT: 17.54 min Scan# 2460
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Tgt Ion:178 Resp: 455442
Ion Ratio Lower Upper
178 100
176 18.5 15.5 23.3
179 16.6 12.9 19.3





#73

Carbazole

Concen: 42.200 ng

RT: 17.81 min Scan# 2506

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:167 Resp: 404617

Ion Ratio Lower Upper

167 100

166 22.9 18.9 28.3




139 12.3 9.8 14.6



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#74

Di-n-butylphthalate

Concen: 40.047 ng

RT: 18.37 min Scan# 2601

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:149 Resp: 453886

Ion Ratio Lower Upper

149 100

150 10.0 8.1 12.1



104 4.5 3.7 5.5




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#75

Fluoranthene

Concen: 46.173 ng

RT: 19.46 min Scan# 2787

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:202 Resp: 607253

Ion Ratio Lower Upper

202 100

101 6.5 0.0 28.3




203 19.0 0.0 39.6



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.69 min Scan# 3166

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:240 Resp: 247140

Ion Ratio Lower Upper

240 100

120 5.7 5.7 8.5#



236 25.8 21.0 31.4



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Benzidine

Concen: 33.949 ng

RT: 19.64 min Scan# 2817

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:184 Resp: 240552

Ion Ratio Lower Upper


184 100

185 15.4 12.2 18.4



183 13.4 10.8 16.2



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 185.00 (184.70 to 185.70): E
Ion 183.00 (182.70 to 183.70): E



Time-->



#78

Pyrene

Concen: 40.590 ng

RT: 19.83 min Scan# 2849

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:202 Resp: 615017

Ion Ratio Lower Upper


202 100

200 21.4 17.8 26.6


203 18.5 15.5 23.3



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E



Time-->





#79

Terphenyl-d14

Concen: 67.659 ng

RT: 20.02 min Scan# 2882

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:244 Resp: 773447

Ion Ratio Lower Upper

244 100

212 6.8 6.0 9.0



122 6.3 6.7 10.1#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 37.216 ng

RT: 20.71 min Scan# 2999

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:149 Resp: 221796

Ion Ratio Lower Upper

149 100

91 54.1 45.9 68.9



206 29.6 22.5 33.7




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BG

Ion 206.00 (205.70 to 206.70): E





#81

Benzo(a)anthracene

Concen: 40.890 ng

RT: 21.67 min Scan# 3162

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:228 Resp: 614745

Ion Ratio Lower Upper

228 100

226 28.0 23.4 35.2




229 20.5 17.4 26.2



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

3,3'-Dichlorobenzidine

Concen: 35.738 ng

RT: 21.58 min Scan# 3147

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:252 Resp: 216088

Ion Ratio Lower Upper

252 100

254 64.6 53.0 79.4



126 6.7 5.7 8.5




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#83

Chrysene

Concen: 40.687 ng

RT: 21.74 min Scan# 3174

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:228 Resp: 589926

Ion Ratio Lower Upper

228 100

226 30.5 24.9 37.3




229 20.9 16.9 25.3



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Bis(2-ethylhexyl)phthalate

Concen: 38.665 ng

RT: 21.57 min Scan# 3145

Delta R.T. -0.01 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:149 Resp: 319933

Ion Ratio Lower Upper

149 100

167 30.8 26.0 39.0



279 6.6 5.4 8.0



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E









#85
Di-n-octyl phthalate
Concen: 39.159 ng
RT: 22.78 min Scan# 3352
Delta R.T. -0.02 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30




Instrument :
BNA_G
ClientSampleId :
PB122665BSD

Tgt Ion	Ratio	Lower	Upper
149	100		
167	1.8	1.6	2.4
43	6.2	5.3	7.9



#86
Indeno(1,2,3-cd)pyrene
Concen: 39.527 ng
RT: 28.63 min Scan# 4348
Delta R.T. -0.02 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Tgt Ion	Ratio	Lower	Upper
276	100		
138	15.2	9.2	13.8#
277	26.3	21.0	31.6









#87
Perylene-d12
Concen: 20.000 ng
RT: 24.93 min Scan# 3717
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Instrument :
BNA_G
ClientSampleId :
PB122665BSD

Tgt Ion: 264 Resp: 266228




Ion	Ratio	Lower	Upper
264	100		
260	22.9	18.5	27.7
265	22.4	17.6	26.4



#88
Benzo(b)fluoranthene
Concen: 46.772 ng
RT: 23.90 min Scan# 3542
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Tgt Ion: 252 Resp: 684560

Ion	Ratio	Lower	Upper
252	100		
253	21.5	18.6	27.8
125	5.7	5.6	8.4









#89
Benzo(k)fluoranthene
Concen: 46.016 ng
RT: 23.96 min Scan# 3553
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30




Instrument :
BNA_G
ClientSampleId :
PB122665BSD


Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	18.6	27.8
125	5.6	5.8	8.6#



#90
Benzo(a)pyrene
Concen: 47.492 ng
RT: 24.77 min Scan# 3691
Delta R.T. -0.01 min
Lab File: BG042598.D
Acq: 30 Aug 2019 13:30

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.5	18.4	27.6
125	6.1	6.0	9.0





#91

Dibenzo(a,h)anthracene

Concen: 45.785 ng

RT: 28.69 min Scan# 4358

Delta R.T. -0.02 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30

Instrument :

BNA_G

ClientSampleId :

PB122665BSD


Tgt Ion:278 Resp: 643840

Ion Ratio Lower Upper




278 100

139 9.8 8.8 13.2

279 24.6 19.1 28.7



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 46.655 ng

RT: 29.79 min Scan# 4545

Delta R.T. -0.02 min

Lab File: BG042598.D

Acq: 30 Aug 2019 13:30


Tgt Ion:276 Resp: 656170

Ion Ratio Lower Upper

276 100

277 23.6 19.8 29.8

138 12.1 11.4 17.0



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

