

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG083118\
 Data File : BG036517.D
 Acq On : 31 Aug 2018 23:02
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 01 01:03:50 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 31 13:03:20 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2	1,4-Dioxane	0.588	0.560	4.8	120	0.00
3	Pyridine	1.652	1.670	-1.1	120	0.00
4	n-Nitrosodimethylamine	0.736	0.677	8.0	111	0.00
5 S	2-Fluorophenol	1.261	1.277	-1.3	122	0.00
6	Aniline	2.291	2.206	3.7	117	0.00
7 S	Phenol-d6	1.805	1.831	-1.4	123	0.00
8	2-Chlorophenol	1.367	1.336	2.3	118	0.00
9	Benzaldehyde	1.243	1.151	7.4	120	0.00
10 C	Phenol	1.889	1.901	-0.6	122	0.00
11	bis(2-Chloroethyl)ether	1.498	1.430	4.5	117	0.00
12	1,3-Dichlorobenzene	1.561	1.536	1.6	122	0.00
13 C	1,4-Dichlorobenzene	1.576	1.563	0.8	122	0.00
14	1,2-Dichlorobenzene	1.505	1.472	2.2	121	0.00
15	Benzyl Alcohol	1.455	1.412	3.0	116	0.00
16	2,2'-oxybis(1-Chloropropane	3.020	2.609	13.6	106	0.00
17	2-Methylphenol	1.254	1.236	1.4	121	0.00
18	Hexachloroethane	0.565	0.552	2.3	118	0.00
19 P	n-Nitroso-di-n-propylamine	1.349	1.231	8.7	112	0.00
20	3+4-Methylphenols	1.751	1.748	0.2	121	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	125	0.00
22	Acetophenone	0.525	0.483	8.0	115	0.00
23 S	Nitrobenzene-d5	0.369	0.392	-6.2	130	0.00
24	Nitrobenzene	0.397	0.392	1.3	120	0.00
25	Isophorone	0.732	0.662	9.6	112	0.00
26 C	2-Nitrophenol	0.158	0.176	-11.4	139	0.00
27	2,4-Dimethylphenol	0.292	0.272	6.8	117	0.00
28	bis(2-Chloroethoxy)methane	0.449	0.416	7.3	115	0.00
29 C	2,4-Dichlorophenol	0.320	0.338	-5.6	129	0.00
30	1,2,4-Trichlorobenzene	0.374	0.367	1.9	125	0.00
31	Naphthalene	1.000	0.941	5.9	117	0.00
32	Benzoic acid	0.202	0.195	3.5	118	0.00
33	4-Chloroaniline	0.458	0.442	3.5	119	0.00
34 C	Hexachlorobutadiene	0.251	0.256	-2.0	125	0.00
35	Caprolactam	0.127	0.124	2.4	116	0.00
36 C	4-Chloro-3-methylphenol	0.371	0.372	-0.3	122	0.00
37	2-Methylnaphthalene	0.732	0.706	3.6	119	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
39	1,2,4,5-Tetrachlorobenzene	0.708	0.684	3.4	124	0.00
40 P	Hexachlorocyclopentadiene	0.368	0.366	0.5	126	0.00
41 S	2,4,6-Tribromophenol	0.239	0.263	-10.0	133	0.00
42 C	2,4,6-Trichlorophenol	0.431	0.446	-3.5	130	0.00
43	2,4,5-Trichlorophenol	0.460	0.480	-4.3	130	0.00
44 S	2-Fluorobiphenyl	1.401	1.369	2.3	127	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG083118\
 Data File : BG036517.D
 Acq On : 31 Aug 2018 23:02
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 01 01:03:50 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 31 13:03:20 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.660	1.526	8.1	121	0.00
46	2-Chloronaphthalene	1.267	1.198	5.4	122	0.00
47	2-Nitroaniline	0.398	0.416	-4.5	127	0.00
48	Acenaphthylene	1.981	1.843	7.0	120	0.00
49	Dimethylphthalate	1.633	1.551	5.0	122	0.00
50	2,6-Dinitrotoluene	0.336	0.339	-0.9	127	0.00
51 C	Acenaphthene	1.269	1.186	6.5	120	0.00
52	3-Nitroaniline	0.362	0.394	-8.8	129	0.00
53 P	2,4-Dinitrophenol	0.127	0.118	7.1	117	0.00
54	Dibenzofuran	1.987	1.882	5.3	123	0.00
55 P	4-Nitrophenol	0.296	0.290	2.0	119	0.00
56	2,4-Dinitrotoluene	0.438	0.494	-12.8	141	0.00
57	Fluorene	1.486	1.412	5.0	121	0.00
58	2,3,4,6-Tetrachlorophenol	0.432	0.460	-6.5	133	0.00
59	Diethylphthalate	1.646	1.545	6.1	119	0.00
60	4-Chlorophenyl-phenylether	0.917	0.895	2.4	127	0.00
61	4-Nitroaniline	0.379	0.405	-6.9	128	0.00
62	Azobenzene	1.675	1.487	11.2	114	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
64	4,6-Dinitro-2-methylphenol	0.088	0.099	-12.5	148	0.00
65 c	n-Nitrosodiphenylamine	0.594	0.548	7.7	125	0.00
66	4-Bromophenyl-phenylether	0.234	0.230	1.7	130	0.00
67	Hexachlorobenzene	0.239	0.230	3.8	127	0.00
68	Atrazine	0.223	0.195	12.6	119	0.00
69 C	Pentachlorophenol	0.163	0.166	-1.8	125	0.00
70	Phenanthrene	1.016	0.953	6.2	125	0.00
71	Anthracene	1.013	0.943	6.9	123	0.00
72	Carbazole	1.012	0.938	7.3	121	0.00
73	Di-n-butylphthalate	1.127	1.034	8.3	118	0.00
74 C	Fluoranthene	1.239	1.166	5.9	123	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	130	0.00
76	Benzidine	0.638	0.543	14.9	118	0.00
77	Pyrene	1.230	1.166	5.2	123	0.00
78 S	Terphenyl-d14	0.856	0.837	2.2	129	0.00
79	Butylbenzylphthalate	0.489	0.464	5.1	119	0.00
80	Benzo(a)anthracene	1.212	1.142	5.8	125	0.00
81	3,3'-Dichlorobenzidine	0.483	0.459	5.0	121	0.00
82	Chrysene	1.148	1.083	5.7	124	0.00
83	Bis(2-ethylhexyl)phthalate	0.685	0.637	7.0	119	0.00
84 c	Di-n-octyl phthalate	1.144	1.077	5.9	118	0.00
85	Indeno(1,2,3-cd)pyrene	1.334	1.235	7.4	120	0.00
86 I	Perylene-d12	1.000	1.000	0.0	131	0.00
87	Benzo(b)fluoranthene	1.111	1.046	5.9	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG083118\
Data File : BG036517.D
Acq On : 31 Aug 2018 23:02
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 01 01:03:50 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 31 13:03:20 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.085	1.051	3.1	127	0.00
89 C	Benzo(a)pyrene	1.067	1.031	3.4	126	0.00
90	Dibenzo(a,h)anthracene	1.030	0.936	9.1	119	0.00
91	Benzo(a,h,i)perylene	1.035	0.965	6.8	122	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0