




Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG090120\
Data File : BG046483.D
Acq On : 1 Sep 2020 13:37
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 01 18:22:11 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 31 16:05:33 2020
Response via : Initial Calibration



(70) Pentachlorophenol (C)




16.962min (-0.001) 30982.79ng

response 61911

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	66.31
264.00	62.70	65.43
0.00	0.00	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG090120\
 Data File : BG046483.D
 Acq On : 1 Sep 2020 13:37
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 01 18:22:11 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 31 16:05:33 2020
 Response via : Initial Calibration



TIC: BG046483.D

(77) Benzidine

19.536min (-0.001) 0.00ng

response 514981

Ion	Exp%	Act%
184.00	100	100
185.00	14.80	14.67
183.00	13.40	12.11
0.00	0.00	0.00