

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090221\
 Data File : BG050139.D
 Acq On : 3 Sep 2021 20:19
 Operator : CG/JU
 Sample : M3549-03MSD
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 EW7A1MSD

Manual Integrations
 APPROVED

mohammad
 9/7/2021 11:34:01 AM

Quant Time: Sep 03 23:47:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG081921.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 30 14:59:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.950	152	35088	20.000	ng/u1	0.00
20) Naphthalene-d8	10.769	136	136363	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.611	164	90821	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.366	188	201972	20.000	ng/u1	#-0.01
79) Chrysene-d12	21.643	240	198270	20.000	ng/u1	0.00
88) Perylene-d12	24.862	264	201018	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.309	96	3741	5.498	ng/uL	0.00
4) Pyridine-d5	3.738	84	24377	11.894	ng/u1	0.00
7) Phenol-d5	7.121	99	26031	8.737	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.268	67	54995	33.223	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.480	132	59456	26.669	ng/u1	-0.01
15) 4-Methylphenol-d8	8.678	113	42867	18.268	ng/u1	0.00
21) Nitrobenzene-d5	9.124	128	37398	35.375	ng/u1	0.00
24) 2-Nitrophenol-d4	9.847	143	41902	33.286	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.405	165	75137	33.464	ng/u1	0.00
31) 4-Chloroaniline-d4	10.910	131	86711	28.373	ng/u1	-0.01
46) Dimethylphthalate-d6	14.018	166	258787	37.232	ng/u1	0.00
49) Acenaphthylene-d8	14.306	160	298770	36.660	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.852	143	9269	9.217	ng/u1	0.00
60) Fluorene-d10	15.610	176	215835	36.625	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	44423	34.188	ng/u1	0.00
73) Anthracene-d10	17.466	188	337411	37.203	ng/u1	-0.01
81) Pyrene-d10	19.757	212	403731	37.681	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.645	264	397365	37.260	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.338	88	5653m	6.815	ng/uL	
5) Pyridine	3.755	79	30545	13.789	ng/u1	85
6) Benzaldehyde	7.080	77	71435	41.663	ng/u1	99
8) Phenol	7.151	94	30399	10.101	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.362	93	77332	37.223	ng/u1	94
12) 2-Chlorophenol	7.515	128	68837	31.247	ng/u1#	91
13) 2-Methylphenol	8.408	108	58140	25.041	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.472	45	81717	37.510	ng/u1	96
16) Acetophenone	8.784	105	140653	36.755	ng/u1#	89
17) N-Nitroso-di-n-propyla...	8.760	70	71509	37.215	ng/u1	97
18) 4-Methylphenol	8.737	108	53999	21.618	ng/u1	86
19) Hexachloroethane	9.030	117	38080	39.497	ng/u1	94
22) Nitrobenzene	9.166	77	113283	39.546	ng/u1	99
23) Isophorone	9.688	82	200992	39.380	ng/u1	97
25) 2-Nitrophenol	9.882	139	47136	38.514	ng/u1	94
26) 2,4-Dimethylphenol	9.947	107	95890	35.254	ng/u1	92
27) Bis(2-Chloroethoxy)met...	10.170	93	107093	40.633	ng/u1	96
29) 2,4-Dichlorophenol	10.429	162	83475	40.506	ng/u1	100
30) Naphthalene	10.822	128	270075	39.625	ng/u1	98
32) 4-Chloroaniline	10.934	127	93647	31.818	ng/u1	91
33) Hexachlorobutadiene	11.098	225	64972	38.973	ng/u1	91
34) Caprolactam	11.727	113	5322m	7.372	ng/u1	
35) 4-Chloro-3-methylphenol	12.073	107	93343	37.899	ng/u1	97
36) 2-Methylnaphthalene	12.432	142	200849	39.642	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090221\
 Data File : BG050139.D
 Acq On : 3 Sep 2021 20:19
 Operator : CG/JU
 Sample : M3549-03MSD
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 EW7A1MSD

Manual Integrations
 APPROVED

mohammad
 9/7/2021 11:34:01 AM

Quant Time: Sep 03 23:47:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG081921.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 30 14:59:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.655	142	196341	38.392	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.802	216	114337	42.862	ng/ul#	92
40) Hexachlorocyclopentadiene	12.772	237	12439	10.419	ng/ul#	95
41) 2,4,6-Trichlorophenol	13.049	196	74298	43.706	ng/ul	97
42) 2,4,5-Trichlorophenol	13.131	196	80216	45.023	ng/ul	97
43) 1,1'-Biphenyl	13.442	154	263664	41.554	ng/ul	97
44) 2-Chloronaphthalene	13.489	162	206969	40.609	ng/ul	96
45) 2-Nitroaniline	13.695	65	75838	43.827	ng/ul	94
47) Dimethylphthalate	14.065	163	288819	41.987	ng/ul	99
48) 2,6-Dinitrotoluene	14.194	165	62638	44.106	ng/ul	93
50) Acenaphthylene	14.335	152	314300	42.196	ng/ul	96
51) 3-Nitroaniline	14.529	138	56454	41.195	ng/ul	97
52) Acenaphthene	14.676	153	231376	42.822	ng/ul	98
53) 2,4-Dinitrophenol	14.752	184	20421	28.153	ng/ul	91
55) 4-Nitrophenol	14.864	109	12229	9.591	ng/ul	96
56) Dibenzofuran	15.011	168	318967	42.628	ng/ul	100
57) 2,4-Dinitrotoluene	14.993	165	93051	45.550	ng/ul#	93
58) 2,3,4,6-Tetrachlorophenol	15.251	232	68892	46.662	ng/ul	95
59) Diethylphthalate	15.428	149	302898	43.044	ng/ul	96
61) Fluorene	15.663	166	273107	43.262	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.651	204	141157	42.261	ng/ul	96
63) 4-Nitroaniline	15.692	138	60428	44.186	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.763	198	46078	39.123	ng/ul	95
67) N-Nitrosodiphenylamine	15.868	169	240923	45.308	ng/ul	96
68) 4-Bromophenyl-phenylether	16.550	248	102584	45.914	ng/ul	99
69) Hexachlorobenzene	16.679	284	114570	44.407	ng/ul	93
70) Atrazine	16.826	200	105695	44.742	ng/ul	95
71) Pentachlorophenol	17.026	266	40879	37.418	ng/ul	96
72) Phenanthrene	17.413	178	450906	44.782	ng/ul	97
74) Anthracene	17.501	178	446154	43.726	ng/ul	94
75) 1,2,3,4-Tetrachloroben...	13.413	216	117046	43.194	ng/ul	92
76) Pentachlorobenzene	14.940	250	125280	43.632	ng/ul	94
77) Carbazole	17.777	167	414154	45.615	ng/ul	99
78) Di-n-butylphthalate	18.318	149	507630	43.716	ng/ul	99
80) Fluoranthene	19.428	202	547111	41.380	ng/ul#	96
82) Pyrene	19.787	202	542059	41.374	ng/ul	98
83) Butylbenzylphthalate	20.662	149	219053	42.369	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.537	252	31338	6.668	ng/ul#	91
85) Benzo(a)anthracene	21.619	228	537315	43.492	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.508	149	312565	44.021	ng/ul	100
87) Chrysene	21.690	228	509397	43.619	ng/ul	95
89) Di-n-octyl phthalate	22.712	149	517043	43.358	ng/ul	100
90) Benzo(b)fluoranthene	23.840	252	553000	43.007	ng/ul#	96
91) Benzo(k)fluoranthene	23.904	252	536380	43.885	ng/ul	99
93) Benzo(a)pyrene	24.709	252	485749	43.499	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	28.557	276	654198	44.411	ng/ul#	98
95) Dibenzo(a,h)anthracene	28.610	278	541761	43.933	ng/ul#	99
96) Benzo(g,h,i)perylene	29.714	276	544328	43.954	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090221\
 Data File : BG050139.D
 Acq On : 3 Sep 2021 20:19
 Operator : CG/JU
 Sample : M3549-03MSD
 Misc :
 ALS Vial : 41 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 EW7A1MSD

Manual Integrations
 APPROVED

mohammad
 9/7/2021 11:34:01 AM

Quant Time: Sep 03 23:47:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG081921.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 30 14:59:55 2021
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.815 ng/uL m
RT: 3.338 min Scan# 51
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Instrument :
BNA_G
ClientSampleId :
EW7A1MSD

Tgt Ion:	88	Resp:	5653
Ion	Ratio	Lower	Upper
88	100		
43	25.4	21.6	32.4
58	57.3	52.0	78.0




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#5
Pyridine
Concen: 13.789 ng/uL
RT: 3.755 min Scan# 122
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19


Tgt Ion:	79	Resp:	30545
Ion	Ratio	Lower	Upper
79	100		
52	62.4	61.1	91.7
51	35.9	34.8	52.2





#6
Benzaldehyde
Concen: 41.663 ng/ul
RT: 7.080 min Scan# 688
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD



Tgt Ion:	77	Resp:	71435
Ion Ratio	100	Lower	Upper
77	100		
105	89.5	72.3	108.5
106	89.9	72.3	108.5




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#8
Phenol
Concen: 10.101 ng/ul
RT: 7.151 min Scan# 700
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19


Tgt Ion:	94	Resp:	30399
Ion Ratio	100	Lower	Upper
94	100		
65	32.2	27.7	41.5
66	49.4	39.0	58.4





#10
 Bis(2-Chloroethyl)ether
 Concen: 37.223 ng/ul
 RT: 7.362 min Scan# 736
 Delta R.T. -0.011 min
 Lab File: BG050139.D
 Acq: 3 Sep 2021 20:19




Instrument :
 BNA_G
 ClientSampleId :
 EW7A1MSD




Tgt Ion:	93	Resp:	77332
Ion	Ratio	Lower	Upper
93	100		
63	81.1	59.9	89.9
95	32.6	27.6	41.4

Manual Integrations
 APPROVED



mohammad
 9/7/2021 11:34:01 AM




#12
 2-Chlorophenol
 Concen: 31.247 ng/ul
 RT: 7.515 min Scan# 762
 Delta R.T. -0.005 min
 Lab File: BG050139.D
 Acq: 3 Sep 2021 20:19




Tgt Ion:	128	Resp:	68837
Ion	Ratio	Lower	Upper
128	100		
64	42.9	28.0	42.0#
130	32.3	24.4	36.6





#13
2-Methylphenol
Concen: 25.041 ng/ul
RT: 8.408 min Scan# 914
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:108 Resp: 58140
Ion Ratio Lower Upper
108 100
107 89.0 71.0 106.4

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#14
2,2'-oxybis(1-Chloropropane)
Concen: 37.510 ng/ul
RT: 8.472 min Scan# 925
Delta R.T. -0.017 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion: 45 Resp: 81717
Ion Ratio Lower Upper
45 100
77 22.8 16.9 25.3
79 15.4 11.0 16.4





#16
Acetophenone
Concen: 36.755 ng/ul
RT: 8.784 min Scan# 978
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:105 Resp: 140653
Ion Ratio Lower Upper
105 100
77 76.4 67.7 101.5
51 23.5 25.3 37.9#

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#17
N-Nitroso-di-n-propylamine
Concen: 37.215 ng/ul
RT: 8.760 min Scan# 974
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19



Tgt Ion: 70 Resp: 71509
Ion Ratio Lower Upper
70 100
42 53.3 44.2 66.4
101 9.2 7.8 11.8
130 17.7 15.3 22.9









#18
4-Methylphenol
Concen: 21.618 ng/ul
RT: 8.737 min Scan# 970
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Instrument :
BNA_G
ClientSampleId :
EW7A1MSD

Tgt Ion:108 Resp: 53999
Ion Ratio Lower Upper
108 100
107 111.5 78.3 117.5




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#19
Hexachloroethane
Concen: 39.497 ng/ul
RT: 9.030 min Scan# 1020
Delta R.T. -0.017 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19


Tgt Ion:117 Resp: 38080
Ion Ratio Lower Upper
117 100
201 101.8 89.3 133.9
199 62.4 50.5 75.7





#22
Nitrobenzene
Concen: 39.546 ng/ul
RT: 9.166 min Scan# 1043
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion	Ratio	Lower	Upper
77	100		
123	37.4	30.4	45.6
65	14.6	11.3	16.9

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#23
Isophorone
Concen: 39.380 ng/ul
RT: 9.688 min Scan# 1132
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion	Ratio	Lower	Upper
82	100		
95	8.7	7.8	11.8
138	15.2	11.4	17.0





#25
2-Nitrophenol
Concen: 38.514 ng/ul
RT: 9.882 min Scan# 1165
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampled :
EW7A1MSD




Tgt Ion	Ratio	Lower	Upper
139	100		
65	64.5	53.1	79.7
109	47.8	44.1	66.1

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#26
2,4-Dimethylphenol
Concen: 35.254 ng/ul
RT: 9.947 min Scan# 1176
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion	Ratio	Lower	Upper
107	100		
121	43.0	32.7	49.1
122	74.7	52.6	79.0





#27
 Bis(2-Chloroethoxy)methane
 Concen: 40.633 ng/ul
 RT: 10.170 min Scan# 1214
 Delta R.T. -0.011 min
 Lab File: BG050139.D
 Acq: 3 Sep 2021 20:19




Instrument :
 BNA_G
 ClientSampleId :
 EW7A1MSD




Tgt Ion: 93 Resp: 107093
 Ion Ratio Lower Upper
 93 100
 95 33.5 24.8 37.2
 123 9.6 8.2 12.2

Manual Integrations
 APPROVED



mohammad
 9/7/2021 11:34:01 AM




#29
 2,4-Dichlorophenol
 Concen: 40.506 ng/ul
 RT: 10.429 min Scan# 1258
 Delta R.T. -0.005 min
 Lab File: BG050139.D
 Acq: 3 Sep 2021 20:19




Tgt Ion: 162 Resp: 83475
 Ion Ratio Lower Upper
 162 100
 164 63.5 51.0 76.6
 98 34.5 27.8 41.8





#30
Naphthalene
Concen: 39.625 ng/ul
RT: 10.822 min Scan# 1325
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.6	9.3	13.9
127	12.7	9.8	14.6

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#32
4-Chloroaniline
Concen: 31.818 ng/ul
RT: 10.934 min Scan# 1344
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion	Ratio	Lower	Upper
127	100		
129	32.4	30.3	45.5





#33
Hexachlorobutadiene
Concen: 38.973 ng/ul
RT: 11.098 min Scan# 1372
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:225 Resp: 64972
Ion Ratio Lower Upper
225 100
223 55.5 51.3 76.9
227 62.9 55.1 82.7

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#34
Caprolactam
Concen: 7.372 ng/ul m
RT: 11.727 min Scan# 1479
Delta R.T. 0.019 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19



Tgt Ion:113 Resp: 5322
Ion Ratio Lower Upper
113 100
55 180.7 114.0 171.0#
56 125.0 107.4 161.0









#35
4-Chloro-3-methylphenol
Concen: 37.899 ng/ul
RT: 12.073 min Scan# 1538
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Tgt Ion:107 Resp: 93343
Ion Ratio Lower Upper
107 100
144 21.3 18.9 28.3
142 71.4 59.2 88.8

Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#36
2-Methylnaphthalene
Concen: 39.642 ng/ul
RT: 12.432 min Scan# 1599
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Tgt Ion:142 Resp: 200849
Ion Ratio Lower Upper
142 100
141 85.5 70.5 105.7









#37
1-Methylnaphthalene
Concen: 38.392 ng/ul
RT: 12.655 min Scan# 1637
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Tgt Ion	Ratio	Lower	Upper
142	100		
141	89.7	75.8	113.8
116	4.8	3.7	5.5

Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM




#39
1,2,4,5-Tetrachlorobenzene
Concen: 42.862 ng/ul
RT: 12.802 min Scan# 1662
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Tgt Ion	Ratio	Lower	Upper
216	100		
214	73.1	64.2	96.2
179	21.7	12.8	19.2#
108	17.3	13.1	19.7





#40
Hexachlorocyclopentadiene
Concen: 10.419 ng/ul
RT: 12.772 min Scan# 1657
Delta R.T. -0.017 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19






Tgt Ion:237 Resp: 12439
Ion Ratio Lower Upper
237 100
235 64.1 53.5 80.3
272 9.6 11.5 17.3#

Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#41
2,4,6-Trichlorophenol
Concen: 43.706 ng/ul
RT: 13.049 min Scan# 1704
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Tgt Ion:196 Resp: 74298
Ion Ratio Lower Upper
196 100
198 101.3 83.7 125.5
200 33.5 25.9 38.9









#42
2,4,5-Trichlorophenol
Concen: 45.023 ng/ul
RT: 13.131 min Scan# 1718
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Instrument :
BNA_G
ClientSampleId :
EW7A1MSD

Tgt Ion:196 Resp: 80216
Ion Ratio Lower Upper
196 100
198 91.1 70.3 105.5
200 31.7 26.6 40.0




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#43
1,1'-Biphenyl
Concen: 41.554 ng/ul
RT: 13.442 min Scan# 1771
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19


Tgt Ion:154 Resp: 263664
Ion Ratio Lower Upper
154 100
153 42.5 32.5 48.7
76 10.4 7.8 11.8





#44
2-Chloronaphthalene
Concen: 40.609 ng/ul
RT: 13.489 min Scan# 1779
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:162 Resp: 206969
Ion Ratio Lower Upper
162 100
164 33.9 25.2 37.8
127 37.5 31.9 47.9

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#45
2-Nitroaniline
Concen: 43.827 ng/ul
RT: 13.695 min Scan# 1814
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19



Tgt Ion: 65 Resp: 75838
Ion Ratio Lower Upper
65 100
92 62.9 55.6 83.4
138 93.2 78.7 118.1









#47
Dimethylphthalate
Concen: 41.987 ng/ul
RT: 14.065 min Scan# 1877
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Tgt Ion:163 Resp: 288819
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.0
164 10.5 8.2 12.4

Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#48
2,6-Dinitrotoluene
Concen: 44.106 ng/ul
RT: 14.194 min Scan# 1899
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19


Tgt Ion:165 Resp: 62638
Ion Ratio Lower Upper
165 100
89 72.6 62.5 93.7
121 23.6 22.2 33.2





#50
Acenaphthylene
Concen: 42.196 ng/ul
RT: 14.335 min Scan# 1923
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:152 Resp: 314300
Ion Ratio Lower Upper
152 100
151 21.5 15.2 22.8
153 12.9 10.7 16.1

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#51
3-Nitroaniline
Concen: 41.195 ng/ul
RT: 14.529 min Scan# 1956
Delta R.T. 0.001 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:138 Resp: 56454
Ion Ratio Lower Upper
138 100
108 10.6 9.0 13.4
92 116.7 96.4 144.6





#52
Acenaphthene
Concen: 42.822 ng/ul
RT: 14.676 min Scan# 1981
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:153 Resp: 231376
Ion Ratio Lower Upper
153 100
152 48.1 37.4 56.2
154 87.3 71.5 107.3

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#53
2,4-Dinitrophenol
Concen: 28.153 ng/ul
RT: 14.752 min Scan# 1994
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:184 Resp: 20421
Ion Ratio Lower Upper
184 100
63 83.7 75.8 113.8
154 86.7 64.6 96.8





#55
4-Nitrophenol
Concen: 9.591 ng/ul
RT: 14.864 min Scan# 2013
Delta R.T. 0.001 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion	Ratio	Lower	Upper
109	100		
139	73.0	61.4	92.0
65	91.8	76.5	114.7

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#56
Dibenzofuran
Concen: 42.628 ng/ul
RT: 15.011 min Scan# 2038
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion	Ratio	Lower	Upper
168	100		
139	42.3	33.8	50.8





#57
2,4-Dinitrotoluene
Concen: 45.550 ng/ul
RT: 14.993 min Scan# 2035
Delta R.T. 0.001 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD



Tgt Ion:165 Resp: 93051
Ion Ratio Lower Upper
165 100
63 47.6 42.3 63.5
182 2.1 2.6 4.0#




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#58
2,3,4,6-Tetrachlorophenol
Concen: 46.662 ng/ul
RT: 15.251 min Scan# 2079
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19


Tgt Ion:232 Resp: 68892
Ion Ratio Lower Upper
232 100
131 46.5 40.0 60.0
130 2.8 1.8 2.8
166 28.3 24.8 37.2





#59
Diethylphthalate
Concen: 43.044 ng/ul
RT: 15.428 min Scan# 2109
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:149 Resp: 302898
Ion Ratio Lower Upper
149 100
177 22.0 15.8 23.6
150 12.6 9.7 14.5

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#61
Fluorene
Concen: 43.262 ng/ul
RT: 15.663 min Scan# 2149
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:166 Resp: 273107
Ion Ratio Lower Upper
166 100
165 101.5 83.3 124.9
167 13.4 11.8 17.6





#62
4-Chlorophenyl-phenylether
Concen: 42.261 ng/ul
RT: 15.651 min Scan# 2147
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion	Ratio	Lower	Upper
204	100		
206	35.4	25.4	38.2
141	62.6	51.8	77.6

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#63
4-Nitroaniline
Concen: 44.186 ng/ul
RT: 15.692 min Scan# 2154
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19



Tgt Ion	Ratio	Lower	Upper
138	100		
92	60.2	50.0	75.0
108	131.4	101.8	152.8









#66
4,6-Dinitro-2-methylphenol
Concen: 39.123 ng/ul
RT: 15.763 min Scan# 2166
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Instrument :
BNA_G
ClientSampled :
EW7A1MSD

Tgt Ion:198 Resp: 46078
Ion Ratio Lower Upper
198 100
182 3.1 3.0 4.6
77 32.3 23.4 35.2




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#67
N-Nitrosodiphenylamine
Concen: 45.308 ng/ul
RT: 15.868 min Scan# 2184
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19


Tgt Ion:169 Resp: 240923
Ion Ratio Lower Upper
169 100
168 65.1 55.0 82.6
167 38.1 29.0 43.4





#68
4-Bromophenyl-phenylether
Concen: 45.914 ng/ul
RT: 16.550 min Scan# 2300
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:248 Resp: 102584
Ion Ratio Lower Upper
248 100
250 101.2 80.2 120.2
141 74.8 60.6 90.8

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#69
Hexachlorobenzene
Concen: 44.407 ng/ul
RT: 16.679 min Scan# 2322
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:284 Resp: 114570
Ion Ratio Lower Upper
284 100
142 30.2 28.7 43.1
249 33.8 25.5 38.3





#70
Atrazine
Concen: 44.742 ng/ul
RT: 16.826 min Scan# 2347
Delta R.T. 0.001 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:200 Resp: 105695
Ion Ratio Lower Upper
200 100
173 23.5 19.3 28.9
215 48.5 34.9 52.3

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#71
Pentachlorophenol
Concen: 37.418 ng/ul
RT: 17.026 min Scan# 2381
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19



Tgt Ion:266 Resp: 40879
Ion Ratio Lower Upper
266 100
264 64.3 54.1 81.1
268 68.4 52.6 78.8









#72
Phenanthrene
Concen: 44.782 ng/ul
RT: 17.413 min Scan# 2447
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19

Instrument :
BNA_G
ClientSampleId :
EW7A1MSD

Tgt Ion:178 Resp: 450906
Ion Ratio Lower Upper
178 100
179 16.3 11.7 17.5
176 19.2 16.2 24.2




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#74
Anthracene
Concen: 43.726 ng/ul
RT: 17.501 min Scan# 2462
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19


Tgt Ion:178 Resp: 446154
Ion Ratio Lower Upper
178 100
179 16.0 11.4 17.0
176 19.1 12.8 19.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 43.194 ng/uL
RT: 13.413 min Scan# 1766
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:216 Resp: 117046
Ion Ratio Lower Upper
216 100
214 78.3 70.9 106.3
218 48.8 40.9 61.3

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#76
Pentachlorobenzene
Concen: 43.632 ng/uL
RT: 14.940 min Scan# 2026
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:250 Resp: 125280
Ion Ratio Lower Upper
250 100
248 67.7 51.5 77.3
252 66.3 47.9 71.9





#77
Carbazole
Concen: 45.615 ng/ul
RT: 17.777 min Scan# 2509
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:167 Resp: 414154
Ion Ratio Lower Upper
167 100
166 23.0 17.8 26.8
139 13.6 11.1 16.7

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#78
Di-n-butylphthalate
Concen: 43.716 ng/ul
RT: 18.318 min Scan# 2601
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:149 Resp: 507630
Ion Ratio Lower Upper
149 100
150 9.5 8.1 12.1
104 6.9 5.3 7.9





#80
Fluoranthene
Concen: 41.380 ng/ul
RT: 19.428 min Scan# 2790
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion	Ratio	Lower	Upper
202	100		
101	6.5	4.3	6.5#
100	5.8	3.5	5.3#

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#82
Pyrene
Concen: 41.374 ng/ul
RT: 19.787 min Scan# 2851
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion	Ratio	Lower	Upper
202	100		
101	7.6	5.4	8.0
100	7.0	5.0	7.6







#83
Butylbenzylphthalate
Concen: 42.369 ng/ul
RT: 20.662 min Scan# 3000
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:149 Resp: 219053
Ion Ratio Lower Upper
149 100
91 76.1 60.6 90.8
206 20.5 16.8 25.2




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#84
3,3'-Dichlorobenzidine
Concen: 6.668 ng/ul
RT: 21.537 min Scan# 3149
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:252 Resp: 31338
Ion Ratio Lower Upper
252 100
254 60.3 54.2 81.2
126 9.2 5.4 8.0#








#85
Benzo(a)anthracene
Concen: 43.492 ng/ul
RT: 21.619 min Scan# 3163
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19


Instrument :
BNA_G
ClientSampleId :
EW7A1MSD





Tgt Ion:228 Resp: 537315
Ion Ratio Lower Upper
228 100
229 20.0 16.2 24.4
226 27.3 21.0 31.4




#86
Bis(2-ethylhexyl)phthalate
Concen: 44.021 ng/ul
RT: 21.508 min Scan# 3144
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:149 Resp: 312565
Ion Ratio Lower Upper
149 100
167 25.4 20.4 30.6
279 4.6 3.7 5.5





#87
Chrysene
Concen: 43.619 ng/ul
RT: 21.690 min Scan# 3175
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD



Tgt Ion:228 Resp: 509397
Ion Ratio Lower Upper
228 100
226 30.8 23.0 34.4
229 19.6 17.8 26.6




Manual Integrations
APPROVED


mohammad
9/7/2021 11:34:01 AM



#89
Di-n-octyl phthalate
Concen: 43.358 ng/ul
RT: 22.712 min Scan# 3349
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19


Tgt Ion:149 Resp: 517043





#90
Benzo(b)fluoranthene
Concen: 43.007 ng/ul
RT: 23.840 min Scan# 3541
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:252 Resp: 553000
Ion Ratio Lower Upper
252 100
253 20.7 18.1 27.1
125 5.4 3.4 5.0#

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#91
Benzo(k)fluoranthene
Concen: 43.885 ng/ul
RT: 23.904 min Scan# 3552
Delta R.T. -0.005 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:252 Resp: 536380
Ion Ratio Lower Upper
252 100
253 22.4 17.7 26.5
125 5.1 4.1 6.1





#93
Benzo(a)pyrene
Concen: 43.499 ng/ul
RT: 24.709 min Scan# 3689
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Instrument :
BNA_G
ClientSampleId :
EW7A1MSD




Tgt Ion:252 Resp: 485749
Ion Ratio Lower Upper
252 100
253 22.7 17.0 25.4
125 5.9 4.4 6.6

Manual Integrations
APPROVED



mohammad
9/7/2021 11:34:01 AM




#94
Indeno(1,2,3-cd)pyrene
Concen: 44.411 ng/ul
RT: 28.557 min Scan# 4344
Delta R.T. -0.011 min
Lab File: BG050139.D
Acq: 3 Sep 2021 20:19




Tgt Ion:276 Resp: 654198
Ion Ratio Lower Upper
276 100
138 9.3 5.5 8.3#
277 23.8 18.6 27.8





#95
 Dibenzo(a,h)anthracene
 Concen: 43.933 ng/ul
 RT: 28.610 min Scan# 4353
 Delta R.T. 0.007 min
 Lab File: BG050139.D
 Acq: 3 Sep 2021 20:19




Instrument :
 BNA_G
 ClientSampleId :
 EW7A1MSD




Tgt Ion:278 Resp: 541761
 Ion Ratio Lower Upper
 278 100
 139 7.9 5.2 7.8#
 279 23.4 19.0 28.4

Manual Integrations
 APPROVED

mohammad
 9/7/2021 11:34:01 AM



#96
 Benzo(g,h,i)perylene
 Concen: 43.954 ng/ul
 RT: 29.714 min Scan# 4541
 Delta R.T. 0.001 min
 Lab File: BG050139.D
 Acq: 3 Sep 2021 20:19



Tgt Ion:276 Resp: 544328
 Ion Ratio Lower Upper
 276 100
 138 10.2 8.6 13.0
 277 22.7 17.4 26.0

