

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090222\
 Data File : BG054838.D
 Acq On : 2 Sep 2022 13:27
 Operator : CG/JU
 Sample : N4376-06DL 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MW-39DL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 09/06/2022
 Supervised By :Jagrut Upadhyay 09/06/2022

Quant Time: Sep 03 00:26:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 25 17:50:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.150	152	35514	20.000	ng	0.00
21) Naphthalene-d8	10.976	136	147557	20.000	ng	0.00
39) Acenaphthene-d10	14.783	164	105714	20.000	ng	0.00
64) Phenanthrene-d10	17.527	188	263701	20.000	ng	0.00
76) Chrysene-d12	21.822	240	264564	20.000	ng	0.00
86) Perylene-d12	25.194	264	278859	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.688	112	23843	11.691	ng	0.00
7) Phenol-d6	7.298	99	19244	6.900	ng	0.00
23) Nitrobenzene-d5	9.319	82	46062	16.388	ng	0.00
42) 2,4,6-Tribromophenol	16.270	330	40438	30.612	ng	0.00
45) 2-Fluorobiphenyl	13.402	172	124705	17.730	ng	0.00
79) Terphenyl-d14	20.124	244	228484	17.112	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	11.029	128	406926	51.401	ng	100
37) 2-Methylnaphthalene	12.621	142	142016	25.594	ng	100
38) 1-Methylnaphthalene	12.838	142	102942m	19.069	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090222\
Data File : BG054838.D
Acq On : 2 Sep 2022 13:27
Operator : CG/JU
Sample : N4376-06DL 5X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
MW-39DL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/06/2022
Supervised By :Jagrut Upadhyay 09/06/2022

Quant Time: Sep 03 00:26:04 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 25 17:50:55 2022
Response via : Initial Calibration

