


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG090419\
 Data File : BG042700.D
 Acq On : 4 Sep 2019 10:19
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG082219.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Aug 22 14:29:15 2019



AutoFind: Scans 2511, 2512, 2513; Background Corrected with Scan 2505

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.8	21202	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	23.2	23597	PASS
70	69	0.00	2	0.3	74	PASS
127	198	10	80	36.3	36978	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	101885	PASS
199	198	5	9	6.8	6896	PASS
275	198	10	60	30.2	30805	PASS
365	198	1	100	3.6	3657	PASS
441	198	0.01	100	13.2	13484	PASS
442	198	50	100	83.3	84850	PASS
443	442	15	24	20.1	17020	PASS