

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG090718\
 Data File : BG036604.D
 Acq On : 7 Sep 2018 19:37
 Operator : JU/SJ
 Sample : J4801-05
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 OILY-WATER

Quant Time: Sep 08 01:28:20 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 31 13:03:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.06	152	61744	20.00	ng	0.00
21) Naphthalene-d8	10.86	136	260181	20.00	ng	0.00
38) Acenaphthene-d10	14.68	164	166055	20.00	ng	0.00
63) Phenanthrene-d10	17.42	188	417579	20.00	ng	0.00
75) Chrysene-d12	21.69	240	428502	20.00	ng	0.00
86) Perylene-d12	24.89	264	446474	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.63	112	319904	82.19	ng	0.00
7) Phenol-d6	7.21	99	306041	54.92	ng	-0.01
23) Nitrobenzene-d5	9.22	82	543553	113.35	ng	0.00
41) 2,4,6-Tribromophenol	16.17	330	368531	185.99	ng	0.00
44) 2-Fluorobiphenyl	13.31	172	1206720	103.76	ng	0.00
78) Terphenyl-d14	20.03	244	1796457	97.99	ng	0.00
Target Compounds						
49) Dimethylphthalate	14.13	163	122319	9.021	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG090718\
Data File : BG036604.D
Acq On : 7 Sep 2018 19:37
Operator : JU/SJ
Sample : J4801-05
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
OILY-WATER

Quant Time: Sep 08 01:28:20 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 31 13:03:20 2018
Response via : Initial Calibration

