

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090823\
 Data File : BG058918.D
 Acq On : 8 Sep 2023 12:44
 Operator : MA/JU
 Sample : SSTD02003
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020402

Quant Time: Sep 08 23:19:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 08 22:56:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.335	152	51704	20.000	ng/u1	0.00
20) Naphthalene-d8	11.179	136	232840	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.963	164	220698	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.712	188	584046	20.000	ng/u1	0.00
79) Chrysene-d12	22.037	240	482998	20.000	ng/u1	0.00
88) Perylene-d12	25.615	264	561891	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.646	96	9114	8.486	ng/uL	0.00
4) Pyridine-d5	4.075	84	78289	21.152	ng/u1	0.00
7) Phenol-d5	7.442	99	108589	21.175	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.654	67	62851	21.665	ng/u1	0.00
11) 2-Chlorophenol-d4	7.847	132	66640	21.785	ng/u1	0.00
15) 4-Methylphenol-d8	9.017	113	85145	21.423	ng/u1	0.00
21) Nitrobenzene-d5	9.534	128	29268	21.340	ng/u1	0.00
24) 2-Nitrophenol-d4	10.250	143	29606	19.706	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.773	165	95836	21.987	ng/u1	0.00
31) 4-Chloroaniline-d4	11.320	131	118993	21.282	ng/u1	0.00
46) Dimethylphthalate-d6	14.357	166	360031	21.133	ng/u1	0.00
49) Acenaphthylene-d8	14.663	160	395818	21.495	ng/u1	0.00
54) 4-Nitrophenol-d4	15.121	143	47774	21.225	ng/u1	0.00
60) Fluorene-d10	15.950	176	336407	20.951	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.050	200	49514	18.621	ng/u1	0.00
73) Anthracene-d10	17.812	188	571102	21.494	ng/u1	0.00
81) Pyrene-d10	20.080	212	594503	21.224	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.362	264	592285	21.398	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.682	88	10769	7.957	ng/uL	100
5) Pyridine	4.099	79	83608	21.202	ng/u1	100
6) Benzaldehyde	7.477	77	75627	25.019	ng/u1	100
8) Phenol	7.471	94	114257	21.365	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.748	93	84969	21.423	ng/u1	100
12) 2-Chlorophenol	7.883	128	67647	21.933	ng/u1	100
13) 2-Methylphenol	8.746	108	86914	21.705	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.840	45	99528	21.804	ng/u1	100
16) Acetophenone	9.175	105	150461	21.231	ng/u1	100
17) N-Nitroso-di-n-propyla...	9.146	70	82727	22.319	ng/u1	100
18) 4-Methylphenol	9.075	108	97268	22.200	ng/u1	100
19) Hexachloroethane	9.422	117	27615	21.937	ng/u1	100
22) Nitrobenzene	9.575	77	100827	21.495	ng/u1	100
23) Isophorone	10.080	82	253911	21.696	ng/u1	100
25) 2-Nitrophenol	10.286	139	34945	21.563	ng/u1	100
26) 2,4-Dimethylphenol	10.303	107	104580	20.756	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.568	93	131089	21.630	ng/u1	100
29) 2,4-Dichlorophenol	10.803	162	92448	22.376	ng/u1	100
30) Naphthalene	11.232	128	268702	21.363	ng/u1	100
32) 4-Chloroaniline	11.343	127	115885	21.852	ng/u1	100
33) Hexachlorobutadiene	11.461	225	89843	22.204	ng/u1	100
34) Caprolactam	12.113	113	16276	11.616	ng/u1	100
35) 4-Chloro-3-methylphenol	12.407	107	109843	22.628	ng/u1	100
36) 2-Methylnaphthalene	12.812	142	203925	21.416	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090823\
 Data File : BG058918.D
 Acq On : 8 Sep 2023 12:44
 Operator : MA/JU
 Sample : SST02003
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST020402

Quant Time: Sep 08 23:19:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 08 22:56:23 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.024	142	203869	20.933	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	13.147	216	177097	21.506	ng/ul	100
40) Hexachlorocyclopentadiene	13.106	237	99896	16.855	ng/ul	100
41) 2,4,6-Trichlorophenol	13.388	196	101776	21.268	ng/ul	100
42) 2,4,5-Trichlorophenol	13.453	196	105794	20.584	ng/ul	100
43) 1,1'-Biphenyl	13.799	154	312452	20.807	ng/ul	100
44) 2-Chloronaphthalene	13.852	162	253737	21.186	ng/ul	100
45) 2-Nitroaniline	14.064	65	53945	20.371	ng/ul	100
47) Dimethylphthalate	14.404	163	373332	21.553	ng/ul	100
48) 2,6-Dinitrotoluene	14.545	165	50053	20.359	ng/ul	100
50) Acenaphthylene	14.692	152	438868	21.357	ng/ul	100
51) 3-Nitroaniline	14.874	138	48188	19.451	ng/ul	100
52) Acenaphthene	15.027	153	283528	21.191	ng/ul	100
53) 2,4-Dinitrophenol	15.068	184	34348	16.516	ng/ul	100
55) 4-Nitrophenol	15.139	109	53887	19.141	ng/ul	100
56) Dibenzofuran	15.356	168	423898	21.223	ng/ul	100
57) 2,4-Dinitrotoluene	15.315	165	70771	19.955	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.562	232	113385	20.916	ng/ul	100
59) Diethylphthalate	15.750	149	325822	20.929	ng/ul	100
61) Fluorene	16.003	166	357703	21.095	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.991	204	215034	21.139	ng/ul	100
63) 4-Nitroaniline	16.038	138	50263	21.415	ng/ul	100
66) 4,6-Dinitro-2-methylph...	16.061	198	54633	19.571	ng/ul	100
67) N-Nitrosodiphenylamine	16.202	169	309633	21.557	ng/ul	100
68) 4-Bromophenyl-phenylether	16.890	248	151824	21.408	ng/ul	100
69) Hexachlorobenzene	16.972	284	176109	21.519	ng/ul	100
70) Atrazine	17.142	200	125873	21.828	ng/ul	100
71) Pentachlorophenol	17.325	266	103040	20.500	ng/ul	100
72) Phenanthrene	17.753	178	608140	21.452	ng/ul	100
74) Anthracene	17.847	178	623244	21.529	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.752	216	184753	21.784	ng/ul	100
76) Pentachlorobenzene	15.251	250	188526	22.057	ng/ul	100
77) Carbazole	18.118	167	479786	21.836	ng/ul	100
78) Di-n-butylphthalate	18.635	149	413848	20.952	ng/ul	100
80) Fluoranthene	19.739	202	721273	21.187	ng/ul	100
82) Pyrene	20.110	202	721845	21.169	ng/ul	100
83) Butylbenzylphthalate	20.979	149	128673	21.763	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.925	252	243710	23.101	ng/ul	100
85) Benzo(a)anthracene	22.013	228	707697	21.521	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.872	149	202185	21.921	ng/ul	100
87) Chrysene	22.090	228	662713	21.513	ng/ul	100
89) Di-n-octyl phthalate	23.224	149	353838	20.164	ng/ul	100
90) Benzo(b)fluoranthene	24.457	252	739244	21.847	ng/ul	100
91) Benzo(k)fluoranthene	24.540	252	738072	21.678	ng/ul	100
93) Benzo(a)pyrene	25.439	252	660283	20.854	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	29.769	276	680211	17.226	ng/ul	100
95) Dibenzo(a,h)anthracene	29.857	278	668022	20.812	ng/ul	100
96) Benzo(g,h,i)perylene	31.073	276	656752	20.921	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090823\
 Data File : BG058918.D
 Acq On : 8 Sep 2023 12:44
 Operator : MA/JU
 Sample : SST02003
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST020402

Quant Time: Sep 08 23:19:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 08 22:56:23 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.957 ng/uL
RT: 3.682 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion: 88 Resp: 10769
Ion Ratio Lower Upper
88 100
43 24.8 19.8 29.8
58 62.4 49.9 74.9



#5
Pyridine
Concen: 21.202 ng/uL
RT: 4.099 min Scan# 104
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion: 79 Resp: 83608
Ion Ratio Lower Upper
79 100
52 57.1 45.7 68.5
51 29.7 23.8 35.6









#6
Benzaldehyde
Concen: 25.019 ng/ul
RT: 7.477 min Scan# 61
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion: 77 Resp: 75627
Ion Ratio Lower Upper
77 100
105 92.9 74.3 111.5
106 84.6 67.7 101.5



#8
Phenol
Concen: 21.365 ng/ul
RT: 7.471 min Scan# 678
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion: 94 Resp: 114257
Ion Ratio Lower Upper
94 100
65 28.0 22.4 33.6
66 47.5 38.0 57.0









#10
Bis(2-Chloroethyl)ether
Concen: 21.423 ng/ul
RT: 7.748 min Scan# 710
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion: 93 Resp: 84969
Ion Ratio Lower Upper
93 100
63 77.2 61.8 92.6
95 29.9 23.9 35.9



#12
2-Chlorophenol
Concen: 21.933 ng/ul
RT: 7.883 min Scan# 748
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:128 Resp: 67647
Ion Ratio Lower Upper
128 100
64 62.5 50.0 75.0
130 36.8 29.4 44.2





#13

2-Methylphenol

Concen: 21.705 ng/ul

RT: 8.746 min Scan# 895

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Instrument :

BNA_G

ClientSampleId :





SSTD020402

Tgt Ion:108 Resp: 86914

Ion Ratio Lower Upper

108 100

107 85.9 68.7 103.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.804 ng/ul

RT: 8.840 min Scan# 911

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44




Tgt Ion: 45 Resp: 99528


Ion Ratio Lower Upper

45 100

77 22.3 17.8 26.8

79 14.9 11.9 17.9









#16
Acetophenone
Concen: 21.231 ng/ul
RT: 9.175 min Scan# 90
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:105 Resp: 150461
Ion Ratio Lower Upper
105 100
77 82.5 66.0 99.0
51 28.7 23.0 34.4



#17
N-Nitroso-di-n-propylamine
Concen: 22.319 ng/ul
RT: 9.146 min Scan# 963
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion: 70 Resp: 82727
Ion Ratio Lower Upper
70 100
42 49.2 39.4 59.0
101 10.7 8.6 12.8
130 17.4 13.9 20.9









#18
4-Methylphenol
Concen: 22.200 ng/ul
RT: 9.075 min Scan# 91
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:108 Resp: 97268
Ion Ratio Lower Upper
108 100
107 108.5 86.8 130.2



#19
Hexachloroethane
Concen: 21.937 ng/ul
RT: 9.422 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:117 Resp: 27615
Ion Ratio Lower Upper
117 100
201 145.3 116.2 174.4
199 102.7 82.2 123.2









#22
Nitrobenzene
Concen: 21.495 ng/ul
RT: 9.575 min Scan# 1104
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion: 77 Resp: 100827
Ion Ratio Lower Upper
77 100
123 36.8 29.4 44.2
65 12.5 10.0 15.0



#23
Isophorone
Concen: 21.696 ng/ul
RT: 10.080 min Scan# 1122
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion: 82 Resp: 253911
Ion Ratio Lower Upper
82 100
95 8.3 6.6 10.0
138 15.4 12.3 18.5









#25
2-Nitrophenol
Concen: 21.563 ng/ul
RT: 10.286 min Scan# 1157
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:139 Resp: 34945
Ion Ratio Lower Upper
139 100
65 59.8 47.8 71.8
109 31.3 25.0 37.6



#26
2,4-Dimethylphenol
Concen: 20.756 ng/ul
RT: 10.303 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:107 Resp: 104580
Ion Ratio Lower Upper
107 100
121 49.3 39.4 59.2
122 84.1 67.3 100.9





#27

Bis(2-Chloroethoxy)methane

Concen: 21.630 ng/ul

RT: 10.568 min Scan# 1205

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Instrument :

BNA_G

ClientSampleId :

SSTD020402





Tgt Ion: 93 Resp: 131089

Ion Ratio Lower Upper

93 100

95 37.2 29.8 44.6

123 11.4 9.1 13.7



#29

2,4-Dichlorophenol

Concen: 22.376 ng/ul

RT: 10.803 min Scan# 1245

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44




Tgt Ion:162 Resp: 92448


Ion Ratio Lower Upper

162 100

164 62.5 50.0 75.0

98 32.9 26.3 39.5









#30
Naphthalene
Concen: 21.363 ng/ul
RT: 11.232 min Scan# 1386
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:128 Resp: 268702
Ion Ratio Lower Upper
128 100
129 9.6 7.7 11.5
127 12.7 10.2 15.2



#32
4-Chloroaniline
Concen: 21.852 ng/ul
RT: 11.343 min Scan# 1337
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:127 Resp: 115885
Ion Ratio Lower Upper
127 100
129 30.7 24.6 36.8









#33
Hexachlorobutadiene
Concen: 22.204 ng/ul
RT: 11.461 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:225 Resp: 89843
Ion Ratio Lower Upper
225 100
223 55.1 44.1 66.1
227 61.2 49.0 73.4



#34
Caprolactam
Concen: 11.616 ng/ul
RT: 12.113 min Scan# 1468
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:113 Resp: 16276
Ion Ratio Lower Upper
113 100
55 130.7 104.6 156.8
56 123.5 98.8 148.2









#35
4-Chloro-3-methylphenol
Concen: 22.628 ng/ul
RT: 12.407 min Scan# 1586
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:107 Resp: 109843
Ion Ratio Lower Upper
107 100
144 21.6 17.3 25.9
142 71.1 56.9 85.3



#36
2-Methylnaphthalene
Concen: 21.416 ng/ul
RT: 12.812 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:142 Resp: 203925
Ion Ratio Lower Upper
142 100
141 89.2 71.4 107.0





#37

1-Methylnaphthalene

Concen: 20.933 ng/ul

RT: 13.024 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Instrument :

BNA_G

ClientSampleId :

SSTD020402





Tgt Ion:142 Resp: 203869

Ion Ratio Lower Upper

142 100

141 97.5 78.0 117.0

116 5.8 4.6 7.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.506 ng/ul

RT: 13.147 min Scan# 1644

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Tgt Ion:216 Resp: 177097




Ion Ratio Lower Upper


216 100

214 73.6 58.9 88.3

179 18.0 14.4 21.6

108 12.6 10.1 15.1





#40

Hexachlorocyclopentadiene

Concen: 16.855 ng/ul

RT: 13.106 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Instrument :

BNA_G

ClientSampleId :

SSTD020402





Tgt Ion:237 Resp: 99896

Ion Ratio Lower Upper

237 100

235 61.3 49.0 73.6

272 13.3 10.6 16.0



#41

2,4,6-Trichlorophenol

Concen: 21.268 ng/ul

RT: 13.388 min Scan# 1685

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44




Tgt Ion:196 Resp: 101776


Ion Ratio Lower Upper

196 100

198 95.1 76.1 114.1

200 32.2 25.8 38.6









#42
2,4,5-Trichlorophenol
Concen: 20.584 ng/ul
RT: 13.453 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:196 Resp: 105794
Ion Ratio Lower Upper
196 100
198 92.9 74.3 111.5
200 30.9 24.7 37.1



#43
1,1'-Biphenyl
Concen: 20.807 ng/ul
RT: 13.799 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:154 Resp: 312452
Ion Ratio Lower Upper
154 100
153 39.8 31.8 47.8
76 10.8 8.6 13.0









#44
2-Chloronaphthalene
Concen: 21.186 ng/ul
RT: 13.852 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:162 Resp: 253737
Ion Ratio Lower Upper
162 100
164 33.9 27.1 40.7
127 34.8 27.8 41.8



#45
2-Nitroaniline
Concen: 20.371 ng/ul
RT: 14.064 min Scan# 1800
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion: 65 Resp: 53945
Ion Ratio Lower Upper
65 100
92 66.0 52.8 79.2
138 95.5 76.4 114.6









#47
Dimethylphthalate
Concen: 21.553 ng/ul
RT: 14.404 min Scan# 1882
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:163 Resp: 373332
Ion Ratio Lower Upper
163 100
194 4.5 3.6 5.4
164 9.2 7.4 11.0



#48
2,6-Dinitrotoluene
Concen: 20.359 ng/ul
RT: 14.545 min Scan# 1882
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:165 Resp: 50053
Ion Ratio Lower Upper
165 100
89 66.3 53.0 79.6
121 17.8 14.2 21.4









#50
Acenaphthylene
Concen: 21.357 ng/ul
RT: 14.692 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:152 Resp: 438868
Ion Ratio Lower Upper
152 100
151 18.7 15.0 22.4
153 11.8 9.4 14.2



#51
3-Nitroaniline
Concen: 19.451 ng/ul
RT: 14.874 min Scan# 1938
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:138 Resp: 48188
Ion Ratio Lower Upper
138 100
108 12.7 10.2 15.2
92 131.5 105.2 157.8









#52
Acenaphthene
Concen: 21.191 ng/ul
RT: 15.027 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:153 Resp: 283528
Ion Ratio Lower Upper
153 100
152 45.8 36.6 55.0
154 82.2 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 16.516 ng/ul
RT: 15.068 min Scan# 1971
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:184 Resp: 34348
Ion Ratio Lower Upper
184 100
63 54.1 43.3 64.9
154 48.2 38.6 57.8









#55
4-Nitrophenol
Concen: 19.141 ng/ul
RT: 15.139 min Scan# 1983
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:109 Resp: 53887
Ion Ratio Lower Upper
109 100
139 88.5 70.8 106.2
65 109.4 87.5 131.3



#56
Dibenzofuran
Concen: 21.223 ng/ul
RT: 15.356 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:168 Resp: 423898
Ion Ratio Lower Upper
168 100
139 36.7 29.4 44.0





#57

2,4-Dinitrotoluene

Concen: 19.955 ng/ul

RT: 15.315 min Scan# 2081

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Instrument :

BNA_G

ClientSampleId :

SSTD020402





Tgt Ion:165 Resp: 70771

Ion Ratio Lower Upper

165 100

63 44.8 35.8 53.8

182 2.1 1.7 2.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.916 ng/ul

RT: 15.562 min Scan# 2055

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Tgt Ion:232 Resp: 113385




Ion Ratio Lower Upper


232 100

131 25.8 20.6 31.0

130 0.0 0.0 0.0

166 21.5 17.2 25.8









#59
Diethylphthalate
Concen: 20.929 ng/ul
RT: 15.750 min Scan# 210
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:149 Resp: 325822
Ion Ratio Lower Upper
149 100
177 26.2 21.0 31.4
150 14.5 11.6 17.4



#61
Fluorene
Concen: 21.095 ng/ul
RT: 16.003 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:166 Resp: 357703
Ion Ratio Lower Upper
166 100
165 98.1 78.5 117.7
167 12.6 10.1 15.1









#62
4-Chlorophenyl-phenylether
Concen: 21.139 ng/ul
RT: 15.991 min Scan# 2196
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:204 Resp: 215034
Ion Ratio Lower Upper
204 100
206 30.2 24.2 36.2
141 41.2 33.0 49.4



#63
4-Nitroaniline
Concen: 21.415 ng/ul
RT: 16.038 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:138 Resp: 50263
Ion Ratio Lower Upper
138 100
92 60.2 48.2 72.2
108 120.8 96.6 145.0









#66
4,6-Dinitro-2-methylphenol
Concen: 19.571 ng/ul
RT: 16.061 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:198 Resp: 54633
Ion Ratio Lower Upper
198 100
182 3.8 3.0 4.6
77 20.8 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 21.557 ng/ul
RT: 16.202 min Scan# 2164
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:169 Resp: 309633
Ion Ratio Lower Upper
169 100
168 65.5 52.4 78.6
167 36.9 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 21.408 ng/ul
RT: 16.890 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:248 Resp: 151824
Ion Ratio Lower Upper
248 100
250 92.4 73.9 110.9
141 45.1 36.1 54.1



#69
Hexachlorobenzene
Concen: 21.519 ng/ul
RT: 16.972 min Scan# 2295
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:284 Resp: 176109
Ion Ratio Lower Upper
284 100
142 20.2 16.2 24.2
249 30.5 24.4 36.6









#70
Atrazine
Concen: 21.828 ng/ul
RT: 17.142 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:200 Resp: 125873
Ion Ratio Lower Upper
200 100
173 20.6 16.5 24.7
215 52.6 42.1 63.1



#71
Pentachlorophenol
Concen: 20.500 ng/ul
RT: 17.325 min Scan# 2355
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:266 Resp: 103040
Ion Ratio Lower Upper
266 100
264 56.7 45.4 68.0
268 62.5 50.0 75.0









#72
Phenanthrene
Concen: 21.452 ng/ul
RT: 17.753 min Scan# 2496
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:178 Resp: 608140
Ion Ratio Lower Upper
178 100
179 16.1 12.9 19.3
176 18.8 15.0 22.6



#74
Anthracene
Concen: 21.529 ng/ul
RT: 17.847 min Scan# 2444
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:178 Resp: 623244
Ion Ratio Lower Upper
178 100
179 13.9 11.1 16.7
176 18.5 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.784 ng/uL

RT: 13.752 min Scan# 11

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Instrument :

BNA_G

ClientSampleId :

SSTD020402





Tgt Ion:216 Resp: 184753

Ion Ratio Lower Upper

216 100

214 79.9 63.9 95.9

218 43.1 34.5 51.7



#76

Pentachlorobenzene

Concen: 22.057 ng/uL

RT: 15.251 min Scan# 2002

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44




Tgt Ion:250 Resp: 188526


Ion Ratio Lower Upper

250 100

248 64.0 51.2 76.8

252 66.8 53.4 80.2









#77
 Carbazole
 Concen: 21.836 ng/ul
 RT: 18.118 min Scan# 2490
 Delta R.T. 0.000 min
 Lab File: BG058918.D
 Acq: 8 Sep 2023 12:44




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020402


Tgt Ion:167 Resp: 479786
 Ion Ratio Lower Upper
 167 100
 166 22.2 17.8 26.6
 139 11.0 8.8 13.2



#78
 Di-n-butylphthalate
 Concen: 20.952 ng/ul
 RT: 18.635 min Scan# 2578
 Delta R.T. 0.000 min
 Lab File: BG058918.D
 Acq: 8 Sep 2023 12:44

Tgt Ion:149 Resp: 413848
 Ion Ratio Lower Upper
 149 100
 150 11.4 9.1 13.7
 104 6.7 5.4 8.0









#80
Fluoranthene
Concen: 21.187 ng/ul
RT: 19.739 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:202 Resp: 721273
Ion Ratio Lower Upper
202 100
101 5.8 4.6 7.0
100 4.7 3.8 5.6



#82
Pyrene
Concen: 21.169 ng/ul
RT: 20.110 min Scan# 2829
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:202 Resp: 721845
Ion Ratio Lower Upper
202 100
101 6.5 5.2 7.8
100 6.2 5.0 7.4









#83
Butylbenzylphthalate
Concen: 21.763 ng/ul
RT: 20.979 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:149 Resp: 128673
Ion Ratio Lower Upper
149 100
91 83.5 66.8 100.2
206 28.4 22.7 34.1



#84
3,3'-Dichlorobenzidine
Concen: 23.101 ng/ul
RT: 21.925 min Scan# 3138
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:252 Resp: 243710
Ion Ratio Lower Upper
252 100
254 63.8 51.0 76.6
126 5.9 4.7 7.1









#85
Benzo(a)anthracene
Concen: 21.521 ng/ul
RT: 22.013 min Scan# 3113
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:228 Resp: 707697
Ion Ratio Lower Upper
228 100
229 20.5 16.4 24.6
226 28.3 22.6 34.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.921 ng/ul
RT: 21.872 min Scan# 3129
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:149 Resp: 202185
Ion Ratio Lower Upper
149 100
167 29.3 23.4 35.2
279 6.7 5.4 8.0





#87

Chrysene

Concen: 21.513 ng/ul

RT: 22.090 min Scan# 3166

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Instrument :

BNA_G

ClientSampleId :

SSTD020402





Tgt Ion:228 Resp: 662713

Ion Ratio Lower Upper

228 100

226 30.9 24.7 37.1

229 18.4 14.7 22.1



#89

Di-n-octyl phthalate

Concen: 20.164 ng/ul




RT: 23.224 min Scan# 3359


Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Tgt Ion:149 Resp: 353838









#90
Benzo(b)fluoranthene
Concen: 21.847 ng/ul
RT: 24.457 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:252 Resp: 739244
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 4.9 3.9 5.9



#91
Benzo(k)fluoranthene
Concen: 21.678 ng/ul
RT: 24.540 min Scan# 3583
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:252 Resp: 738072
Ion Ratio Lower Upper
252 100
253 20.8 16.6 25.0
125 4.6 3.7 5.5









#93
Benzo(a)pyrene
Concen: 20.854 ng/ul
RT: 25.439 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44




Instrument :
BNA_G
ClientSampleId :
SSTD020402


Tgt Ion:252 Resp: 660283
Ion Ratio Lower Upper
252 100
253 20.5 16.4 24.6
125 5.4 4.3 6.5



#94
Indeno(1,2,3-cd)pyrene
Concen: 17.226 ng/ul
RT: 29.769 min Scan# 4473
Delta R.T. 0.000 min
Lab File: BG058918.D
Acq: 8 Sep 2023 12:44

Tgt Ion:276 Resp: 680211
Ion Ratio Lower Upper
276 100
138 7.7 6.2 9.2
277 22.2 17.8 26.6





#95

Dibenzo(a,h)anthracene

Concen: 20.812 ng/ul

RT: 29.857 min Scan# 4488

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Instrument :

BNA_G

ClientSampleId :

SSTD020402





Tgt Ion:278 Resp: 668022

Ion Ratio Lower Upper

278 100

139 6.9 5.5 8.3

279 24.4 19.5 29.3



#96

Benzo(g,h,i)perylene

Concen: 20.921 ng/ul

RT: 31.073 min Scan# 4695

Delta R.T. 0.000 min

Lab File: BG058918.D

Acq: 8 Sep 2023 12:44

Tgt Ion:276 Resp: 656752

Ion Ratio Lower Upper

276 100

138 10.9 8.7 13.1

277 24.6 19.7 29.5

