

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090823\
 Data File : BG058919.D
 Acq On : 8 Sep 2023 14:06
 Operator : MA/JU
 Sample : SST04004
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST040403

Quant Time: Sep 08 23:20:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 08 22:56:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.336	152	58047	20.000	ng/ul	0.00
20) Naphthalene-d8	11.180	136	244138	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.963	164	221628	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.713	188	581267	20.000	ng/ul	0.00
79) Chrysene-d12	22.043	240	477369	20.000	ng/ul	0.00
88) Perylene-d12	25.622	264	553907	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.700	96	260	0.216	ng/uL	0.05
4) Pyridine-d5	4.076	84	164455	39.577	ng/ul	0.00
7) Phenol-d5	7.443	99	226282	39.304	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.654	67	128305	39.394	ng/ul	0.00
11) 2-Chlorophenol-d4	7.848	132	140962	41.047	ng/ul	0.00
15) 4-Methylphenol-d8	9.023	113	172301	38.614	ng/ul	0.00
21) Nitrobenzene-d5	9.535	128	70433	48.979	ng/ul	0.00
24) 2-Nitrophenol-d4	10.251	143	75417	47.875	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.774	165	200027	43.766	ng/ul	0.00
31) 4-Chloroaniline-d4	11.321	131	246752	42.089	ng/ul	0.00
46) Dimethylphthalate-d6	14.364	166	702817	41.080	ng/ul	0.00
49) Acenaphthylene-d8	14.664	160	770165	41.648	ng/ul	0.00
54) 4-Nitrophenol-d4	15.128	143	97213	43.008	ng/ul	0.00
60) Fluorene-d10	15.951	176	658273	40.824	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.056	200	112416	42.479	ng/ul	0.00
73) Anthracene-d10	17.813	188	1072399	40.554	ng/ul	0.00
81) Pyrene-d10	20.081	212	1078217	38.946	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.375	264	1130276	41.423	ng/ul	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.683	88	23108	15.208	ng/uL#	91
5) Pyridine	4.100	79	173941	39.289	ng/ul	95
6) Benzaldehyde	7.478	77	144196	42.490	ng/ul	94
8) Phenol	7.472	94	231123	38.495	ng/ul	96
10) Bis(2-Chloroethyl)ether	7.754	93	176491	39.636	ng/ul	96
12) 2-Chlorophenol	7.884	128	138959	40.131	ng/ul	99
13) 2-Methylphenol	8.753	108	178023	39.600	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.841	45	200483	39.121	ng/ul#	92
16) Acetophenone	9.182	105	314359	39.511	ng/ul	98
17) N-Nitroso-di-n-propyla...	9.153	70	169664	40.772	ng/ul	99
18) 4-Methylphenol	9.082	108	182786	37.159	ng/ul	93
19) Hexachloroethane	9.423	117	57727	40.846	ng/ul	93
22) Nitrobenzene	9.576	77	225063	45.761	ng/ul	93
23) Isophorone	10.093	82	505410	41.187	ng/ul	96
25) 2-Nitrophenol	10.287	139	78296	46.078	ng/ul#	70
26) 2,4-Dimethylphenol	10.304	107	215540	40.799	ng/ul	94
27) Bis(2-Chloroethoxy)met...	10.574	93	259558	40.845	ng/ul	94
29) 2,4-Dichlorophenol	10.804	162	190107	43.883	ng/ul	95
30) Naphthalene	11.233	128	543866	41.238	ng/ul#	97
32) 4-Chloroaniline	11.350	127	232341	41.784	ng/ul	100
33) Hexachlorobutadiene	11.462	225	176278	41.549	ng/ul	94
34) Caprolactam	12.184	113	21510	14.642	ng/ul	83
35) 4-Chloro-3-methylphenol	12.414	107	211745	41.601	ng/ul	97
36) 2-Methylnaphthalene	12.813	142	409501	41.016	ng/ul	96


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090823\
 Data File : BG058919.D
 Acq On : 8 Sep 2023 14:06
 Operator : MA/JU
 Sample : SST04004
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST040403

Quant Time: Sep 08 23:20:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 08 22:56:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.030	142	413464	40.489	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	13.154	216	343168	41.497	ng/ul	95
40) Hexachlorocyclopentadiene	13.107	237	210901	35.435	ng/ul	99
41) 2,4,6-Trichlorophenol	13.389	196	206064	42.880	ng/ul	94
42) 2,4,5-Trichlorophenol	13.459	196	225238	43.640	ng/ul	97
43) 1,1'-Biphenyl	13.800	154	620947	41.177	ng/ul	99
44) 2-Chloronaphthalene	13.853	162	497716	41.384	ng/ul	98
45) 2-Nitroaniline	14.065	65	129251	48.604	ng/ul	92
47) Dimethylphthalate	14.411	163	714727	41.089	ng/ul	99
48) 2,6-Dinitrotoluene	14.552	165	114532	46.391	ng/ul#	94
50) Acenaphthylene	14.693	152	850458	41.212	ng/ul	98
51) 3-Nitroaniline	14.887	138	107361	43.154	ng/ul	98
52) Acenaphthene	15.028	153	559792	41.664	ng/ul	93
53) 2,4-Dinitrophenol	15.075	184	74166	35.512	ng/ul	97
55) 4-Nitrophenol	15.146	109	119889	42.407	ng/ul	97
56) Dibenzofuran	15.363	168	807160	40.243	ng/ul	97
57) 2,4-Dinitrotoluene	15.322	165	161206	45.263	ng/ul#	91
58) 2,3,4,6-Tetrachlorophenol	15.563	232	232463	42.703	ng/ul#	96
59) Diethylphthalate	15.757	149	636132	40.690	ng/ul	99
61) Fluorene	16.009	166	677837	39.807	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.992	204	418873	41.005	ng/ul	92
63) 4-Nitroaniline	16.045	138	101232	42.950	ng/ul	92
66) 4,6-Dinitro-2-methylph...	16.074	198	114250	41.123	ng/ul#	94
67) N-Nitrosodiphenylamine	16.209	169	603475	42.216	ng/ul	98
68) 4-Bromophenyl-phenylether	16.891	248	292117	41.386	ng/ul	94
69) Hexachlorobenzene	16.979	284	330704	40.603	ng/ul	99
70) Atrazine	17.149	200	238001	41.470	ng/ul	94
71) Pentachlorophenol	17.325	266	202953	40.572	ng/ul	95
72) Phenanthrene	17.754	178	1138078	40.337	ng/ul	100
74) Anthracene	17.848	178	1162546	40.351	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.759	216	372478	44.129	ng/uL	95
76) Pentachlorobenzene	15.251	250	367823	43.239	ng/uL	95
77) Carbazole	18.119	167	896243	40.984	ng/ul	100
78) Di-n-butylphthalate	18.642	149	828173	42.128	ng/ul	96
80) Fluoranthene	19.746	202	1296710	38.539	ng/ul	99
82) Pyrene	20.110	202	1314579	39.006	ng/ul	98
83) Butylbenzylphthalate	20.980	149	266752	45.648	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.932	252	444805	42.659	ng/ul#	97
85) Benzo(a)anthracene	22.020	228	1322568	40.693	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.879	149	412605	45.263	ng/ul	98
87) Chrysene	22.096	228	1238589	40.682	ng/ul	96
89) Di-n-octyl phthalate	23.230	149	717113	41.456	ng/ul	100
90) Benzo(b)fluoranthene	24.470	252	1346928	40.379	ng/ul	99
91) Benzo(k)fluoranthene	24.546	252	1365675	40.689	ng/ul	98
93) Benzo(a)pyrene	25.463	252	1273735	40.810	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	29.787	276	1367682	35.134	ng/ul	98
95) Dibenzo(a,h)anthracene	29.881	278	1291506	40.816	ng/ul	98
96) Benzo(g,h,i)perylene	31.092	276	694097	22.429	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 15.208 ng/uL
RT: 3.683 min Scan# 31
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion: 88 Resp: 23108
Ion Ratio Lower Upper
88 100
43 18.5 19.8 29.8#
58 68.2 49.9 74.9



#5
Pyridine
Concen: 39.289 ng/uL
RT: 4.100 min Scan# 104
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion: 79 Resp: 173941
Ion Ratio Lower Upper
79 100
52 60.6 45.7 68.5
51 33.4 23.8 35.6









#6
Benzaldehyde
Concen: 42.490 ng/ul
RT: 7.478 min Scan# 61
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion: 77 Resp: 144196
Ion Ratio Lower Upper
77 100
105 94.2 74.3 111.5
106 95.4 67.7 101.5



#8
Phenol
Concen: 38.495 ng/ul
RT: 7.472 min Scan# 678
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion: 94 Resp: 231123
Ion Ratio Lower Upper
94 100
65 32.0 22.4 33.6
66 49.1 38.0 57.0









#10
Bis(2-Chloroethyl)ether
Concen: 39.636 ng/ul
RT: 7.754 min Scan# 71
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion: 93 Resp: 176491
Ion Ratio Lower Upper
93 100
63 72.4 61.8 92.6
95 29.5 23.9 35.9



#12
2-Chlorophenol
Concen: 40.131 ng/ul
RT: 7.884 min Scan# 748
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:128 Resp: 138959
Ion Ratio Lower Upper
128 100
64 62.0 50.0 75.0
130 37.8 29.4 44.2





#13

2-Methylphenol

Concen: 39.600 ng/ul

RT: 8.753 min Scan# 896

Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Instrument :

BNA_G

ClientSampleId :





SSTD040403

Tgt Ion:108 Resp: 178023

Ion Ratio Lower Upper

108 100

107 84.7 68.7 103.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 39.121 ng/ul

RT: 8.841 min Scan# 911

Delta R.T. 0.001 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06




Tgt Ion: 45 Resp: 200483


Ion Ratio Lower Upper

45 100

77 16.5 17.8 26.8#

79 13.8 11.9 17.9









#16
Acetophenone
Concen: 39.511 ng/ul
RT: 9.182 min Scan# 90
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:105 Resp: 314359
Ion Ratio Lower Upper
105 100
77 84.0 66.0 99.0
51 27.5 23.0 34.4



#17
N-Nitroso-di-n-propylamine
Concen: 40.772 ng/ul
RT: 9.153 min Scan# 964
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion: 70 Resp: 169664
Ion Ratio Lower Upper
70 100
42 49.5 39.4 59.0
101 8.8 8.6 12.8
130 17.6 13.9 20.9









#18
4-Methylphenol
Concen: 37.159 ng/ul
RT: 9.082 min Scan# 91
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:108 Resp: 182786
Ion Ratio Lower Upper
108 100
107 116.1 86.8 130.2



#19
Hexachloroethane
Concen: 40.846 ng/ul
RT: 9.423 min Scan# 1010
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:117 Resp: 57727
Ion Ratio Lower Upper
117 100
201 149.1 116.2 174.4
199 89.3 82.2 123.2









#22
Nitrobenzene
Concen: 45.761 ng/ul
RT: 9.576 min Scan# 1104
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion: 77 Resp: 225063
Ion Ratio Lower Upper
77 100
123 32.2 29.4 44.2
65 14.1 10.0 15.0



#23
Isophorone
Concen: 41.187 ng/ul
RT: 10.093 min Scan# 1124
Delta R.T. 0.013 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion: 82 Resp: 505410
Ion Ratio Lower Upper
82 100
95 8.4 6.6 10.0
138 13.0 12.3 18.5









#25
2-Nitrophenol
Concen: 46.078 ng/ul
RT: 10.287 min Scan# 1157
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:139 Resp: 78296
Ion Ratio Lower Upper
139 100
65 80.8 47.8 71.8#
109 49.6 25.0 37.6#



#26
2,4-Dimethylphenol
Concen: 40.799 ng/ul
RT: 10.304 min Scan# 1160
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:107 Resp: 215540
Ion Ratio Lower Upper
107 100
121 45.4 39.4 59.2
122 79.0 67.3 100.9









#27
 Bis(2-Chloroethoxy)methane
 Concen: 40.845 ng/ul
 RT: 10.574 min Scan# 11
 Delta R.T. 0.007 min
 Lab File: BG058919.D
 Acq: 8 Sep 2023 14:06




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040403


Tgt Ion:	93	Resp:	259558
Ion	Ratio	Lower	Upper
93	100		
95	33.0	29.8	44.6
123	9.9	9.1	13.7



#29
 2,4-Dichlorophenol
 Concen: 43.883 ng/ul
 RT: 10.804 min Scan# 1245
 Delta R.T. 0.001 min
 Lab File: BG058919.D
 Acq: 8 Sep 2023 14:06

Tgt Ion:	162	Resp:	190107
Ion	Ratio	Lower	Upper
162	100		
164	65.3	50.0	75.0
98	36.7	26.3	39.5









#30
Naphthalene
Concen: 41.238 ng/ul
RT: 11.233 min Scan# 11
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:128 Resp: 543866
Ion Ratio Lower Upper
128 100
129 11.6 7.7 11.5#
127 12.2 10.2 15.2



#32
4-Chloroaniline
Concen: 41.784 ng/ul
RT: 11.350 min Scan# 1338
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:127 Resp: 232341
Ion Ratio Lower Upper
127 100
129 30.5 24.6 36.8









#33
Hexachlorobutadiene
Concen: 41.549 ng/ul
RT: 11.462 min Scan# 11
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:225 Resp: 176278
Ion Ratio Lower Upper
225 100
223 60.2 44.1 66.1
227 64.5 49.0 73.4



#34
Caprolactam
Concen: 14.642 ng/ul
RT: 12.184 min Scan# 1480
Delta R.T. 0.071 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:113 Resp: 21510
Ion Ratio Lower Upper
113 100
55 156.4 104.6 156.8
56 110.8 98.8 148.2









#35
4-Chloro-3-methylphenol
Concen: 41.601 ng/ul
RT: 12.414 min Scan# 1586
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:107 Resp: 211745
Ion Ratio Lower Upper
107 100
144 24.4 17.3 25.9
142 72.5 56.9 85.3



#36
2-Methylnaphthalene
Concen: 41.016 ng/ul
RT: 12.813 min Scan# 1587
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:142 Resp: 409501
Ion Ratio Lower Upper
142 100
141 85.2 71.4 107.0





#37

1-Methylnaphthalene

Concen: 40.489 ng/ul

RT: 13.030 min Scan# 1

Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD040403





Tgt Ion:142 Resp: 413464

Ion Ratio Lower Upper

142 100

141 94.7 78.0 117.0

116 5.0 4.6 7.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 41.497 ng/ul

RT: 13.154 min Scan# 1645

Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Tgt Ion:216 Resp: 343168




Ion Ratio Lower Upper


216 100

214 79.0 58.9 88.3

179 17.7 14.4 21.6

108 14.0 10.1 15.1









#40
Hexachlorocyclopentadiene
Concen: 35.435 ng/ul
RT: 13.107 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:237 Resp: 210901
Ion Ratio Lower Upper
237 100
235 62.4 49.0 73.6
272 12.6 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 42.880 ng/ul
RT: 13.389 min Scan# 1685
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:196 Resp: 206064
Ion Ratio Lower Upper
196 100
198 89.1 76.1 114.1
200 29.2 25.8 38.6









#42
2,4,5-Trichlorophenol
Concen: 43.640 ng/ul
RT: 13.459 min Scan# 1755
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:196 Resp: 225238
Ion Ratio Lower Upper
196 100
198 94.4 74.3 111.5
200 27.8 24.7 37.1



#43
1,1'-Biphenyl
Concen: 41.177 ng/ul
RT: 13.800 min Scan# 1755
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:154 Resp: 620947
Ion Ratio Lower Upper
154 100
153 40.4 31.8 47.8
76 11.5 8.6 13.0









#44
2-Chloronaphthalene
Concen: 41.384 ng/ul
RT: 13.853 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:162 Resp: 497716
Ion Ratio Lower Upper
162 100
164 33.1 27.1 40.7
127 32.7 27.8 41.8



#45
2-Nitroaniline
Concen: 48.604 ng/ul
RT: 14.065 min Scan# 1800
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion: 65 Resp: 129251
Ion Ratio Lower Upper
65 100
92 77.4 52.8 79.2
138 92.8 76.4 114.6









#47
Dimethylphthalate
Concen: 41.089 ng/ul
RT: 14.411 min Scan# 1859
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:163 Resp: 714727
Ion Ratio Lower Upper
163 100
194 4.3 3.6 5.4
164 9.8 7.4 11.0



#48
2,6-Dinitrotoluene
Concen: 46.391 ng/ul
RT: 14.552 min Scan# 1883
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:165 Resp: 114532
Ion Ratio Lower Upper
165 100
89 63.0 53.0 79.6
121 22.8 14.2 21.4#









#50
Acenaphthylene
Concen: 41.212 ng/ul
RT: 14.693 min Scan# 1975
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:152 Resp: 850458
Ion Ratio Lower Upper
152 100
151 19.2 15.0 22.4
153 13.2 9.4 14.2



#51
3-Nitroaniline
Concen: 43.154 ng/ul
RT: 14.887 min Scan# 1940
Delta R.T. 0.013 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:138 Resp: 107361
Ion Ratio Lower Upper
138 100
108 12.7 10.2 15.2
92 134.6 105.2 157.8









#52
Acenaphthene
Concen: 41.664 ng/ul
RT: 15.028 min Scan# 1964
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:153 Resp: 559792
Ion Ratio Lower Upper
153 100
152 50.6 36.6 55.0
154 88.3 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 35.512 ng/ul
RT: 15.075 min Scan# 1972
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:184 Resp: 74166
Ion Ratio Lower Upper
184 100
63 55.4 43.3 64.9
154 51.2 38.6 57.8









#55
4-Nitrophenol
Concen: 42.407 ng/ul
RT: 15.146 min Scan# 1984
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:109 Resp: 119889
Ion Ratio Lower Upper
109 100
139 87.2 70.8 106.2
65 105.2 87.5 131.3



#56
Dibenzofuran
Concen: 40.243 ng/ul
RT: 15.363 min Scan# 2021
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:168 Resp: 807160
Ion Ratio Lower Upper
168 100
139 34.8 29.4 44.0









#57
2,4-Dinitrotoluene
Concen: 45.263 ng/ul
RT: 15.322 min Scan# 2081
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:165 Resp: 161206
Ion Ratio Lower Upper
165 100
63 50.9 35.8 53.8
182 1.5 1.7 2.5#



#58
2,3,4,6-Tetrachlorophenol
Concen: 42.703 ng/ul
RT: 15.563 min Scan# 2055
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:232 Resp: 232463
Ion Ratio Lower Upper
232 100
131 28.8 20.6 31.0
130 1.7 0.0 0.0#
166 21.8 17.2 25.8









#59
Diethylphthalate
Concen: 40.690 ng/ul
RT: 15.757 min Scan# 2088
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:149 Resp: 636132
Ion Ratio Lower Upper
149 100
177 26.6 21.0 31.4
150 13.8 11.6 17.4



#61
Fluorene
Concen: 39.807 ng/ul
RT: 16.009 min Scan# 2131
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:166 Resp: 677837
Ion Ratio Lower Upper
166 100
165 97.0 78.5 117.7
167 13.2 10.1 15.1









#62
4-Chlorophenyl-phenylether
Concen: 41.005 ng/ul
RT: 15.992 min Scan# 2196
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:204 Resp: 418873
Ion Ratio Lower Upper
204 100
206 33.0 24.2 36.2
141 47.2 33.0 49.4



#63
4-Nitroaniline
Concen: 42.950 ng/ul
RT: 16.045 min Scan# 2137
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:138 Resp: 101232
Ion Ratio Lower Upper
138 100
92 63.5 48.2 72.2
108 132.3 96.6 145.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 41.123 ng/ul
 RT: 16.074 min Scan# 2165
 Delta R.T. 0.013 min
 Lab File: BG058919.D
 Acq: 8 Sep 2023 14:06




Instrument :
 BNA_G
 ClientSampleId :
 SST040403


Tgt Ion:198 Resp: 114250
 Ion Ratio Lower Upper
 198 100
 182 2.2 3.0 4.6#
 77 17.8 16.6 25.0



#67
 N-Nitrosodiphenylamine
 Concen: 42.216 ng/ul
 RT: 16.209 min Scan# 2165
 Delta R.T. 0.007 min
 Lab File: BG058919.D
 Acq: 8 Sep 2023 14:06

Tgt Ion:169 Resp: 603475
 Ion Ratio Lower Upper
 169 100
 168 66.4 52.4 78.6
 167 34.4 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 41.386 ng/ul
RT: 16.891 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:248 Resp: 292117
Ion Ratio Lower Upper
248 100
250 99.0 73.9 110.9
141 48.5 36.1 54.1



#69
Hexachlorobenzene
Concen: 40.603 ng/ul
RT: 16.979 min Scan# 2296
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:284 Resp: 330704
Ion Ratio Lower Upper
284 100
142 20.9 16.2 24.2
249 31.2 24.4 36.6





#70

Atrazine

Concen: 41.470 ng/ul

RT: 17.149 min Scan# 21

Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD040403





Tgt Ion:200 Resp: 238001

Ion Ratio Lower Upper

200 100

173 22.2 16.5 24.7

215 47.6 42.1 63.1



#71

Pentachlorophenol

Concen: 40.572 ng/ul

RT: 17.325 min Scan# 2355

Delta R.T. 0.001 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06




Tgt Ion:266 Resp: 202953


Ion Ratio Lower Upper

266 100

264 63.6 45.4 68.0

268 62.8 50.0 75.0





#72

Phenanthrene

Concen: 40.337 ng/ul

RT: 17.754 min Scan# 2496

Delta R.T. 0.001 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD040403





Tgt Ion:178 Resp: 1138078

Ion Ratio Lower Upper

178 100

179 15.9 12.9 19.3

176 19.0 15.0 22.6



#74

Anthracene

Concen: 40.351 ng/ul

RT: 17.848 min Scan# 2444

Delta R.T. 0.001 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06




Tgt Ion:178 Resp: 1162546


Ion Ratio Lower Upper

178 100

179 15.3 11.1 16.7

176 19.0 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 44.129 ng/uL

RT: 13.759 min Scan# 1

Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD040403





Tgt Ion:216 Resp: 372478

Ion Ratio Lower Upper

216 100

214 77.9 63.9 95.9

218 50.4 34.5 51.7



#76

Pentachlorobenzene

Concen: 43.239 ng/uL

RT: 15.251 min Scan# 2002

Delta R.T. 0.001 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06




Tgt Ion:250 Resp: 367823


Ion Ratio Lower Upper

250 100

248 59.7 51.2 76.8

252 62.5 53.4 80.2





#77

Carbazole

Concen: 40.984 ng/ul

RT: 18.119 min Scan# 2490

Delta R.T. 0.001 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD040403





Tgt Ion:167 Resp: 896243

Ion Ratio Lower Upper

167 100

166 22.1 17.8 26.6

139 11.3 8.8 13.2



#78

Di-n-butylphthalate

Concen: 42.128 ng/ul

RT: 18.642 min Scan# 2579

Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06




Tgt Ion:149 Resp: 828173


Ion Ratio Lower Upper

149 100

150 9.6 9.1 13.7

104 7.2 5.4 8.0









#80
Fluoranthene
Concen: 38.539 ng/ul
RT: 19.746 min Scan# 21
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:202 Resp: 1296710
Ion Ratio Lower Upper
202 100
101 6.1 4.6 7.0
100 5.1 3.8 5.6



#82
Pyrene
Concen: 39.006 ng/ul
RT: 20.110 min Scan# 2829
Delta R.T. 0.001 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:202 Resp: 1314579
Ion Ratio Lower Upper
202 100
101 7.3 5.2 7.8
100 5.8 5.0 7.4





#83

Butylbenzylphthalate

Concen: 45.648 ng/ul

RT: 20.980 min Scan# 20

Delta R.T. 0.001 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD040403





Tgt Ion:149 Resp: 266752

Ion Ratio Lower Upper

149 100

91 88.5 66.8 100.2

206 30.8 22.7 34.1



#84

3,3'-Dichlorobenzidine

Concen: 42.659 ng/ul

RT: 21.932 min Scan# 3139

Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06




Tgt Ion:252 Resp: 444805


Ion Ratio Lower Upper

252 100

254 65.9 51.0 76.6

126 4.7 4.7 7.1#





#85

Benzo(a)anthracene

Concen: 40.693 ng/ul

RT: 22.020 min Scan# 3115

Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD040403





Tgt Ion:228 Resp: 1322568

Ion Ratio Lower Upper

228 100

229 19.6 16.4 24.6

226 27.7 22.6 34.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 45.263 ng/ul

RT: 21.879 min Scan# 3130

Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06




Tgt Ion:149 Resp: 412605


Ion Ratio Lower Upper

149 100

167 28.5 23.4 35.2

279 5.8 5.4 8.0





#87

Chrysene

Concen: 40.682 ng/ul

RT: 22.096 min Scan# 3167

Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD040403





Tgt Ion:228 Resp: 1238589

Ion Ratio Lower Upper

228 100

226 27.9 24.7 37.1

229 19.0 14.7 22.1



#89

Di-n-octyl phthalate

Concen: 41.456 ng/ul




RT: 23.230 min Scan# 3360


Delta R.T. 0.007 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Tgt Ion:149 Resp: 717113









#90
Benzo(b)fluoranthene
Concen: 40.379 ng/ul
RT: 24.470 min Scan# 31
Delta R.T. 0.013 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD040403


Tgt Ion:252 Resp: 1346928
Ion Ratio Lower Upper
252 100
253 21.4 17.6 26.4
125 4.3 3.9 5.9



#91
Benzo(k)fluoranthene
Concen: 40.689 ng/ul
RT: 24.546 min Scan# 3584
Delta R.T. 0.007 min
Lab File: BG058919.D
Acq: 8 Sep 2023 14:06

Tgt Ion:252 Resp: 1365675
Ion Ratio Lower Upper
252 100
253 22.2 16.6 25.0
125 4.7 3.7 5.5





#93

Benzo(a)pyrene

Concen: 40.810 ng/ul

RT: 25.463 min Scan# 31

Delta R.T. 0.024 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD040403





Tgt Ion:252 Resp: 1273735

Ion Ratio Lower Upper

252 100

253 21.5 16.4 24.6

125 4.2 4.3 6.5#



#94

Indeno(1,2,3-cd)pyrene

Concen: 35.134 ng/ul

RT: 29.787 min Scan# 4476

Delta R.T. 0.018 min

Lab File: BG058919.D

Acq: 8 Sep 2023 14:06




Tgt Ion:276 Resp: 1367682


Ion Ratio Lower Upper

276 100

138 8.5 6.2 9.2

277 23.2 17.8 26.6









#95
 Dibenzo(a,h)anthracene
 Concen: 40.816 ng/ul
 RT: 29.881 min Scan# 4492
 Delta R.T. 0.024 min
 Lab File: BG058919.D
 Acq: 8 Sep 2023 14:06

Instrument :
 BNA_G
 ClientSampleId :
 SST040403

Tgt Ion:278 Resp: 1291506
 Ion Ratio Lower Upper
 278 100
 139 6.9 5.5 8.3
 279 23.1 19.5 29.3



#96
 Benzo(g,h,i)perylene
 Concen: 22.429 ng/ul
 RT: 31.092 min Scan# 4698
 Delta R.T. 0.018 min
 Lab File: BG058919.D
 Acq: 8 Sep 2023 14:06

Tgt Ion:276 Resp: 694097
 Ion Ratio Lower Upper
 276 100
 138 9.4 8.7 13.1
 277 23.3 19.7 29.5

