

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090823\
 Data File : BG058922.D
 Acq On : 8 Sep 2023 17:58
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV406

Quant Time: Sep 09 00:21:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 09 00:11:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.332	152	44811	20.000	ng/ul	0.00
20) Naphthalene-d8	11.176	136	208102	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.960	164	199942	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.709	188	565877	20.000	ng/ul	0.00
79) Chrysene-d12	22.034	240	478740	20.000	ng/ul	0.00
88) Perylene-d12	25.612	264	529545	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.638	96	9628	10.306	ng/uL	0.00
4) Pyridine-d5	4.072	84	65200	20.325	ng/ul	0.00
7) Phenol-d5	7.439	99	90226	20.301	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.651	67	51364	20.429	ng/ul	0.00
11) 2-Chlorophenol-d4	7.845	132	57639	21.741	ng/ul	0.00
15) 4-Methylphenol-d8	9.014	113	72185	20.956	ng/ul	0.00
21) Nitrobenzene-d5	9.531	128	28945	23.614	ng/ul	0.00
24) 2-Nitrophenol-d4	10.248	143	32547	24.239	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.771	165	84501	21.691	ng/ul	0.00
31) 4-Chloroaniline-d4	11.317	131	103093	20.630	ng/ul	0.00
46) Dimethylphthalate-d6	14.355	166	329576	21.353	ng/ul	0.00
49) Acenaphthylene-d8	14.660	160	348904	20.914	ng/ul	0.00
54) 4-Nitrophenol-d4	15.118	143	46345	22.727	ng/ul	0.00
60) Fluorene-d10	15.941	176	302144	20.771	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.047	200	57082	22.157	ng/ul	0.00
73) Anthracene-d10	17.809	188	527879	20.507	ng/ul	0.00
81) Pyrene-d10	20.077	212	568483	20.475	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.359	264	552795	21.191	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.679	88	8063	7.087	ng/uL#	90
5) Pyridine	4.090	79	69438	20.563	ng/ul#	89
6) Benzaldehyde	7.474	77	61097	23.321	ng/ul	90
8) Phenol	7.463	94	94977	20.492	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.745	93	68390	19.895	ng/ul#	91
12) 2-Chlorophenol	7.874	128	57964	21.684	ng/ul	89
13) 2-Methylphenol	8.743	108	69299	19.968	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.855	45	79788	20.168	ng/ul#	88
16) Acetophenone	9.172	105	128055	20.849	ng/ul	97
17) N-Nitroso-di-n-propyla...	9.137	70	69481	21.629	ng/ul#	97
18) 4-Methylphenol	9.073	108	78862	20.766	ng/ul	98
19) Hexachloroethane	9.419	117	23301	21.357	ng/ul	92
22) Nitrobenzene	9.572	77	91684	21.870	ng/ul	99
23) Isophorone	10.083	82	214578	20.515	ng/ul	96
25) 2-Nitrophenol	10.277	139	34174	23.594	ng/ul#	91
26) 2,4-Dimethylphenol	10.300	107	94094	20.895	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.565	93	110350	20.372	ng/ul#	88
29) 2,4-Dichlorophenol	10.794	162	76851	20.812	ng/ul	99
30) Naphthalene	11.229	128	230346	20.490	ng/ul	98
32) 4-Chloroaniline	11.340	127	99673	21.029	ng/ul	100
33) Hexachlorobutadiene	11.458	225	74595	20.627	ng/ul	92
34) Caprolactam	12.181	113	2635	2.157	ng/ul	96
35) 4-Chloro-3-methylphenol	12.404	107	90983	20.971	ng/ul	93
36) 2-Methylnaphthalene	12.803	142	176107	20.694	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090823\
 Data File : BG058922.D
 Acq On : 8 Sep 2023 17:58
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV406

Quant Time: Sep 09 00:21:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 09 00:11:02 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.021	142	177984	20.447	ng/ul	93
39) 1,2,4,5-Tetrachloroben...	13.144	216	147999	19.838	ng/ul	96
40) Hexachlorocyclopentadiene	13.109	237	94963	17.686	ng/ul	98
41) 2,4,6-Trichlorophenol	13.385	196	91071	21.007	ng/ul	98
42) 2,4,5-Trichlorophenol	13.450	196	98468	21.184	ng/ul	96
43) 1,1'-Biphenyl	13.796	154	273207	20.082	ng/ul	99
44) 2-Chloronaphthalene	13.849	162	215013	19.817	ng/ul	97
45) 2-Nitroaniline	14.061	65	60924	25.395	ng/ul#	87
47) Dimethylphthalate	14.402	163	323825	20.636	ng/ul	98
48) 2,6-Dinitrotoluene	14.543	165	54502	24.470	ng/ul#	96
50) Acenaphthylene	14.689	152	384903	20.675	ng/ul	100
51) 3-Nitroaniline	14.872	138	51593	22.987	ng/ul	96
52) Acenaphthene	15.018	153	245428	20.248	ng/ul	97
53) 2,4-Dinitrophenol	15.065	184	36229	19.228	ng/ul	92
55) 4-Nitrophenol	15.136	109	61186	23.990	ng/ul	92
56) Dibenzofuran	15.353	168	365918	20.222	ng/ul	95
57) 2,4-Dinitrotoluene	15.318	165	78541	24.445	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.559	232	110288	22.457	ng/ul#	97
59) Diethylphthalate	15.753	149	289207	20.506	ng/ul	99
61) Fluorene	16.000	166	312718	20.357	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.988	204	194153	21.068	ng/ul	92
63) 4-Nitroaniline	16.035	138	55092	26.381	ng/ul	89
66) 4,6-Dinitro-2-methylph...	16.064	198	59478	21.991	ng/ul	96
67) N-Nitrosodiphenylamine	16.205	169	285217	20.495	ng/ul	97
68) 4-Bromophenyl-phenylether	16.887	248	143361	20.863	ng/ul	97
69) Hexachlorobenzene	16.969	284	163026	20.560	ng/ul	98
70) Atrazine	17.140	200	122272	21.884	ng/ul	96
71) Pentachlorophenol	17.322	266	99482	20.428	ng/ul	93
72) Phenanthrene	17.751	178	561400	20.436	ng/ul	98
74) Anthracene	17.845	178	584596	20.843	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.755	216	160514	19.534	ng/ul	97
76) Pentachlorobenzene	15.248	250	163009	19.684	ng/ul	96
77) Carbazole	18.115	167	455055	21.375	ng/ul	99
78) Di-n-butylphthalate	18.632	149	416180	21.746	ng/ul	97
80) Fluoranthene	19.742	202	688156	20.394	ng/ul	99
82) Pyrene	20.107	202	693059	20.505	ng/ul	99
83) Butylbenzylphthalate	20.976	149	135749	23.164	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.928	252	223943	21.416	ng/ul	98
85) Benzo(a)anthracene	22.010	228	656318	20.136	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.869	149	202585	22.160	ng/ul	99
87) Chrysene	22.087	228	595642	19.508	ng/ul	98
89) Di-n-octyl phthalate	23.221	149	358410	21.673	ng/ul	100
90) Benzo(b)fluoranthene	24.454	252	659976	20.695	ng/ul	98
91) Benzo(k)fluoranthene	24.537	252	660407	20.581	ng/ul	97
93) Benzo(a)pyrene	25.447	252	620668	20.801	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	29.760	276	633717	17.035	ng/ul	97
95) Dibenzo(a,h)anthracene	29.860	278	630808	20.853	ng/ul	97
96) Benzo(g,h,i)perylene	31.041	276	198607	6.724	ng/ul#	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090823\
Data File : BG058922.D
Acq On : 8 Sep 2023 17:58
Operator : MA/JU
Sample : SSTDICV020
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SICV406

Quant Time: Sep 09 00:21:36 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090823.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 09 00:11:02 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.087 ng/uL
RT: 3.679 min Scan# 31
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion: 88 Resp: 8063
Ion Ratio Lower Upper
88 100
43 33.7 19.8 29.8#
58 68.1 49.9 74.9



#5
Pyridine
Concen: 20.563 ng/uL
RT: 4.090 min Scan# 102
Delta R.T. -0.009 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion: 79 Resp: 69438
Ion Ratio Lower Upper
79 100
52 69.3 45.7 68.5#
51 29.0 23.8 35.6









#6
Benzaldehyde
Concen: 23.321 ng/ul
RT: 7.474 min Scan# 61
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion: 77 Resp: 61097
Ion Ratio Lower Upper
77 100
105 85.4 74.3 111.5
106 95.0 67.7 101.5



#8
Phenol
Concen: 20.492 ng/ul
RT: 7.463 min Scan# 676
Delta R.T. -0.009 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion: 94 Resp: 94977
Ion Ratio Lower Upper
94 100
65 28.2 22.4 33.6
66 49.8 38.0 57.0









#10
Bis(2-Chloroethyl)ether
Concen: 19.895 ng/ul
RT: 7.745 min Scan# 71
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion: 93 Resp: 68390
Ion Ratio Lower Upper
93 100
63 84.2 61.8 92.6
95 36.1 23.9 35.9#



#12
2-Chlorophenol
Concen: 21.684 ng/ul
RT: 7.874 min Scan# 746
Delta R.T. -0.009 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:128 Resp: 57964
Ion Ratio Lower Upper
128 100
64 52.1 50.0 75.0
130 33.5 29.4 44.2









#13
2-Methylphenol
Concen: 19.968 ng/ul
RT: 8.743 min Scan# 894
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:108 Resp: 69299
Ion Ratio Lower Upper
108 100
107 88.7 68.7 103.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.168 ng/ul
RT: 8.855 min Scan# 913
Delta R.T. 0.015 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion: 45 Resp: 79788
Ion Ratio Lower Upper
45 100
77 13.2 17.8 26.8#
79 14.3 11.9 17.9









#16
Acetophenone
Concen: 20.849 ng/ul
RT: 9.172 min Scan# 90
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:105 Resp: 128055
Ion Ratio Lower Upper
105 100
77 79.2 66.0 99.0
51 28.2 23.0 34.4



#17
N-Nitroso-di-n-propylamine
Concen: 21.629 ng/ul
RT: 9.137 min Scan# 961
Delta R.T. -0.009 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion: 70 Resp: 69481
Ion Ratio Lower Upper
70 100
42 48.7 39.4 59.0
101 7.3 8.6 12.8#
130 15.8 13.9 20.9









#18
4-Methylphenol
Concen: 20.766 ng/ul
RT: 9.073 min Scan# 91
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:108 Resp: 78862
Ion Ratio Lower Upper
108 100
107 106.1 86.8 130.2



#19
Hexachloroethane
Concen: 21.357 ng/ul
RT: 9.419 min Scan# 1009
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:117 Resp: 23301
Ion Ratio Lower Upper
117 100
201 134.1 116.2 174.4
199 96.2 82.2 123.2









#22
Nitrobenzene
Concen: 21.870 ng/ul
RT: 9.572 min Scan# 1104
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion: 77 Resp: 91684
Ion Ratio Lower Upper
77 100
123 35.8 29.4 44.2
65 12.5 10.0 15.0



#23
Isophorone
Concen: 20.515 ng/ul
RT: 10.083 min Scan# 1122
Delta R.T. 0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion: 82 Resp: 214578
Ion Ratio Lower Upper
82 100
95 9.4 6.6 10.0
138 13.7 12.3 18.5









#25
2-Nitrophenol
Concen: 23.594 ng/ul
RT: 10.277 min Scan# 1155
Delta R.T. -0.009 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:139 Resp: 34174
Ion Ratio Lower Upper
139 100
65 59.7 47.8 71.8
109 46.3 25.0 37.6#



#26
2,4-Dimethylphenol
Concen: 20.895 ng/ul
RT: 10.300 min Scan# 1159
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:107 Resp: 94094
Ion Ratio Lower Upper
107 100
121 47.7 39.4 59.2
122 80.4 67.3 100.9









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.372 ng/ul
 RT: 10.565 min Scan# 11
 Delta R.T. -0.003 min
 Lab File: BG058922.D
 Acq: 8 Sep 2023 17:58




Instrument :
 BNA_G
 ClientSampleId :
 SICV406


Tgt Ion: 93 Resp: 110350
 Ion Ratio Lower Upper
 93 100
 95 28.2 29.8 44.6#
 123 10.2 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 20.812 ng/ul
 RT: 10.794 min Scan# 1243
 Delta R.T. -0.009 min
 Lab File: BG058922.D
 Acq: 8 Sep 2023 17:58

Tgt Ion:162 Resp: 76851
 Ion Ratio Lower Upper
 162 100
 164 63.1 50.0 75.0
 98 34.3 26.3 39.5









#30
Naphthalene
Concen: 20.490 ng/ul
RT: 11.229 min Scan# 1317
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:128 Resp: 230346
Ion Ratio Lower Upper
128 100
129 11.0 7.7 11.5
127 12.3 10.2 15.2



#32
4-Chloroaniline
Concen: 21.029 ng/ul
RT: 11.340 min Scan# 1336
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:127 Resp: 99673
Ion Ratio Lower Upper
127 100
129 30.8 24.6 36.8









#33
Hexachlorobutadiene
Concen: 20.627 ng/ul
RT: 11.458 min Scan# 11
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:225 Resp: 74595
Ion Ratio Lower Upper
225 100
223 61.0 44.1 66.1
227 66.9 49.0 73.4



#34
Caprolactam
Concen: 2.157 ng/ul
RT: 12.181 min Scan# 1479
Delta R.T. 0.068 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:113 Resp: 2635
Ion Ratio Lower Upper
113 100
55 126.4 104.6 156.8
56 119.7 98.8 148.2









#35
4-Chloro-3-methylphenol
Concen: 20.971 ng/ul
RT: 12.404 min Scan# 1586
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:107 Resp: 90983
Ion Ratio Lower Upper
107 100
144 25.1 17.3 25.9
142 76.4 56.9 85.3



#36
2-Methylnaphthalene
Concen: 20.694 ng/ul
RT: 12.803 min Scan# 1585
Delta R.T. -0.009 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:142 Resp: 176107
Ion Ratio Lower Upper
142 100
141 86.7 71.4 107.0





#37

1-Methylnaphthalene

Concen: 20.447 ng/ul

RT: 13.021 min Scan# 1

Delta R.T. -0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:142 Resp: 177984

Ion Ratio Lower Upper

142 100

141 90.3 78.0 117.0

116 4.7 4.6 7.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.838 ng/ul

RT: 13.144 min Scan# 1643

Delta R.T. -0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Tgt Ion:216 Resp: 147999




Ion Ratio Lower Upper


216 100

214 69.9 58.9 88.3

179 16.4 14.4 21.6

108 10.8 10.1 15.1









#40
Hexachlorocyclopentadiene
Concen: 17.686 ng/ul
RT: 13.109 min Scan# 1
Delta R.T. 0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:237 Resp: 94963
Ion Ratio Lower Upper
237 100
235 60.1 49.0 73.6
272 14.6 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 21.007 ng/ul
RT: 13.385 min Scan# 1684
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:196 Resp: 91071
Ion Ratio Lower Upper
196 100
198 94.6 76.1 114.1
200 27.5 25.8 38.6









#42
2,4,5-Trichlorophenol
Concen: 21.184 ng/ul
RT: 13.450 min Scan# 1754
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:196 Resp: 98468
Ion Ratio Lower Upper
196 100
198 97.5 74.3 111.5
200 32.4 24.7 37.1



#43
1,1'-Biphenyl
Concen: 20.082 ng/ul
RT: 13.796 min Scan# 1754
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:154 Resp: 273207
Ion Ratio Lower Upper
154 100
153 40.3 31.8 47.8
76 10.6 8.6 13.0









#44
2-Chloronaphthalene
Concen: 19.817 ng/ul
RT: 13.849 min Scan# 1763
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:162 Resp: 215013
Ion Ratio Lower Upper
162 100
164 31.7 27.1 40.7
127 36.0 27.8 41.8



#45
2-Nitroaniline
Concen: 25.395 ng/ul
RT: 14.061 min Scan# 1799
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion: 65 Resp: 60924
Ion Ratio Lower Upper
65 100
92 81.8 52.8 79.2#
138 103.1 76.4 114.6









#47
Dimethylphthalate
Concen: 20.636 ng/ul
RT: 14.402 min Scan# 1881
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:163 Resp: 323825
Ion Ratio Lower Upper
163 100
194 3.8 3.6 5.4
164 10.1 7.4 11.0



#48
2,6-Dinitrotoluene
Concen: 24.470 ng/ul
RT: 14.543 min Scan# 1881
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:165 Resp: 54502
Ion Ratio Lower Upper
165 100
89 65.2 53.0 79.6
121 24.8 14.2 21.4#









#50
Acenaphthylene
Concen: 20.675 ng/ul
RT: 14.689 min Scan# 1906
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:152 Resp: 384903
Ion Ratio Lower Upper
152 100
151 18.8 15.0 22.4
153 11.9 9.4 14.2



#51
3-Nitroaniline
Concen: 22.987 ng/ul
RT: 14.872 min Scan# 1937
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:138 Resp: 51593
Ion Ratio Lower Upper
138 100
108 10.9 10.2 15.2
92 126.9 105.2 157.8









#52
Acenaphthene
Concen: 20.248 ng/ul
RT: 15.018 min Scan# 1962
Delta R.T. -0.009 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:153 Resp: 245428
Ion Ratio Lower Upper
153 100
152 46.8 36.6 55.0
154 85.8 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 19.228 ng/ul
RT: 15.065 min Scan# 1970
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:184 Resp: 36229
Ion Ratio Lower Upper
184 100
63 45.4 43.3 64.9
154 46.3 38.6 57.8









#55
4-Nitrophenol
Concen: 23.990 ng/ul
RT: 15.136 min Scan# 1982
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:109 Resp: 61186
Ion Ratio Lower Upper
109 100
139 77.0 70.8 106.2
65 104.3 87.5 131.3



#56
Dibenzofuran
Concen: 20.222 ng/ul
RT: 15.353 min Scan# 2019
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:168 Resp: 365918
Ion Ratio Lower Upper
168 100
139 34.0 29.4 44.0





#57

2,4-Dinitrotoluene

Concen: 24.445 ng/ul

RT: 15.318 min Scan# 2081

Delta R.T. 0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:165 Resp: 78541

Ion Ratio Lower Upper

165 100

63 41.2 35.8 53.8

182 2.4 1.7 2.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.457 ng/ul

RT: 15.559 min Scan# 2054

Delta R.T. -0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Tgt Ion:232 Resp: 110288




Ion Ratio Lower Upper


232 100

131 26.6 20.6 31.0

130 1.3 0.0 0.0#

166 19.2 17.2 25.8









#59
Diethylphthalate
Concen: 20.506 ng/ul
RT: 15.753 min Scan# 20
Delta R.T. 0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:149 Resp: 289207
Ion Ratio Lower Upper
149 100
177 26.4 21.0 31.4
150 13.1 11.6 17.4



#61
Fluorene
Concen: 20.357 ng/ul
RT: 16.000 min Scan# 2129
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:166 Resp: 312718
Ion Ratio Lower Upper
166 100
165 94.8 78.5 117.7
167 14.6 10.1 15.1









#62
4-Chlorophenyl-phenylether
Concen: 21.068 ng/ul
RT: 15.988 min Scan# 2196
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:204 Resp: 194153
Ion Ratio Lower Upper
204 100
206 32.8 24.2 36.2
141 47.7 33.0 49.4



#63
4-Nitroaniline
Concen: 26.381 ng/ul
RT: 16.035 min Scan# 2135
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:138 Resp: 55092
Ion Ratio Lower Upper
138 100
92 63.9 48.2 72.2
108 135.8 96.6 145.0









#66
4,6-Dinitro-2-methylphenol
Concen: 21.991 ng/ul
RT: 16.064 min Scan# 2164
Delta R.T. 0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:198 Resp: 59478
Ion Ratio Lower Upper
198 100
182 3.2 3.0 4.6
77 18.8 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 20.495 ng/ul
RT: 16.205 min Scan# 2164
Delta R.T. 0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:169 Resp: 285217
Ion Ratio Lower Upper
169 100
168 67.7 52.4 78.6
167 35.2 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 20.863 ng/ul
RT: 16.887 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:248 Resp: 143361
Ion Ratio Lower Upper
248 100
250 90.4 73.9 110.9
141 49.0 36.1 54.1



#69
Hexachlorobenzene
Concen: 20.560 ng/ul
RT: 16.969 min Scan# 2294
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:284 Resp: 163026
Ion Ratio Lower Upper
284 100
142 20.8 16.2 24.2
249 31.7 24.4 36.6





#70

Atrazine

Concen: 21.884 ng/ul

RT: 17.140 min Scan# 21

Delta R.T. -0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:200 Resp: 122272

Ion Ratio Lower Upper

200 100

173 22.8 16.5 24.7

215 50.4 42.1 63.1



#71

Pentachlorophenol

Concen: 20.428 ng/ul

RT: 17.322 min Scan# 2354

Delta R.T. -0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58




Tgt Ion:266 Resp: 99482


Ion Ratio Lower Upper

266 100

264 65.3 45.4 68.0

268 64.2 50.0 75.0









#72
Phenanthrene
Concen: 20.436 ng/ul
RT: 17.751 min Scan# 24
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:178 Resp: 561400
Ion Ratio Lower Upper
178 100
179 15.3 12.9 19.3
176 19.8 15.0 22.6



#74
Anthracene
Concen: 20.843 ng/ul
RT: 17.845 min Scan# 2443
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:178 Resp: 584596
Ion Ratio Lower Upper
178 100
179 14.8 11.1 16.7
176 18.2 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.534 ng/uL
RT: 13.755 min Scan# 11
Delta R.T. 0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:216 Resp: 160514
Ion Ratio Lower Upper
216 100
214 81.7 63.9 95.9
218 46.4 34.5 51.7



#76
Pentachlorobenzene
Concen: 19.684 ng/uL
RT: 15.248 min Scan# 2001
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:250 Resp: 163009
Ion Ratio Lower Upper
250 100
248 64.9 51.2 76.8
252 61.3 53.4 80.2









#77
 Carbazole
 Concen: 21.375 ng/ul
 RT: 18.115 min Scan# 24
 Delta R.T. -0.003 min
 Lab File: BG058922.D
 Acq: 8 Sep 2023 17:58




Instrument :
 BNA_G
 ClientSampleId :
 SICV406


Tgt Ion:167 Resp: 455055
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.8 26.6
 139 11.6 8.8 13.2



#78
 Di-n-butylphthalate
 Concen: 21.746 ng/ul
 RT: 18.632 min Scan# 2577
 Delta R.T. -0.003 min
 Lab File: BG058922.D
 Acq: 8 Sep 2023 17:58

Tgt Ion:149 Resp: 416180
 Ion Ratio Lower Upper
 149 100
 150 9.6 9.1 13.7
 104 7.1 5.4 8.0









#80
Fluoranthene
Concen: 20.394 ng/ul
RT: 19.742 min Scan# 21
Delta R.T. 0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:202 Resp: 688156
Ion Ratio Lower Upper
202 100
101 5.8 4.6 7.0
100 5.4 3.8 5.6



#82
Pyrene
Concen: 20.505 ng/ul
RT: 20.107 min Scan# 2828
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:202 Resp: 693059
Ion Ratio Lower Upper
202 100
101 6.1 5.2 7.8
100 6.4 5.0 7.4









#83
Butylbenzylphthalate
Concen: 23.164 ng/ul
RT: 20.976 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:149 Resp: 135749
Ion Ratio Lower Upper
149 100
91 78.9 66.8 100.2
206 24.8 22.7 34.1



#84
3,3'-Dichlorobenzidine
Concen: 21.416 ng/ul
RT: 21.928 min Scan# 3138
Delta R.T. 0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:252 Resp: 223943
Ion Ratio Lower Upper
252 100
254 65.6 51.0 76.6
126 5.4 4.7 7.1





#85

Benzo(a)anthracene

Concen: 20.136 ng/ul

RT: 22.010 min Scan# 3152

Delta R.T. -0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:228 Resp: 656318

Ion Ratio Lower Upper

228 100

229 19.2 16.4 24.6

226 27.8 22.6 34.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.160 ng/ul

RT: 21.869 min Scan# 3128

Delta R.T. -0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58




Tgt Ion:149 Resp: 202585


Ion Ratio Lower Upper

149 100

167 29.0 23.4 35.2

279 6.2 5.4 8.0





#87

Chrysene

Concen: 19.508 ng/ul

RT: 22.087 min Scan# 3165

Delta R.T. -0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:228 Resp: 595642

Ion Ratio Lower Upper

228 100

226 29.4 24.7 37.1

229 19.1 14.7 22.1



#89

Di-n-octyl phthalate

Concen: 21.673 ng/ul




RT: 23.221 min Scan# 3358


Delta R.T. -0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Tgt Ion:149 Resp: 358410









#90
Benzo(b)fluoranthene
Concen: 20.695 ng/ul
RT: 24.454 min Scan# 3568
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:252 Resp: 659976
Ion Ratio Lower Upper
252 100
253 23.0 17.6 26.4
125 4.8 3.9 5.9



#91
Benzo(k)fluoranthene
Concen: 20.581 ng/ul
RT: 24.537 min Scan# 3582
Delta R.T. -0.003 min
Lab File: BG058922.D
Acq: 8 Sep 2023 17:58

Tgt Ion:252 Resp: 660407
Ion Ratio Lower Upper
252 100
253 22.2 16.6 25.0
125 4.0 3.7 5.5





#93

Benzo(a)pyrene

Concen: 20.801 ng/ul

RT: 25.447 min Scan# 31

Delta R.T. 0.009 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:252 Resp: 620668

Ion Ratio Lower Upper

252 100

253 20.8 16.4 24.6

125 4.9 4.3 6.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.035 ng/ul

RT: 29.760 min Scan# 4471

Delta R.T. -0.009 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58




Tgt Ion:276 Resp: 633717


Ion Ratio Lower Upper

276 100

138 9.2 6.2 9.2

277 23.6 17.8 26.6





#95

Dibenzo(a,h)anthracene

Concen: 20.853 ng/ul

RT: 29.860 min Scan# 4

Delta R.T. 0.003 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:278 Resp: 630808

Ion Ratio Lower Upper

278 100

139 6.4 5.5 8.3

279 22.4 19.5 29.3



#96

Benzo(g,h,i)perylene

Concen: 6.724 ng/ul

RT: 31.041 min Scan# 4689

Delta R.T. -0.032 min

Lab File: BG058922.D

Acq: 8 Sep 2023 17:58

Tgt Ion:276 Resp: 198607

Ion Ratio Lower Upper

276 100

138 8.3 8.7 13.1#

277 18.9 19.7 29.5#

