

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090823\
 Data File : BG058926.D
 Acq On : 8 Sep 2023 20:58
 Operator : MA/JU
 Sample : 04227-21
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
ClientSampleId :
 BBJ90

Manual Integrations
APPROVED
 Reviewed By :Yogesh Patel 09/09/2023
 Supervised By :mohammad ahmed 09/10/2023

Quant Time: Sep 09 03:54:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 09 00:11:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.333	152	62132	20.000	ng/u1	0.00
20) Naphthalene-d8	11.176	136	270654	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.960	164	255502	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.704	188	721469m	20.000	ng/u1	0.00
79) Chrysene-d12	22.034	240	606373	20.000	ng/u1	0.00
88) Perylene-d12	25.612	264	699789	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.638	96	6378m	4.924	ng/uL	0.00
4) Pyridine-d5	4.073	84	47737	10.733	ng/u1	0.00
7) Phenol-d5	7.434	99	44823	7.274	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.651	67	116117	33.308	ng/u1	0.00
11) 2-Chlorophenol-d4	7.845	132	94371	25.673	ng/u1	0.00
15) 4-Methylphenol-d8	9.008	113	74617	15.623	ng/u1	0.00
21) Nitrobenzene-d5	9.531	128	67474	42.324	ng/u1	0.00
24) 2-Nitrophenol-d4	10.242	143	70544	40.394	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.765	165	162459	32.064	ng/u1	0.00
31) 4-Chloroaniline-d4	11.317	131	177955	27.380	ng/u1	0.00
46) Dimethylphthalate-d6	14.355	166	775007	39.294	ng/u1	0.00
49) Acenaphthylene-d8	14.660	160	771669	36.197	ng/u1	0.00
54) 4-Nitrophenol-d4	15.113	143	19671	7.549	ng/u1	0.00
60) Fluorene-d10	15.947	176	710415	38.217	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.047	200	126786	38.599	ng/u1	0.00
73) Anthracene-d10	17.810	188	1296603	39.508	ng/u1	0.00
81) Pyrene-d10	20.078	212	1389916	39.524	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.366	264	1387976	40.264	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090823\
 Data File : BG058926.D
 Acq On : 8 Sep 2023 20:58
 Operator : MA/JU
 Sample : 04227-21
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
ClientSampleId :
 BBJ90

Quant Time: Sep 09 03:54:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 09 00:11:02 2023
 Response via : Initial Calibration

Manual Integrations
APPROVED
 Reviewed By :Yogesh Patel 09/09/2023
 Supervised By :mohammad ahmed 09/10/2023

