

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091023\
 Data File : BG058955.D
 Acq On : 10 Sep 2023 16:06
 Operator : MA/JU
 Sample : SSTD00507
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005407

Quant Time: Sep 10 23:05:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Sep 10 22:58:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.329	152	52707	20.000	ng/ul	0.00
20) Naphthalene-d8	11.167	136	253808	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.951	164	194483	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.700	188	477131	20.000	ng/ul	# 0.00
79) Chrysene-d12	22.019	240	445945	20.000	ng/ul	0.00
88) Perylene-d12	25.591	264	513469	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.635	96	2215	2.016	ng/uL	0.00
4) Pyridine-d5	4.075	84	16696	4.425	ng/ul	0.00
7) Phenol-d5	7.430	99	24973	4.777	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.648	67	15701	5.309	ng/ul	0.00
11) 2-Chlorophenol-d4	7.841	132	14978	4.803	ng/ul	0.00
15) 4-Methylphenol-d8	9.005	113	20821	5.139	ng/ul	0.00
21) Nitrobenzene-d5	9.516	128	8666	5.797	ng/ul	0.00
24) 2-Nitrophenol-d4	10.239	143	8759	5.348	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.762	165	20300	4.272	ng/ul	0.00
31) 4-Chloroaniline-d4	11.302	131	20599	3.380	ng/ul	0.00
46) Dimethylphthalate-d6	14.346	166	73214	4.877	ng/ul	0.00
49) Acenaphthylene-d8	14.651	160	79160	4.878	ng/ul	0.00
54) 4-Nitrophenol-d4	15.109	143	9037	4.556	ng/ul	0.00
60) Fluorene-d10	15.932	176	66223	4.680	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.038	200	8623	3.970	ng/ul	0.00
73) Anthracene-d10	17.700	188	477131	21.983	ng/ul	-0.09
81) Pyrene-d10	20.068	212	116670	4.511	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.333	264	114919	4.543	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.687	88	3664	2.738	ng/uL#	57
5) Pyridine	4.093	79	21195	5.336	ng/ul	92
6) Benzaldehyde	7.471	77	18072	5.865	ng/ul#	74
8) Phenol	7.459	94	26893	4.933	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.742	93	21104	5.220	ng/ul#	84
12) 2-Chlorophenol	7.877	128	18493	5.882	ng/ul	95
13) 2-Methylphenol	8.740	108	20966	5.136	ng/ul	93
14) 2,2'-oxybis(1-Chloropr...	8.828	45	11976	2.574	ng/ul#	90
16) Acetophenone	9.163	105	36742	5.086	ng/ul	92
17) N-Nitroso-di-n-propyla...	9.128	70	19650	5.201	ng/ul#	92
18) 4-Methylphenol	9.063	108	21772	4.874	ng/ul	95
19) Hexachloroethane	9.410	117	7285	5.677	ng/ul	89
22) Nitrobenzene	9.563	77	24964	4.882	ng/ul#	93
23) Isophorone	10.068	82	56712	4.446	ng/ul#	97
25) 2-Nitrophenol	10.268	139	9602	5.436	ng/ul#	77
26) 2,4-Dimethylphenol	10.286	107	24664	4.491	ng/ul	95
27) Bis(2-Chloroethoxy)met...	10.556	93	30344	4.593	ng/ul	97
29) 2,4-Dichlorophenol	10.797	162	19417	4.311	ng/ul#	89
30) Naphthalene	11.226	128	64228	4.684	ng/ul	97
32) 4-Chloroaniline	11.337	127	25423	4.398	ng/ul	99
33) Hexachlorobutadiene	11.449	225	16006	3.629	ng/ul	93
34) Caprolactam	12.101	113	7525	5.051	ng/ul#	68
35) 4-Chloro-3-methylphenol	12.395	107	23632	4.466	ng/ul#	82
36) 2-Methylnaphthalene	12.800	142	50168	4.833	ng/ul	87

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091023\
 Data File : BG058955.D
 Acq On : 10 Sep 2023 16:06
 Operator : MA/JU
 Sample : SSTD00507
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005407

Quant Time: Sep 10 23:05:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Sep 10 22:58:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.018	142	47868	4.509	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	13.135	216	29528	4.069	ng/ul#	90
40) Hexachlorocyclopentadiene	13.094	237	14436	2.764	ng/ul#	94
41) 2,4,6-Trichlorophenol	13.376	196	17779	4.216	ng/ul#	93
42) 2,4,5-Trichlorophenol	13.441	196	20052	4.435	ng/ul	98
43) 1,1'-Biphenyl	13.787	154	68529	5.179	ng/ul	95
44) 2-Chloronaphthalene	13.840	162	52237	4.950	ng/ul#	89
45) 2-Nitroaniline	14.052	65	14747	6.320	ng/ul	95
47) Dimethylphthalate	14.393	163	70906	4.645	ng/ul	97
48) 2,6-Dinitrotoluene	14.528	165	11532	5.323	ng/ul	96
50) Acenaphthylene	14.680	152	86167	4.758	ng/ul	96
51) 3-Nitroaniline	14.868	138	11329	5.189	ng/ul#	93
52) Acenaphthene	15.015	153	61041	5.177	ng/ul	95
53) 2,4-Dinitrophenol	15.056	184	4952	2.702	ng/ul#	82
55) 4-Nitrophenol	15.127	109	11636	4.690	ng/ul#	75
56) Dibenzofuran	15.344	168	90126	5.121	ng/ul	99
57) 2,4-Dinitrotoluene	15.303	165	15689	5.020	ng/ul#	94
58) 2,3,4,6-Tetrachlorophenol	15.550	232	19112	4.001	ng/ul#	97
59) Diethylphthalate	15.738	149	69909	5.096	ng/ul#	94
61) Fluorene	15.991	166	71288	4.771	ng/ul#	95
62) 4-Chlorophenyl-phenyle...	15.979	204	38270	4.269	ng/ul	89
63) 4-Nitroaniline	16.026	138	10946	5.389	ng/ul#	68
66) 4,6-Dinitro-2-methylph...	16.049	198	8782	3.851	ng/ul#	97
67) N-Nitrosodiphenylamine	16.196	169	61672	5.256	ng/ul	96
68) 4-Bromophenyl-phenylether	16.872	248	25663	4.429	ng/ul	84
69) Hexachlorobenzene	16.960	284	29020	4.341	ng/ul	98
70) Atrazine	17.125	200	22638	4.805	ng/ul#	85
71) Pentachlorophenol	17.313	266	14174	3.452	ng/ul#	86
72) Phenanthrene	17.742	178	120967	5.222	ng/ul	99
74) Anthracene	17.836	178	122981	5.200	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.746	216	30849	4.453	ng/ul	92
76) Pentachlorobenzene	15.239	250	30597	4.382	ng/ul	94
77) Carbazole	18.106	167	102167	5.692	ng/ul	93
78) Di-n-butylphthalate	18.623	149	110235	6.831	ng/ul	98
80) Fluoranthene	19.733	202	138473	4.406	ng/ul#	95
82) Pyrene	20.098	202	147160	4.674	ng/ul#	95
83) Butylbenzylphthalate	20.961	149	39347	7.208	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.913	252	49845	5.117	ng/ul#	87
85) Benzo(a)anthracene	22.001	228	149381	4.920	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.849	149	61829	7.261	ng/ul	97
87) Chrysene	22.072	228	138530	4.871	ng/ul	97
89) Di-n-octyl phthalate	23.182	149	100858	6.290	ng/ul	100
90) Benzo(b)fluoranthene	24.434	252	147445	4.768	ng/ul	97
91) Benzo(k)fluoranthene	24.504	252	152631	4.906	ng/ul	96
93) Benzo(a)pyrene	25.421	252	136371	4.713	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	29.692	276	53944	1.495	ng/ul#	95
95) Dibenzo(a,h)anthracene	29.792	278	16601	0.566	ng/ul	97
96) Benzo(g,h,i)perylene	31.002	276	51533	1.799	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA G

BNA G

ClientSampleId :

SSTD005407


Quant Time: Sep 10 23:05:09 2023


Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091023.MA.M

Quant Title : SVOA CALIBRATION

QLast Update : Sun Sep 10 22:58:52 2023

Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 2.738 ng/uL
RT: 3.687 min Scan# 34
Delta R.T. 0.011 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion: 88 Resp: 3664
Ion Ratio Lower Upper
88 100
43 0.0 18.6 28.0#
58 34.4 53.8 80.8#



#5
Pyridine
Concen: 5.336 ng/uL
RT: 4.093 min Scan# 103
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion: 79 Resp: 21195
Ion Ratio Lower Upper
79 100
52 56.4 51.4 77.0
51 32.9 25.2 37.8









#6
Benzaldehyde
Concen: 5.865 ng/ul
RT: 7.471 min Scan# 61
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion: 77 Resp: 18072
Ion Ratio Lower Upper
77 100
105 80.1 81.6 122.4#
106 75.0 84.5 126.7#



#8
Phenol
Concen: 4.933 ng/ul
RT: 7.459 min Scan# 676
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion: 94 Resp: 26893
Ion Ratio Lower Upper
94 100
65 29.4 22.4 33.6
66 41.5 35.0 52.6









#10
Bis(2-Chloroethyl)ether
Concen: 5.220 ng/ul
RT: 7.742 min Scan# 71
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion: 93 Resp: 21104
Ion Ratio Lower Upper
93 100
63 87.5 56.2 84.4#
95 30.6 26.4 39.6



#12
2-Chlorophenol
Concen: 5.882 ng/ul
RT: 7.877 min Scan# 747
Delta R.T. 0.005 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:128 Resp: 18493
Ion Ratio Lower Upper
128 100
64 52.5 43.1 64.7
130 23.3 23.2 34.8





#13

2-Methylphenol

Concen: 5.136 ng/ul

RT: 8.740 min Scan# 894

Delta R.T. -0.001 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Instrument :

BNA_G

ClientSampleId :





SSTD005407

Tgt Ion:108 Resp: 20966

Ion Ratio Lower Upper

108 100

107 82.0 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 2.574 ng/ul

RT: 8.828 min Scan# 909

Delta R.T. -0.007 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06




Tgt Ion: 45 Resp: 11976


Ion Ratio Lower Upper

45 100

77 20.8 13.2 19.8#

79 8.1 9.8 14.6#









#16
Acetophenone
Concen: 5.086 ng/ul
RT: 9.163 min Scan# 916
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:105 Resp: 36742
Ion Ratio Lower Upper
105 100
77 76.5 67.3 100.9
51 21.3 19.3 28.9



#17
N-Nitroso-di-n-propylamine
Concen: 5.201 ng/ul
RT: 9.128 min Scan# 960
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion: 70 Resp: 19650
Ion Ratio Lower Upper
70 100
42 43.6 37.5 56.3
101 8.8 5.0 7.6#
130 9.4 13.9 20.9#









#18
4-Methylphenol
Concen: 4.874 ng/ul
RT: 9.063 min Scan# 94
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:108 Resp: 21772
Ion Ratio Lower Upper
108 100
107 116.4 88.9 133.3



#19
Hexachloroethane
Concen: 5.677 ng/ul
RT: 9.410 min Scan# 1008
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:117 Resp: 7285
Ion Ratio Lower Upper
117 100
201 122.7 108.2 162.4
199 79.3 56.0 84.0





#22

Nitrobenzene

Concen: 4.882 ng/ul

RT: 9.563 min Scan# 1034

Delta R.T. -0.001 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Instrument :

BNA_G

ClientSampleId :

SSTD005407





Tgt Ion: 77 Resp: 24964

Ion Ratio Lower Upper

77 100

123 39.4 28.5 42.7

65 11.4 11.5 17.3#



#23

Isophorone

Concen: 4.446 ng/ul

RT: 10.068 min Scan# 1120

Delta R.T. -0.007 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06




Tgt Ion: 82 Resp: 56712


Ion Ratio Lower Upper

82 100

95 6.5 6.6 9.8#

138 13.5 10.4 15.6









#25
2-Nitrophenol
Concen: 5.436 ng/ul
RT: 10.268 min Scan# 1154
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:139 Resp: 9602
Ion Ratio Lower Upper
139 100
65 42.9 53.5 80.3#
109 33.0 32.0 48.0



#26
2,4-Dimethylphenol
Concen: 4.491 ng/ul
RT: 10.286 min Scan# 1157
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:107 Resp: 24664
Ion Ratio Lower Upper
107 100
121 45.8 36.4 54.6
122 79.7 69.4 104.0





#27

Bis(2-Chloroethoxy)methane

Concen: 4.593 ng/ul

RT: 10.556 min Scan# 1203

Delta R.T. -0.001 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Instrument :

BNA_G

ClientSampleId :

SSTD005407





Tgt Ion: 93 Resp: 30344

Ion Ratio Lower Upper

93 100

95 36.3 27.8 41.6

123 7.9 7.3 10.9



#29

2,4-Dichlorophenol

Concen: 4.311 ng/ul

RT: 10.797 min Scan# 1244

Delta R.T. 0.005 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06




Tgt Ion:162 Resp: 19417


Ion Ratio Lower Upper

162 100

164 58.2 51.4 77.0

98 24.2 27.2 40.8#









#30
Naphthalene
Concen: 4.684 ng/ul
RT: 11.226 min Scan# 11
Delta R.T. 0.005 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:128 Resp: 64228
Ion Ratio Lower Upper
128 100
129 8.7 8.2 12.2
127 14.4 10.9 16.3



#32
4-Chloroaniline
Concen: 4.398 ng/ul
RT: 11.337 min Scan# 1336
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:127 Resp: 25423
Ion Ratio Lower Upper
127 100
129 32.5 26.6 39.8









#33
Hexachlorobutadiene
Concen: 3.629 ng/ul
RT: 11.449 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:225 Resp: 16006
Ion Ratio Lower Upper
225 100
223 61.6 45.5 68.3
227 56.3 50.2 75.4



#34
Caprolactam
Concen: 5.051 ng/ul
RT: 12.101 min Scan# 1466
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:113 Resp: 7525
Ion Ratio Lower Upper
113 100
55 126.0 140.5 210.7#
56 83.5 90.4 135.6#









#35
4-Chloro-3-methylphenol
Concen: 4.466 ng/ul
RT: 12.395 min Scan# 1516
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:107 Resp: 23632
Ion Ratio Lower Upper
107 100
144 28.8 20.2 30.2
142 82.0 51.6 77.4#



#36
2-Methylnaphthalene
Concen: 4.833 ng/ul
RT: 12.800 min Scan# 1585
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:142 Resp: 50168
Ion Ratio Lower Upper
142 100
141 93.6 65.7 98.5





#37

1-Methylnaphthalene

Concen: 4.509 ng/ul

RT: 13.018 min Scan# 1

Delta R.T. -0.001 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Instrument :

BNA_G

ClientSampleId :

SSTD005407





Tgt Ion:142 Resp: 47868

Ion Ratio Lower Upper

142 100

141 87.1 73.1 109.7

116 7.2 3.4 5.0#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.069 ng/ul

RT: 13.135 min Scan# 1642

Delta R.T. -0.007 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Tgt Ion:216 Resp: 29528




Ion Ratio Lower Upper


216 100

214 63.6 59.6 89.4

179 17.3 13.6 20.4

108 13.7 13.8 20.6#









#40
Hexachlorocyclopentadiene
Concen: 2.764 ng/ul
RT: 13.094 min Scan# 1635
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:237 Resp: 14436
Ion Ratio Lower Upper
237 100
235 61.4 52.6 78.8
272 15.0 8.6 13.0#



#41
2,4,6-Trichlorophenol
Concen: 4.216 ng/ul
RT: 13.376 min Scan# 1683
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:196 Resp: 17779
Ion Ratio Lower Upper
196 100
198 93.1 78.3 117.5
200 37.9 23.7 35.5#





#42

2,4,5-Trichlorophenol

Concen: 4.435 ng/ul

RT: 13.441 min Scan# 1

Delta R.T. -0.007 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Instrument :

BNA_G

ClientSampleId :

SSTD005407





Tgt Ion:196 Resp: 20052

Ion Ratio Lower Upper

196 100

198 108.0 84.8 127.2

200 35.5 27.7 41.5



#43

1,1'-Biphenyl

Concen: 5.179 ng/ul

RT: 13.787 min Scan# 1753

Delta R.T. -0.001 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06




Tgt Ion:154 Resp: 68529


Ion Ratio Lower Upper

154 100

153 43.0 31.6 47.4

76 13.0 9.0 13.4









#44
2-Chloronaphthalene
Concen: 4.950 ng/ul
RT: 13.840 min Scan# 1762
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:162 Resp: 52237
Ion Ratio Lower Upper
162 100
164 24.4 27.6 41.4#
127 39.2 29.4 44.0



#45
2-Nitroaniline
Concen: 6.320 ng/ul
RT: 14.052 min Scan# 1798
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion: 65 Resp: 14747
Ion Ratio Lower Upper
65 100
92 58.5 54.0 81.0
138 80.5 64.8 97.2









#47
Dimethylphthalate
Concen: 4.645 ng/ul
RT: 14.393 min Scan# 1856
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:163 Resp: 70906
Ion Ratio Lower Upper
163 100
194 3.4 2.8 4.2
164 8.4 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 5.323 ng/ul
RT: 14.528 min Scan# 1879
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:165 Resp: 11532
Ion Ratio Lower Upper
165 100
89 73.4 55.8 83.6
121 21.9 16.3 24.5









#50
Acenaphthylene
Concen: 4.758 ng/ul
RT: 14.680 min Scan# 1905
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:152 Resp: 86167
Ion Ratio Lower Upper
152 100
151 19.8 14.6 21.8
153 14.3 10.4 15.6



#51
3-Nitroaniline
Concen: 5.189 ng/ul
RT: 14.868 min Scan# 1937
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:138 Resp: 11329
Ion Ratio Lower Upper
138 100
108 22.8 11.8 17.6#
92 145.1 122.2 183.2









#52
Acenaphthene
Concen: 5.177 ng/ul
RT: 15.015 min Scan# 1962
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:153 Resp: 61041
Ion Ratio Lower Upper
153 100
152 49.2 36.5 54.7
154 88.5 74.1 111.1



#53
2,4-Dinitrophenol
Concen: 2.702 ng/ul
RT: 15.056 min Scan# 1969
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:184 Resp: 4952
Ion Ratio Lower Upper
184 100
63 68.0 57.7 86.5
154 50.3 60.7 91.1#









#55
4-Nitrophenol
Concen: 4.690 ng/ul
RT: 15.127 min Scan# 1981
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:109 Resp: 11636
Ion Ratio Lower Upper
109 100
139 62.1 74.3 111.5#
65 96.1 93.4 140.0



#56
Dibenzofuran
Concen: 5.121 ng/ul
RT: 15.344 min Scan# 2018
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:168 Resp: 90126
Ion Ratio Lower Upper
168 100
139 33.5 27.4 41.2





#57

2,4-Dinitrotoluene

Concen: 5.020 ng/ul

RT: 15.303 min Scan# 2012

Delta R.T. -0.007 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Instrument :

BNA_G

ClientSampleId :

SSTD005407





Tgt Ion:165 Resp: 15689

Ion Ratio Lower Upper

165 100

63 37.6 33.3 49.9

182 0.0 1.4 2.0#



#58

2,3,4,6-Tetrachlorophenol

Concen: 4.001 ng/ul

RT: 15.550 min Scan# 2053

Delta R.T. -0.001 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Tgt Ion:232 Resp: 19112




Ion Ratio Lower Upper


232 100

131 31.7 27.1 40.7

130 2.8 0.7 1.1#

166 25.3 19.5 29.3









#59
Diethylphthalate
Concen: 5.096 ng/ul
RT: 15.738 min Scan# 2085
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:149 Resp: 69909
Ion Ratio Lower Upper
149 100
177 24.0 16.9 25.3
150 14.5 9.7 14.5#



#61
Fluorene
Concen: 4.771 ng/ul
RT: 15.991 min Scan# 2128
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:166 Resp: 71288
Ion Ratio Lower Upper
166 100
165 100.5 76.4 114.6
167 14.1 9.3 13.9#









#62
4-Chlorophenyl-phenylether
Concen: 4.269 ng/ul
RT: 15.979 min Scan# 2126
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:204 Resp: 38270
Ion Ratio Lower Upper
204 100
206 28.2 27.7 41.5
141 46.6 43.2 64.8



#63
4-Nitroaniline
Concen: 5.389 ng/ul
RT: 16.026 min Scan# 2134
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:138 Resp: 10946
Ion Ratio Lower Upper
138 100
92 70.2 49.9 74.9
108 100.0 121.4 182.2#









#66
4,6-Dinitro-2-methylphenol
Concen: 3.851 ng/ul
RT: 16.049 min Scan# 2138
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:198 Resp: 8782
Ion Ratio Lower Upper
198 100
182 4.6 3.0 4.6#
77 23.6 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 5.256 ng/ul
RT: 16.196 min Scan# 2163
Delta R.T. 0.005 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:169 Resp: 61672
Ion Ratio Lower Upper
169 100
168 64.9 53.0 79.6
167 32.9 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 4.429 ng/ul
RT: 16.872 min Scan# 211
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:248 Resp: 25663
Ion Ratio Lower Upper
248 100
250 109.4 73.9 110.9
141 66.3 45.4 68.0



#69
Hexachlorobenzene
Concen: 4.341 ng/ul
RT: 16.960 min Scan# 2293
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:284 Resp: 29020
Ion Ratio Lower Upper
284 100
142 28.8 24.6 37.0
249 30.0 24.3 36.5





#70

Atrazine

Concen: 4.805 ng/ul

RT: 17.125 min Scan# 21

Delta R.T. -0.007 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Instrument :

BNA_G

ClientSampleId :

SSTD005407





Tgt Ion:200 Resp: 22638

Ion Ratio Lower Upper

200 100

173 22.0 17.4 26.2

215 58.5 35.4 53.0#



#71

Pentachlorophenol

Concen: 3.452 ng/ul

RT: 17.313 min Scan# 2353

Delta R.T. -0.001 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06




Tgt Ion:266 Resp: 14174


Ion Ratio Lower Upper

266 100

264 88.5 53.1 79.7#

268 65.0 51.6 77.4









#72
Phenanthrene
Concen: 5.222 ng/ul
RT: 17.742 min Scan# 2426
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:178 Resp: 120967
Ion Ratio Lower Upper
178 100
179 15.3 11.8 17.8
176 18.6 15.1 22.7



#74
Anthracene
Concen: 5.200 ng/ul
RT: 17.836 min Scan# 2442
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:178 Resp: 122981
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.4
176 18.3 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.453 ng/uL
RT: 13.746 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:216 Resp: 30849
Ion Ratio Lower Upper
216 100
214 68.5 60.6 91.0
218 46.4 33.6 50.4



#76
Pentachlorobenzene
Concen: 4.382 ng/uL
RT: 15.239 min Scan# 2000
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:250 Resp: 30597
Ion Ratio Lower Upper
250 100
248 58.3 50.3 75.5
252 63.9 47.6 71.4









#77
 Carbazole
 Concen: 5.692 ng/ul
 RT: 18.106 min Scan# 2488
 Delta R.T. -0.001 min
 Lab File: BG058955.D
 Acq: 10 Sep 2023 16:06




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005407


Tgt Ion:167 Resp: 102167
 Ion Ratio Lower Upper
 167 100
 166 16.9 16.8 25.2
 139 10.7 9.4 14.2



#78
 Di-n-butylphthalate
 Concen: 6.831 ng/ul
 RT: 18.623 min Scan# 2576
 Delta R.T. -0.001 min
 Lab File: BG058955.D
 Acq: 10 Sep 2023 16:06

Tgt Ion:149 Resp: 110235
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.7 11.5
 104 5.0 4.6 6.8









#80
Fluoranthene
Concen: 4.406 ng/ul
RT: 19.733 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:202 Resp: 138473
Ion Ratio Lower Upper
202 100
101 10.4 6.1 9.1#
100 6.8 5.9 8.9



#82
Pyrene
Concen: 4.674 ng/ul
RT: 20.098 min Scan# 2827
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:202 Resp: 147160
Ion Ratio Lower Upper
202 100
101 12.7 7.5 11.3#
100 7.6 6.3 9.5









#83
Butylbenzylphthalate
Concen: 7.208 ng/ul
RT: 20.961 min Scan# 2974
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:149 Resp: 39347
Ion Ratio Lower Upper
149 100
91 78.2 68.7 103.1
206 24.4 17.3 25.9



#84
3,3'-Dichlorobenzidine
Concen: 5.117 ng/ul
RT: 21.913 min Scan# 3136
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:252 Resp: 49845
Ion Ratio Lower Upper
252 100
254 60.6 57.6 86.4
126 11.9 7.8 11.8#









#85
Benzo(a)anthracene
Concen: 4.920 ng/ul
RT: 22.001 min Scan# 3151
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:228 Resp: 149381
Ion Ratio Lower Upper
228 100
229 17.6 16.0 24.0
226 24.9 20.8 31.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 7.261 ng/ul
RT: 21.849 min Scan# 3125
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:149 Resp: 61829
Ion Ratio Lower Upper
149 100
167 28.0 21.0 31.6
279 3.9 3.8 5.8









#87
Chrysene
Concen: 4.871 ng/ul
RT: 22.072 min Scan# 3163
Delta R.T. -0.001 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:228 Resp: 138530
Ion Ratio Lower Upper
228 100
226 28.5 23.2 34.8
229 22.3 15.3 22.9



#89
Di-n-octyl phthalate
Concen: 6.290 ng/ul
RT: 23.182 min Scan# 3352
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:149 Resp: 100858





#90

Benzo(b)fluoranthene

Concen: 4.768 ng/ul

RT: 24.434 min Scan# 3565

Delta R.T. -0.001 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Instrument :

BNA_G

ClientSampleId :

SSTD005407





Tgt Ion:252 Resp: 147445

Ion Ratio Lower Upper

252 100

253 20.3 17.8 26.8

125 7.7 6.2 9.2



#91

Benzo(k)fluoranthene

Concen: 4.906 ng/ul

RT: 24.504 min Scan# 3577

Delta R.T. -0.007 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06




Tgt Ion:252 Resp: 152631


Ion Ratio Lower Upper

252 100

253 22.1 19.1 28.7

125 7.2 5.0 7.4









#93
Benzo(a)pyrene
Concen: 4.713 ng/ul
RT: 25.421 min Scan# 31
Delta R.T. -0.007 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06




Instrument :
BNA_G
ClientSampleId :
SSTD005407


Tgt Ion:252 Resp: 136371
Ion Ratio Lower Upper
252 100
253 20.5 17.1 25.7
125 8.4 5.3 7.9#



#94
Indeno(1,2,3-cd)pyrene
Concen: 1.495 ng/ul
RT: 29.692 min Scan# 4460
Delta R.T. -0.024 min
Lab File: BG058955.D
Acq: 10 Sep 2023 16:06

Tgt Ion:276 Resp: 53944
Ion Ratio Lower Upper
276 100
138 11.5 11.8 17.8#
277 24.3 20.7 31.1





#95

Dibenzo(a,h)anthracene

Concen: 0.566 ng/ul

RT: 29.792 min Scan# 44

Delta R.T. -0.018 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Instrument :

BNA_G

ClientSampleId :

SSTD005407





Tgt Ion:278 Resp: 16601

Ion Ratio Lower Upper

278 100

139 12.3 9.8 14.8

279 21.0 18.5 27.7



#96

Benzo(g,h,i)perylene

Concen: 1.799 ng/ul

RT: 31.002 min Scan# 4683

Delta R.T. -0.036 min

Lab File: BG058955.D

Acq: 10 Sep 2023 16:06

Tgt Ion:276 Resp: 51533

Ion Ratio Lower Upper

276 100

138 16.5 11.4 17.2

277 26.9 17.4 26.2#

