

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091023\
 Data File : BG058987.D
 Acq On : 11 Sep 2023 17:32
 Operator : MA/JU
 Sample : 04234-03MS
 Misc :
 ALS Vial : 86 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 YBY14MS

Quant Time: Sep 12 00:51:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 00:07:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.328	152	82139	20.000	ng/u1	0.00
20) Naphthalene-d8	11.172	136	418801	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.950	164	300058	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.700	188	744248	20.000	ng/u1	# 0.00
79) Chrysene-d12	22.024	240	701251	20.000	ng/u1	0.00
88) Perylene-d12	25.591	264	772993	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.640	96	12814	6.631	ng/uL	0.00
4) Pyridine-d5	4.069	84	33482	5.473	ng/u1	0.00
7) Phenol-d5	7.435	99	276345	33.363	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.647	67	149774	30.837	ng/u1	0.00
11) 2-Chlorophenol-d4	7.835	132	173451	32.650	ng/u1	0.00
15) 4-Methylphenol-d8	9.010	113	197582	30.876	ng/u1	0.00
21) Nitrobenzene-d5	9.521	128	90372	29.749	ng/u1	0.00
24) 2-Nitrophenol-d4	10.238	143	110723	33.217	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.761	165	215882	30.474	ng/u1	0.00
31) 4-Chloroaniline-d4	11.307	131	123924	13.008	ng/u1	0.00
46) Dimethylphthalate-d6	14.345	166	682781	29.335	ng/u1	0.00
49) Acenaphthylene-d8	14.650	160	753398	29.068	ng/u1	0.00
54) 4-Nitrophenol-d4	15.121	143	119420	30.888	ng/u1	0.00
60) Fluorene-d10	15.937	176	624048	30.212	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.043	200	107523	29.432	ng/u1	0.00
73) Anthracene-d10	17.800	188	953717	28.866	ng/u1	0.00
81) Pyrene-d10	20.068	212	1105794	29.768	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.344	264	1151380	30.386	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.675	88	36066	14.485	ng/uL	99
5) Pyridine	4.092	79	239822	36.710	ng/u1	97
6) Benzaldehyde	7.471	77	187585	37.033	ng/u1#	79
8) Phenol	7.465	94	382457	43.420	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.741	93	253986	39.453	ng/u1	96
12) 2-Chlorophenol	7.876	128	238586	43.450	ng/u1	98
13) 2-Methylphenol	8.740	108	270959	40.855	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.834	45	180520	23.352	ng/u1	97
16) Acetophenone	9.169	105	439706	41.445	ng/u1	95
17) N-Nitroso-di-n-propyla...	9.133	70	237084	39.865	ng/u1#	97
18) 4-Methylphenol	9.075	108	307582	42.465	ng/u1	94
19) Hexachloroethane	9.410	117	94912	41.242	ng/u1	86
22) Nitrobenzene	9.568	77	345351	37.974	ng/u1	96
23) Isophorone	10.073	82	720958	37.320	ng/u1	99
25) 2-Nitrophenol	10.273	139	141782	41.715	ng/u1#	94
26) 2,4-Dimethylphenol	10.297	107	254248	30.184	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.561	93	386156	37.296	ng/u1	95
29) 2,4-Dichlorophenol	10.790	162	259587	38.680	ng/u1	95
30) Naphthalene	11.225	128	833061	37.736	ng/u1	99
32) 4-Chloroaniline	11.337	127	266968	29.188	ng/u1	97
33) Hexachlorobutadiene	11.454	225	190683	36.314	ng/u1	98
34) Caprolactam	12.124	113	68751	28.866	ng/u1	88
35) 4-Chloro-3-methylphenol	12.400	107	323743	39.263	ng/u1	95
36) 2-Methylnaphthalene	12.800	142	603871	36.658	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091023\
 Data File : BG058987.D
 Acq On : 11 Sep 2023 17:32
 Operator : MA/JU
 Sample : 04234-03MS
 Misc :
 ALS Vial : 86 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 YBY14MS

Quant Time: Sep 12 00:51:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 00:07:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.017	142	601706	37.032	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	13.140	216	355381	36.283	ng/ul#	94
40) Hexachlorocyclopentadiene	13.099	237	201717	27.989	ng/ul	95
41) 2,4,6-Trichlorophenol	13.381	196	231790	37.277	ng/ul	95
42) 2,4,5-Trichlorophenol	13.446	196	256437	38.094	ng/ul	92
43) 1,1'-Biphenyl	13.787	154	857275	38.221	ng/ul	97
44) 2-Chloronaphthalene	13.840	162	658659	38.449	ng/ul	96
45) 2-Nitroaniline	14.057	65	215621	41.041	ng/ul	86
47) Dimethylphthalate	14.398	163	855794	38.887	ng/ul	99
48) 2,6-Dinitrotoluene	14.539	165	183455	43.823	ng/ul	95
50) Acenaphthylene	14.680	152	1057198	37.413	ng/ul	98
51) 3-Nitroaniline	14.874	138	174377	43.406	ng/ul#	85
52) Acenaphthene	15.015	153	740455	39.049	ng/ul	97
53) 2,4-Dinitrophenol	15.068	184	106466	39.457	ng/ul	92
55) 4-Nitrophenol	15.138	109	174715	38.302	ng/ul	96
56) Dibenzofuran	15.350	168	1056524	38.467	ng/ul	99
57) 2,4-Dinitrotoluene	15.309	165	250498	43.366	ng/ul#	83
58) 2,3,4,6-Tetrachlorophenol	15.549	232	238739	37.128	ng/ul#	94
59) Diethylphthalate	15.743	149	844527	38.925	ng/ul	95
61) Fluorene	15.996	166	866263	38.375	ng/ul	94
62) 4-Chlorophenyl-phenyle...	15.978	204	448093	37.514	ng/ul	94
63) 4-Nitroaniline	16.037	138	184459	48.448	ng/ul	86
66) 4,6-Dinitro-2-methylph...	16.061	198	156788	39.323	ng/ul#	96
67) N-Nitrosodiphenylamine	16.196	169	739213	37.681	ng/ul	99
68) 4-Bromophenyl-phenylether	16.877	248	294886	36.118	ng/ul	90
69) Hexachlorobenzene	16.965	284	336617	35.914	ng/ul	95
70) Atrazine	17.130	200	163683	21.321	ng/ul	96
71) Pentachlorophenol	17.312	266	202434	35.038	ng/ul	91
72) Phenanthrene	17.741	178	1416891	37.511	ng/ul	99
74) Anthracene	17.835	178	1404932	36.595	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.746	216	39605	3.688	ng/uL#	91
76) Pentachlorobenzene	15.238	250	392944	37.861	ng/uL	96
77) Carbazole	18.105	167	1274302	39.638	ng/ul	99
78) Di-n-butylphthalate	18.616	149	1413533	39.663	ng/ul	99
80) Fluoranthene	19.733	202	1680639	38.960	ng/ul#	97
82) Pyrene	20.097	202	1732609	38.635	ng/ul#	95
83) Butylbenzylphthalate	20.961	149	613653	43.302	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.913	252	539215	34.658	ng/ul	92
85) Benzo(a)anthracene	22.001	228	1709938	36.360	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.842	149	901345	42.468	ng/ul	100
87) Chrysene	22.077	228	1629303	37.268	ng/ul	98
89) Di-n-octyl phthalate	23.182	149	1538340	43.262	ng/ul	100
90) Benzo(b)fluoranthene	24.439	252	1815023	39.417	ng/ul	98
91) Benzo(k)fluoranthene	24.521	252	1807202	38.405	ng/ul	96
93) Benzo(a)pyrene	25.426	252	1604477	37.346	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	29.721	276	981286	18.077	ng/ul	98
95) Dibenzo(a,h)anthracene	29.821	278	1732391	38.514	ng/ul	97
96) Benzo(g,h,i)perylene	31.043	276	1746843	40.655	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 14.485 ng/uL
RT: 3.675 min Scan# 31
Delta R.T. -0.002 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion: 88 Resp: 36066
Ion Ratio Lower Upper
88 100
43 23.0 18.6 28.0
58 66.7 53.8 80.8



#5
Pyridine
Concen: 36.710 ng/uL
RT: 4.092 min Scan# 103
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion: 79 Resp: 239822
Ion Ratio Lower Upper
79 100
52 61.4 51.4 77.0
51 30.1 25.2 37.8









#6
Benzaldehyde
Concen: 37.033 ng/ul
RT: 7.471 min Scan# 61
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion: 77 Resp: 187585
Ion Ratio Lower Upper
77 100
105 85.2 81.6 122.4
106 80.6 84.5 126.7#



#8
Phenol
Concen: 43.420 ng/ul
RT: 7.465 min Scan# 677
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion: 94 Resp: 382457
Ion Ratio Lower Upper
94 100
65 28.0 22.4 33.6
66 41.7 35.0 52.6









#10
Bis(2-Chloroethyl)ether
Concen: 39.453 ng/ul
RT: 7.741 min Scan# 71
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion: 93 Resp: 253986
Ion Ratio Lower Upper
93 100
63 75.1 56.2 84.4
95 32.1 26.4 39.6



#12
2-Chlorophenol
Concen: 43.450 ng/ul
RT: 7.876 min Scan# 747
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:128 Resp: 238586
Ion Ratio Lower Upper
128 100
64 52.4 43.1 64.7
130 30.4 23.2 34.8





#13

2-Methylphenol

Concen: 40.855 ng/ul

RT: 8.740 min Scan# 894

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :





YBY14MS

Tgt Ion:108 Resp: 270959

Ion Ratio Lower Upper

108 100

107 91.8 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 23.352 ng/ul

RT: 8.834 min Scan# 910

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32




Tgt Ion: 45 Resp: 180520


Ion Ratio Lower Upper

45 100

77 17.4 13.2 19.8

79 13.7 9.8 14.6









#16
Acetophenone
Concen: 41.445 ng/ul
RT: 9.169 min Scan# 90
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:105 Resp: 439706
Ion Ratio Lower Upper
105 100
77 78.8 67.3 100.9
51 24.2 19.3 28.9



#17
N-Nitroso-di-n-propylamine
Concen: 39.865 ng/ul
RT: 9.133 min Scan# 961
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion: 70 Resp: 237084
Ion Ratio Lower Upper
70 100
42 48.1 37.5 56.3
101 8.2 5.0 7.6#
130 15.8 13.9 20.9









#18
4-Methylphenol
Concen: 42.465 ng/ul
RT: 9.075 min Scan# 91
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:108 Resp: 307582
Ion Ratio Lower Upper
108 100
107 105.2 88.9 133.3



#19
Hexachloroethane
Concen: 41.242 ng/ul
RT: 9.410 min Scan# 1008
Delta R.T. -0.007 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:117 Resp: 94912
Ion Ratio Lower Upper
117 100
201 110.8 108.2 162.4
199 72.2 56.0 84.0









#22
Nitrobenzene
Concen: 37.974 ng/ul
RT: 9.568 min Scan# 1035
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion: 77 Resp: 345351
Ion Ratio Lower Upper
77 100
123 38.6 28.5 42.7
65 14.5 11.5 17.3



#23
Isophorone
Concen: 37.320 ng/ul
RT: 10.073 min Scan# 1121
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion: 82 Resp: 720958
Ion Ratio Lower Upper
82 100
95 7.7 6.6 9.8
138 13.1 10.4 15.6









#25
2-Nitrophenol
Concen: 41.715 ng/ul
RT: 10.273 min Scan# 1155
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:139 Resp: 141782
Ion Ratio Lower Upper
139 100
65 68.8 53.5 80.3
109 48.2 32.0 48.0#



#26
2,4-Dimethylphenol
Concen: 30.184 ng/ul
RT: 10.297 min Scan# 1159
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:107 Resp: 254248
Ion Ratio Lower Upper
107 100
121 43.4 36.4 54.6
122 82.6 69.4 104.0





#27

Bis(2-Chloroethoxy)methane

Concen: 37.296 ng/ul

RT: 10.561 min Scan# 1204

Delta R.T. 0.004 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :

YBY14MS





Tgt Ion: 93 Resp: 386156

Ion Ratio Lower Upper

93 100

95 30.8 27.8 41.6

123 9.0 7.3 10.9



#29

2,4-Dichlorophenol

Concen: 38.680 ng/ul

RT: 10.790 min Scan# 1243

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32




Tgt Ion:162 Resp: 259587


Ion Ratio Lower Upper

162 100

164 67.4 51.4 77.0

98 37.8 27.2 40.8









#30
Naphthalene
Concen: 37.736 ng/ul
RT: 11.225 min Scan# 11
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:128 Resp: 833061
Ion Ratio Lower Upper
128 100
129 10.4 8.2 12.2
127 12.9 10.9 16.3



#32
4-Chloroaniline
Concen: 29.188 ng/ul
RT: 11.337 min Scan# 1336
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:127 Resp: 266968
Ion Ratio Lower Upper
127 100
129 34.7 26.6 39.8









#33
Hexachlorobutadiene
Concen: 36.314 ng/ul
RT: 11.454 min Scan# 11
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:225 Resp: 190683
Ion Ratio Lower Upper
225 100
223 57.4 45.5 68.3
227 60.9 50.2 75.4



#34
Caprolactam
Concen: 28.866 ng/ul
RT: 12.124 min Scan# 1470
Delta R.T. 0.016 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:113 Resp: 68751
Ion Ratio Lower Upper
113 100
55 165.2 140.5 210.7
56 133.5 90.4 135.6









#35
4-Chloro-3-methylphenol
Concen: 39.263 ng/ul
RT: 12.400 min Scan# 1516
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:107 Resp: 323743
Ion Ratio Lower Upper
107 100
144 21.6 20.2 30.2
142 68.2 51.6 77.4



#36
2-Methylnaphthalene
Concen: 36.658 ng/ul
RT: 12.800 min Scan# 1585
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:142 Resp: 603871
Ion Ratio Lower Upper
142 100
141 83.3 65.7 98.5





#37

1-Methylnaphthalene

Concen: 37.032 ng/ul

RT: 13.017 min Scan# 1

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :

YBY14MS





Tgt Ion:142 Resp: 601706

Ion Ratio Lower Upper

142 100

141 89.6 73.1 109.7

116 4.6 3.4 5.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 36.283 ng/ul

RT: 13.140 min Scan# 1643

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Tgt Ion:216 Resp: 355381




Ion Ratio Lower Upper


216 100

214 79.9 59.6 89.4

179 21.4 13.6 20.4#

108 18.2 13.8 20.6









#40
Hexachlorocyclopentadiene
Concen: 27.989 ng/ul
RT: 13.099 min Scan# 1
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:237 Resp: 201717
Ion Ratio Lower Upper
237 100
235 61.3 52.6 78.8
272 11.5 8.6 13.0



#41
2,4,6-Trichlorophenol
Concen: 37.277 ng/ul
RT: 13.381 min Scan# 1684
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:196 Resp: 231790
Ion Ratio Lower Upper
196 100
198 101.7 78.3 117.5
200 34.2 23.7 35.5





#42

2,4,5-Trichlorophenol

Concen: 38.094 ng/ul

RT: 13.446 min Scan# 1

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :

YBY14MS





Tgt Ion:196 Resp: 256437

Ion Ratio Lower Upper

196 100

198 96.9 84.8 127.2

200 30.2 27.7 41.5



#43

1,1'-Biphenyl

Concen: 38.221 ng/ul

RT: 13.787 min Scan# 1753

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32




Tgt Ion:154 Resp: 857275


Ion Ratio Lower Upper

154 100

153 38.5 31.6 47.4

76 13.3 9.0 13.4









#44
2-Chloronaphthalene
Concen: 38.449 ng/ul
RT: 13.840 min Scan# 1762
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:162 Resp: 658659
Ion Ratio Lower Upper
162 100
164 32.0 27.6 41.4
127 38.5 29.4 44.0



#45
2-Nitroaniline
Concen: 41.041 ng/ul
RT: 14.057 min Scan# 1799
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion: 65 Resp: 215621
Ion Ratio Lower Upper
65 100
92 75.4 54.0 81.0
138 95.9 64.8 97.2









#47
Dimethylphthalate
Concen: 38.887 ng/ul
RT: 14.398 min Scan# 1857
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:163 Resp: 855794
Ion Ratio Lower Upper
163 100
194 3.8 2.8 4.2
164 10.5 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 43.823 ng/ul
RT: 14.539 min Scan# 1881
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:165 Resp: 183455
Ion Ratio Lower Upper
165 100
89 65.5 55.8 83.6
121 23.5 16.3 24.5









#50
Acenaphthylene
Concen: 37.413 ng/ul
RT: 14.680 min Scan# 1905
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:152 Resp: 1057198
Ion Ratio Lower Upper
152 100
151 19.1 14.6 21.8
153 13.4 10.4 15.6



#51
3-Nitroaniline
Concen: 43.406 ng/ul
RT: 14.874 min Scan# 1938
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:138 Resp: 174377
Ion Ratio Lower Upper
138 100
108 10.7 11.8 17.6#
92 133.0 122.2 183.2









#52
Acenaphthene
Concen: 39.049 ng/ul
RT: 15.015 min Scan# 1962
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:153 Resp: 740455
Ion Ratio Lower Upper
153 100
152 45.8 36.5 54.7
154 88.1 74.1 111.1



#53
2,4-Dinitrophenol
Concen: 39.457 ng/ul
RT: 15.068 min Scan# 1971
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:184 Resp: 106466
Ion Ratio Lower Upper
184 100
63 64.2 57.7 86.5
154 70.5 60.7 91.1









#55
4-Nitrophenol
Concen: 38.302 ng/ul
RT: 15.138 min Scan# 1981
Delta R.T. 0.010 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:109 Resp: 174715
Ion Ratio Lower Upper
109 100
139 92.6 74.3 111.5
65 108.6 93.4 140.0



#56
Dibenzofuran
Concen: 38.467 ng/ul
RT: 15.350 min Scan# 2019
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:168 Resp: 1056524
Ion Ratio Lower Upper
168 100
139 33.7 27.4 41.2





#57

2,4-Dinitrotoluene

Concen: 43.366 ng/ul

RT: 15.309 min Scan# 2012

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :

YBY14MS





Tgt Ion:165 Resp: 250498

Ion Ratio Lower Upper

165 100

63 52.5 33.3 49.9#

182 2.5 1.4 2.0#



#58

2,3,4,6-Tetrachlorophenol

Concen: 37.128 ng/ul

RT: 15.549 min Scan# 2053

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Tgt Ion:232 Resp: 238739




Ion Ratio Lower Upper


232 100

131 36.8 27.1 40.7

130 2.3 0.7 1.1#

166 27.7 19.5 29.3





#59

Diethylphthalate

Concen: 38.925 ng/ul

RT: 15.743 min Scan# 2086

Delta R.T. 0.004 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :

YBY14MS





Tgt Ion:149 Resp: 844527

Ion Ratio Lower Upper

149 100

177 24.3 16.9 25.3

150 11.3 9.7 14.5



#61

Fluorene

Concen: 38.375 ng/ul

RT: 15.996 min Scan# 2129

Delta R.T. 0.004 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32




Tgt Ion:166 Resp: 866263


Ion Ratio Lower Upper

166 100

165 101.2 76.4 114.6

167 13.5 9.3 13.9









#62
4-Chlorophenyl-phenylether
Concen: 37.514 ng/ul
RT: 15.978 min Scan# 2126
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:204 Resp: 448093
Ion Ratio Lower Upper
204 100
206 29.9 27.7 41.5
141 50.3 43.2 64.8



#63
4-Nitroaniline
Concen: 48.448 ng/ul
RT: 16.037 min Scan# 2136
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:138 Resp: 184459
Ion Ratio Lower Upper
138 100
92 59.2 49.9 74.9
108 128.8 121.4 182.2









#66
4,6-Dinitro-2-methylphenol
Concen: 39.323 ng/ul
RT: 16.061 min Scan# 2139
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:198 Resp: 156788
Ion Ratio Lower Upper
198 100
182 2.4 3.0 4.6#
77 26.8 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 37.681 ng/ul
RT: 16.196 min Scan# 2163
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:169 Resp: 739213
Ion Ratio Lower Upper
169 100
168 66.7 53.0 79.6
167 38.3 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 36.118 ng/ul
RT: 16.877 min Scan# 21
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:248 Resp: 294886
Ion Ratio Lower Upper
248 100
250 99.8 73.9 110.9
141 66.2 45.4 68.0



#69
Hexachlorobenzene
Concen: 35.914 ng/ul
RT: 16.965 min Scan# 2294
Delta R.T. 0.004 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:284 Resp: 336617
Ion Ratio Lower Upper
284 100
142 26.6 24.6 37.0
249 31.6 24.3 36.5









#70
 Atrazine
 Concen: 21.321 ng/ul
 RT: 17.130 min Scan# 21
 Delta R.T. -0.001 min
 Lab File: BG058987.D
 Acq: 11 Sep 2023 17:32




Instrument :
 BNA_G
 ClientSampleId :
 YBY14MS


Tgt Ion:200 Resp: 163683
 Ion Ratio Lower Upper
 200 100
 173 22.7 17.4 26.2
 215 47.1 35.4 53.0



#71
 Pentachlorophenol
 Concen: 35.038 ng/ul
 RT: 17.312 min Scan# 2353
 Delta R.T. -0.001 min
 Lab File: BG058987.D
 Acq: 11 Sep 2023 17:32

Tgt Ion:266 Resp: 202434
 Ion Ratio Lower Upper
 266 100
 264 60.2 53.1 79.7
 268 55.8 51.6 77.4









#72
Phenanthrene
Concen: 37.511 ng/ul
RT: 17.741 min Scan# 2426
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:178 Resp: 1416891
Ion Ratio Lower Upper
178 100
179 15.3 11.8 17.8
176 19.1 15.1 22.7



#74
Anthracene
Concen: 36.595 ng/ul
RT: 17.835 min Scan# 2442
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:178 Resp: 1404932
Ion Ratio Lower Upper
178 100
179 14.9 12.2 18.4
176 18.7 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 3.688 ng/uL
RT: 13.746 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:216 Resp: 39605
Ion Ratio Lower Upper
216 100
214 74.2 60.6 91.0
218 56.1 33.6 50.4#



#76
Pentachlorobenzene
Concen: 37.861 ng/uL
RT: 15.238 min Scan# 2000
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:250 Resp: 392944
Ion Ratio Lower Upper
250 100
248 62.7 50.3 75.5
252 65.6 47.6 71.4









#77
Carbazole
Concen: 39.638 ng/ul
RT: 18.105 min Scan# 2488
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:167 Resp: 1274302
Ion Ratio Lower Upper
167 100
166 21.1 16.8 25.2
139 12.6 9.4 14.2



#78
Di-n-butylphthalate
Concen: 39.663 ng/ul
RT: 18.616 min Scan# 2575
Delta R.T. -0.007 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:149 Resp: 1413533
Ion Ratio Lower Upper
149 100
150 9.3 7.7 11.5
104 6.5 4.6 6.8









#80
Fluoranthene
Concen: 38.960 ng/ul
RT: 19.733 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32




Instrument :
BNA_G
ClientSampleId :
YBY14MS


Tgt Ion:202 Resp: 1680639
Ion Ratio Lower Upper
202 100
101 9.8 6.1 9.1#
100 7.4 5.9 8.9



#82
Pyrene
Concen: 38.635 ng/ul
RT: 20.097 min Scan# 2827
Delta R.T. -0.001 min
Lab File: BG058987.D
Acq: 11 Sep 2023 17:32

Tgt Ion:202 Resp: 1732609
Ion Ratio Lower Upper
202 100
101 11.2 7.5 11.3
100 9.6 6.3 9.5#





#83

Butylbenzylphthalate

Concen: 43.302 ng/ul

RT: 20.961 min Scan# 2974

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :

YBY14MS





Tgt Ion:149 Resp: 613653

Ion Ratio Lower Upper

149 100

91 86.8 68.7 103.1

206 21.1 17.3 25.9



#84

3,3'-Dichlorobenzidine

Concen: 34.658 ng/ul

RT: 21.913 min Scan# 3136

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32




Tgt Ion:252 Resp: 539215


Ion Ratio Lower Upper

252 100

254 64.5 57.6 86.4

126 8.2 7.8 11.8





#85

Benzo(a)anthracene

Concen: 36.360 ng/ul

RT: 22.001 min Scan# 3151

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :

YBY14MS





Tgt Ion:228 Resp: 1709938

Ion Ratio Lower Upper

228 100

229 20.1 16.0 24.0

226 26.4 20.8 31.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 42.468 ng/ul

RT: 21.842 min Scan# 3124

Delta R.T. -0.007 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32




Tgt Ion:149 Resp: 901345


Ion Ratio Lower Upper

149 100

167 26.4 21.0 31.6

279 4.3 3.8 5.8





#87

Chrysene

Concen: 37.268 ng/ul

RT: 22.077 min Scan# 3163

Delta R.T. 0.004 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :

YBY14MS





Tgt Ion:228 Resp: 1629303

Ion Ratio Lower Upper

228 100

226 29.5 23.2 34.8

229 20.7 15.3 22.9



#89

Di-n-octyl phthalate

Concen: 43.262 ng/ul




RT: 23.182 min Scan# 3352


Delta R.T. -0.007 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Tgt Ion:149 Resp: 1538340





#90

Benzo(b)fluoranthene

Concen: 39.417 ng/ul

RT: 24.439 min Scan# 3565

Delta R.T. 0.004 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :

YBY14MS





Tgt Ion:252 Resp: 1815023

Ion Ratio Lower Upper

252 100

253 21.3 17.8 26.8

125 7.5 6.2 9.2



#91

Benzo(k)fluoranthene

Concen: 38.405 ng/ul

RT: 24.521 min Scan# 3580

Delta R.T. 0.010 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32




Tgt Ion:252 Resp: 1807202


Ion Ratio Lower Upper

252 100

253 21.4 19.1 28.7

125 6.8 5.0 7.4





#93

Benzo(a)pyrene

Concen: 37.346 ng/ul

RT: 25.426 min Scan# 31

Delta R.T. -0.001 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32

Instrument :

BNA_G

ClientSampleId :

YBY14MS





Tgt Ion:252 Resp: 1604477

Ion Ratio Lower Upper

252 100

253 20.9 17.1 25.7

125 8.3 5.3 7.9#



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.077 ng/ul

RT: 29.721 min Scan# 4465

Delta R.T. 0.004 min

Lab File: BG058987.D

Acq: 11 Sep 2023 17:32




Tgt Ion:276 Resp: 981286


Ion Ratio Lower Upper

276 100

138 13.4 11.8 17.8

277 25.3 20.7 31.1









#95
 Dibenzo(a,h)anthracene
 Concen: 38.514 ng/ul
 RT: 29.821 min Scan# 4482
 Delta R.T. 0.010 min
 Lab File: BG058987.D
 Acq: 11 Sep 2023 17:32

Instrument :
 BNA_G
 ClientSampleId :
 YBY14MS

Tgt Ion:278 Resp: 1732391
 Ion Ratio Lower Upper
 278 100
 139 10.6 9.8 14.8
 279 24.4 18.5 27.7



#96
 Benzo(g,h,i)perylene
 Concen: 40.655 ng/ul
 RT: 31.043 min Scan# 4690
 Delta R.T. 0.004 min
 Lab File: BG058987.D
 Acq: 11 Sep 2023 17:32

Tgt Ion:276 Resp: 1746843
 Ion Ratio Lower Upper
 276 100
 138 16.0 11.4 17.2
 277 23.8 17.4 26.2

