



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091023\
Data File : BG058953.D
Acq On : 10 Sep 2023 14:44
Operator : MA/JU
Sample : DFTPP527
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP527

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091023.MA.M
Title : SVOA CALIBRATION
Last Update : Mon Sep 11 00:07:05 2023



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2473

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.3	44507	PASS
68	69	0.00	2	0.7	339	PASS
69	198	0.00	100	45.1	47536	PASS
70	69	0.00	2	0.3	136	PASS
127	198	10	80	39.6	41701	PASS
197	198	0.00	2	0.1	122	PASS
198	198	100	100	100.0	105307	PASS
199	198	5	9	8.2	8618	PASS
275	198	10	60	31.4	33024	PASS
365	198	1	100	5.2	5477	PASS
441	198	0.01	100	15.1	15912	PASS
442	198	50	100	93.5	98435	PASS
443	442	15	24	20.6	20258	PASS

m/z	Abundance
44.00	333.0
44.90	234.0
61.70	187.0
71.80	158.0
123.30	322.0
134.50	210.0
161.30	177.0
165.00	298.0
189.00	222.0
207.10	545.0
225.20	380.0
235.90	296.0
238.10	253.0
252.80	176.0
303.60	165.0
344.00	249.0
377.90	171.0
397.20	210.0
420.90	168.0
438.00	186.0
491.90	203.0
512.90	181.0
523.30	326.0

Instrument :
BNA_G
ClientSampleId :
DFTPP527

m/z	Abundance
36.70	153.0
38.00	575.0
39.10	1744.0
40.10	234.0
43.90	285.0
45.10	229.0
49.00	476.0
50.10	13291.0
51.10	36784.0
52.10	2117.0
53.20	326.0
55.10	212.0
56.20	1226.0
56.90	2784.0
60.90	464.0
62.30	685.0
63.10	1992.0
63.90	421.0
65.00	637.0
69.00	44384.0
70.20	198.0
73.00	517.0
74.00	4879.0
75.00	6789.0
76.00	1648.0
77.10	48536.0
78.00	2771.0
79.00	3048.0
80.10	2740.0
81.10	3042.0
81.90	609.0
83.00	522.0
85.10	440.0
86.00	1286.0
87.10	240.0
88.10	155.0
88.90	372.0
90.90	234.0
92.00	997.0
93.00	6616.0
94.00	339.0
96.20	329.0
98.10	5691.0
99.10	3202.0
100.10	491.0
101.00	2134.0
102.80	496.0
104.10	1584.0
105.10	989.0
107.10	13803.0
108.00	2300.0
110.00	21464.0
111.00	2440.0
111.90	527.0
113.10	180.0
116.10	596.0
117.00	14050.0
118.10	875.0
118.90	269.0
120.10	369.0
120.60	176.0
121.90	1096.0
123.00	1563.0
123.80	458.0
124.20	287.0
125.10	471.0
127.10	34088.0
128.10	3180.0
129.00	13966.0
130.00	2252.0
132.10	194.0
134.00	484.0
135.10	1025.0
136.10	809.0
137.00	838.0
138.00	408.0
139.30	151.0
140.20	247.0
141.00	2347.0
142.00	1007.0
142.90	422.0
145.90	316.0
147.00	1739.0
148.00	3053.0
149.10	690.0
150.10	237.0
153.20	440.0
153.90	608.0
155.10	1075.0
156.10	1546.0
157.00	735.0
158.00	600.0

Instrument :
BNA_G
ClientSampleId :
DFTPP527

159.10	789.0
160.10	847.0
161.00	1503.0
162.00	559.0
163.00	274.0
165.00	469.0
165.90	740.0
167.00	5463.0
168.00	3067.0
168.90	405.0
171.90	196.0
172.80	215.0
173.20	312.0
173.90	883.0
175.00	1777.0
176.00	1723.0
177.00	869.0
178.10	618.0
179.00	4366.0
180.10	2486.0
181.10	873.0
181.70	216.0
184.00	487.0
185.10	2020.0
186.10	13440.0
187.10	2842.0
187.90	287.0
189.10	1257.0
190.00	226.0
190.90	451.0
192.10	1381.0
193.10	847.0
194.00	438.0
194.80	246.0
196.00	2685.0
197.00	365.0
198.00	81368.0
199.00	7097.0
200.00	708.0
200.70	169.0
201.40	165.0
203.10	776.0
204.10	3114.0
205.10	7274.0
206.10	24936.0
207.10	2983.0
208.00	1006.0
209.10	312.0
210.10	448.0
210.90	1042.0
215.20	171.0
215.90	558.0
217.00	7767.0
218.10	1325.0
219.00	183.0
221.00	2849.0
221.90	990.0
223.00	1472.0
224.10	16720.0
225.00	3724.0
227.00	7698.0
228.10	1092.0
229.00	1793.0
230.90	272.0
233.10	240.0
234.00	1058.0
235.00	840.0
235.90	207.0
237.10	841.0
239.00	380.0
239.80	272.0
241.20	322.0
242.00	958.0
242.80	804.0
244.10	11935.0
245.00	1614.0
246.10	2594.0
247.00	732.0
247.50	313.0
249.00	703.0
250.50	299.0
252.00	224.0
253.70	231.0
255.00	57976.0
256.00	10186.0
257.00	1444.0
258.00	4352.0
259.10	1165.0
265.00	1695.0
271.20	336.0
273.10	1507.0
274.10	4811.0
275.00	23144.0
276.00	3331.0
277.10	1849.0
278.00	1089.0

Instrument :
BNA_G
ClientSampleId :
DFTPP527

278.90	211.0
282.60	312.0
284.10	309.0
285.00	314.0
286.20	375.0
290.10	175.0
293.10	320.0
295.00	203.0
296.00	7320.0
297.00	1974.0
303.20	469.0
304.10	219.0
308.90	223.0
314.10	270.0
315.00	683.0
315.90	300.0
317.00	278.0
322.20	302.0
323.10	2941.0
324.00	553.0
326.80	470.0
328.10	180.0
332.10	546.0
332.90	336.0
334.20	1389.0
335.00	288.0
340.90	228.0
342.00	151.0
346.10	551.0
347.10	163.0
351.90	808.0
353.00	242.0
354.00	430.0
354.90	188.0
365.00	3711.0
367.40	219.0
372.00	808.0
372.70	223.0
373.70	588.0
383.10	663.0
389.80	169.0
402.00	799.0
403.00	888.0
404.00	172.0
405.20	264.0
407.80	157.0
420.90	444.0
421.80	374.0
423.00	4154.0
424.00	754.0
424.80	333.0
441.00	9689.0
442.00	62080.0
443.00	12014.0
444.10	1556.0

Instrument :
BNA_G
ClientSampleId :
DFTPP527

m/z	Abundance
37.10	205.0
38.00	562.0
39.10	4227.0
41.10	484.0
44.10	343.0
45.10	179.0
46.90	151.0
49.10	618.0
50.10	15316.0
51.10	56656.0
52.10	2975.0
55.00	378.0
56.00	1209.0
57.00	2962.0
60.70	370.0
61.10	436.0
62.10	949.0
63.00	2476.0
64.10	442.0
65.00	1183.0
67.10	261.0
68.00	1017.0
69.00	55464.0
73.10	803.0
74.00	5886.0
75.10	10626.0
76.10	3615.0
77.10	68776.0
78.10	3611.0
79.10	4527.0
80.00	3289.0
81.10	4473.0
82.00	1750.0
83.00	1098.0
85.00	1127.0
86.10	1510.0
87.30	539.0
91.00	642.0
92.10	1577.0
93.00	9738.0
94.10	517.0
96.10	330.0
96.90	198.0
98.10	7080.0
99.10	4793.0
99.90	229.0
101.10	3090.0
101.90	197.0
103.00	1333.0
104.00	2638.0
105.00	1656.0
107.00	20424.0
108.10	2888.0
110.00	30432.0
111.10	5327.0
112.10	575.0
115.20	208.0
116.00	1127.0
117.00	21336.0
118.00	1322.0
120.10	233.0
121.90	1137.0
123.00	1609.0
124.10	1210.0
125.10	545.0
127.00	49608.0
128.10	5220.0
129.00	22976.0
130.00	1790.0
130.90	457.0
131.80	182.0
132.80	206.0
134.00	665.0
135.10	2520.0
136.00	1235.0
137.10	1609.0
138.00	472.0
139.90	474.0
141.00	3880.0
142.00	789.0
143.00	852.0
144.10	212.0
145.10	212.0
146.00	352.0
147.00	2500.0
148.10	3990.0
149.10	851.0
150.90	431.0
152.00	350.0
152.90	974.0
154.00	517.0
155.10	2218.0

Instrument :
BNA_G
ClientSampleId :
DFTPP527

156.00	2625.0
157.10	593.0
158.00	254.0
159.00	386.0
159.80	1167.0
161.10	1427.0
161.90	250.0
163.10	291.0
164.00	233.0
165.00	1348.0
166.00	849.0
167.00	6914.0
168.00	2716.0
168.90	187.0
169.20	240.0
170.00	303.0
170.90	459.0
171.90	758.0
173.00	1099.0
174.00	1246.0
175.10	3094.0
176.00	614.0
177.10	2439.0
178.10	673.0
179.00	6234.0
180.10	2553.0
180.90	1913.0
182.00	264.0
182.90	191.0
184.10	671.0
185.00	3125.0
186.10	19112.0
187.00	5451.0
187.90	691.0
189.00	1940.0
190.00	306.0
190.90	765.0
192.10	1674.0
193.10	1449.0
194.20	527.0
194.90	351.0
196.10	3557.0
198.00	125872.0
199.00	9827.0
200.00	1142.0
200.70	200.0
201.60	466.0
203.00	1731.0
204.10	5767.0
205.00	9516.0
206.10	35056.0
207.00	4277.0
208.00	2105.0
209.00	951.0
210.20	1159.0
211.10	2006.0
213.20	233.0
215.00	462.0
216.00	1061.0
217.00	10712.0
217.80	1628.0
219.10	305.0
220.00	567.0
221.00	4425.0
221.90	1473.0
223.00	2723.0
224.00	23632.0
225.10	7029.0
226.00	536.0
227.00	12250.0
228.00	915.0
229.10	2773.0
229.80	201.0
231.10	759.0
233.00	214.0
233.90	583.0
234.90	607.0
236.20	752.0
236.80	677.0
238.00	199.0
239.00	548.0
239.90	301.0
241.00	537.0
242.10	1711.0
243.10	1831.0
244.10	17448.0
244.90	1982.0
245.90	5479.0
247.00	973.0
248.20	361.0
249.10	590.0
250.00	268.0
252.10	416.0
252.90	687.0
255.00	98312.0
256.00	15939.0

Instrument :
BNA_G
ClientSampleId :
DFTPP527

257.10	742.0
258.00	5610.0
258.90	890.0
265.00	3032.0
265.90	586.0
267.10	154.0
270.00	219.0
271.00	230.0
271.90	330.0
273.00	3233.0
274.00	8373.0
275.00	40472.0
276.10	4960.0
277.00	3944.0
278.00	1007.0
280.90	155.0
283.00	508.0
284.40	230.0
285.00	774.0
285.60	281.0
286.20	317.0
289.10	228.0
290.10	190.0
292.20	198.0
293.00	804.0
295.10	348.0
296.00	11490.0
297.00	1524.0
298.10	236.0
301.80	183.0
303.00	1240.0
303.90	331.0
304.20	533.0
307.80	336.0
308.80	312.0
310.40	203.0
313.10	403.0
314.00	802.0
315.00	1614.0
316.00	453.0
317.00	251.0
321.00	233.0
322.20	310.0
323.00	4684.0
324.00	399.0
326.00	235.0
327.00	477.0
327.90	452.0
332.00	415.0
333.00	562.0
333.90	2168.0
335.20	1135.0
340.90	557.0
346.10	714.0
347.00	199.0
352.10	1136.0
352.90	550.0
353.90	1428.0
355.10	304.0
356.50	242.0
365.00	6107.0
366.00	1132.0
368.90	195.0
371.20	479.0
371.90	1453.0
373.00	521.0
382.90	780.0
383.90	469.0
390.00	258.0
391.00	659.0
400.60	292.0
402.00	788.0
403.00	1114.0
403.90	703.0
404.90	580.0
415.00	231.0
420.30	283.0
420.90	980.0
422.00	907.0
423.00	5945.0
424.00	1542.0
425.30	203.0
434.50	154.0
435.30	310.0
441.00	17112.0
442.00	112896.0
443.00	24448.0
443.90	2712.0
445.10	252.0
462.50	152.0
490.20	163.0
511.20	248.0

Instrument :
BNA_G
ClientSampleId :
DFTPP527

m/z	Abundance
38.10	355.0
39.10	2115.0
49.00	452.0
50.10	11907.0
51.10	40080.0
52.10	1731.0
56.10	924.0
57.00	2481.0
61.00	624.0
62.20	842.0
63.10	1301.0
64.10	167.0
65.10	1229.0
66.00	182.0
69.00	42760.0
70.20	211.0
73.20	416.0
74.10	6166.0
75.00	8805.0
76.10	3087.0
77.10	54488.0
78.20	3056.0
79.00	4313.0
80.00	2221.0
81.10	4116.0
82.20	1057.0
82.90	388.0
84.10	186.0
84.90	490.0
86.00	901.0
87.00	394.0
88.70	283.0
90.90	498.0
93.00	7514.0
94.00	515.0
95.10	505.0
96.10	252.0
98.10	5557.0
99.00	3952.0
100.00	160.0
101.00	2612.0
103.00	967.0
104.10	2218.0
105.00	1693.0
107.10	16330.0
108.10	2953.0
109.00	650.0
110.00	27648.0
111.00	3696.0
112.10	324.0
113.10	813.0
114.50	438.0
116.20	1143.0
117.10	17856.0
118.00	1214.0
120.10	169.0
121.90	573.0
122.20	585.0
123.00	1373.0
124.00	1195.0
125.10	771.0
127.10	41408.0
128.00	2724.0
129.00	19176.0
130.00	1961.0
131.10	554.0
132.10	245.0
134.00	691.0
135.00	1492.0
136.10	387.0
137.10	608.0
140.20	325.0
141.00	2536.0
142.10	955.0
143.00	560.0
145.90	268.0
147.00	1794.0
148.00	3967.0
149.10	561.0
150.10	326.0
150.90	440.0
152.30	222.0
153.10	502.0
153.80	425.0
154.00	438.0
155.10	1575.0
156.10	2085.0
157.10	243.0
157.70	243.0
159.00	247.0
160.00	1401.0
161.00	1530.0

Instrument :

BNA_G

ClientSampleId :

DFTPP527

162.10	201.0
165.00	1175.0
166.10	760.0
167.00	7825.0
168.10	2676.0
169.10	548.0
170.70	252.0
171.90	594.0
173.00	618.0
173.90	1010.0
175.10	2426.0
176.00	936.0
177.00	1447.0
177.80	190.0
179.00	4714.0
180.10	3008.0
181.00	2096.0
182.20	242.0
183.30	210.0
184.00	446.0
185.10	2355.0
186.10	16146.0
187.00	4689.0
188.00	728.0
189.00	1385.0
189.90	390.0
190.60	153.0
191.30	215.0
192.00	1998.0
193.00	1829.0
193.90	436.0
194.10	386.0
196.00	4878.0
198.00	108680.0
199.00	8929.0
200.10	1101.0
200.90	211.0
201.10	227.0
201.70	906.0
203.20	984.0
204.00	4613.0
205.10	9218.0
206.10	34576.0
207.10	4127.0
208.00	1349.0
209.00	656.0
210.10	655.0
211.00	1344.0
211.80	240.0
215.00	705.0
216.00	925.0
217.00	11301.0
218.10	1340.0
219.90	234.0
221.10	3164.0
223.00	2075.0
224.00	21432.0
225.10	5545.0
227.00	9952.0
227.90	1708.0
229.00	2110.0
230.10	376.0
231.10	685.0
233.10	671.0
234.10	772.0
234.90	757.0
235.80	527.0
237.00	576.0
238.00	216.0
239.30	426.0
240.00	333.0
241.00	535.0
242.00	1280.0
242.90	1604.0
244.10	15990.0
245.00	2174.0
246.00	3034.0
247.10	1147.0
248.20	252.0
248.90	414.0
249.20	460.0
250.90	624.0
251.80	212.0
253.00	469.0
255.00	88064.0
256.00	12601.0
257.00	1000.0
258.00	7218.0
259.00	926.0
263.10	197.0
264.00	276.0
265.10	1831.0
265.90	296.0
272.10	167.0
273.00	2780.0
274.00	8191.0

Instrument :
BNA_G
ClientSampleId :
DFTPP527

275.00	35456.0
276.00	4223.0
276.90	3087.0
277.80	320.0
278.10	345.0
279.20	236.0
282.70	329.0
284.00	362.0
285.10	459.0
285.90	289.0
289.10	249.0
289.90	204.0
292.90	870.0
296.00	14671.0
297.00	1654.0
301.10	169.0
302.10	221.0
303.10	1431.0
303.90	351.0
306.20	153.0
308.20	184.0
310.10	259.0
311.90	212.0
314.20	618.0
314.90	1514.0
316.00	1194.0
316.80	395.0
321.00	756.0
323.00	3031.0
324.10	822.0
325.00	209.0
326.90	862.0
327.90	547.0
329.10	173.0
332.00	654.0
332.90	334.0
334.10	2426.0
335.10	503.0
341.20	423.0
342.10	221.0
343.80	277.0
346.00	1402.0
351.90	1443.0
353.10	777.0
353.90	1509.0
355.10	341.0
365.00	6613.0
366.00	1286.0
371.20	275.0
372.00	2097.0
373.30	346.0
379.20	270.0
381.60	157.0
382.80	572.0
383.30	175.0
384.90	184.0
389.90	167.0
390.90	276.0
401.80	798.0
403.00	1351.0
404.00	508.0
415.00	167.0
421.10	1190.0
422.00	855.0
423.00	5745.0
424.00	941.0
424.90	529.0
441.00	20936.0
442.00	120328.0
443.00	24312.0
444.10	2732.0
444.90	275.0
515.00	187.0
520.90	151.0
526.00	180.0

Instrument :
BNA_G
ClientSampleId :
DFTPP527