

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091323\
 Data File : BG059049.D
 Acq On : 13 Sep 2023 21:29
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020560

Quant Time: Sep 14 01:07:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 12 23:54:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.307	152	96234	20.000	ng/ul	0.00
20) Naphthalene-d8	11.145	136	480957	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.935	164	370403	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.684	188	980413	20.000	ng/ul	# 0.00
79) Chrysene-d12	22.009	240	869902	20.000	ng/ul	0.00
88) Perylene-d12	25.563	264	1014937	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.671	96	369	0.163	ng/uL	0.04
4) Pyridine-d5	4.041	84	140465	19.596	ng/ul	-0.02
7) Phenol-d5	7.414	99	198899	20.496	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.625	67	115293	20.261	ng/ul	0.00
11) 2-Chlorophenol-d4	7.813	132	140895	22.638	ng/ul	-0.01
15) 4-Methylphenol-d8	8.983	113	175274	23.378	ng/ul	-0.01
21) Nitrobenzene-d5	9.500	128	71455	20.482	ng/ul	-0.01
24) 2-Nitrophenol-d4	10.222	143	95558	24.963	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.745	165	172368	21.187	ng/ul	0.00
31) 4-Chloroaniline-d4	11.292	131	242831	22.195	ng/ul	0.00
46) Dimethylphthalate-d6	14.323	166	639207	22.247	ng/ul	-0.01
49) Acenaphthylene-d8	14.635	160	693356	21.671	ng/ul	0.00
54) 4-Nitrophenol-d4	15.105	143	115080	24.113	ng/ul	0.00
60) Fluorene-d10	15.922	176	572658	22.459	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.027	200	123272	25.615	ng/ul	0.00
73) Anthracene-d10	17.784	188	942812	21.662	ng/ul	0.00
81) Pyrene-d10	20.052	212	1065333	23.119	ng/ul	-0.01
92) Benzo(a)pyrene-d12	25.311	264	1030185	20.706	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.648	88	21830	7.483	ng/uL#	96
5) Pyridine	4.065	79	151821	19.836	ng/ul	93
6) Benzaldehyde	7.449	77	135213	22.784	ng/ul#	85
8) Phenol	7.443	94	212923	20.633	ng/ul	95
10) Bis(2-Chloroethyl)ether	7.719	93	157920	20.937	ng/ul	95
12) 2-Chlorophenol	7.849	128	135079	20.997	ng/ul#	94
13) 2-Methylphenol	8.724	108	171080	22.017	ng/ul	92
14) 2,2'-oxybis(1-Chloropr...	8.806	45	194665	21.494	ng/ul	98
16) Acetophenone	9.147	105	279911	22.519	ng/ul	97
17) N-Nitroso-di-n-propyla...	9.106	70	157137	22.552	ng/ul#	91
18) 4-Methylphenol	9.053	108	189442	22.324	ng/ul	98
19) Hexachloroethane	9.382	117	61900	22.958	ng/ul#	79
22) Nitrobenzene	9.541	77	220725	21.134	ng/ul	100
23) Isophorone	10.052	82	476319	21.470	ng/ul#	95
25) 2-Nitrophenol	10.252	139	91552	23.455	ng/ul#	92
26) 2,4-Dimethylphenol	10.275	107	215430	22.270	ng/ul	91
27) Bis(2-Chloroethoxy)met...	10.534	93	239296	20.125	ng/ul	94
29) 2,4-Dichlorophenol	10.769	162	166369	21.586	ng/ul	95
30) Naphthalene	11.198	128	539538	21.282	ng/ul	98
32) 4-Chloroaniline	11.315	127	230725	21.966	ng/ul	96
33) Hexachlorobutadiene	11.427	225	124223	20.600	ng/ul	92
34) Caprolactam	12.085	113	67862	24.810	ng/ul#	90
35) 4-Chloro-3-methylphenol	12.379	107	221653	23.408	ng/ul	91
36) 2-Methylnaphthalene	12.778	142	407080	21.518	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091323\
 Data File : BG059049.D
 Acq On : 13 Sep 2023 21:29
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020560

Quant Time: Sep 14 01:07:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 12 23:54:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.996	142	404732	21.690	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	13.125	216	251214	20.777	ng/ul#	97
40) Hexachlorocyclopentadiene	13.072	237	177053	19.901	ng/ul#	90
41) 2,4,6-Trichlorophenol	13.360	196	168836	21.996	ng/ul	98
42) 2,4,5-Trichlorophenol	13.430	196	178862	21.524	ng/ul	89
43) 1,1'-Biphenyl	13.771	154	579843	20.942	ng/ul	98
44) 2-Chloronaphthalene	13.824	162	458208	21.668	ng/ul	95
45) 2-Nitroaniline	14.036	65	152514	23.516	ng/ul	92
47) Dimethylphthalate	14.376	163	599225	22.058	ng/ul	98
48) 2,6-Dinitrotoluene	14.517	165	124589	24.109	ng/ul	96
50) Acenaphthylene	14.664	152	764872	21.927	ng/ul	96
51) 3-Nitroaniline	14.852	138	117976	23.790	ng/ul#	95
52) Acenaphthene	14.999	153	501298	21.416	ng/ul	96
53) 2,4-Dinitrophenol	15.046	184	75107	22.549	ng/ul	93
55) 4-Nitrophenol	15.123	109	132979	23.616	ng/ul	86
56) Dibenzofuran	15.328	168	735795	21.702	ng/ul	98
57) 2,4-Dinitrotoluene	15.293	165	185900	26.071	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.534	232	190313	23.976	ng/ul#	95
59) Diethylphthalate	15.722	149	617319	23.049	ng/ul	96
61) Fluorene	15.974	166	644393	23.125	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.957	204	335063	22.724	ng/ul	95
63) 4-Nitroaniline	16.016	138	113863	24.226	ng/ul	95
66) 4,6-Dinitro-2-methylph...	16.045	198	134419	25.592	ng/ul	98
67) N-Nitrosodiphenylamine	16.174	169	532311	20.598	ng/ul	97
68) 4-Bromophenyl-phenylether	16.856	248	233279	21.690	ng/ul	88
69) Hexachlorobenzene	16.950	284	256798	20.798	ng/ul	95
70) Atrazine	17.114	200	234005	23.138	ng/ul	96
71) Pentachlorophenol	17.302	266	166066	21.820	ng/ul	95
72) Phenanthrene	17.725	178	1060285	21.309	ng/ul	98
74) Anthracene	17.819	178	1100979	21.770	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.730	216	273064	19.302	ng/uL	98
76) Pentachlorobenzene	15.222	250	276985	20.260	ng/uL	95
77) Carbazole	18.090	167	941040	22.221	ng/ul	96
78) Di-n-butylphthalate	18.601	149	1060391	22.586	ng/ul	97
80) Fluoranthene	19.717	202	1265673	23.652	ng/ul	98
82) Pyrene	20.081	202	1278864	22.988	ng/ul	98
83) Butylbenzylphthalate	20.945	149	438341	24.935	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.891	252	456988	23.678	ng/ul#	89
85) Benzo(a)anthracene	21.985	228	1266219	21.705	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.821	149	634987	24.118	ng/ul	97
87) Chrysene	22.056	228	1174068	21.649	ng/ul	98
89) Di-n-octyl phthalate	23.148	149	1100311	23.567	ng/ul	100
90) Benzo(b)fluoranthene	24.412	252	1247302	20.630	ng/ul#	96
91) Benzo(k)fluoranthene	24.482	252	1255717	20.324	ng/ul	93
93) Benzo(a)pyrene	25.393	252	1176501	20.856	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	29.682	276	1252967	17.579	ng/ul	97
95) Dibenzo(a,h)anthracene	29.758	278	1175909	19.910	ng/ul#	94
96) Benzo(g,h,i)perylene	30.980	276	771335	13.672	ng/ul#	94





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.483 ng/uL
RT: 3.648 min Scan# 21
Delta R.T. -0.016 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion: 88 Resp: 21830
Ion Ratio Lower Upper
88 100
43 29.9 18.6 28.0#
58 66.6 53.8 80.8



#5
Pyridine
Concen: 19.836 ng/uL
RT: 4.065 min Scan# 98
Delta R.T. -0.016 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion: 79 Resp: 151821
Ion Ratio Lower Upper
79 100
52 58.7 51.4 77.0
51 34.8 25.2 37.8









#6
Benzaldehyde
Concen: 22.784 ng/ul
RT: 7.449 min Scan# 61
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion: 77 Resp: 135213
Ion Ratio Lower Upper
77 100
105 97.8 81.6 122.4
106 79.6 84.5 126.7#



#8
Phenol
Concen: 20.633 ng/ul
RT: 7.443 min Scan# 673
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion: 94 Resp: 212923
Ion Ratio Lower Upper
94 100
65 33.6 22.4 33.6
66 45.1 35.0 52.6









#10
Bis(2-Chloroethyl)ether
Concen: 20.937 ng/ul
RT: 7.719 min Scan# 719
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion: 93 Resp: 157920
Ion Ratio Lower Upper
93 100
63 74.8 56.2 84.4
95 35.1 26.4 39.6



#12
2-Chlorophenol
Concen: 20.997 ng/ul
RT: 7.849 min Scan# 742
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:128 Resp: 135079
Ion Ratio Lower Upper
128 100
64 56.2 43.1 64.7
130 35.7 23.2 34.8#





#13

2-Methylphenol

Concen: 22.017 ng/ul

RT: 8.724 min Scan# 891

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :





SSTD020560

Tgt Ion:108 Resp: 171080

Ion Ratio Lower Upper

108 100

107 81.8 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.494 ng/ul

RT: 8.806 min Scan# 905

Delta R.T. -0.010 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29




Tgt Ion: 45 Resp: 194665


Ion Ratio Lower Upper

45 100

77 16.5 13.2 19.8

79 13.7 9.8 14.6









#16
Acetophenone
Concen: 22.519 ng/ul
RT: 9.147 min Scan# 90
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:105 Resp: 279911
Ion Ratio Lower Upper
105 100
77 80.9 67.3 100.9
51 25.4 19.3 28.9



#17
N-Nitroso-di-n-propylamine
Concen: 22.552 ng/ul
RT: 9.106 min Scan# 956
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion: 70 Resp: 157137
Ion Ratio Lower Upper
70 100
42 52.6 37.5 56.3
101 11.0 5.0 7.6#
130 14.2 13.9 20.9









#18
4-Methylphenol
Concen: 22.324 ng/ul
RT: 9.053 min Scan# 94
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:108 Resp: 189442
Ion Ratio Lower Upper
108 100
107 113.2 88.9 133.3



#19
Hexachloroethane
Concen: 22.958 ng/ul
RT: 9.382 min Scan# 1003
Delta R.T. -0.016 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:117 Resp: 61900
Ion Ratio Lower Upper
117 100
201 106.2 108.2 162.4#
199 58.7 56.0 84.0









#22
Nitrobenzene
Concen: 21.134 ng/ul
RT: 9.541 min Scan# 1031
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion: 77 Resp: 220725
Ion Ratio Lower Upper
77 100
123 35.8 28.5 42.7
65 14.2 11.5 17.3



#23
Isophorone
Concen: 21.470 ng/ul
RT: 10.052 min Scan# 1117
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion: 82 Resp: 476319
Ion Ratio Lower Upper
82 100
95 7.3 6.6 9.8
138 15.8 10.4 15.6#









#25
2-Nitrophenol
Concen: 23.455 ng/ul
RT: 10.252 min Scan# 1151
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:139 Resp: 91552
Ion Ratio Lower Upper
139 100
65 65.4 53.5 80.3
109 51.0 32.0 48.0#



#26
2,4-Dimethylphenol
Concen: 22.270 ng/ul
RT: 10.275 min Scan# 1155
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:107 Resp: 215430
Ion Ratio Lower Upper
107 100
121 41.4 36.4 54.6
122 76.5 69.4 104.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.125 ng/ul
 RT: 10.534 min Scan# 1199
 Delta R.T. -0.010 min
 Lab File: BG059049.D
 Acq: 13 Sep 2023 21:29




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020560


Tgt Ion: 93 Resp: 239296
 Ion Ratio Lower Upper
 93 100
 95 31.0 27.8 41.6
 123 10.5 7.3 10.9



#29
 2,4-Dichlorophenol
 Concen: 21.586 ng/ul
 RT: 10.769 min Scan# 1239
 Delta R.T. -0.010 min
 Lab File: BG059049.D
 Acq: 13 Sep 2023 21:29

Tgt Ion:162 Resp: 166369
 Ion Ratio Lower Upper
 162 100
 164 65.3 51.4 77.0
 98 40.4 27.2 40.8









#30
Naphthalene
Concen: 21.282 ng/ul
RT: 11.198 min Scan# 11
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:128 Resp: 539538
Ion Ratio Lower Upper
128 100
129 11.0 8.2 12.2
127 13.0 10.9 16.3



#32
4-Chloroaniline
Concen: 21.966 ng/ul
RT: 11.315 min Scan# 1332
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:127 Resp: 230725
Ion Ratio Lower Upper
127 100
129 30.7 26.6 39.8









#33
Hexachlorobutadiene
Concen: 20.600 ng/ul
RT: 11.427 min Scan# 11
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:225 Resp: 124223
Ion Ratio Lower Upper
225 100
223 61.4 45.5 68.3
227 70.9 50.2 75.4



#34
Caprolactam
Concen: 24.810 ng/ul
RT: 12.085 min Scan# 1463
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:113 Resp: 67862
Ion Ratio Lower Upper
113 100
55 177.5 140.5 210.7
56 138.7 90.4 135.6#









#35
4-Chloro-3-methylphenol
Concen: 23.408 ng/ul
RT: 12.379 min Scan# 1513
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:107 Resp: 221653
Ion Ratio Lower Upper
107 100
144 20.7 20.2 30.2
142 71.8 51.6 77.4



#36
2-Methylnaphthalene
Concen: 21.518 ng/ul
RT: 12.778 min Scan# 1581
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:142 Resp: 407080
Ion Ratio Lower Upper
142 100
141 83.1 65.7 98.5





#37

1-Methylnaphthalene

Concen: 21.690 ng/ul

RT: 12.996 min Scan# 1

Delta R.T. -0.010 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :

SSTD020560





Tgt Ion:142 Resp: 404732

Ion Ratio Lower Upper

142 100

141 89.6 73.1 109.7

116 4.8 3.4 5.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.777 ng/ul

RT: 13.125 min Scan# 1640

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Tgt Ion:216 Resp: 251214




Ion Ratio Lower Upper


216 100

214 76.3 59.6 89.4

179 20.4 13.6 20.4#

108 17.3 13.8 20.6





#40

Hexachlorocyclopentadiene

Concen: 19.901 ng/ul

RT: 13.072 min Scan# 1

Delta R.T. -0.010 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :

SSTD020560





Tgt Ion:237 Resp: 177053

Ion Ratio Lower Upper

237 100

235 57.6 52.6 78.8

272 14.1 8.6 13.0#



#41

2,4,6-Trichlorophenol

Concen: 21.996 ng/ul

RT: 13.360 min Scan# 1680

Delta R.T. -0.010 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29




Tgt Ion:196 Resp: 168836


Ion Ratio Lower Upper

196 100

198 96.2 78.3 117.5

200 31.6 23.7 35.5





#42

2,4,5-Trichlorophenol

Concen: 21.524 ng/ul

RT: 13.430 min Scan# 1

Delta R.T. -0.010 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :

SSTD020560





Tgt Ion:196 Resp: 178862

Ion Ratio Lower Upper

196 100

198 92.2 84.8 127.2

200 32.1 27.7 41.5



#43

1,1'-Biphenyl

Concen: 20.942 ng/ul

RT: 13.771 min Scan# 1750

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29




Tgt Ion:154 Resp: 579843


Ion Ratio Lower Upper

154 100

153 38.7 31.6 47.4

76 12.9 9.0 13.4









#44
2-Chloronaphthalene
Concen: 21.668 ng/ul
RT: 13.824 min Scan# 1759
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:162 Resp: 458208
Ion Ratio Lower Upper
162 100
164 28.9 27.6 41.4
127 36.5 29.4 44.0



#45
2-Nitroaniline
Concen: 23.516 ng/ul
RT: 14.036 min Scan# 1795
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion: 65 Resp: 152514
Ion Ratio Lower Upper
65 100
92 71.4 54.0 81.0
138 91.1 64.8 97.2









#47
Dimethylphthalate
Concen: 22.058 ng/ul
RT: 14.376 min Scan# 1853
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:163 Resp: 599225
Ion Ratio Lower Upper
163 100
194 3.6 2.8 4.2
164 11.1 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 24.109 ng/ul
RT: 14.517 min Scan# 1877
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:165 Resp: 124589
Ion Ratio Lower Upper
165 100
89 73.5 55.8 83.6
121 21.4 16.3 24.5









#50
Acenaphthylene
Concen: 21.927 ng/ul
RT: 14.664 min Scan# 1903
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:152 Resp: 764872
Ion Ratio Lower Upper
152 100
151 20.0 14.6 21.8
153 14.1 10.4 15.6



#51
3-Nitroaniline
Concen: 23.790 ng/ul
RT: 14.852 min Scan# 1934
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:138 Resp: 117976
Ion Ratio Lower Upper
138 100
108 11.4 11.8 17.6#
92 146.1 122.2 183.2









#52
Acenaphthene
Concen: 21.416 ng/ul
RT: 14.999 min Scan# 1959
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:153 Resp: 501298
Ion Ratio Lower Upper
153 100
152 44.0 36.5 54.7
154 87.4 74.1 111.1



#53
2,4-Dinitrophenol
Concen: 22.549 ng/ul
RT: 15.046 min Scan# 1967
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:184 Resp: 75107
Ion Ratio Lower Upper
184 100
63 78.6 57.7 86.5
154 80.8 60.7 91.1









#55
4-Nitrophenol
Concen: 23.616 ng/ul
RT: 15.123 min Scan# 1981
Delta R.T. 0.002 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:109 Resp: 132979
Ion Ratio Lower Upper
109 100
139 82.0 74.3 111.5
65 100.0 93.4 140.0



#56
Dibenzofuran
Concen: 21.702 ng/ul
RT: 15.328 min Scan# 2015
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:168 Resp: 735795
Ion Ratio Lower Upper
168 100
139 35.6 27.4 41.2





#57

2,4-Dinitrotoluene

Concen: 26.071 ng/ul

RT: 15.293 min Scan# 2011

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :

SSTD020560





Tgt Ion:165 Resp: 185900

Ion Ratio Lower Upper

165 100

63 55.3 33.3 49.9#

182 2.4 1.4 2.0#



#58

2,3,4,6-Tetrachlorophenol

Concen: 23.976 ng/ul

RT: 15.534 min Scan# 2050

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Tgt Ion:232 Resp: 190313




Ion Ratio Lower Upper


232 100

131 37.7 27.1 40.7

130 2.3 0.7 1.1#

166 25.5 19.5 29.3









#59
Diethylphthalate
Concen: 23.049 ng/ul
RT: 15.722 min Scan# 2083
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:149 Resp: 617319
Ion Ratio Lower Upper
149 100
177 24.3 16.9 25.3
150 12.1 9.7 14.5



#61
Fluorene
Concen: 23.125 ng/ul
RT: 15.974 min Scan# 2125
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:166 Resp: 644393
Ion Ratio Lower Upper
166 100
165 98.3 76.4 114.6
167 11.2 9.3 13.9









#62
4-Chlorophenyl-phenylether
Concen: 22.724 ng/ul
RT: 15.957 min Scan# 2122
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:204 Resp: 335063
Ion Ratio Lower Upper
204 100
206 32.1 27.7 41.5
141 57.6 43.2 64.8



#63
4-Nitroaniline
Concen: 24.226 ng/ul
RT: 16.016 min Scan# 2132
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:138 Resp: 113863
Ion Ratio Lower Upper
138 100
92 70.1 49.9 74.9
108 147.8 121.4 182.2









#66
4,6-Dinitro-2-methylphenol
Concen: 25.592 ng/ul
RT: 16.045 min Scan# 2139
Delta R.T. 0.002 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:198 Resp: 134419
Ion Ratio Lower Upper
198 100
182 4.2 3.0 4.6
77 25.8 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 20.598 ng/ul
RT: 16.174 min Scan# 2159
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:169 Resp: 532311
Ion Ratio Lower Upper
169 100
168 69.0 53.0 79.6
167 36.1 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 21.690 ng/ul
RT: 16.856 min Scan# 21
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:248 Resp: 233279
Ion Ratio Lower Upper
248 100
250 106.3 73.9 110.9
141 62.5 45.4 68.0



#69
Hexachlorobenzene
Concen: 20.798 ng/ul
RT: 16.950 min Scan# 2291
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:284 Resp: 256798
Ion Ratio Lower Upper
284 100
142 25.7 24.6 37.0
249 30.1 24.3 36.5





#70

Atrazine

Concen: 23.138 ng/ul

RT: 17.114 min Scan# 21

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :

SSTD020560





Tgt Ion:200 Resp: 234005

Ion Ratio Lower Upper

200 100

173 22.7 17.4 26.2

215 47.2 35.4 53.0



#71

Pentachlorophenol

Concen: 21.820 ng/ul

RT: 17.302 min Scan# 2351

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29




Tgt Ion:266 Resp: 166066


Ion Ratio Lower Upper

266 100

264 64.3 53.1 79.7

268 58.8 51.6 77.4









#72
Phenanthrene
Concen: 21.309 ng/ul
RT: 17.725 min Scan# 2423
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:178 Resp: 1060285
Ion Ratio Lower Upper
178 100
179 15.3 11.8 17.8
176 20.0 15.1 22.7



#74
Anthracene
Concen: 21.770 ng/ul
RT: 17.819 min Scan# 2439
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:178 Resp: 1100979
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.4
176 18.7 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.302 ng/uL

RT: 13.730 min Scan# 1

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :

SSTD020560





Tgt Ion:216 Resp: 273064

Ion Ratio Lower Upper

216 100

214 73.6 60.6 91.0

218 42.6 33.6 50.4



#76

Pentachlorobenzene

Concen: 20.260 ng/uL

RT: 15.222 min Scan# 1997

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29




Tgt Ion:250 Resp: 276985


Ion Ratio Lower Upper

250 100

248 60.1 50.3 75.5

252 63.9 47.6 71.4





#77

Carbazole

Concen: 22.221 ng/ul

RT: 18.090 min Scan# 2485

Delta R.T. -0.010 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :

SSTD020560





Tgt Ion:167 Resp: 941040

Ion Ratio Lower Upper

167 100

166 23.5 16.8 25.2

139 12.1 9.4 14.2



#78

Di-n-butylphthalate

Concen: 22.586 ng/ul

RT: 18.601 min Scan# 2572

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29




Tgt Ion:149 Resp: 1060391


Ion Ratio Lower Upper

149 100

150 8.5 7.7 11.5

104 6.4 4.6 6.8









#80
Fluoranthene
Concen: 23.652 ng/ul
RT: 19.717 min Scan# 21
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:202 Resp: 1265673
Ion Ratio Lower Upper
202 100
101 8.5 6.1 9.1
100 6.7 5.9 8.9



#82
Pyrene
Concen: 22.988 ng/ul
RT: 20.081 min Scan# 2824
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:202 Resp: 1278864
Ion Ratio Lower Upper
202 100
101 9.8 7.5 11.3
100 9.4 6.3 9.5









#83
Butylbenzylphthalate
Concen: 24.935 ng/ul
RT: 20.945 min Scan# 2971
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:149 Resp: 438341
Ion Ratio Lower Upper
149 100
91 83.3 68.7 103.1
206 23.6 17.3 25.9



#84
3,3'-Dichlorobenzidine
Concen: 23.678 ng/ul
RT: 21.891 min Scan# 3132
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:252 Resp: 456988
Ion Ratio Lower Upper
252 100
254 62.2 57.6 86.4
126 7.3 7.8 11.8#









#85
Benzo(a)anthracene
Concen: 21.705 ng/ul
RT: 21.985 min Scan# 3150
Delta R.T. -0.004 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29




Instrument :
BNA_G
ClientSampleId :
SSTD020560


Tgt Ion:228 Resp: 1266219
Ion Ratio Lower Upper
228 100
229 20.8 16.0 24.0
226 26.7 20.8 31.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 24.118 ng/ul
RT: 21.821 min Scan# 3120
Delta R.T. -0.010 min
Lab File: BG059049.D
Acq: 13 Sep 2023 21:29

Tgt Ion:149 Resp: 634987
Ion Ratio Lower Upper
149 100
167 27.9 21.0 31.6
279 5.0 3.8 5.8





#87

Chrysene

Concen: 21.649 ng/ul

RT: 22.056 min Scan# 3160

Delta R.T. -0.004 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :

SSTD020560





Tgt Ion:228 Resp: 1174068

Ion Ratio Lower Upper

228 100

226 30.8 23.2 34.8

229 18.8 15.3 22.9



#89

Di-n-octyl phthalate

Concen: 23.567 ng/ul




RT: 23.148 min Scan# 3346


Delta R.T. -0.010 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Tgt Ion:149 Resp: 1100311









#90
 Benzo(b)fluoranthene
 Concen: 20.630 ng/ul
 RT: 24.412 min Scan# 3561
 Delta R.T. -0.004 min
 Lab File: BG059049.D
 Acq: 13 Sep 2023 21:29

Instrument : BNA_G
 ClientSampleId : SSTD020560

Tgt Ion:252 Resp: 1247302




Ion	Ratio	Lower	Upper
252	100		
253	21.1	17.8	26.8
125	5.3	6.2	9.2




#91
 Benzo(k)fluoranthene
 Concen: 20.324 ng/ul
 RT: 24.482 min Scan# 3573
 Delta R.T. -0.016 min
 Lab File: BG059049.D
 Acq: 13 Sep 2023 21:29

Tgt Ion:252 Resp: 1255717

Ion	Ratio	Lower	Upper
252	100		
253	19.7	19.1	28.7
125	6.1	5.0	7.4





#93

Benzo(a)pyrene

Concen: 20.856 ng/ul

RT: 25.393 min Scan# 31

Delta R.T. -0.010 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :

SSTD020560





Tgt Ion:252 Resp: 1176501

Ion Ratio Lower Upper

252 100

253 22.4 17.1 25.7

125 7.4 5.3 7.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.579 ng/ul

RT: 29.682 min Scan# 4458

Delta R.T. 0.002 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29




Tgt Ion:276 Resp: 1252967


Ion Ratio Lower Upper

276 100

138 11.9 11.8 17.8

277 25.6 20.7 31.1





#95

Dibenzo(a,h)anthracene

Concen: 19.910 ng/ul

RT: 29.758 min Scan# 4471

Delta R.T. -0.016 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Instrument :

BNA_G

ClientSampleId :

SSTD020560





Tgt Ion:278 Resp: 1175909

Ion Ratio Lower Upper

278 100

139 8.0 9.8 14.8#

279 24.6 18.5 27.7



#96

Benzo(g,h,i)perylene

Concen: 13.672 ng/ul

RT: 30.980 min Scan# 4679

Delta R.T. -0.010 min

Lab File: BG059049.D

Acq: 13 Sep 2023 21:29

Tgt Ion:276 Resp: 771335

Ion Ratio Lower Upper

276 100

138 10.3 11.4 17.2#

277 20.5 17.4 26.2

