

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG091418\
 Data File : BG036713.D
 Acq On : 14 Sep 2018 20:00
 Operator : JU/SJ
 Sample : J4906-17
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SP-PIT-9

Quant Time: Sep 15 02:34:48 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 31 13:03:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	49754	20.00	ng	-0.01
21) Naphthalene-d8	10.86	136	205574	20.00	ng	-0.01
38) Acenaphthene-d10	14.67	164	145025	20.00	ng	-0.01
63) Phenanthrene-d10	17.42	188	352548	20.00	ng	-0.01
75) Chrysene-d12	21.68	240	338747	20.00	ng	-0.01
86) Perylene-d12	24.89	264	347518	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	253331	80.77	ng	0.00
7) Phenol-d6	7.21	99	344716	76.76	ng	-0.01
23) Nitrobenzene-d5	9.21	82	224370	59.22	ng	-0.01
41) 2,4,6-Tribromophenol	16.16	330	159389	92.11	ng	-0.01
44) 2-Fluorobiphenyl	13.30	172	507078	49.92	ng	-0.01
78) Terphenyl-d14	20.02	244	806296	55.63	ng	-0.01
Target Compounds						
10) Phenol	7.23	94	10076	2.144	ng	Qvalue 87
37) 2-Methylnaphthalene	12.51	142	64598	8.591	ng	95
49) Dimethylphthalate	14.12	163	52581	4.440	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG091418\
Data File : BG036713.D
Acq On : 14 Sep 2018 20:00
Operator : JU/SJ
Sample : J4906-17
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SP-PIT-9

Quant Time: Sep 15 02:34:48 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 31 13:03:20 2018
Response via : Initial Calibration

