

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091523\
 Data File : BG059076.D
 Acq On : 15 Sep 2023 21:02
 Operator : MA/JU
 Sample : PB155287BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK287

Quant Time: Sep 16 04:37:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 16 00:35:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.297	152	32011	20.000	ng/ul	0.00
20) Naphthalene-d8	11.140	136	154380	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.924	164	110256	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.680	188	256175	20.000	ng/ul	0.00
79) Chrysene-d12	21.998	240	231255	20.000	ng/ul	0.00
88) Perylene-d12	25.547	264	270971	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.661	96	166	0.245	ng/uL	0.05
4) Pyridine-d5	4.043	84	61108	25.791	ng/ul	0.00
7) Phenol-d5	7.409	99	117271	36.799	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.615	67	72915	37.409	ng/ul	0.00
11) 2-Chlorophenol-d4	7.809	132	89398	37.680	ng/ul	0.00
15) 4-Methylphenol-d8	8.978	113	95587	36.921	ng/ul	0.00
21) Nitrobenzene-d5	9.495	128	47600	38.093	ng/ul	0.00
24) 2-Nitrophenol-d4	10.212	143	51278	40.599	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.735	165	93112	38.349	ng/ul	0.00
31) 4-Chloroaniline-d4	11.281	131	150969	39.874	ng/ul	0.00
46) Dimethylphthalate-d6	14.319	166	363513	42.977	ng/ul	0.00
49) Acenaphthylene-d8	14.630	160	363582	39.849	ng/ul	0.00
54) 4-Nitrophenol-d4	15.095	143	72390	43.405	ng/ul	0.00
60) Fluorene-d10	15.917	176	297991	40.867	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.017	200	55401	37.576	ng/ul	0.00
73) Anthracene-d10	17.774	188	505170	43.216	ng/ul	0.00
81) Pyrene-d10	20.048	212	596839	45.443	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.300	264	594669	44.468	ng/ul	0.00
Target Compounds						
17) N-Nitroso-di-n-propyla...	8.978	70	2371	1.071	ng/ul#	Qvalue 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091523\
Data File : BG059076.D
Acq On : 15 Sep 2023 21:02
Operator : MA/JU
Sample : PB155287BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK287


Quant Time: Sep 16 04:37:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 16 00:35:53 2023
Response via : Initial Calibration





#17
 N-Nitroso-di-n-propylamine
 Concen: 1.071 ng/ul
 RT: 8.978 min Scan# 91
 Delta R.T. -0.126 min
 Lab File: BG059076.D
 Acq: 15 Sep 2023 21:02

Instrument :
 BNA_G
 ClientSampleId :
 SBLK287



Tgt Ion:	70	Resp:	2371
Ion	Ratio	Lower	Upper
70	100		
42	548.0	63.0	94.4#
101	0.0	9.1	13.7#
130	0.0	15.7	23.5#

