

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092023\
 Data File : BG059157.D
 Acq On : 22 Sep 2023 9:45
 Operator : MA/JU
 Sample : 04386-12
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0AG0

Quant Time: Sep 22 23:44:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 20 04:25:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.296	152	24099	20.000	ng/ul	0.00
20) Naphthalene-d8	11.140	136	117621	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.924	164	81739	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.673	188	184426	20.000	ng/ul	0.00
79) Chrysene-d12	21.992	240	155776	20.000	ng/ul	# 0.00
88) Perylene-d12	25.535	264	184515	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.607	96	3667	5.956	ng/uL	0.00
4) Pyridine-d5	4.036	84	55701	30.838	ng/ul	0.00
7) Phenol-d5	7.409	99	85881	36.242	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.615	67	44715	30.750	ng/ul	0.00
11) 2-Chlorophenol-d4	7.808	132	58354	33.370	ng/ul	0.00
15) 4-Methylphenol-d8	8.978	113	63402	33.596	ng/ul	0.00
21) Nitrobenzene-d5	9.495	128	28796	31.727	ng/ul	0.00
24) 2-Nitrophenol-d4	10.211	143	33965	37.044	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.734	165	61643	35.413	ng/ul	0.00
31) 4-Chloroaniline-d4	11.281	131	86926	31.996	ng/ul	0.00
46) Dimethylphthalate-d6	14.318	166	203793	32.418	ng/ul	0.00
49) Acenaphthylene-d8	14.624	160	244137	31.895	ng/ul	0.00
54) 4-Nitrophenol-d4	15.100	143	23824	20.772	ng/ul	0.00
60) Fluorene-d10	15.911	176	192258	33.412	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.016	200	19929	19.007	ng/ul	0.00
73) Anthracene-d10	17.773	188	291143	34.695	ng/ul	0.00
81) Pyrene-d10	20.047	212	313359	34.416	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.282	264	314221	34.876	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092023\
Data File : BG059157.D
Acq On : 22 Sep 2023 9:45
Operator : MA/JU
Sample : 04386-12
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0AG0

Quant Time: Sep 22 23:44:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091923.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 20 04:25:24 2023
Response via : Initial Calibration

