



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092023\
Data File : BG059100.D
Acq On : 20 Sep 2023 8:13
Operator : MA/JU
Sample : DFTPP533
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP533

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091923.MA.M
Title : SVOA CALIBRATION
Last Update : Wed Sep 20 04:25:24 2023



AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.9	50368	PASS
68	69	0.00	2	0.7	243	PASS
69	198	0.00	100	36.0	34285	PASS
70	69	0.00	2	0.5	166	PASS
127	198	10	80	44.6	42416	PASS
197	198	0.00	2	0.2	168	PASS
198	198	100	100	100.0	95195	PASS
199	198	5	9	7.0	6698	PASS
275	198	10	60	24.1	22965	PASS
365	198	1	100	3.5	3318	PASS
441	198	0.01	100	17.1	16251	PASS
442	198	50	100	94.2	89669	PASS
443	442	15	24	18.7	16778	PASS

m/z	Abundance
131.40	232.0
207.10	228.0

Instrument :

BNA_G

ClientSampleId :

DFTPP533

m/z	Abundance
38.20	646.0
39.10	4117.0
39.90	208.0
41.10	298.0
44.10	224.0
45.20	178.0
49.10	460.0
50.10	15391.0
51.10	57312.0
52.10	3458.0
53.00	222.0
54.90	181.0
56.00	1441.0
57.00	3412.0
58.00	157.0
61.10	678.0
62.10	376.0
63.10	2219.0
64.20	244.0
65.10	813.0
67.90	589.0
69.00	38800.0
70.00	240.0
73.10	257.0
74.10	4040.0
75.10	6053.0
76.10	2095.0
77.10	45336.0
78.10	3241.0
79.00	2726.0
80.10	2122.0
81.00	2900.0
82.00	840.0
83.10	887.0
85.10	647.0
86.10	1080.0
86.90	616.0
88.20	153.0
91.00	726.0
92.00	730.0
93.00	4767.0
94.00	214.0
95.00	222.0
96.10	363.0
98.00	3825.0
99.00	3709.0
99.90	551.0
101.00	2177.0
103.10	763.0
104.00	1046.0
105.10	869.0
107.00	12949.0
108.10	1819.0
109.00	211.0
110.00	24112.0
111.00	4438.0
112.00	489.0
116.10	810.0
117.10	10666.0
118.10	866.0
120.00	182.0
122.00	1021.0
123.10	1178.0
124.00	703.0
125.10	723.0
127.00	47712.0
128.10	3642.0
129.00	15168.0
130.00	1518.0
131.10	205.0
132.40	194.0
134.10	269.0
135.00	1201.0
136.10	497.0
137.10	600.0
141.00	2764.0
142.00	510.0
143.00	1038.0
146.00	307.0
147.00	1153.0
148.10	2042.0
149.10	429.0
150.00	213.0
151.10	559.0
152.10	256.0
152.90	524.0
154.00	661.0
155.00	1442.0
156.00	2246.0
157.00	368.0
158.00	409.0
159.00	343.0

Instrument :
BNA_G
ClientSampleId :
DFTPP533

160.00	621.0
161.00	740.0
162.10	277.0
164.10	180.0
165.10	809.0
166.00	864.0
167.10	3953.0
168.00	2639.0
169.10	461.0
171.90	300.0
172.80	406.0
174.00	952.0
175.10	1387.0
176.20	930.0
177.10	689.0
177.90	215.0
179.00	3972.0
180.00	2270.0
181.10	1406.0
183.80	280.0
185.00	1548.0
186.00	12487.0
187.00	3658.0
188.10	320.0
189.10	569.0
191.00	615.0
192.00	913.0
193.10	1028.0
194.00	312.0
196.10	2858.0
198.00	99352.0
199.00	6290.0
199.80	529.0
201.40	727.0
203.00	641.0
204.00	3991.0
205.10	6277.0
206.00	20768.0
207.10	2752.0
208.00	822.0
208.80	179.0
209.00	184.0
209.90	418.0
210.90	957.0
211.70	233.0
213.10	192.0
214.70	216.0
216.00	463.0
217.00	6417.0
218.00	882.0
221.00	4028.0
221.60	1127.0
223.00	1465.0
224.00	14229.0
225.10	3602.0
225.90	397.0
227.00	5871.0
228.00	847.0
229.00	1025.0
231.10	650.0
232.10	202.0
233.10	169.0
233.70	199.0
234.10	279.0
235.00	448.0
236.20	289.0
237.00	383.0
238.70	201.0
239.90	153.0
240.90	237.0
242.10	896.0
243.00	937.0
244.10	9688.0
245.00	1144.0
246.00	1939.0
247.00	590.0
249.00	492.0
253.00	244.0
255.00	58912.0
256.00	8621.0
256.90	775.0
258.00	3237.0
259.00	402.0
261.20	216.0
265.00	1212.0
265.90	190.0
267.00	158.0
272.00	254.0
273.00	1711.0
274.00	3742.0
275.00	21024.0
276.00	3170.0
277.00	1476.0
278.00	511.0
283.10	491.0
285.00	323.0

Instrument :
BNA_G
ClientSampleId :
DFTPP533

292.10	289.0
293.10	611.0
296.00	6755.0
297.00	1037.0
302.90	1094.0
303.80	354.0
308.20	157.0
309.00	168.0
313.90	257.0
314.10	246.0
315.00	548.0
316.00	248.0
320.90	185.0
322.90	2247.0
324.00	474.0
326.90	539.0
327.90	475.0
332.00	183.0
332.70	335.0
334.00	1342.0
335.00	422.0
341.00	311.0
346.00	427.0
351.90	665.0
353.00	693.0
354.10	630.0
364.90	3234.0
365.90	546.0
370.80	314.0
372.00	1103.0
372.90	285.0
383.00	443.0
389.70	151.0
390.10	191.0
401.90	436.0
403.00	848.0
421.00	838.0
421.90	618.0
423.00	4273.0
424.00	791.0
441.00	13305.0
442.00	72104.0
443.00	13852.0
444.10	1360.0

Instrument :
BNA_G
ClientSampleId :
DFTPP533

m/z	Abundance
38.10	723.0
39.10	3545.0
41.10	171.0
44.00	283.0
49.00	427.0
50.10	12745.0
51.10	59824.0
52.10	2425.0
53.10	150.0
54.90	245.0
56.00	1578.0
57.00	4646.0
61.00	543.0
62.10	611.0
63.10	2163.0
64.20	225.0
65.10	650.0
67.00	167.0
68.10	729.0
69.00	40064.0
70.10	258.0
73.00	270.0
74.10	3391.0
75.10	5733.0
76.10	2517.0
77.10	46472.0
78.10	3606.0
79.10	2978.0
80.00	2502.0
81.00	2646.0
82.00	432.0
83.00	730.0
85.00	452.0
86.20	625.0
87.00	381.0
88.00	324.0
91.00	875.0
92.10	769.0
93.00	4961.0
94.10	300.0
95.10	216.0
96.00	171.0
98.00	4581.0
99.00	3874.0
101.10	1725.0
103.10	340.0
103.90	1159.0
105.00	1185.0
106.00	410.0
107.00	13819.0
108.00	1984.0
110.00	24120.0
111.00	3676.0
112.00	412.0
116.00	879.0
117.00	11677.0
118.00	759.0
120.10	195.0
121.90	564.0
123.00	1280.0
124.00	699.0
125.00	560.0
127.00	48792.0
128.00	4236.0
129.00	18904.0
130.00	1388.0
132.00	221.0
134.00	601.0
135.00	1218.0
136.10	882.0
137.00	993.0
137.70	214.0
139.00	286.0
141.00	1903.0
142.00	515.0
142.90	367.0
145.00	268.0
145.80	374.0
147.10	1146.0
148.00	2126.0
148.90	540.0
151.30	360.0
152.00	262.0
153.10	689.0
154.00	504.0
155.00	1015.0
156.00	1996.0
157.10	323.0
157.90	456.0
159.10	396.0
160.00	666.0
160.90	936.0

Instrument :
BNA_G
ClientSampleId :
DFTPP533

165.00	1010.0
166.00	720.0
167.10	4945.0
168.10	2529.0
168.90	384.0
170.90	241.0
172.00	565.0
173.00	547.0
173.90	693.0
175.00	1875.0
176.00	645.0
176.90	1260.0
177.90	350.0
179.00	3648.0
180.00	2304.0
181.00	1617.0
182.00	212.0
182.90	204.0
184.00	287.0
185.00	2123.0
186.00	14896.0
187.00	4532.0
188.10	405.0
188.90	833.0
190.10	293.0
191.00	626.0
192.00	1444.0
193.00	1362.0
194.20	491.0
194.90	304.0
196.00	3270.0
198.00	112448.0
199.00	8663.0
199.90	955.0
201.40	790.0
203.00	928.0
204.00	2837.0
205.00	6389.0
206.10	29392.0
207.10	3754.0
208.00	1081.0
209.00	296.0
209.80	324.0
211.00	952.0
215.10	465.0
215.90	528.0
217.00	8085.0
218.00	966.0
218.90	313.0
221.00	4775.0
221.90	1271.0
223.00	1522.0
224.00	15685.0
225.00	4208.0
226.10	479.0
227.00	6900.0
228.00	1032.0
229.00	1182.0
230.00	196.0
231.00	519.0
233.80	353.0
234.80	329.0
236.00	363.0
237.00	637.0
238.90	224.0
240.10	220.0
241.10	602.0
241.90	813.0
243.00	1008.0
244.00	13041.0
245.00	1757.0
246.00	1951.0
246.90	462.0
248.10	236.0
248.90	610.0
251.00	173.0
252.00	298.0
255.00	71640.0
256.00	10116.0
256.90	1071.0
258.00	4437.0
258.90	495.0
259.70	156.0
264.00	197.0
265.00	1499.0
266.10	268.0
271.00	220.0
271.90	257.0
273.00	2345.0
274.00	4832.0
275.00	27912.0
276.00	3949.0
276.90	2709.0
277.90	419.0
279.00	341.0
283.00	236.0

Instrument :
BNA_G
ClientSampleId :
DFTPP533

284.00	279.0
285.00	439.0
293.00	700.0
293.90	309.0
296.00	9514.0
297.00	1127.0
301.90	389.0
303.10	940.0
303.90	301.0
313.90	545.0
314.90	936.0
316.10	367.0
317.00	366.0
322.10	231.0
323.00	2904.0
323.70	303.0
324.00	402.0
327.00	649.0
328.00	375.0
332.20	250.0
333.10	326.0
334.00	1766.0
334.90	270.0
336.00	184.0
340.90	407.0
345.90	430.0
352.10	689.0
353.10	457.0
354.00	801.0
354.90	221.0
358.50	153.0
364.90	3827.0
365.80	491.0
370.90	348.0
372.00	2158.0
373.00	435.0
382.90	568.0
384.00	152.0
390.10	395.0
390.70	203.0
400.80	234.0
401.80	345.0
402.90	1152.0
403.90	363.0
421.00	1008.0
422.00	994.0
422.20	944.0
422.90	5290.0
424.00	890.0
430.70	213.0
441.00	17856.0
442.00	105072.0
442.90	20392.0
443.90	1464.0
445.00	223.0

Instrument :
BNA_G
ClientSampleId :
DFTPP533

m/z	Abundance
38.10	310.0
39.10	1847.0
41.10	202.0
49.10	390.0
50.10	7964.0
51.10	33968.0
52.10	1570.0
56.00	706.0
57.10	2272.0
61.10	393.0
62.10	462.0
63.10	686.0
65.00	349.0
69.00	23992.0
72.90	381.0
74.00	2466.0
75.00	2935.0
76.10	1416.0
77.10	27880.0
78.10	1978.0
79.00	1392.0
80.00	1333.0
81.00	1973.0
82.10	328.0
83.00	607.0
85.00	321.0
86.00	692.0
87.10	243.0
91.00	347.0
92.00	414.0
93.00	3379.0
96.10	167.0
98.00	1974.0
99.10	1661.0
100.10	176.0
101.00	1127.0
103.00	349.0
104.10	837.0
105.00	562.0
105.80	208.0
106.00	206.0
107.00	8836.0
108.10	1095.0
109.10	266.0
110.00	15350.0
111.10	1961.0
112.10	208.0
116.10	388.0
117.00	5861.0
118.00	379.0
122.00	515.0
123.10	696.0
123.90	254.0
125.00	619.0
125.70	203.0
127.00	30744.0
128.00	2250.0
129.00	12254.0
130.00	1039.0
131.00	267.0
134.00	561.0
135.00	1206.0
136.00	448.0
137.00	507.0
141.00	1891.0
142.00	393.0
143.00	312.0
145.80	182.0
147.00	522.0
148.10	1616.0
148.90	228.0
151.00	154.0
153.10	597.0
154.00	369.0
155.00	789.0
156.10	1484.0
157.00	183.0
157.40	266.0
157.90	217.0
160.00	626.0
161.10	785.0
161.80	246.0
165.00	333.0
165.90	285.0
167.00	3567.0
168.00	1231.0
169.10	326.0
171.10	226.0
171.90	455.0
172.90	404.0
174.00	464.0
175.00	1216.0

Instrument :
BNA_G
ClientSampleId :
DFTPP533

176.00	695.0
176.80	374.0
179.00	2674.0
180.00	1587.0
181.00	570.0
184.00	266.0
184.90	1231.0
186.00	10725.0
187.00	2561.0
188.10	278.0
188.90	549.0
191.00	295.0
192.10	956.0
193.10	992.0
194.00	216.0
196.10	2282.0
196.80	503.0
198.00	73784.0
199.00	5140.0
200.10	388.0
201.30	559.0
203.00	706.0
204.00	2408.0
205.00	3921.0
206.00	21168.0
207.00	2885.0
208.00	424.0
209.00	215.0
209.90	396.0
211.00	965.0
211.80	176.0
215.00	253.0
216.10	358.0
216.90	4776.0
218.00	685.0
221.00	4005.0
222.00	489.0
223.00	1409.0
224.00	10420.0
225.00	3232.0
227.10	4307.0
228.00	574.0
228.90	1056.0
230.10	267.0
231.00	522.0
234.00	428.0
235.10	359.0
235.90	426.0
237.10	396.0
239.00	308.0
240.00	238.0
241.00	237.0
241.80	620.0
243.00	493.0
244.00	10171.0
245.10	989.0
246.00	1839.0
247.00	396.0
249.00	198.0
250.80	152.0
252.00	263.0
253.10	179.0
253.80	432.0
255.00	52272.0
256.00	6443.0
257.00	522.0
257.90	2364.0
259.00	476.0
265.10	1400.0
265.90	271.0
270.90	160.0
272.00	151.0
272.90	1190.0
274.00	3948.0
275.00	19960.0
276.00	3042.0
277.00	1927.0
277.90	287.0
282.00	154.0
282.90	247.0
285.00	632.0
289.10	206.0
292.90	661.0
294.00	234.0
296.00	6405.0
297.00	963.0
298.10	165.0
301.10	165.0
301.90	180.0
303.00	973.0
304.10	183.0
309.10	235.0
309.90	175.0
313.90	230.0
314.10	221.0
315.00	781.0

Instrument :
BNA_G
ClientSampleId :
DFTPP533

316.00	578.0
320.70	198.0
322.00	158.0
323.00	1959.0
323.80	440.0
326.80	597.0
327.70	298.0
332.00	220.0
332.90	308.0
334.00	1279.0
334.90	481.0
345.80	515.0
352.00	537.0
353.00	409.0
353.80	667.0
365.00	2894.0
365.80	379.0
372.00	1410.0
372.90	294.0
382.80	341.0
383.10	326.0
390.10	324.0
391.00	245.0
401.60	433.0
402.10	533.0
402.90	825.0
403.90	456.0
420.80	462.0
421.90	788.0
423.00	4317.0
423.90	1137.0
441.00	17592.0
442.00	91832.0
443.00	16089.0
443.90	1879.0

Instrument :
BNA_G
ClientSampleId :
DFTPP533