


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG092118\  
 Data File : BG036862.D  
 Acq On : 21 Sep 2018 17:27  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG092018.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Sep 20 20:07:11 2018



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2471

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 36.8      | 16508   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.0       | 150     | PASS             |
| 69          | 198          | 0.00         | 100          | 34.1      | 15321   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 40.1      | 18007   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 44914   | PASS             |
| 199         | 198          | 5            | 9            | 7.0       | 3139    | PASS             |
| 275         | 198          | 10           | 60           | 28.0      | 12584   | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 1506    | PASS             |
| 441         | 198          | 0.01         | 100          | 9.5       | 4284    | PASS             |
| 442         | 198          | 50           | 100          | 60.6      | 27208   | PASS             |
| 443         | 442          | 15           | 24           | 19.7      | 5358    | PASS             |