


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG092119\
 Data File : BG042904.D
 Acq On : 21 Sep 2019 10:00
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG092119.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 23 06:01:34 2019



AutoFind: Scans 2517, 2518, 2519; Background Corrected with Scan 2510

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.6	152640	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.5	156032	PASS
70	69	0.00	2	0.5	823	PASS
127	198	10	80	46.1	182186	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	395456	PASS
199	198	5	9	7.5	29629	PASS
275	198	10	60	32.3	127765	PASS
365	198	1	100	4.5	17975	PASS
441	198	0.01	100	11.8	46696	PASS
442	198	50	100	74.7	295296	PASS
443	442	15	24	20.4	60221	PASS