

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092521\
 Data File : BG050269.D
 Acq On : 25 Sep 2021 18:41
 Operator : CG/JU
 Sample : SSTD16040
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160438

Quant Time: Sep 27 02:05:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG092521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 27 01:56:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.292	152	34728	20.000	ng/ul	0.00
20) Naphthalene-d8	11.124	136	144321	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.908	164	103657	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.652	188	251644	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.970	240	237407	20.000	ng/ul	0.01
88) Perylene-d12	25.413	264	242236	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.662	96	44907	66.681	ng/uL	0.00
4) Pyridine-d5	4.097	84	358144	176.553	ng/ul	0.00
7) Phenol-d5	7.435	99	453178	153.673	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.617	67	246407	150.400	ng/ul	0.00
11) 2-Chlorophenol-d4	7.822	132	332853	150.848	ng/ul	0.00
15) 4-Methylphenol-d8	8.997	113	349238	150.374	ng/ul	0.01
21) Nitrobenzene-d5	9.473	128	156367	139.752	ng/ul	0.00
24) 2-Nitrophenol-d4	10.196	143	184439	138.436	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.737	165	366260	154.129	ng/ul	0.01
31) 4-Chloroaniline-d4	11.254	131	469031	145.013	ng/ul	0.00
46) Dimethylphthalate-d6	14.315	166	1067337	134.544	ng/ul	0.00
49) Acenaphthylene-d8	14.608	160	1314185	141.288	ng/ul	0.00
54) 4-Nitrophenol-d4	15.102	143	207323	180.633	ng/ul	0.03
60) Fluorene-d10	15.901	176	972793	144.631	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.019	200	207306	128.050	ng/ul	0.02
73) Anthracene-d10	17.758	188	1500065	132.748	ng/ul	0.00
81) Pyrene-d10	20.032	212	1728995	134.770	ng/ul	0.01
92) Benzo(a)pyrene-d12	25.196	264	1844752	143.546	ng/ul	0.03
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.698	88	52815	64.330	ng/uL	98
5) Pyridine	4.115	79	359944	164.176	ng/ul	97
6) Benzaldehyde	7.429	77	173433	102.199	ng/ul	97
8) Phenol	7.464	94	449459	150.894	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.711	93	319859	155.556	ng/ul	97
12) 2-Chlorophenol	7.858	128	327714	150.300	ng/ul#	81
13) 2-Methylphenol	8.727	108	355974	154.905	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.815	45	411325	190.763	ng/ul#	92
16) Acetophenone	9.133	105	508898	134.360	ng/ul	98
17) N-Nitroso-di-n-propyla...	9.133	70	284915	149.812	ng/ul	96
18) 4-Methylphenol	9.068	108	357761	144.713	ng/ul	83
19) Hexachloroethane	9.391	117	141013	147.777	ng/ul#	87
22) Nitrobenzene	9.514	77	428549	141.352	ng/ul	97
23) Isophorone	10.061	82	811008	150.136	ng/ul	95
25) 2-Nitrophenol	10.225	139	185406	143.137	ng/ul#	88
26) 2,4-Dimethylphenol	10.272	107	412103	143.156	ng/ul#	84
27) Bis(2-Chloroethoxy)met...	10.519	93	434713	155.843	ng/ul	97
29) 2,4-Dichlorophenol	10.760	162	347922	159.520	ng/ul	100
30) Naphthalene	11.177	128	1097504	152.144	ng/ul	97
32) 4-Chloroaniline	11.283	127	462124	148.357	ng/ul	93
33) Hexachlorobutadiene	11.447	225	287096	162.715	ng/ul	95
34) Caprolactam	12.111	113	68696	89.916	ng/ul	88
35) 4-Chloro-3-methylphenol	12.382	107	382165	146.611	ng/ul	99
36) 2-Methylnaphthalene	12.758	142	759105	141.565	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092521\
 Data File : BG050269.D
 Acq On : 25 Sep 2021 18:41
 Operator : CG/JU
 Sample : SSTD16040
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160438


Quant Time: Sep 27 02:05:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG092521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 27 01:56:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.975	142	759223	140.270	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	13.116	216	492745	161.844	ng/ul#	98
40) Hexachlorocyclopentadiene	13.093	237	373530	274.134	ng/ul	95
41) 2,4,6-Trichlorophenol	13.351	196	328817	169.475	ng/ul	95
42) 2,4,5-Trichlorophenol	13.422	196	344630	169.479	ng/ul	91
43) 1,1'-Biphenyl	13.751	154	1018463	140.636	ng/ul#	93
44) 2-Chloronaphthalene	13.798	162	821956	141.305	ng/ul	97
45) 2-Nitroaniline	13.997	65	272937	138.199	ng/ul	99
47) Dimethylphthalate	14.368	163	1034946	131.824	ng/ul	98
48) 2,6-Dinitrotoluene	14.485	165	222558	137.305	ng/ul	91
50) Acenaphthylene	14.644	152	1266545	148.984	ng/ul#	92
51) 3-Nitroaniline	14.820	138	197012	125.959	ng/ul	99
52) Acenaphthene	14.979	153	880448	142.771	ng/ul	97
53) 2,4-Dinitrophenol	15.014	184	144182	174.162	ng/ul#	72
55) 4-Nitrophenol	15.173	109	225	0.155	ng/ul	84
56) Dibenzofuran	15.314	168	1258941	147.416	ng/ul	99
57) 2,4-Dinitrotoluene	15.267	165	318404	136.564	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	15.525	232	326868	193.980	ng/ul	90
59) Diethylphthalate	15.719	149	1093718	136.179	ng/ul#	92
61) Fluorene	15.960	166	996286	138.275	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.942	204	586122	153.748	ng/ul	92
63) 4-Nitroaniline	15.989	138	184690	118.327	ng/ul	89
66) 4,6-Dinitro-2-methylph...	16.030	198	206746	140.891	ng/ul	99
67) N-Nitrosodiphenylamine	16.154	169	919848	138.841	ng/ul	95
68) 4-Bromophenyl-phenylether	16.835	248	404576	145.334	ng/ul	95
69) Hexachlorobenzene	16.953	284	449469	139.825	ng/ul	95
70) Atrazine	17.106	200	428956	145.741	ng/ul	94
71) Pentachlorophenol	17.294	266	315691	231.926	ng/ul	96
72) Phenanthrene	17.699	178	1725773	137.565	ng/ul#	93
74) Anthracene	17.793	178	1672536	131.562	ng/ul#	86
75) 1,2,3,4-Tetrachloroben...	13.715	216	545024	161.432	ng/ul	92
76) Pentachlorobenzene	15.225	250	531388	148.538	ng/ul	97
77) Carbazole	18.057	167	1575854	139.304	ng/ul	96
78) Di-n-butylphthalate	18.598	149	1765164	122.006	ng/ul#	96
80) Fluoranthene	19.697	202	2060273	130.139	ng/ul#	89
82) Pyrene	20.061	202	1973098	125.775	ng/ul#	88
83) Butylbenzylphthalate	20.930	149	817555	132.061	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.853	252	689646	122.545	ng/ul#	97
85) Benzo(a)anthracene	21.947	228	1986259	134.271	ng/ul#	89
86) Bis(2-ethylhexyl)phtha...	21.835	149	1115920	131.256	ng/ul#	91
87) Chrysene	21.947	228	1986259	142.042	ng/ul	93
89) Di-n-octyl phthalate	23.134	149	1744236	121.379	ng/ul	100
90) Benzo(b)fluoranthene	24.326	252	2087197	134.702	ng/ul#	96
91) Benzo(k)fluoranthene	24.403	252	1859657	126.262	ng/ul#	94
93) Benzo(a)pyrene	25.278	252	1980147	147.149	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	29.426	276	2109699	118.850	ng/ul#	96
95) Dibenzo(a,h)anthracene	29.509	278	1980949	133.308	ng/ul#	98
96) Benzo(g,h,i)perylene	30.690	276	2025420	135.720	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 27 02:05:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG092521.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 27 01:56:44 2021
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 64.330 ng/uL
RT: 3.698 min Scan# 112
Delta R.T. -0.004 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion	Ratio	Lower	Upper
88	100		
43	30.4	21.6	32.4
58	65.3	52.0	78.0



#5
Pyridine
Concen: 164.176 ng/uL
RT: 4.115 min Scan# 183
Delta R.T. -0.004 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41


Tgt Ion	Ratio	Lower	Upper
79	100		
52	76.0	61.1	91.7
51	39.0	34.8	52.2








#6
Benzaldehyde
Concen: 102.199 ng/ul
RT: 7.429 min Scan# 747
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41


Instrument :
BNA_G
ClientSampleId :
SSTD160438





Tgt Ion: 77 Resp: 173433
Ion Ratio Lower Upper
77 100
105 91.5 72.3 108.5
106 85.5 72.3 108.5




#8
Phenol
Concen: 150.894 ng/ul
RT: 7.464 min Scan# 753
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41



Tgt Ion: 94 Resp: 449459
Ion Ratio Lower Upper
94 100
65 32.9 27.7 41.5
66 46.7 39.0 58.4









#10
Bis(2-Chloroethyl)ether
Concen: 155.556 ng/ul
RT: 7.711 min Scan# 795
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion: 93 Resp: 319859
Ion Ratio Lower Upper
93 100
63 78.1 59.9 89.9
95 33.4 27.6 41.4



#12
2-Chlorophenol
Concen: 150.300 ng/ul
RT: 7.858 min Scan# 820
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:128 Resp: 327714
Ion Ratio Lower Upper
128 100
64 52.5 28.0 42.0#
130 33.7 24.4 36.6









#13
2-Methylphenol
Concen: 154.905 ng/ul
RT: 8.727 min Scan# 968
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:108 Resp: 355974
Ion Ratio Lower Upper
108 100
107 92.1 71.0 106.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 190.763 ng/ul
RT: 8.815 min Scan# 983
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion: 45 Resp: 411325
Ion Ratio Lower Upper
45 100
77 16.2 16.9 25.3#
79 12.0 11.0 16.4









#16
Acetophenone
Concen: 134.360 ng/ul
RT: 9.133 min Scan# 1037
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:105 Resp: 508898
Ion Ratio Lower Upper
105 100
77 86.4 67.7 101.5
51 30.7 25.3 37.9



#17
N-Nitroso-di-n-propylamine
Concen: 149.812 ng/ul
RT: 9.133 min Scan# 1037
Delta R.T. 0.031 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion: 70 Resp: 284915
Ion Ratio Lower Upper
70 100
42 59.3 44.2 66.4
101 9.1 7.8 11.8
130 19.9 15.3 22.9









#18
4-Methylphenol
Concen: 144.713 ng/ul
RT: 9.068 min Scan# 1026
Delta R.T. 0.019 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:108 Resp: 357761
Ion Ratio Lower Upper
108 100
107 114.6 78.3 117.5



#19
Hexachloroethane
Concen: 147.777 ng/ul
RT: 9.391 min Scan# 1081
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:117 Resp: 141013
Ion Ratio Lower Upper
117 100
201 121.6 89.3 133.9
199 78.2 50.5 75.7#









#22
Nitrobenzene
Concen: 141.352 ng/ul
RT: 9.514 min Scan# 1102
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion: 77 Resp: 428549
Ion Ratio Lower Upper
77 100
123 40.3 30.4 45.6
65 14.3 11.3 16.9



#23
Isophorone
Concen: 150.136 ng/ul
RT: 10.061 min Scan# 1195
Delta R.T. 0.019 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion: 82 Resp: 811008
Ion Ratio Lower Upper
82 100
95 8.2 7.8 11.8
138 16.6 11.4 17.0









#25
2-Nitrophenol
Concen: 143.137 ng/ul
RT: 10.225 min Scan# 1223
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:139 Resp: 185406
Ion Ratio Lower Upper
139 100
65 62.5 53.1 79.7
109 39.5 44.1 66.1#



#26
2,4-Dimethylphenol
Concen: 143.156 ng/ul
RT: 10.272 min Scan# 1231
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:107 Resp: 412103
Ion Ratio Lower Upper
107 100
121 49.7 32.7 49.1#
122 79.7 52.6 79.0#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 155.843 ng/ul
 RT: 10.519 min Scan# 1273
 Delta R.T. 0.007 min
 Lab File: BG050269.D
 Acq: 25 Sep 2021 18:41




Instrument :
 BNA_G
 ClientSampleId :
 SSTD160438


Tgt Ion: 93 Resp: 434713
 Ion Ratio Lower Upper
 93 100
 95 33.2 24.8 37.2
 123 11.0 8.2 12.2



#29
 2,4-Dichlorophenol
 Concen: 159.520 ng/ul
 RT: 10.760 min Scan# 1314
 Delta R.T. 0.007 min
 Lab File: BG050269.D
 Acq: 25 Sep 2021 18:41

Tgt Ion:162 Resp: 347922
 Ion Ratio Lower Upper
 162 100
 164 63.9 51.0 76.6
 98 34.6 27.8 41.8









#30
Naphthalene
Concen: 152.144 ng/ul
RT: 11.177 min Scan# 1385
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:128 Resp: 1097504
Ion Ratio Lower Upper
128 100
129 11.7 9.3 13.9
127 14.2 9.8 14.6



#32
4-Chloroaniline
Concen: 148.357 ng/ul
RT: 11.283 min Scan# 1403
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:127 Resp: 462124
Ion Ratio Lower Upper
127 100
129 33.5 30.3 45.5









#33
Hexachlorobutadiene
Concen: 162.715 ng/ul
RT: 11.447 min Scan# 1431
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:	225	Resp:	287096
Ion Ratio	Lower	Upper	
225	100		
223	63.0	51.3	76.9
227	62.5	55.1	82.7



#34
Caprolactam
Concen: 89.916 ng/ul
RT: 12.111 min Scan# 1544
Delta R.T. 0.048 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:	113	Resp:	68696
Ion Ratio	Lower	Upper	
113	100		
55	164.5	114.0	171.0
56	127.4	107.4	161.0









#35
4-Chloro-3-methylphenol
Concen: 146.611 ng/ul
RT: 12.382 min Scan# 1590
Delta R.T. 0.019 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:107 Resp: 382165
Ion Ratio Lower Upper
107 100
144 24.6 18.9 28.3
142 73.3 59.2 88.8



#36
2-Methylnaphthalene
Concen: 141.565 ng/ul
RT: 12.758 min Scan# 1654
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:142 Resp: 759105
Ion Ratio Lower Upper
142 100
141 90.2 70.5 105.7









#37
1-Methylnaphthalene
Concen: 140.270 ng/ul
RT: 12.975 min Scan# 1691
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:142 Resp: 759223
Ion Ratio Lower Upper
142 100
141 96.6 75.8 113.8
116 5.1 3.7 5.5



#39
1,2,4,5-Tetrachlorobenzene
Concen: 161.844 ng/ul
RT: 13.116 min Scan# 1715
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:216 Resp: 492745
Ion Ratio Lower Upper
216 100
214 80.1 64.2 96.2
179 20.7 12.8 19.2#
108 16.2 13.1 19.7









#40
Hexachlorocyclopentadiene
Concen: 274.134 ng/ul
RT: 13.093 min Scan# 1711
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:237	Resp:	373530
Ion Ratio	Lower	Upper
237	100	
235	62.7	53.5 80.3
272	12.8	11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 169.475 ng/ul
RT: 13.351 min Scan# 1755
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:196	Resp:	328817
Ion Ratio	Lower	Upper
196	100	
198	98.1	83.7 125.5
200	31.5	25.9 38.9









#42
2,4,5-Trichlorophenol
Concen: 169.479 ng/ul
RT: 13.422 min Scan# 1767
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:196 Resp: 344630
Ion Ratio Lower Upper
196 100
198 97.3 70.3 105.5
200 30.8 26.6 40.0



#43
1,1'-Biphenyl
Concen: 140.636 ng/ul
RT: 13.751 min Scan# 1823
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:154 Resp: 1018463
Ion Ratio Lower Upper
154 100
153 44.1 32.5 48.7
76 13.8 7.8 11.8#









#44
2-Chloronaphthalene
Concen: 141.305 ng/ul
RT: 13.798 min Scan# 1831
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:	162	Resp:	821956
Ion Ratio	Lower	Upper	
162	100		
164	34.6	25.2	37.8
127	40.2	31.9	47.9



#45
2-Nitroaniline
Concen: 138.199 ng/ul
RT: 13.997 min Scan# 1865
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:	65	Resp:	272937
Ion Ratio	Lower	Upper	
65	100		
92	70.4	55.6	83.4
138	98.1	78.7	118.1









#47
Dimethylphthalate
Concen: 131.824 ng/ul
RT: 14.368 min Scan# 1928
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:163 Resp: 1034946
Ion Ratio Lower Upper
163 100
194 4.7 3.4 5.0
164 11.0 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 137.305 ng/ul
RT: 14.485 min Scan# 1948
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:165 Resp: 222558
Ion Ratio Lower Upper
165 100
89 70.9 62.5 93.7
121 22.3 22.2 33.2









#50
Acenaphthylene
Concen: 148.984 ng/ul
RT: 14.644 min Scan# 1975
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:152 Resp: 1266545
Ion Ratio Lower Upper
152 100
151 23.7 15.2 22.8#
153 15.0 10.7 16.1



#51
3-Nitroaniline
Concen: 125.959 ng/ul
RT: 14.820 min Scan# 2005
Delta R.T. 0.019 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:138 Resp: 197012
Ion Ratio Lower Upper
138 100
108 10.8 9.0 13.4
92 122.0 96.4 144.6









#52
Acenaphthene
Concen: 142.771 ng/ul
RT: 14.979 min Scan# 2032
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:153 Resp: 880448
Ion Ratio Lower Upper
153 100
152 50.7 37.4 56.2
154 87.6 71.5 107.3



#53
2,4-Dinitrophenol
Concen: 174.162 ng/ul
RT: 15.014 min Scan# 2038
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41


Tgt Ion:184 Resp: 144182
Ion Ratio Lower Upper
184 100
63 63.7 75.8 113.8#
154 60.7 64.6 96.8#








#55
4-Nitrophenol
Concen: 0.155 ng/ul
RT: 15.173 min Scan# 2065
Delta R.T. 0.090 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41


Instrument :
BNA_G
ClientSampleId :
SSTD160438





Tgt Ion:109 Resp: 225
Ion Ratio Lower Upper
109 100
139 65.4 61.4 92.0
65 77.3 76.5 114.7




#56
Dibenzofuran
Concen: 147.416 ng/ul
RT: 15.314 min Scan# 2089
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41



Tgt Ion:168 Resp: 1258941
Ion Ratio Lower Upper
168 100
139 41.7 33.8 50.8









#57
2,4-Dinitrotoluene
Concen: 136.564 ng/ul
RT: 15.267 min Scan# 2081
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:165 Resp: 318404
Ion Ratio Lower Upper
165 100
63 50.9 42.3 63.5
182 2.5 2.6 4.0#



#58
2,3,4,6-Tetrachlorophenol
Concen: 193.980 ng/ul
RT: 15.525 min Scan# 2125
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:232 Resp: 326868
Ion Ratio Lower Upper
232 100
131 41.3 40.0 60.0
130 2.7 1.8 2.8
166 27.2 24.8 37.2









#59
Diethylphthalate
Concen: 136.179 ng/ul
RT: 15.719 min Scan# 2158
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:149 Resp: 1093718
Ion Ratio Lower Upper
149 100
177 24.3 15.8 23.6#
150 13.5 9.7 14.5



#61
Fluorene
Concen: 138.275 ng/ul
RT: 15.960 min Scan# 2199
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:166 Resp: 996286
Ion Ratio Lower Upper
166 100
165 102.4 83.3 124.9
167 15.4 11.8 17.6









#62
4-Chlorophenyl-phenylether
Concen: 153.748 ng/ul
RT: 15.942 min Scan# 2196
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:204 Resp: 586122
Ion Ratio Lower Upper
204 100
206 34.2 25.4 38.2
141 56.6 51.8 77.6



#63
4-Nitroaniline
Concen: 118.327 ng/ul
RT: 15.989 min Scan# 2204
Delta R.T. 0.031 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:138 Resp: 184690
Ion Ratio Lower Upper
138 100
92 56.1 50.0 75.0
108 113.3 101.8 152.8









#66
4,6-Dinitro-2-methylphenol
Concen: 140.891 ng/ul
RT: 16.030 min Scan# 2211
Delta R.T. 0.019 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:198 Resp: 206746
Ion Ratio Lower Upper
198 100
182 3.4 3.0 4.6
77 28.5 23.4 35.2



#67
N-Nitrosodiphenylamine
Concen: 138.841 ng/ul
RT: 16.154 min Scan# 2232
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:169 Resp: 919848
Ion Ratio Lower Upper
169 100
168 72.2 55.0 82.6
167 40.2 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 145.334 ng/ul
RT: 16.835 min Scan# 2348
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:248	Resp:	404576
Ion Ratio	Lower	Upper
248	100	
250	96.5	80.2 120.2
141	69.5	60.6 90.8



#69
Hexachlorobenzene
Concen: 139.825 ng/ul
RT: 16.953 min Scan# 2368
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:284	Resp:	449469
Ion Ratio	Lower	Upper
284	100	
142	30.6	28.7 43.1
249	32.4	25.5 38.3









#70
Atrazine
Concen: 145.741 ng/ul
RT: 17.106 min Scan# 2394
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:200 Resp: 428956
Ion Ratio Lower Upper
200 100
173 24.4 19.3 28.9
215 49.3 34.9 52.3



#71
Pentachlorophenol
Concen: 231.926 ng/ul
RT: 17.294 min Scan# 2426
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:266 Resp: 315691
Ion Ratio Lower Upper
266 100
264 63.6 54.1 81.1
268 63.2 52.6 78.8









#72
Phenanthrene
Concen: 137.565 ng/ul
RT: 17.699 min Scan# 2495
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:178 Resp: 1725773
Ion Ratio Lower Upper
178 100
179 18.5 11.7 17.5#
176 22.7 16.2 24.2



#74
Anthracene
Concen: 131.562 ng/ul
RT: 17.793 min Scan# 2511
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:178 Resp: 1672536
Ion Ratio Lower Upper
178 100
179 19.2 11.4 17.0#
176 22.6 12.8 19.2#









#75
1,2,3,4-Tetrachlorobenzene
Concen: 161.432 ng/uL
RT: 13.715 min Scan# 1817
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:216 Resp: 545024
Ion Ratio Lower Upper
216 100
214 78.9 70.9 106.3
218 47.7 40.9 61.3



#76
Pentachlorobenzene
Concen: 148.538 ng/uL
RT: 15.225 min Scan# 2074
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:250 Resp: 531388
Ion Ratio Lower Upper
250 100
248 63.7 51.5 77.3
252 64.0 47.9 71.9









#77
 Carbazole
 Concen: 139.304 ng/ul
 RT: 18.057 min Scan# 2556
 Delta R.T. 0.007 min
 Lab File: BG050269.D
 Acq: 25 Sep 2021 18:41




Instrument :
 BNA_G
 ClientSampleId :
 SSTD160438


Tgt Ion:167 Resp: 1575854
 Ion Ratio Lower Upper
 167 100
 166 24.8 17.8 26.8
 139 14.9 11.1 16.7



#78
 Di-n-butylphthalate
 Concen: 122.006 ng/ul
 RT: 18.598 min Scan# 2648
 Delta R.T. 0.001 min
 Lab File: BG050269.D
 Acq: 25 Sep 2021 18:41

Tgt Ion:149 Resp: 1765164
 Ion Ratio Lower Upper
 149 100
 150 11.4 8.1 12.1
 104 8.0 5.3 7.9#









#80
Fluoranthene
Concen: 130.139 ng/ul
RT: 19.697 min Scan# 2835
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:202 Resp: 2060273
Ion Ratio Lower Upper
202 100
101 9.4 4.3 6.5#
100 7.9 3.5 5.3#



#82
Pyrene
Concen: 125.775 ng/ul
RT: 20.061 min Scan# 2897
Delta R.T. 0.013 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:202 Resp: 1973098
Ion Ratio Lower Upper
202 100
101 11.1 5.4 8.0#
100 10.1 5.0 7.6#









#83
Butylbenzylphthalate
Concen: 132.061 ng/ul
RT: 20.930 min Scan# 3045
Delta R.T. 0.001 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:149 Resp: 817555
Ion Ratio Lower Upper
149 100
91 79.6 60.6 90.8
206 24.8 16.8 25.2



#84
3,3'-Dichlorobenzidine
Concen: 122.545 ng/ul
RT: 21.853 min Scan# 3202
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:252 Resp: 689646
Ion Ratio Lower Upper
252 100
254 65.8 54.2 81.2
126 8.1 5.4 8.0#









#85
Benzo(a)anthracene
Concen: 134.271 ng/ul
RT: 21.947 min Scan# 3218
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:228 Resp: 1986259
Ion Ratio Lower Upper
228 100
229 24.0 16.2 24.4
226 33.3 21.0 31.4#



#86
Bis(2-ethylhexyl)phthalate
Concen: 131.256 ng/ul
RT: 21.835 min Scan# 3199
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:149 Resp: 1115920
Ion Ratio Lower Upper
149 100
167 30.4 20.4 30.6
279 5.8 3.7 5.5#









#87
Chrysene
Concen: 142.042 ng/ul
RT: 21.947 min Scan# 3218
Delta R.T. -0.063 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Instrument :
BNA_G
ClientSampleId :
SSTD160438




Tgt Ion:228 Resp: 1986259


Ion	Ratio	Lower	Upper
228	100		
226	33.3	23.0	34.4
229	24.0	17.8	26.6



#89
Di-n-octyl phthalate
Concen: 121.379 ng/ul
RT: 23.134 min Scan# 3420
Delta R.T. 0.007 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:149 Resp: 1744236









#90
Benzo(b)fluoranthene
Concen: 134.702 ng/ul
RT: 24.326 min Scan# 3623
Delta R.T. 0.025 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41




Instrument :
BNA_G
ClientSampleId :
SSTD160438


Tgt Ion:252 Resp: 2087197
Ion Ratio Lower Upper
252 100
253 24.2 18.1 27.1
125 6.3 3.4 5.0#



#91
Benzo(k)fluoranthene
Concen: 126.262 ng/ul
RT: 24.403 min Scan# 3636
Delta R.T. 0.031 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41


Tgt Ion:252 Resp: 1859657
Ion Ratio Lower Upper
252 100
253 24.9 17.7 26.5
125 6.7 4.1 6.1#








#93
Benzo(a)pyrene
Concen: 147.149 ng/ul
RT: 25.278 min Scan# 3785
Delta R.T. 0.037 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Instrument :
BNA_G
ClientSampleId :
SSTD160438







Tgt Ion:252 Resp: 1980147
Ion Ratio Lower Upper
252 100
253 23.6 17.0 25.4
125 7.1 4.4 6.6#



#94
Indeno(1,2,3-cd)pyrene
Concen: 118.850 ng/ul
RT: 29.426 min Scan# 4491
Delta R.T. 0.072 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:276 Resp: 2109699
Ion Ratio Lower Upper
276 100
138 10.7 5.5 8.3#
277 24.2 18.6 27.8









#95
Dibenzo(a,h)anthracene
Concen: 133.308 ng/ul
RT: 29.509 min Scan# 4505
Delta R.T. 0.066 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Instrument :
BNA_G
ClientSampleId :
SSTD160438

Tgt Ion:278 Resp: 1980949
Ion Ratio Lower Upper
278 100
139 8.2 5.2 7.8#
279 24.4 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 135.720 ng/ul
RT: 30.690 min Scan# 4706
Delta R.T. 0.084 min
Lab File: BG050269.D
Acq: 25 Sep 2021 18:41

Tgt Ion:276 Resp: 2025420
Ion Ratio Lower Upper
276 100
138 11.8 8.6 13.0
277 24.2 17.4 26.0

