




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092622\
Data File : BG055022.D
Acq On : 26 Sep 2022 9:33
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 27 06:38:43 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG091922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 19 17:41:40 2022
Response via : Initial Calibration



TIC: BG055022.D\data.ms

(70) Pentachlorophenol (C)




17.102min (-0.007) 3592.01 ng

response 93230

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.00	63.42
264.00	61.70	62.93
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092622\
Data File : BG055022.D
Acq On : 26 Sep 2022 9:33
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 27 06:38:43 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG091922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 19 17:41:40 2022
Response via : Initial Calibration



TIC: BG055022.D\data.ms

(77) Benzidine

19.675min (-0.007) 0.00 ng

response 529045

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.84
183.00	12.30	12.59
0.00	0.00	0.00