

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092721\
 Data File : BG050277.D
 Acq On : 27 Sep 2021 14:30
 Operator : CG/JU
 Sample : M3919-14DL 10X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 GB6Z3DL

Manual Integrations
 APPROVED

mohammad
 9/28/2021 1:39:43 PM

Quant Time: Sep 27 15:11:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG092521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 27 03:10:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.288	152	34934	20.000	ng/u1	0.00
20) Naphthalene-d8	11.120	136	133613	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.910	164	93105	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.653	188	213378	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.960	240	239977	20.000	ng/u1	0.00
88) Perylene-d12	25.403	264	239970	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0	0.000	ng/uL	
4) Pyridine-d5	4.105	84	3369m	1.547	ng/u1	0.00
7) Phenol-d5	7.418	99	2540	0.916	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.601	67	5885	3.655	ng/u1	0.00
11) 2-Chlorophenol-d4	7.812	132	5513	2.735	ng/u1	0.00
15) 4-Methylphenol-d8	8.981	113	4528	2.072	ng/u1	0.00
21) Nitrobenzene-d5	9.457	128	3210	3.889	ng/u1	0.00
24) 2-Nitrophenol-d4	10.186	143	3186	3.495	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.720	165	6220	2.925	ng/u1	0.00
31) 4-Chloroaniline-d4	11.237	131	8824	3.158	ng/u1	0.00
46) Dimethylphthalate-d6	14.304	166	23694	3.744	ng/u1	0.00
49) Acenaphthylene-d8	14.598	160	29279	3.681	ng/u1	0.00
54) 4-Nitrophenol-d4	0.000	143	0d	0.000	ng/u1	
60) Fluorene-d10	15.891	176	21890	3.778	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.991	200	3295	3.645	ng/u1	0.00
73) Anthracene-d10	17.747	188	38877m	4.278	ng/u1	0.00
81) Pyrene-d10	20.021	212	48758	3.753	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.156	264	46430	3.795	ng/u1	0.00
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.693	88	28100	30.833	ng/uL#	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092721\
Data File : BG050277.D
Acq On : 27 Sep 2021 14:30
Operator : CG/JU
Sample : M3919-14DL 10X
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
GB6Z3DL

Manual Integrations
APPROVED

mohammad
9/28/2021 1:39:43 PM


Quant Time: Sep 27 15:11:06 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG092521.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 27 03:10:48 2021
Response via : Initial Calibration





1,4-Dioxane
 Concen: 30.833 ng/uL
 RT: 3.693 min Scan# 35
 Delta R.T. -0.009 min
 Lab File: BG050277.D
 Acq: 27 Sep 2021 14:30

Instrument :
 BNA_G
 ClientSampleID :
 GB6Z3DL



Tgt Ion: 88 Resp: 28100
 Ion Ratio Lower Upper
 88 100
 43 35.5 21.6 32.4#
 58 65.4 52.0 78.0

Manual Integrations
 APPROVED

mohammad
 9/28/2021 1:39:43 PM

