

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092724\
 Data File : BG063075.D
 Acq On : 27 Sep 2024 15:50
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020512

Quant Time: Sep 27 22:55:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG092524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 25 13:16:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.850	152	87735	20.000	ng/u1	0.00
20) Naphthalene-d8	10.635	136	391225	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.483	164	358003	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.233	188	1006724	20.000	ng/u1	0.00
79) Chrysene-d12	21.487	240	1108188	20.000	ng/u1	# 0.00
88) Perylene-d12	24.524	264	1222354	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.314	96	12016	7.895	ng/uL	0.00
4) Pyridine-d5	3.719	84	86465	23.235	ng/u1	0.00
7) Phenol-d5	7.021	99	117789	21.736	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.180	67	49906	23.574	ng/u1	0.00
11) 2-Chlorophenol-d4	7.386	132	119273	22.401	ng/u1	0.00
15) 4-Methylphenol-d8	8.555	113	111363	22.765	ng/u1	0.00
21) Nitrobenzene-d5	9.001	128	58143	20.211	ng/u1	0.00
24) 2-Nitrophenol-d4	9.718	143	74501	20.278	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.259	165	158877	20.719	ng/u1	0.00
31) 4-Chloroaniline-d4	10.776	131	186230	20.697	ng/u1	0.00
46) Dimethylphthalate-d6	13.890	166	540435	18.986	ng/u1	0.00
49) Acenaphthylene-d8	14.178	160	594287	19.702	ng/u1	0.00
54) 4-Nitrophenol-d4	14.689	143	82111	19.127	ng/u1	0.00
60) Fluorene-d10	15.476	176	505880	19.396	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.600	200	117011	16.632	ng/u1	0.00
73) Anthracene-d10	17.233	188	1006724	20.843	ng/u1	-0.10
81) Pyrene-d10	19.630	212	1233861	19.773	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.319	264	1304614	19.928	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.343	88	14053	7.822	ng/uL#	74
5) Pyridine	3.737	79	86945	23.735	ng/u1#	87
6) Benzaldehyde	6.986	77	67479	26.812	ng/u1	93
8) Phenol	7.045	94	119646	21.942	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.274	93	76751	22.621	ng/u1	96
12) 2-Chlorophenol	7.415	128	111947	21.657	ng/u1	94
13) 2-Methylphenol	8.290	108	106815	22.284	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.367	45	62603	24.220	ng/u1	96
16) Acetophenone	8.667	105	164968	21.685	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.649	70	69059	22.199	ng/u1#	93
18) 4-Methylphenol	8.620	108	121649	23.168	ng/u1	98
19) Hexachloroethane	8.931	117	40661	19.877	ng/u1	95
22) Nitrobenzene	9.048	77	108053	19.895	ng/u1#	86
23) Isophorone	9.565	82	211024	20.121	ng/u1#	97
25) 2-Nitrophenol	9.753	139	76123	20.250	ng/u1	93
26) 2,4-Dimethylphenol	9.812	107	131218	19.396	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.053	93	113455	20.697	ng/u1	98
29) 2,4-Dichlorophenol	10.288	162	150223	20.470	ng/u1	97
30) Naphthalene	10.688	128	398645	20.047	ng/u1	99
32) 4-Chloroaniline	10.793	127	164900	19.930	ng/u1	94
33) Hexachlorobutadiene	10.976	225	134540	18.853	ng/u1	96
34) Caprolactam	11.545	113	29656	14.448	ng/u1#	85
35) 4-Chloro-3-methylphenol	11.927	107	137275	21.286	ng/u1	94
36) 2-Methylnaphthalene	12.303	142	320854	19.891	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092724\
 Data File : BG063075.D
 Acq On : 27 Sep 2024 15:50
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020512

Quant Time: Sep 27 22:55:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG092524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 25 13:16:45 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.521	142	325038	19.935	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.674	216	264123	18.992	ng/ul	99
40) Hexachlorocyclopentadiene	12.650	237	126904	16.057	ng/ul	97
41) 2,4,6-Trichlorophenol	12.914	196	163359	20.419	ng/ul	97
42) 2,4,5-Trichlorophenol	12.985	196	179461	20.411	ng/ul	96
43) 1,1'-Biphenyl	13.314	154	465604	19.643	ng/ul	98
44) 2-Chloronaphthalene	13.355	162	384961	20.169	ng/ul	98
45) 2-Nitroaniline	13.561	65	70880	20.634	ng/ul	92
47) Dimethylphthalate	13.937	163	518821	19.006	ng/ul	99
48) 2,6-Dinitrotoluene	14.060	165	113017	20.106	ng/ul	98
50) Acenaphthylene	14.207	152	605618	20.035	ng/ul	98
51) 3-Nitroaniline	14.389	138	103770	21.874	ng/ul	97
52) Acenaphthene	14.548	153	428916	20.371	ng/ul	98
53) 2,4-Dinitrophenol	14.601	184	65982	16.604	ng/ul	95
55) 4-Nitrophenol	14.707	109	92832	17.863	ng/ul	91
56) Dibenzofuran	14.883	168	661878	19.646	ng/ul	97
57) 2,4-Dinitrotoluene	14.848	165	171840	19.188	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.112	232	180553	19.219	ng/ul	97
59) Diethylphthalate	15.306	149	506960	18.772	ng/ul	99
61) Fluorene	15.535	166	551067	19.848	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.529	204	331352	18.993	ng/ul	98
63) 4-Nitroaniline	15.553	138	107443	20.885	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.617	198	120599	16.167	ng/ul#	90
67) N-Nitrosodiphenylamine	15.741	169	490692	19.552	ng/ul	100
68) 4-Bromophenyl-phenylether	16.422	248	244093	19.757	ng/ul	92
69) Hexachlorobenzene	16.546	284	264455	19.211	ng/ul	100
70) Atrazine	16.692	200	231577	18.765	ng/ul	98
71) Pentachlorophenol	16.886	266	124863	14.577	ng/ul	91
72) Phenanthrene	17.274	178	1023418	19.721	ng/ul	99
74) Anthracene	17.368	178	1067918	20.011	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.279	216	282582	19.670	ng/ul	93
76) Pentachlorobenzene	14.806	250	312494	19.965	ng/ul	94
77) Carbazole	17.638	167	869305	20.355	ng/ul	99
78) Di-n-butylphthalate	18.202	149	937238	20.208	ng/ul	99
80) Fluoranthene	19.295	202	1398154	19.990	ng/ul#	99
82) Pyrene	19.660	202	1455565	19.868	ng/ul#	98
83) Butylbenzylphthalate	20.558	149	421200	20.829	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.387	252	527552	20.141	ng/ul#	95
85) Benzo(a)anthracene	21.469	228	1514126	19.672	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.387	149	597068	20.648	ng/ul	94
87) Chrysene	21.534	228	1438675	20.136	ng/ul	98
89) Di-n-octyl phthalate	22.533	149	997136	21.514	ng/ul	100
90) Benzo(b)fluoranthene	23.567	252	1496112	19.681	ng/ul	98
91) Benzo(k)fluoranthene	23.567	252	1496112	19.812	ng/ul	99
93) Benzo(a)pyrene	24.383	252	1380879	19.847	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	27.956	276	1751089	20.443	ng/ul	94
95) Dibenzo(a,h)anthracene	28.003	278	1414388	20.262	ng/ul	99
96) Benzo(g,h,i)perylene	29.025	276	1333574	20.556	ng/ul	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092724\
 Data File : BG063075.D
 Acq On : 27 Sep 2024 15:50
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020512


Quant Time: Sep 27 22:55:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG092524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 25 13:16:45 2024
 Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 7.822 ng/uL
RT: 3.343 min Scan# 41
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Instrument :
BNA_G
ClientSampleId :
SSTD020512







Tgt Ion	Ratio	Lower	Upper
88	100		
43	18.1	16.6	25.0
58	66.5	34.8	52.2



#5
Pyridine
Concen: 23.735 ng/uL
RT: 3.737 min Scan# 110
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion	Ratio	Lower	Upper
79	100		
52	53.7	55.3	82.9
51	39.4	30.5	45.7









#6
Benzaldehyde
Concen: 26.812 ng/ul
RT: 6.986 min Scan# 60
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion: 77 Resp: 67479
Ion Ratio Lower Upper
77 100
105 117.1 97.2 145.8
106 101.0 89.8 134.8



#8
Phenol
Concen: 21.942 ng/ul
RT: 7.045 min Scan# 673
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion: 94 Resp: 119646
Ion Ratio Lower Upper
94 100
65 25.5 20.1 30.1
66 41.8 34.6 51.8









#10
Bis(2-Chloroethyl)ether
Concen: 22.621 ng/ul
RT: 7.274 min Scan# 712
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion: 93	Resp:	76751
Ion	Ratio	Lower Upper
93	100	
63	60.4	47.2 70.8
95	36.1	25.3 37.9



#12
2-Chlorophenol
Concen: 21.657 ng/ul
RT: 7.415 min Scan# 736
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion: 128	Resp:	111947
Ion	Ratio	Lower Upper
128	100	
64	29.2	26.2 39.4
130	29.9	26.6 40.0









#13
2-Methylphenol
Concen: 22.284 ng/ul
RT: 8.290 min Scan# 886
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:108 Resp: 106815
Ion Ratio Lower Upper
108 100
107 85.1 72.6 109.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 24.220 ng/ul
RT: 8.367 min Scan# 898
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion: 45 Resp: 62603
Ion Ratio Lower Upper
45 100
77 25.3 21.7 32.5
79 18.1 13.0 19.6









#16
Acetophenone
Concen: 21.685 ng/ul
RT: 8.667 min Scan# 949
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:105 Resp: 164968
Ion Ratio Lower Upper
105 100
77 68.3 54.2 81.4
51 21.1 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 22.199 ng/ul
RT: 8.649 min Scan# 946
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion: 70 Resp: 69059
Ion Ratio Lower Upper
70 100
42 52.5 37.0 55.4
101 14.7 7.6 11.4#
130 36.2 27.6 41.4









#18
4-Methylphenol
Concen: 23.168 ng/ul
RT: 8.620 min Scan# 941
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:108 Resp: 121649
Ion Ratio Lower Upper
108 100
107 101.9 83.0 124.4



#19
Hexachloroethane
Concen: 19.877 ng/ul
RT: 8.931 min Scan# 994
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:117 Resp: 40661
Ion Ratio Lower Upper
117 100
201 187.0 143.8 215.8
199 116.7 89.1 133.7









#22
Nitrobenzene
Concen: 19.895 ng/ul
RT: 9.048 min Scan# 1013
Delta R.T. 0.008 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion: 77 Resp: 108053
Ion Ratio Lower Upper
77 100
123 59.0 37.6 56.4#
65 14.6 12.0 18.0



#23
Isophorone
Concen: 20.121 ng/ul
RT: 9.565 min Scan# 1102
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion: 82 Resp: 211024
Ion Ratio Lower Upper
82 100
95 9.4 10.0 15.0#
138 24.7 19.5 29.3









#25
2-Nitrophenol
Concen: 20.250 ng/ul
RT: 9.753 min Scan# 1134
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:139 Resp: 76123
Ion Ratio Lower Upper
139 100
65 36.5 25.0 37.4
109 39.4 33.9 50.9



#26
2,4-Dimethylphenol
Concen: 19.396 ng/ul
RT: 9.812 min Scan# 1144
Delta R.T. -0.009 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:107 Resp: 131218
Ion Ratio Lower Upper
107 100
121 46.4 38.3 57.5
122 78.8 63.7 95.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.697 ng/ul
 RT: 10.053 min Scan# 1185
 Delta R.T. 0.003 min
 Lab File: BG063075.D
 Acq: 27 Sep 2024 15:50




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020512


Tgt Ion: 93 Resp: 113455
 Ion Ratio Lower Upper
 93 100
 95 33.3 26.2 39.4
 123 12.7 8.7 13.1



#29
 2,4-Dichlorophenol
 Concen: 20.470 ng/ul
 RT: 10.288 min Scan# 1225
 Delta R.T. -0.003 min
 Lab File: BG063075.D
 Acq: 27 Sep 2024 15:50

Tgt Ion:162 Resp: 150223
 Ion Ratio Lower Upper
 162 100
 164 64.5 52.6 78.8
 98 21.8 20.0 30.0









#30
Naphthalene
Concen: 20.047 ng/ul
RT: 10.688 min Scan# 1293
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:128 Resp: 398645
Ion Ratio Lower Upper
128 100
129 11.3 9.3 13.9
127 13.1 10.7 16.1



#32
4-Chloroaniline
Concen: 19.930 ng/ul
RT: 10.793 min Scan# 1311
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:127 Resp: 164900
Ion Ratio Lower Upper
127 100
129 32.9 29.4 44.0









#33
Hexachlorobutadiene
Concen: 18.853 ng/ul
RT: 10.976 min Scan# 11
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:225 Resp: 134540
Ion Ratio Lower Upper
225 100
223 59.1 47.0 70.6
227 59.7 52.6 79.0



#34
Caprolactam
Concen: 14.448 ng/ul
RT: 11.545 min Scan# 1439
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:113 Resp: 29656
Ion Ratio Lower Upper
113 100
55 93.3 85.4 128.2
56 96.0 63.5 95.3#









#35
4-Chloro-3-methylphenol
Concen: 21.286 ng/ul
RT: 11.927 min Scan# 11
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:107 Resp: 137275
Ion Ratio Lower Upper
107 100
144 30.1 22.2 33.4
142 95.8 71.8 107.6



#36
2-Methylnaphthalene
Concen: 19.891 ng/ul
RT: 12.303 min Scan# 1568
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:142 Resp: 320854
Ion Ratio Lower Upper
142 100
141 92.0 75.8 113.8





#37

1-Methylnaphthalene

Concen: 19.935 ng/ul

RT: 12.521 min Scan# 1

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:142 Resp: 325038

Ion Ratio Lower Upper

142 100

141 91.0 73.4 110.0

116 3.8 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.992 ng/ul

RT: 12.674 min Scan# 1631

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Tgt Ion:216 Resp: 264123




Ion Ratio Lower Upper


216 100

214 76.1 60.4 90.6

179 16.8 14.2 21.4

108 9.0 8.3 12.5





#40

Hexachlorocyclopentadiene

Concen: 16.057 ng/ul

RT: 12.650 min Scan# 1

Delta R.T. -0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:237 Resp: 126904

Ion Ratio Lower Upper

237 100

235 59.5 49.0 73.6

272 12.2 8.7 13.1



#41

2,4,6-Trichlorophenol

Concen: 20.419 ng/ul

RT: 12.914 min Scan# 1672

Delta R.T. -0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50




Tgt Ion:196 Resp: 163359


Ion Ratio Lower Upper

196 100

198 93.5 77.6 116.4

200 30.9 25.1 37.7





#42

2,4,5-Trichlorophenol

Concen: 20.411 ng/ul

RT: 12.985 min Scan# 1

Delta R.T. -0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:196 Resp: 179461

Ion Ratio Lower Upper

196 100

198 89.4 74.0 111.0

200 30.3 22.1 33.1



#43

1,1'-Biphenyl

Concen: 19.643 ng/ul

RT: 13.314 min Scan# 1740

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50




Tgt Ion:154 Resp: 465604


Ion Ratio Lower Upper

154 100

153 39.4 32.6 48.8

76 5.9 4.1 6.1









#44
2-Chloronaphthalene
Concen: 20.169 ng/ul
RT: 13.355 min Scan# 11
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:162 Resp: 384961
Ion Ratio Lower Upper
162 100
164 34.0 27.0 40.6
127 35.9 26.6 40.0



#45
2-Nitroaniline
Concen: 20.634 ng/ul
RT: 13.561 min Scan# 1782
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion: 65 Resp: 70880
Ion Ratio Lower Upper
65 100
92 82.2 67.4 101.0
138 183.3 135.4 203.2









#47
Dimethylphthalate
Concen: 19.006 ng/ul
RT: 13.937 min Scan# 1846
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:163 Resp: 518821
Ion Ratio Lower Upper
163 100
194 4.1 3.5 5.3
164 10.5 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 20.106 ng/ul
RT: 14.060 min Scan# 1867
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:165 Resp: 113017
Ion Ratio Lower Upper
165 100
89 44.7 34.9 52.3
121 17.7 13.6 20.4









#50
Acenaphthylene
Concen: 20.035 ng/ul
RT: 14.207 min Scan# 1892
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:152 Resp: 605618
Ion Ratio Lower Upper
152 100
151 20.0 16.9 25.3
153 12.8 9.4 14.2



#51
3-Nitroaniline
Concen: 21.874 ng/ul
RT: 14.389 min Scan# 1923
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:138 Resp: 103770
Ion Ratio Lower Upper
138 100
108 7.6 5.8 8.8
92 77.6 64.2 96.4









#52
Acenaphthene
Concen: 20.371 ng/ul
RT: 14.548 min Scan# 1950
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:153 Resp: 428916
Ion Ratio Lower Upper
153 100
152 47.2 36.9 55.3
154 89.4 70.2 105.4



#53
2,4-Dinitrophenol
Concen: 16.604 ng/ul
RT: 14.601 min Scan# 1959
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:184 Resp: 65982
Ion Ratio Lower Upper
184 100
63 38.3 35.8 53.8
154 71.4 55.4 83.2









#55
4-Nitrophenol
Concen: 17.863 ng/ul
RT: 14.707 min Scan# 1977
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:109 Resp: 92832
Ion Ratio Lower Upper
109 100
139 91.9 66.8 100.2
65 62.6 55.6 83.4



#56
Dibenzofuran
Concen: 19.646 ng/ul
RT: 14.883 min Scan# 2007
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:168 Resp: 661878
Ion Ratio Lower Upper
168 100
139 35.1 26.8 40.2





#57

2,4-Dinitrotoluene

Concen: 19.188 ng/ul

RT: 14.848 min Scan# 2001

Delta R.T. -0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:165 Resp: 171840

Ion Ratio Lower Upper

165 100

63 29.1 25.0 37.4

182 2.3 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.219 ng/ul

RT: 15.112 min Scan# 2046

Delta R.T. -0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Tgt Ion:232 Resp: 180553




Ion Ratio Lower Upper


232 100

131 33.3 25.8 38.6

130 2.2 1.9 2.9

166 23.7 17.6 26.4









#59
Diethylphthalate
Concen: 18.772 ng/ul
RT: 15.306 min Scan# 2079
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:149 Resp: 506960
Ion Ratio Lower Upper
149 100
177 26.9 20.9 31.3
150 14.0 11.4 17.2



#61
Fluorene
Concen: 19.848 ng/ul
RT: 15.535 min Scan# 2118
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:166 Resp: 551067
Ion Ratio Lower Upper
166 100
165 99.9 80.2 120.4
167 12.7 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 18.993 ng/ul
RT: 15.529 min Scan# 2117
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:204 Resp: 331352
Ion Ratio Lower Upper
204 100
206 34.8 27.0 40.6
141 49.9 38.3 57.5



#63
4-Nitroaniline
Concen: 20.885 ng/ul
RT: 15.553 min Scan# 2121
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:138 Resp: 107443
Ion Ratio Lower Upper
138 100
92 34.0 33.5 50.3
108 113.6 85.0 127.6









#66
4,6-Dinitro-2-methylphenol
Concen: 16.167 ng/ul
RT: 15.617 min Scan# 2132
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:198 Resp: 120599
Ion Ratio Lower Upper
198 100
182 3.3 2.2 3.2#
77 18.8 11.3 16.9#



#67
N-Nitrosodiphenylamine
Concen: 19.552 ng/ul
RT: 15.741 min Scan# 2153
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:169 Resp: 490692
Ion Ratio Lower Upper
169 100
168 70.0 55.8 83.8
167 36.2 28.6 42.8









#68
4-Bromophenyl-phenylether
Concen: 19.757 ng/ul
RT: 16.422 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:248 Resp: 244093
Ion Ratio Lower Upper
248 100
250 104.2 75.1 112.7
141 59.1 44.6 66.8



#69
Hexachlorobenzene
Concen: 19.211 ng/ul
RT: 16.546 min Scan# 2290
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:284 Resp: 264455
Ion Ratio Lower Upper
284 100
142 23.6 19.0 28.4
249 33.3 26.3 39.5





#70

Atrazine

Concen: 18.765 ng/ul

RT: 16.692 min Scan# 21

Delta R.T. -0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:200 Resp: 231577

Ion Ratio Lower Upper

200 100

173 22.1 17.9 26.9

215 49.6 40.8 61.2



#71

Pentachlorophenol

Concen: 14.577 ng/ul

RT: 16.886 min Scan# 2348

Delta R.T. -0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50




Tgt Ion:266 Resp: 124863


Ion Ratio Lower Upper

266 100

264 65.3 57.7 86.5

268 62.7 57.3 85.9





#72

Phenanthrene

Concen: 19.721 ng/ul

RT: 17.274 min Scan# 2414

Delta R.T. -0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:178 Resp: 1023418

Ion Ratio Lower Upper

178 100

179 14.9 12.2 18.4

176 19.1 14.9 22.3



#74

Anthracene

Concen: 20.011 ng/ul

RT: 17.368 min Scan# 2430

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50




Tgt Ion:178 Resp: 1067918


Ion Ratio Lower Upper

178 100

179 15.6 12.1 18.1

176 18.4 14.1 21.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.670 ng/ul
RT: 13.279 min Scan# 1734
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:216 Resp: 282582
Ion Ratio Lower Upper
216 100
214 78.5 55.3 82.9
218 46.4 37.7 56.5



#76
Pentachlorobenzene
Concen: 19.965 ng/ul
RT: 14.806 min Scan# 1994
Delta R.T. -0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:250 Resp: 312494
Ion Ratio Lower Upper
250 100
248 58.2 52.7 79.1
252 61.3 47.5 71.3









#77
 Carbazole
 Concen: 20.355 ng/ul
 RT: 17.638 min Scan# 2476
 Delta R.T. -0.003 min
 Lab File: BG063075.D
 Acq: 27 Sep 2024 15:50




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020512


Tgt Ion:167 Resp: 869305
 Ion Ratio Lower Upper
 167 100
 166 23.3 18.5 27.7
 139 11.5 9.7 14.5



#78
 Di-n-butylphthalate
 Concen: 20.208 ng/ul
 RT: 18.202 min Scan# 2572
 Delta R.T. 0.003 min
 Lab File: BG063075.D
 Acq: 27 Sep 2024 15:50

Tgt Ion:149 Resp: 937238
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.1 10.7
 104 5.5 4.3 6.5





#80

Fluoranthene

Concen: 19.990 ng/ul

RT: 19.295 min Scan# 21

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:202 Resp: 1398154

Ion Ratio Lower Upper

202 100

101 3.6 2.7 4.1

100 3.0 1.9 2.9#



#82

Pyrene

Concen: 19.868 ng/ul

RT: 19.660 min Scan# 2820

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50




Tgt Ion:202 Resp: 1455565


Ion Ratio Lower Upper

202 100

101 4.5 3.1 4.7

100 4.0 2.4 3.6#





#83

Butylbenzylphthalate

Concen: 20.829 ng/ul

RT: 20.558 min Scan# 2972

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:149 Resp: 421200

Ion Ratio Lower Upper

149 100

91 53.6 42.2 63.2

206 29.7 24.4 36.6



#84

3,3'-Dichlorobenzidine

Concen: 20.141 ng/ul

RT: 21.387 min Scan# 3114

Delta R.T. -0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50




Tgt Ion:252 Resp: 527552


Ion Ratio Lower Upper

252 100

254 65.4 49.3 73.9

126 6.0 3.7 5.5#





#85

Benzo(a)anthracene

Concen: 19.672 ng/ul

RT: 21.469 min Scan# 3128

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:228 Resp: 1514126

Ion Ratio Lower Upper

228 100

229 19.9 15.5 23.3

226 27.8 22.6 34.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.648 ng/ul

RT: 21.387 min Scan# 3114

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50




Tgt Ion:149 Resp: 597068


Ion Ratio Lower Upper

149 100

167 29.9 20.8 31.2

279 5.7 4.4 6.6





#87

Chrysene

Concen: 20.136 ng/ul

RT: 21.534 min Scan# 3139

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:228 Resp: 1438675

Ion Ratio Lower Upper

228 100

226 30.7 23.9 35.9

229 18.6 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 21.514 ng/ul




RT: 22.533 min Scan# 3309


Delta R.T. -0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Tgt Ion:149 Resp: 997136









#90
Benzo(b)fluoranthene
Concen: 19.681 ng/ul
RT: 23.567 min Scan# 3485
Delta R.T. 0.003 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50




Instrument :
BNA_G
ClientSampleId :
SSTD020512


Tgt Ion:252 Resp: 1496112
Ion Ratio Lower Upper
252 100
253 21.4 17.9 26.9
125 3.1 2.7 4.1



#91
Benzo(k)fluoranthene
Concen: 19.812 ng/ul
RT: 23.567 min Scan# 3485
Delta R.T. -0.062 min
Lab File: BG063075.D
Acq: 27 Sep 2024 15:50

Tgt Ion:252 Resp: 1496112
Ion Ratio Lower Upper
252 100
253 21.4 17.7 26.5
125 3.1 2.6 3.8





#93

Benzo(a)pyrene

Concen: 19.847 ng/ul

RT: 24.383 min Scan# 3

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:252 Resp: 1380879

Ion Ratio Lower Upper

252 100

253 21.8 16.7 25.1

125 4.2 2.6 3.8#



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.443 ng/ul

RT: 27.956 min Scan# 4232

Delta R.T. 0.014 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50




Tgt Ion:276 Resp: 1751089


Ion Ratio Lower Upper

276 100

138 8.3 5.7 8.5

277 24.3 16.9 25.3





#95

Dibenzo(a,h)anthracene

Concen: 20.262 ng/ul

RT: 28.003 min Scan# 41

Delta R.T. 0.003 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Instrument :

BNA_G

ClientSampleId :

SSTD020512





Tgt Ion:278 Resp: 1414388

Ion Ratio Lower Upper

278 100

139 6.4 4.5 6.7

279 22.9 18.6 28.0



#96

Benzo(g,h,i)perylene

Concen: 20.556 ng/ul

RT: 29.025 min Scan# 4414

Delta R.T. 0.008 min

Lab File: BG063075.D

Acq: 27 Sep 2024 15:50

Tgt Ion:276 Resp: 1333574

Ion Ratio Lower Upper

276 100

138 8.2 5.6 8.4

277 21.2 19.2 28.8

