


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG093019\  
 Data File : BG042954.D  
 Acq On : 30 Sep 2019 23:24  
 Operator : JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG092519.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Sep 25 15:35:52 2019



AutoFind: Scans 2514, 2515, 2516; Background Corrected with Scan 2508

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 36.5         | 18298      | PASS                |
| 68             | 69              | 0.00            | 2               | 0.6          | 114        | PASS                |
| 69             | 198             | 0.00            | 100             | 35.4         | 17763      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.0          | 0          | PASS                |
| 127            | 198             | 10              | 80              | 41.3         | 20722      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.3          | 150        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 50138      | PASS                |
| 199            | 198             | 5               | 9               | 7.6          | 3829       | PASS                |
| 275            | 198             | 10              | 60              | 29.8         | 14954      | PASS                |
| 365            | 198             | 1               | 100             | 4.7          | 2340       | PASS                |
| 441            | 198             | 0.01            | 100             | 12.5         | 6267       | PASS                |
| 442            | 198             | 50              | 100             | 75.1         | 37640      | PASS                |
| 443            | 442             | 15              | 24              | 19.6         | 7363       | PASS                |