

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100323\
 Data File : BG059272.D
 Acq On : 3 Oct 2023 20:37
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020582

Quant Time: Oct 03 23:35:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG092723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 28 04:33:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.231	152	57091	20.000	ng/ul	0.00
20) Naphthalene-d8	11.068	136	269960	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.864	164	172186	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.614	188	383714	20.000	ng/ul	0.00
79) Chrysene-d12	21.909	240	349446	20.000	ng/ul	-0.01
88) Perylene-d12	25.381	264	404013	20.000	ng/ul	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.530	96	11071	7.319	ng/uL	0.00
4) Pyridine-d5	3.965	84	82403	18.670	ng/ul	0.00
7) Phenol-d5	7.355	99	112162	19.627	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.555	67	68266	18.700	ng/ul	0.00
11) 2-Chlorophenol-d4	7.749	132	77747	20.239	ng/ul	0.00
15) 4-Methylphenol-d8	8.924	113	84043	19.025	ng/ul	0.00
21) Nitrobenzene-d5	9.429	128	43346	22.754	ng/ul	0.00
24) 2-Nitrophenol-d4	10.146	143	46073	24.636	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.675	165	85949	22.372	ng/ul	0.00
31) 4-Chloroaniline-d4	11.215	131	126158	20.297	ng/ul	0.00
46) Dimethylphthalate-d6	14.259	166	252023	20.244	ng/ul	0.00
49) Acenaphthylene-d8	14.564	160	316742	20.271	ng/ul	0.00
54) 4-Nitrophenol-d4	15.058	143	45566	20.960	ng/ul	0.00
60) Fluorene-d10	15.851	176	232190	19.596	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.963	200	39721	20.139	ng/ul	0.00
73) Anthracene-d10	17.708	188	356825	20.409	ng/ul	-0.01
81) Pyrene-d10	19.981	212	387138	19.783	ng/ul	-0.01
92) Benzo(a)pyrene-d12	25.140	264	394686	20.329	ng/ul	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.571	88	13504	7.598	ng/uL#	68
5) Pyridine	3.988	79	83012	18.521	ng/ul#	90
6) Benzaldehyde	7.379	77	52607	20.962	ng/ul	92
8) Phenol	7.384	94	109221	19.213	ng/ul#	92
10) Bis(2-Chloroethyl)ether	7.643	93	89138	19.118	ng/ul	91
12) 2-Chlorophenol	7.778	128	81962	20.236	ng/ul	95
13) 2-Methylphenol	8.659	108	82360	19.175	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.736	45	187187	18.161	ng/ul	97
16) Acetophenone	9.071	105	130174	19.027	ng/ul	98
17) N-Nitroso-di-n-propyla...	9.035	70	66036	19.029	ng/ul	92
18) 4-Methylphenol	8.994	108	90760	19.944	ng/ul	99
19) Hexachloroethane	9.312	117	30587	19.169	ng/ul	96
22) Nitrobenzene	9.470	77	93493	19.881	ng/ul	96
23) Isophorone	9.976	82	194071	19.191	ng/ul	99
25) 2-Nitrophenol	10.175	139	48952	23.757	ng/ul	97
26) 2,4-Dimethylphenol	10.211	107	94839	21.525	ng/ul	94
27) Bis(2-Chloroethoxy)met...	10.463	93	130783	20.497	ng/ul	94
29) 2,4-Dichlorophenol	10.704	162	83092	21.758	ng/ul	99
30) Naphthalene	11.121	128	289283	19.986	ng/ul#	93
32) 4-Chloroaniline	11.245	127	121596	20.458	ng/ul	99
33) Hexachlorobutadiene	11.350	225	51882	20.300	ng/ul	92
34) Caprolactam	12.026	113	26355	19.946	ng/ul#	72
35) 4-Chloro-3-methylphenol	12.326	107	84236	20.673	ng/ul	89
36) 2-Methylnaphthalene	12.708	142	186636	19.683	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100323\
 Data File : BG059272.D
 Acq On : 3 Oct 2023 20:37
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 100 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD020582


Quant Time: Oct 03 23:35:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG092723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 28 04:33:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.925	142	196547	20.423	ng/ul#	88
39) 1,2,4,5-Tetrachloroben...	13.054	216	101617	19.712	ng/ul	96
40) Hexachlorocyclopentadiene	13.001	237	57841	18.190	ng/ul	98
41) 2,4,6-Trichlorophenol	13.301	196	65552	20.770	ng/ul	91
42) 2,4,5-Trichlorophenol	13.372	196	75823	22.895	ng/ul	92
43) 1,1'-Biphenyl	13.701	154	272066	19.565	ng/ul	97
44) 2-Chloronaphthalene	13.753	162	208506	19.915	ng/ul	95
45) 2-Nitroaniline	13.971	65	65773	22.271	ng/ul	95
47) Dimethylphthalate	14.306	163	253087	19.848	ng/ul	99
48) 2,6-Dinitrotoluene	14.453	165	51313	23.660	ng/ul	86
50) Acenaphthylene	14.594	152	338535	20.297	ng/ul	98
51) 3-Nitroaniline	14.793	138	55418	23.671	ng/ul#	89
52) Acenaphthene	14.929	153	230716	20.153	ng/ul	95
53) 2,4-Dinitrophenol	14.987	184	34805	28.870	ng/ul#	84
55) 4-Nitrophenol	15.070	109	31749	20.704	ng/ul	94
56) Dibenzofuran	15.258	168	303593	20.006	ng/ul	99
57) 2,4-Dinitrotoluene	15.228	165	74199	23.106	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	15.469	232	63777	21.273	ng/ul#	97
59) Diethylphthalate	15.651	149	247022	20.039	ng/ul	97
61) Fluorene	15.910	166	253066	19.861	ng/ul#	97
62) 4-Chlorophenyl-phenyle...	15.892	204	126704	19.948	ng/ul	96
63) 4-Nitroaniline	15.951	138	46285	21.483	ng/ul	87
66) 4,6-Dinitro-2-methylph...	15.974	198	47019	22.609	ng/ul#	88
67) N-Nitrosodiphenylamine	16.110	169	209978	20.529	ng/ul	98
68) 4-Bromophenyl-phenylether	16.785	248	78477	21.111	ng/ul	94
69) Hexachlorobenzene	16.879	284	87771	20.900	ng/ul	96
70) Atrazine	17.044	200	82189	21.132	ng/ul	98
71) Pentachlorophenol	17.232	266	50062	20.713	ng/ul	92
72) Phenanthrene	17.655	178	428111	20.330	ng/ul	99
74) Anthracene	17.743	178	439278	20.203	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.659	216	100852	19.945	ng/ul	98
76) Pentachlorobenzene	15.152	250	104835	20.997	ng/ul	97
77) Carbazole	18.025	167	362589	20.460	ng/ul	97
78) Di-n-butylphthalate	18.524	149	432713	20.598	ng/ul	99
80) Fluoranthene	19.647	202	482663	20.021	ng/ul	98
82) Pyrene	20.011	202	503826	19.731	ng/ul	98
83) Butylbenzylphthalate	20.869	149	192586	21.387	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.803	252	177790	21.969	ng/ul	99
85) Benzo(a)anthracene	21.891	228	495169	20.129	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.721	149	283076	20.702	ng/ul	98
87) Chrysene	21.961	228	460114	19.804	ng/ul	98
89) Di-n-octyl phthalate	23.013	149	480217	20.437	ng/ul	100
90) Benzo(b)fluoranthene	24.259	252	496768	20.075	ng/ul	99
91) Benzo(k)fluoranthene	24.329	252	516527	20.601	ng/ul	99
93) Benzo(a)pyrene	25.216	252	476534	20.300	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	29.394	276	446872	16.817	ng/ul	97
95) Dibenzo(a,h)anthracene	29.476	278	426095	19.718	ng/ul	97
96) Benzo(g,h,i)perylene	30.675	276	428651	19.839	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 03 23:35:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG092723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 28 04:33:25 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.598 ng/uL
RT: 3.571 min Scan# 31
Delta R.T. 0.005 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion: 88 Resp: 13504
Ion Ratio Lower Upper
88 100
43 39.7 34.5 51.7
58 74.9 96.6 145.0#



#5
Pyridine
Concen: 18.521 ng/uL
RT: 3.988 min Scan# 102
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion: 79 Resp: 83012
Ion Ratio Lower Upper
79 100
52 112.4 94.8 142.2
51 46.7 48.2 72.2#









#6
Benzaldehyde
Concen: 20.962 ng/ul
RT: 7.379 min Scan# 61
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion: 77 Resp: 52607
Ion Ratio Lower Upper
77 100
105 109.9 81.4 122.2
106 104.8 78.3 117.5



#8
Phenol
Concen: 19.213 ng/ul
RT: 7.384 min Scan# 680
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion: 94 Resp: 109221
Ion Ratio Lower Upper
94 100
65 28.7 23.6 35.4
66 38.7 24.6 37.0#









#10
Bis(2-Chloroethyl)ether
Concen: 19.118 ng/ul
RT: 7.643 min Scan# 71
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion: 93 Resp: 89138
Ion Ratio Lower Upper
93 100
63 83.3 76.0 114.0
95 33.0 26.6 39.8



#12
2-Chlorophenol
Concen: 20.236 ng/ul
RT: 7.778 min Scan# 747
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:128 Resp: 81962
Ion Ratio Lower Upper
128 100
64 50.5 42.6 64.0
130 33.4 23.3 34.9









#13
2-Methylphenol
Concen: 19.175 ng/ul
RT: 8.659 min Scan# 897
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:108 Resp: 82360
Ion Ratio Lower Upper
108 100
107 86.8 67.3 100.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.161 ng/ul
RT: 8.736 min Scan# 910
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion: 45 Resp: 187187
Ion Ratio Lower Upper
45 100
77 8.8 6.6 9.8
79 6.2 5.9 8.9









#16
Acetophenone
Concen: 19.027 ng/ul
RT: 9.071 min Scan# 90
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:105 Resp: 130174
Ion Ratio Lower Upper
105 100
77 72.3 58.8 88.2
51 42.5 32.6 48.8



#17
N-Nitroso-di-n-propylamine
Concen: 19.029 ng/ul
RT: 9.035 min Scan# 961
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion: 70 Resp: 66036
Ion Ratio Lower Upper
70 100
42 78.1 69.6 104.4
101 8.9 6.7 10.1
130 23.6 16.8 25.2









#18
4-Methylphenol
Concen: 19.944 ng/ul
RT: 8.994 min Scan# 91
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:108 Resp: 90760
Ion Ratio Lower Upper
108 100
107 103.3 83.8 125.8



#19
Hexachloroethane
Concen: 19.169 ng/ul
RT: 9.312 min Scan# 1008
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:117 Resp: 30587
Ion Ratio Lower Upper
117 100
201 98.6 78.2 117.4
199 60.4 53.5 80.3









#22
Nitrobenzene
Concen: 19.881 ng/ul
RT: 9.470 min Scan# 1035
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion: 77 Resp: 93493
Ion Ratio Lower Upper
77 100
123 45.4 34.6 52.0
65 13.4 13.0 19.4



#23
Isophorone
Concen: 19.191 ng/ul
RT: 9.976 min Scan# 1121
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion: 82 Resp: 194071
Ion Ratio Lower Upper
82 100
95 5.7 4.7 7.1
138 16.7 13.5 20.3









#25
2-Nitrophenol
Concen: 23.757 ng/ul
RT: 10.175 min Scan# 1155
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:139 Resp: 48952
Ion Ratio Lower Upper
139 100
65 54.4 41.2 61.8
109 28.0 22.6 33.8



#26
2,4-Dimethylphenol
Concen: 21.525 ng/ul
RT: 10.211 min Scan# 1161
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:107 Resp: 94839
Ion Ratio Lower Upper
107 100
121 47.2 42.3 63.5
122 89.3 75.2 112.8





#27

Bis(2-Chloroethoxy)methane

Concen: 20.497 ng/ul

RT: 10.463 min Scan# 1204

Delta R.T. -0.001 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion: 93 Resp: 130783

Ion Ratio Lower Upper

93 100

95 29.5 26.5 39.7

123 10.3 7.3 10.9



#29

2,4-Dichlorophenol

Concen: 21.758 ng/ul

RT: 10.704 min Scan# 1245

Delta R.T. -0.001 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37




Tgt Ion:162 Resp: 83092


Ion Ratio Lower Upper

162 100

164 66.1 52.5 78.7

98 36.6 28.9 43.3





#30

Naphthalene

Concen: 19.986 ng/ul

RT: 11.121 min Scan# 11

Delta R.T. -0.007 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion:128 Resp: 289283

Ion Ratio Lower Upper

128 100

129 12.1 8.0 12.0#

127 10.9 11.1 16.7#



#32

4-Chloroaniline

Concen: 20.458 ng/ul

RT: 11.245 min Scan# 1337

Delta R.T. -0.001 min

Lab File: BG059272.D




Acq: 3 Oct 2023 20:37


Tgt Ion:127 Resp: 121596

Ion Ratio Lower Upper

127 100

129 28.8 23.6 35.4









#33
Hexachlorobutadiene
Concen: 20.300 ng/ul
RT: 11.350 min Scan# 11
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:225 Resp: 51882
Ion Ratio Lower Upper
225 100
223 59.4 54.1 81.1
227 62.3 53.5 80.3



#34
Caprolactam
Concen: 19.946 ng/ul
RT: 12.026 min Scan# 1470
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:113 Resp: 26355
Ion Ratio Lower Upper
113 100
55 222.5 211.3 316.9
56 155.3 167.9 251.9#









#35
4-Chloro-3-methylphenol
Concen: 20.673 ng/ul
RT: 12.326 min Scan# 1521
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:107 Resp: 84236
Ion Ratio Lower Upper
107 100
144 28.6 20.2 30.2
142 84.0 58.8 88.2



#36
2-Methylnaphthalene
Concen: 19.683 ng/ul
RT: 12.708 min Scan# 1586
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:142 Resp: 186636
Ion Ratio Lower Upper
142 100
141 89.9 70.2 105.2





#37

1-Methylnaphthalene

Concen: 20.423 ng/ul

RT: 12.925 min Scan# 1

Delta R.T. -0.007 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion:142 Resp: 196547

Ion Ratio Lower Upper

142 100

141 100.7 71.6 107.4

116 3.8 2.3 3.5#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.712 ng/ul

RT: 13.054 min Scan# 1645

Delta R.T. -0.001 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Tgt Ion:216 Resp: 101617




Ion Ratio Lower Upper


216 100

214 73.7 62.4 93.6

179 19.9 16.3 24.5

108 14.7 13.0 19.6





#40

Hexachlorocyclopentadiene

Concen: 18.190 ng/ul

RT: 13.001 min Scan# 1

Delta R.T. -0.007 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion:237 Resp: 57841

Ion Ratio Lower Upper

237 100

235 64.8 50.6 76.0

272 13.1 11.7 17.5



#41

2,4,6-Trichlorophenol

Concen: 20.770 ng/ul

RT: 13.301 min Scan# 1687

Delta R.T. -0.001 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37




Tgt Ion:196 Resp: 65552


Ion Ratio Lower Upper

196 100

198 99.9 72.4 108.6

200 34.2 23.8 35.8









#42
2,4,5-Trichlorophenol
Concen: 22.895 ng/ul
RT: 13.372 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:196 Resp: 75823
Ion Ratio Lower Upper
196 100
198 95.1 82.2 123.4
200 29.3 28.2 42.2



#43
1,1'-Biphenyl
Concen: 19.565 ng/ul
RT: 13.701 min Scan# 1755
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:154 Resp: 272066
Ion Ratio Lower Upper
154 100
153 34.8 29.2 43.8
76 11.6 10.1 15.1









#44
2-Chloronaphthalene
Concen: 19.915 ng/ul
RT: 13.753 min Scan# 11
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:162 Resp: 208506
Ion Ratio Lower Upper
162 100
164 36.7 24.6 37.0
127 35.4 28.1 42.1



#45
2-Nitroaniline
Concen: 22.271 ng/ul
RT: 13.971 min Scan# 1801
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion: 65 Resp: 65773
Ion Ratio Lower Upper
65 100
92 70.7 51.0 76.4
138 103.6 80.5 120.7









#47
Dimethylphthalate
Concen: 19.848 ng/ul
RT: 14.306 min Scan# 1858
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:163 Resp: 253087
Ion Ratio Lower Upper
163 100
194 3.6 2.6 4.0
164 9.7 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 23.660 ng/ul
RT: 14.453 min Scan# 1883
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:165 Resp: 51313
Ion Ratio Lower Upper
165 100
89 53.7 53.0 79.4
121 19.6 19.2 28.8





#50

Acenaphthylene

Concen: 20.297 ng/ul

RT: 14.594 min Scan# 1908

Delta R.T. -0.007 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion:152 Resp: 338535

Ion Ratio Lower Upper

152 100

151 19.3 14.6 21.8

153 14.0 10.9 16.3



#51

3-Nitroaniline

Concen: 23.671 ng/ul

RT: 14.793 min Scan# 1941

Delta R.T. -0.001 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37




Tgt Ion:138 Resp: 55418


Ion Ratio Lower Upper

138 100

108 7.9 10.0 15.0#

92 106.8 95.0 142.6









#52
Acenaphthene
Concen: 20.153 ng/ul
RT: 14.929 min Scan# 1964
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:153 Resp: 230716
Ion Ratio Lower Upper
153 100
152 45.1 38.5 57.7
154 90.8 76.3 114.5



#53
2,4-Dinitrophenol
Concen: 28.870 ng/ul
RT: 14.987 min Scan# 1974
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:184 Resp: 34805
Ion Ratio Lower Upper
184 100
63 80.6 69.9 104.9
154 69.5 38.6 58.0#









#55
4-Nitrophenol
Concen: 20.704 ng/ul
RT: 15.070 min Scan# 1989
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:109 Resp: 31749
Ion Ratio Lower Upper
109 100
139 137.4 115.5 173.3
65 162.6 122.8 184.2



#56
Dibenzofuran
Concen: 20.006 ng/ul
RT: 15.258 min Scan# 2020
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:168 Resp: 303593
Ion Ratio Lower Upper
168 100
139 33.9 26.6 39.8









#57
2,4-Dinitrotoluene
Concen: 23.106 ng/ul
RT: 15.228 min Scan# 2015
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:165 Resp: 74199
Ion Ratio Lower Upper
165 100
63 46.1 43.5 65.3
182 2.6 1.8 2.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.273 ng/ul
RT: 15.469 min Scan# 2056
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:232 Resp: 63777
Ion Ratio Lower Upper
232 100
131 42.0 33.9 50.9
130 3.2 1.5 2.3#
166 28.9 25.4 38.0









#59
Diethylphthalate
Concen: 20.039 ng/ul
RT: 15.651 min Scan# 2087
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:149 Resp: 247022
Ion Ratio Lower Upper
149 100
177 23.7 17.3 25.9
150 12.0 10.0 15.0



#61
Fluorene
Concen: 19.861 ng/ul
RT: 15.910 min Scan# 2131
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:166 Resp: 253066
Ion Ratio Lower Upper
166 100
165 100.7 78.6 118.0
167 11.7 11.8 17.8#









#62
4-Chlorophenyl-phenylether
Concen: 19.948 ng/ul
RT: 15.892 min Scan# 2128
Delta R.T. -0.001 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:204 Resp: 126704
Ion Ratio Lower Upper
204 100
206 35.3 23.7 35.5
141 52.4 41.6 62.4



#63
4-Nitroaniline
Concen: 21.483 ng/ul
RT: 15.951 min Scan# 2138
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:138 Resp: 46285
Ion Ratio Lower Upper
138 100
92 59.1 41.0 61.6
108 83.0 57.0 85.4









#66
4,6-Dinitro-2-methylphenol
Concen: 22.609 ng/ul
RT: 15.974 min Scan# 2142
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:198 Resp: 47019
Ion Ratio Lower Upper
198 100
182 3.6 4.4 6.6#
77 22.0 23.2 34.8#



#67
N-Nitrosodiphenylamine
Concen: 20.529 ng/ul
RT: 16.110 min Scan# 2165
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:169 Resp: 209978
Ion Ratio Lower Upper
169 100
168 69.3 57.0 85.6
167 36.8 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 21.111 ng/ul
RT: 16.785 min Scan# 2111
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:248 Resp: 78477
Ion Ratio Lower Upper
248 100
250 91.1 78.9 118.3
141 64.2 49.6 74.4



#69
Hexachlorobenzene
Concen: 20.900 ng/ul
RT: 16.879 min Scan# 2296
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:284 Resp: 87771
Ion Ratio Lower Upper
284 100
142 29.5 26.8 40.2
249 29.4 24.2 36.4





#70

Atrazine

Concen: 21.132 ng/ul

RT: 17.044 min Scan# 21

Delta R.T. -0.007 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion:200 Resp: 82189

Ion Ratio Lower Upper

200 100

173 23.1 17.4 26.0

215 45.7 35.8 53.6



#71

Pentachlorophenol

Concen: 20.713 ng/ul

RT: 17.232 min Scan# 2356

Delta R.T. -0.007 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37




Tgt Ion:266 Resp: 50062


Ion Ratio Lower Upper

266 100

264 66.6 46.4 69.6

268 60.2 46.0 69.0









#72
Phenanthrene
Concen: 20.330 ng/ul
RT: 17.655 min Scan# 2428
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:178 Resp: 428111
Ion Ratio Lower Upper
178 100
179 16.0 12.3 18.5
176 18.0 14.6 21.8



#74
Anthracene
Concen: 20.203 ng/ul
RT: 17.743 min Scan# 2443
Delta R.T. -0.012 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:178 Resp: 439278
Ion Ratio Lower Upper
178 100
179 14.8 12.0 18.0
176 17.9 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.945 ng/uL

RT: 13.659 min Scan# 1

Delta R.T. -0.007 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion:216 Resp: 100852

Ion Ratio Lower Upper

216 100

214 77.6 64.3 96.5

218 48.1 38.6 57.8



#76

Pentachlorobenzene

Concen: 20.997 ng/uL

RT: 15.152 min Scan# 2002

Delta R.T. -0.007 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37




Tgt Ion:250 Resp: 104835


Ion Ratio Lower Upper

250 100

248 59.1 50.4 75.6

252 68.9 55.8 83.6









#77
 Carbazole
 Concen: 20.460 ng/ul
 RT: 18.025 min Scan# 2491
 Delta R.T. -0.007 min
 Lab File: BG059272.D
 Acq: 3 Oct 2023 20:37




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020582


Tgt Ion:167 Resp: 362589
 Ion Ratio Lower Upper
 167 100
 166 20.0 17.4 26.0
 139 12.0 9.0 13.4



#78
 Di-n-butylphthalate
 Concen: 20.598 ng/ul
 RT: 18.524 min Scan# 2576
 Delta R.T. -0.012 min
 Lab File: BG059272.D
 Acq: 3 Oct 2023 20:37

Tgt Ion:149 Resp: 432713
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.2 10.8
 104 4.8 3.6 5.4





#80

Fluoranthene

Concen: 20.021 ng/ul

RT: 19.647 min Scan# 21

Delta R.T. -0.012 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion:202 Resp: 482663

Ion Ratio Lower Upper

202 100

101 11.4 8.2 12.4

100 10.5 8.2 12.2



#82

Pyrene

Concen: 19.731 ng/ul

RT: 20.011 min Scan# 2829

Delta R.T. -0.012 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37




Tgt Ion:202 Resp: 503826


Ion Ratio Lower Upper

202 100

101 13.8 10.2 15.4

100 12.5 9.6 14.4









#83
Butylbenzylphthalate
Concen: 21.387 ng/ul
RT: 20.869 min Scan# 2975
Delta R.T. -0.007 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:149 Resp: 192586
Ion Ratio Lower Upper
149 100
91 76.8 64.5 96.7
206 20.7 17.0 25.6



#84
3,3'-Dichlorobenzidine
Concen: 21.969 ng/ul
RT: 21.803 min Scan# 3134
Delta R.T. -0.012 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:252 Resp: 177790
Ion Ratio Lower Upper
252 100
254 62.1 50.2 75.2
126 11.0 7.9 11.9









#85
Benzo(a)anthracene
Concen: 20.129 ng/ul
RT: 21.891 min Scan# 3149
Delta R.T. -0.012 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37




Instrument :
BNA_G
ClientSampleId :
SSTD020582


Tgt Ion:228 Resp: 495169
Ion Ratio Lower Upper
228 100
229 19.6 15.5 23.3
226 26.2 21.4 32.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.702 ng/ul
RT: 21.721 min Scan# 3120
Delta R.T. -0.018 min
Lab File: BG059272.D
Acq: 3 Oct 2023 20:37

Tgt Ion:149 Resp: 283076
Ion Ratio Lower Upper
149 100
167 25.1 19.1 28.7
279 3.4 2.9 4.3





#87

Chrysene

Concen: 19.804 ng/ul

RT: 21.961 min Scan# 3161

Delta R.T. -0.012 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion:228 Resp: 460114

Ion Ratio Lower Upper

228 100

226 31.8 24.2 36.4

229 19.4 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 20.437 ng/ul




RT: 23.013 min Scan# 3340


Delta R.T. -0.018 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Tgt Ion:149 Resp: 480217





#90

Benzo(b)fluoranthene

Concen: 20.075 ng/ul

RT: 24.259 min Scan# 3552

Delta R.T. -0.024 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion:252 Resp: 496768

Ion Ratio Lower Upper

252 100

253 22.1 18.0 27.0

125 9.3 8.1 12.1



#91

Benzo(k)fluoranthene

Concen: 20.601 ng/ul

RT: 24.329 min Scan# 3564

Delta R.T. -0.030 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37




Tgt Ion:252 Resp: 516527


Ion Ratio Lower Upper

252 100

253 21.7 17.6 26.4

125 9.7 8.1 12.1





#93

Benzo(a)pyrene

Concen: 20.300 ng/ul

RT: 25.216 min Scan# 3715

Delta R.T. -0.024 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37

Instrument :

BNA_G

ClientSampleId :

SSTD020582





Tgt Ion:252 Resp: 476534

Ion Ratio Lower Upper

252 100

253 21.7 17.2 25.8

125 10.9 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.817 ng/ul

RT: 29.394 min Scan# 4426

Delta R.T. -0.042 min

Lab File: BG059272.D

Acq: 3 Oct 2023 20:37




Tgt Ion:276 Resp: 446872


Ion Ratio Lower Upper

276 100

138 22.2 17.0 25.6

277 21.2 18.6 28.0









#95
 Dibenzo(a,h)anthracene
 Concen: 19.718 ng/ul
 RT: 29.476 min Scan# 4440
 Delta R.T. -0.036 min
 Lab File: BG059272.D
 Acq: 3 Oct 2023 20:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020582

Tgt Ion:278 Resp: 426095
 Ion Ratio Lower Upper
 278 100
 139 15.7 12.1 18.1
 279 21.4 18.6 27.8



#96
 Benzo(g,h,i)perylene
 Concen: 19.839 ng/ul
 RT: 30.675 min Scan# 4644
 Delta R.T. -0.036 min
 Lab File: BG059272.D
 Acq: 3 Oct 2023 20:37

Tgt Ion:276 Resp: 428651
 Ion Ratio Lower Upper
 276 100
 138 20.1 16.6 25.0
 277 23.4 20.0 30.0

