

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG100419\
 Data File : BG043041.D
 Acq On : 4 Oct 2019 23:23
 Operator : JU
 Sample : K5154-14
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MW-103D

Manual Integrations
 APPROVED

mohammad
 10/9/2019 8:31:59 AM

Quant Time: Oct 05 06:43:36 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 25 15:35:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.06	152	76146	20.00	ng	-0.02
21) Naphthalene-d8	10.86	136	286618	20.00	ng	-0.02
39) Acenaphthene-d10	14.68	164	205730	20.00	ng	-0.02
64) Phenanthrene-d10	17.42	188	470328	20.00	ng	-0.02
76) Chrysene-d12	21.69	240	519933	20.00	ng	-0.02
87) Perylene-d12	24.91	264	604630	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.65	112	333328	78.12	ng	-0.01
7) Phenol-d6	7.23	99	317190	51.62	ng	-0.01
23) Nitrobenzene-d5	9.22	82	431285	73.14	ng	-0.01
42) 2,4,6-Tribromophenol	16.17	330	411832	116.41	ng	-0.01
45) 2-Fluorobiphenyl	13.31	172	1012441	71.34	ng	-0.02
79) Terphenyl-d14	20.04	244	1949858	79.91	ng	-0.01
Target Compounds						
10) Phenol	7.26	94	43050	7.637	ng	Qvalue 100
20) 3+4-Methylphenols	8.83	107	21973	4.065	ng	# 82
32) Benzoic acid	10.11	122	2850m	3.885	ng	
35) Caprolactam	11.79	113	7788	4.829	ng	# 63
50) Dimethylphthalate	14.13	163	156750	10.169	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG100419\
Data File : BG043041.D
Acq On : 4 Oct 2019 23:23
Operator : JU
Sample : K5154-14
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
MW-103D

Manual Integrations
APPROVED

mohammad
10/9/2019 8:31:59 AM

Quant Time: Oct 05 06:43:36 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 25 15:35:52 2019
Response via : Initial Calibration

