

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG100419\
 Data File : BG043045.D
 Acq On : 5 Oct 2019 3:35
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Oct 05 06:26:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 25 15:35:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.06	152	63796	20.00	ng	-0.02
21) Naphthalene-d8	10.86	136	237906	20.00	ng	-0.02
39) Acenaphthene-d10	14.68	164	170317	20.00	ng	-0.01
64) Phenanthrene-d10	17.42	188	417972	20.00	ng	-0.01
76) Chrysene-d12	21.69	240	462625	20.00	ng	-0.02
87) Perylene-d12	24.92	264	535904	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.64	112	332431	92.99	ng	-0.02
7) Phenol-d6	7.23	99	312363	60.68	ng	-0.01
23) Nitrobenzene-d5	9.22	82	424136	86.65	ng	-0.01
42) 2,4,6-Tribromophenol	16.17	330	424906	145.08	ng	-0.01
45) 2-Fluorobiphenyl	13.31	172	1025419	87.28	ng	-0.02
79) Terphenyl-d14	20.03	244	2076087	95.63	ng	-0.01
Target Compounds						
10) Phenol	7.25	94	44271	9.373	ng	90
19) n-Nitroso-di-n-propylamine	9.22	70	59955	17.737	ng	# 73
35) Caprolactam	11.77	113	8138	6.079	ng	# 69
50) Dimethylphthalate	14.13	163	178880	14.017	ng	# 96
51) 2,6-Dinitrotoluene	14.22	165	10305	3.792	ng	# 21
57) 2,4-Dinitrotoluene	14.68	165	22896	5.753	ng	# 25
67) 4-Bromophenyl-phenylether	16.17	248	27668	5.274	ng	# 1
77) Benzidine	20.03	184	33220	2.868	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG100419\
Data File : BG043045.D
Acq On : 5 Oct 2019 3:35
Operator : JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040EC

Quant Time: Oct 05 06:26:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 25 15:35:52 2019
Response via : Initial Calibration

