

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100523\
 Data File : BG059327.D
 Acq On : 6 Oct 2023 13:18
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020588

Quant Time: Oct 06 22:29:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG092723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 28 04:33:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.230	152	67402	20.000	ng/u1	0.00
20) Naphthalene-d8	11.074	136	306995	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.863	164	192740	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.613	188	442182	20.000	ng/u1	0.00
79) Chrysene-d12	21.914	240	409970	20.000	ng/u1	0.00
88) Perylene-d12	25.398	264	483035	20.000	ng/u1	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.536	96	13336	7.468	ng/uL	0.01
4) Pyridine-d5	3.964	84	107411	20.613	ng/u1	0.00
7) Phenol-d5	7.360	99	147059	21.797	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.554	67	84387	19.580	ng/u1	0.00
11) 2-Chlorophenol-d4	7.748	132	112596	24.827	ng/u1	0.00
15) 4-Methylphenol-d8	8.929	113	109823	21.058	ng/u1	0.00
21) Nitrobenzene-d5	9.435	128	56462	26.063	ng/u1	0.00
24) 2-Nitrophenol-d4	10.145	143	63760	29.981	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.674	165	116438	26.652	ng/u1	0.00
31) 4-Chloroaniline-d4	11.221	131	166490	23.554	ng/u1	0.00
46) Dimethylphthalate-d6	14.258	166	320015	22.965	ng/u1	0.00
49) Acenaphthylene-d8	14.570	160	403594	23.074	ng/u1	0.00
54) 4-Nitrophenol-d4	15.057	143	60404	24.822	ng/u1	0.00
60) Fluorene-d10	15.851	176	309189	23.312	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.968	200	56227	24.738	ng/u1	0.00
73) Anthracene-d10	17.713	188	483049	23.975	ng/u1	0.00
81) Pyrene-d10	19.987	212	543134	23.657	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.151	264	548311	23.622	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.571	88	15392	7.336	ng/uL#	81
5) Pyridine	3.988	79	104329	19.717	ng/u1	96
6) Benzaldehyde	7.378	77	73834	24.920	ng/u1	97
8) Phenol	7.390	94	144882	21.587	ng/u1	92
10) Bis(2-Chloroethyl)ether	7.648	93	114689	20.836	ng/u1	93
12) 2-Chlorophenol	7.784	128	116549	24.373	ng/u1	91
13) 2-Methylphenol	8.659	108	106344	20.972	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.735	45	230662	18.955	ng/u1#	95
16) Acetophenone	9.070	105	163380	20.228	ng/u1	95
17) N-Nitroso-di-n-propyla...	9.041	70	80098	19.550	ng/u1	92
18) 4-Methylphenol	8.994	108	112911	21.016	ng/u1	97
19) Hexachloroethane	9.311	117	43700	23.197	ng/u1	87
22) Nitrobenzene	9.476	77	123193	23.037	ng/u1	96
23) Isophorone	9.981	82	235648	20.492	ng/u1	99
25) 2-Nitrophenol	10.181	139	65757	28.063	ng/u1	95
26) 2,4-Dimethylphenol	10.204	107	117426	23.437	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.463	93	152553	21.025	ng/u1	99
29) 2,4-Dichlorophenol	10.704	162	110659	25.481	ng/u1	98
30) Naphthalene	11.127	128	383793	23.316	ng/u1	98
32) 4-Chloroaniline	11.244	127	156262	23.118	ng/u1	96
33) Hexachlorobutadiene	11.350	225	72552	24.963	ng/u1	96
34) Caprolactam	12.020	113	33179	22.081	ng/u1	85
35) 4-Chloro-3-methylphenol	12.325	107	110481	23.844	ng/u1	87
36) 2-Methylnaphthalene	12.707	142	249835	23.169	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100523\
 Data File : BG059327.D
 Acq On : 6 Oct 2023 13:18
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020588

Quant Time: Oct 06 22:29:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG092723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 28 04:33:25 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.925	142	254316	23.238	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	13.054	216	131942	22.866	ng/ul	96
40) Hexachlorocyclopentadiene	13.001	237	59486	16.712	ng/ul	93
41) 2,4,6-Trichlorophenol	13.295	196	85770	24.278	ng/ul	94
42) 2,4,5-Trichlorophenol	13.371	196	88982	24.003	ng/ul	96
43) 1,1'-Biphenyl	13.700	154	348956	22.418	ng/ul	98
44) 2-Chloronaphthalene	13.753	162	270461	23.077	ng/ul	98
45) 2-Nitroaniline	13.970	65	83902	25.380	ng/ul	97
47) Dimethylphthalate	14.305	163	314323	22.022	ng/ul	98
48) 2,6-Dinitrotoluene	14.452	165	67573	27.835	ng/ul	88
50) Acenaphthylene	14.599	152	433622	23.226	ng/ul	97
51) 3-Nitroaniline	14.793	138	71867	27.424	ng/ul#	97
52) Acenaphthene	14.928	153	290272	22.651	ng/ul	96
53) 2,4-Dinitrophenol	14.993	184	26323	19.506	ng/ul	88
55) 4-Nitrophenol	15.069	109	39548	23.039	ng/ul	96
56) Dibenzofuran	15.263	168	403524	23.755	ng/ul	100
57) 2,4-Dinitrotoluene	15.234	165	98310	27.350	ng/ul#	86
58) 2,3,4,6-Tetrachlorophenol	15.475	232	83163	24.781	ng/ul#	92
59) Diethylphthalate	15.651	149	306546	22.216	ng/ul	95
61) Fluorene	15.909	166	332112	23.285	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.892	204	169524	23.843	ng/ul	98
63) 4-Nitroaniline	15.956	138	64154	26.601	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.986	198	56548	23.596	ng/ul#	87
67) N-Nitrosodiphenylamine	16.109	169	273756	23.225	ng/ul	99
68) 4-Bromophenyl-phenylether	16.791	248	100005	23.345	ng/ul	98
69) Hexachlorobenzene	16.885	284	116546	24.082	ng/ul	96
70) Atrazine	17.043	200	107956	24.087	ng/ul#	95
71) Pentachlorophenol	17.237	266	68008	24.417	ng/ul#	90
72) Phenanthrene	17.660	178	566663	23.352	ng/ul	99
74) Anthracene	17.748	178	589103	23.511	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.659	216	137107	23.529	ng/ul	95
76) Pentachlorobenzene	15.157	250	133776	23.251	ng/ul	97
77) Carbazole	18.024	167	498318	24.401	ng/ul	99
78) Di-n-butylphthalate	18.530	149	569844	23.539	ng/ul	99
80) Fluoranthene	19.652	202	678571	23.992	ng/ul	98
82) Pyrene	20.016	202	697813	23.293	ng/ul	99
83) Butylbenzylphthalate	20.868	149	267987	25.367	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.808	252	246527	25.966	ng/ul	96
85) Benzo(a)anthracene	21.891	228	695706	24.106	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.720	149	395247	24.638	ng/ul	98
87) Chrysene	21.967	228	631722	23.176	ng/ul	100
89) Di-n-octyl phthalate	23.013	149	668532	23.797	ng/ul	100
90) Benzo(b)fluoranthene	24.270	252	684130	23.124	ng/ul	97
91) Benzo(k)fluoranthene	24.346	252	688835	22.979	ng/ul	99
93) Benzo(a)pyrene	25.228	252	648612	23.111	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	29.423	276	741712	23.346	ng/ul	97
95) Dibenzo(a,h)anthracene	29.505	278	596126	23.074	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100523\
 Data File : BG059327.D
 Acq On : 6 Oct 2023 13:18
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020588

Quant Time: Oct 06 22:29:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG092723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 28 04:33:25 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.336 ng/uL
RT: 3.571 min Scan# 31
Delta R.T. 0.005 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion: 88 Resp: 15392
Ion Ratio Lower Upper
88 100
43 30.0 34.5 51.7#
58 100.3 96.6 145.0



#5
Pyridine
Concen: 19.717 ng/uL
RT: 3.988 min Scan# 102
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion: 79 Resp: 104329
Ion Ratio Lower Upper
79 100
52 115.1 94.8 142.2
51 55.2 48.2 72.2









#6
Benzaldehyde
Concen: 24.920 ng/ul
RT: 7.378 min Scan# 61
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion: 77 Resp: 73834
Ion Ratio Lower Upper
77 100
105 100.1 81.4 122.2
106 93.2 78.3 117.5



#8
Phenol
Concen: 21.587 ng/ul
RT: 7.390 min Scan# 681
Delta R.T. 0.005 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion: 94 Resp: 144882
Ion Ratio Lower Upper
94 100
65 26.4 23.6 35.4
66 36.8 24.6 37.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.836 ng/ul
 RT: 7.648 min Scan# 725
 Delta R.T. -0.001 min
 Lab File: BG059327.D
 Acq: 6 Oct 2023 13:18

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020588

Tgt Ion: 93 Resp: 114689




Ion	Ratio	Lower	Upper
93	100		
63	86.2	76.0	114.0
95	32.3	26.6	39.8




#12
 2-Chlorophenol
 Concen: 24.373 ng/ul
 RT: 7.784 min Scan# 748
 Delta R.T. -0.001 min
 Lab File: BG059327.D
 Acq: 6 Oct 2023 13:18

Tgt Ion:128 Resp: 116549

Ion	Ratio	Lower	Upper
128	100		
64	44.7	42.6	64.0
130	27.5	23.3	34.9









#13
2-Methylphenol
Concen: 20.972 ng/ul
RT: 8.659 min Scan# 897
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:108 Resp: 106344
Ion Ratio Lower Upper
108 100
107 84.1 67.3 100.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.955 ng/ul
RT: 8.735 min Scan# 910
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion: 45 Resp: 230662
Ion Ratio Lower Upper
45 100
77 10.8 6.6 9.8#
79 6.9 5.9 8.9









#16
Acetophenone
Concen: 20.228 ng/ul
RT: 9.070 min Scan# 968
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:105 Resp: 163380
Ion Ratio Lower Upper
105 100
77 70.2 58.8 88.2
51 35.9 32.6 48.8



#17
N-Nitroso-di-n-propylamine
Concen: 19.550 ng/ul
RT: 9.041 min Scan# 962
Delta R.T. 0.005 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion: 70 Resp: 80098
Ion Ratio Lower Upper
70 100
42 77.8 69.6 104.4
101 8.7 6.7 10.1
130 19.9 16.8 25.2









#18
4-Methylphenol
Concen: 21.016 ng/ul
RT: 8.994 min Scan# 91
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:108 Resp: 112911
Ion Ratio Lower Upper
108 100
107 107.8 83.8 125.8



#19
Hexachloroethane
Concen: 23.197 ng/ul
RT: 9.311 min Scan# 1008
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:117 Resp: 43700
Ion Ratio Lower Upper
117 100
201 115.8 78.2 117.4
199 71.6 53.5 80.3









#22
Nitrobenzene
Concen: 23.037 ng/ul
RT: 9.476 min Scan# 1036
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion: 77 Resp: 123193
Ion Ratio Lower Upper
77 100
123 46.3 34.6 52.0
65 14.6 13.0 19.4



#23
Isophorone
Concen: 20.492 ng/ul
RT: 9.981 min Scan# 1122
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion: 82 Resp: 235648
Ion Ratio Lower Upper
82 100
95 6.7 4.7 7.1
138 17.4 13.5 20.3









#25
2-Nitrophenol
Concen: 28.063 ng/ul
RT: 10.181 min Scan# 1156
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:139 Resp: 65757
Ion Ratio Lower Upper
139 100
65 56.7 41.2 61.8
109 27.4 22.6 33.8



#26
2,4-Dimethylphenol
Concen: 23.437 ng/ul
RT: 10.204 min Scan# 1160
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:107 Resp: 117426
Ion Ratio Lower Upper
107 100
121 50.2 42.3 63.5
122 94.7 75.2 112.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.025 ng/ul
 RT: 10.463 min Scan# 1204
 Delta R.T. -0.001 min
 Lab File: BG059327.D
 Acq: 6 Oct 2023 13:18




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020588


Tgt Ion: 93 Resp: 152553
 Ion Ratio Lower Upper
 93 100
 95 32.6 26.5 39.7
 123 9.6 7.3 10.9



#29
 2,4-Dichlorophenol
 Concen: 25.481 ng/ul
 RT: 10.704 min Scan# 1245
 Delta R.T. -0.001 min
 Lab File: BG059327.D
 Acq: 6 Oct 2023 13:18

Tgt Ion:162 Resp: 110659
 Ion Ratio Lower Upper
 162 100
 164 67.3 52.5 78.7
 98 36.5 28.9 43.3









#30
Naphthalene
Concen: 23.316 ng/ul
RT: 11.127 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:128 Resp: 383793
Ion Ratio Lower Upper
128 100
129 9.8 8.0 12.0
127 12.5 11.1 16.7



#32
4-Chloroaniline
Concen: 23.118 ng/ul
RT: 11.244 min Scan# 1337
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:127 Resp: 156262
Ion Ratio Lower Upper
127 100
129 31.9 23.6 35.4





#33

Hexachlorobutadiene

Concen: 24.963 ng/ul

RT: 11.350 min Scan# 11

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Instrument :

BNA_G

ClientSampleId :

SSTD020588





Tgt Ion:225 Resp: 72552

Ion Ratio Lower Upper

225 100

223 67.3 54.1 81.1

227 61.0 53.5 80.3



#34

Caprolactam

Concen: 22.081 ng/ul

RT: 12.020 min Scan# 1469

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18




Tgt Ion:113 Resp: 33179


Ion Ratio Lower Upper

113 100

55 251.9 211.3 316.9

56 170.9 167.9 251.9









#35
4-Chloro-3-methylphenol
Concen: 23.844 ng/ul
RT: 12.325 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:107 Resp: 110481
Ion Ratio Lower Upper
107 100
144 29.7 20.2 30.2
142 86.1 58.8 88.2



#36
2-Methylnaphthalene
Concen: 23.169 ng/ul
RT: 12.707 min Scan# 1586
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:142 Resp: 249835
Ion Ratio Lower Upper
142 100
141 88.6 70.2 105.2









#37
1-Methylnaphthalene
Concen: 23.238 ng/ul
RT: 12.925 min Scan# 1623
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:142 Resp: 254316
Ion Ratio Lower Upper
142 100
141 91.1 71.6 107.4
116 3.2 2.3 3.5



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.866 ng/ul
RT: 13.054 min Scan# 1645
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:216 Resp: 131942
Ion Ratio Lower Upper
216 100
214 81.5 62.4 93.6
179 21.3 16.3 24.5
108 13.7 13.0 19.6





#40

Hexachlorocyclopentadiene

Concen: 16.712 ng/ul

RT: 13.001 min Scan# 1

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Instrument :

BNA_G

ClientSampleId :

SSTD020588





Tgt Ion:237 Resp: 59486

Ion Ratio Lower Upper

237 100

235 57.0 50.6 76.0

272 13.1 11.7 17.5



#41

2,4,6-Trichlorophenol

Concen: 24.278 ng/ul

RT: 13.295 min Scan# 1686

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18




Tgt Ion:196 Resp: 85770


Ion Ratio Lower Upper

196 100

198 95.5 72.4 108.6

200 33.5 23.8 35.8









#42
2,4,5-Trichlorophenol
Concen: 24.003 ng/ul
RT: 13.371 min Scan# 1699
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:196 Resp: 88982
Ion Ratio Lower Upper
196 100
198 98.9 82.2 123.4
200 33.1 28.2 42.2



#43
1,1'-Biphenyl
Concen: 22.418 ng/ul
RT: 13.700 min Scan# 1755
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:154 Resp: 348956
Ion Ratio Lower Upper
154 100
153 35.5 29.2 43.8
76 11.5 10.1 15.1





#44

2-Chloronaphthalene

Concen: 23.077 ng/ul

RT: 13.753 min Scan# 1

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Instrument :

BNA_G

ClientSampleId :

SSTD020588





Tgt Ion:162 Resp: 270461

Ion Ratio Lower Upper

162 100

164 30.4 24.6 37.0

127 36.6 28.1 42.1



#45

2-Nitroaniline

Concen: 25.380 ng/ul

RT: 13.970 min Scan# 1801

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18




Tgt Ion: 65 Resp: 83902


Ion Ratio Lower Upper

65 100

92 69.2 51.0 76.4

138 100.6 80.5 120.7









#47
Dimethylphthalate
Concen: 22.022 ng/ul
RT: 14.305 min Scan# 1858
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:163 Resp: 314323
Ion Ratio Lower Upper
163 100
194 3.3 2.6 4.0
164 11.1 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 27.835 ng/ul
RT: 14.452 min Scan# 1883
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:165 Resp: 67573
Ion Ratio Lower Upper
165 100
89 54.9 53.0 79.4
121 21.6 19.2 28.8





#50

Acenaphthylene

Concen: 23.226 ng/ul

RT: 14.599 min Scan# 1908

Delta R.T. -0.001 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Instrument :

BNA_G

ClientSampleId :

SSTD020588





Tgt Ion:152 Resp: 433622

Ion Ratio Lower Upper

152 100

151 20.1 14.6 21.8

153 13.7 10.9 16.3



#51

3-Nitroaniline

Concen: 27.424 ng/ul

RT: 14.793 min Scan# 1941

Delta R.T. -0.001 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18




Tgt Ion:138 Resp: 71867


Ion Ratio Lower Upper

138 100

108 8.7 10.0 15.0#

92 116.4 95.0 142.6









#52
Acenaphthene
Concen: 22.651 ng/ul
RT: 14.928 min Scan# 1964
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:153 Resp: 290272
Ion Ratio Lower Upper
153 100
152 46.0 38.5 57.7
154 91.7 76.3 114.5



#53
2,4-Dinitrophenol
Concen: 19.506 ng/ul
RT: 14.993 min Scan# 1975
Delta R.T. 0.005 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:184 Resp: 26323
Ion Ratio Lower Upper
184 100
63 75.7 69.9 104.9
154 55.4 38.6 58.0









#55
4-Nitrophenol
Concen: 23.039 ng/ul
RT: 15.069 min Scan# 1989
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:109 Resp: 39548
Ion Ratio Lower Upper
109 100
139 138.9 115.5 173.3
65 149.2 122.8 184.2



#56
Dibenzofuran
Concen: 23.755 ng/ul
RT: 15.263 min Scan# 2021
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:168 Resp: 403524
Ion Ratio Lower Upper
168 100
139 33.0 26.6 39.8









#57
2,4-Dinitrotoluene
Concen: 27.350 ng/ul
RT: 15.234 min Scan# 2016
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:165 Resp: 98310
Ion Ratio Lower Upper
165 100
63 43.7 43.5 65.3
182 1.5 1.8 2.6#



#58
2,3,4,6-Tetrachlorophenol
Concen: 24.781 ng/ul
RT: 15.475 min Scan# 2057
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:232 Resp: 83163
Ion Ratio Lower Upper
232 100
131 39.7 33.9 50.9
130 2.8 1.5 2.3#
166 24.5 25.4 38.0#









#59
Diethylphthalate
Concen: 22.216 ng/ul
RT: 15.651 min Scan# 2087
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:149 Resp: 306546
Ion Ratio Lower Upper
149 100
177 24.9 17.3 25.9
150 12.6 10.0 15.0



#61
Fluorene
Concen: 23.285 ng/ul
RT: 15.909 min Scan# 2131
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:166 Resp: 332112
Ion Ratio Lower Upper
166 100
165 100.4 78.6 118.0
167 12.7 11.8 17.8









#62
4-Chlorophenyl-phenylether
Concen: 23.843 ng/ul
RT: 15.892 min Scan# 2128
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:204 Resp: 169524
Ion Ratio Lower Upper
204 100
206 30.3 23.7 35.5
141 50.5 41.6 62.4



#63
4-Nitroaniline
Concen: 26.601 ng/ul
RT: 15.956 min Scan# 2139
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:138 Resp: 64154
Ion Ratio Lower Upper
138 100
92 53.2 41.0 61.6
108 82.6 57.0 85.4









#66
4,6-Dinitro-2-methylphenol
Concen: 23.596 ng/ul
RT: 15.986 min Scan# 2143
Delta R.T. 0.005 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:198 Resp: 56548
Ion Ratio Lower Upper
198 100
182 4.6 4.4 6.6
77 21.2 23.2 34.8#



#67
N-Nitrosodiphenylamine
Concen: 23.225 ng/ul
RT: 16.109 min Scan# 2165
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:169 Resp: 273756
Ion Ratio Lower Upper
169 100
168 71.2 57.0 85.6
167 39.2 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 23.345 ng/ul
RT: 16.791 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:248 Resp: 100005
Ion Ratio Lower Upper
248 100
250 98.9 78.9 118.3
141 66.5 49.6 74.4



#69
Hexachlorobenzene
Concen: 24.082 ng/ul
RT: 16.885 min Scan# 2297
Delta R.T. -0.001 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:284 Resp: 116546
Ion Ratio Lower Upper
284 100
142 29.2 26.8 40.2
249 29.8 24.2 36.4





#70

Atrazine

Concen: 24.087 ng/ul

RT: 17.043 min Scan# 21

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Instrument :

BNA_G

ClientSampleId :

SSTD020588





Tgt Ion:200 Resp: 107956

Ion Ratio Lower Upper

200 100

173 26.5 17.4 26.0#

215 46.6 35.8 53.6



#71

Pentachlorophenol

Concen: 24.417 ng/ul

RT: 17.237 min Scan# 2357

Delta R.T. -0.001 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18




Tgt Ion:266 Resp: 68008


Ion Ratio Lower Upper

266 100

264 57.4 46.4 69.6

268 71.2 46.0 69.0#





#72

Phenanthrene

Concen: 23.352 ng/ul

RT: 17.660 min Scan# 2429

Delta R.T. -0.001 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Instrument :

BNA_G

ClientSampleId :

SSTD020588





Tgt Ion:178 Resp: 566663

Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 19.3 14.6 21.8



#74

Anthracene

Concen: 23.511 ng/ul

RT: 17.748 min Scan# 2444

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18




Tgt Ion:178 Resp: 589103


Ion Ratio Lower Upper

178 100

179 16.1 12.0 18.0

176 18.5 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 23.529 ng/uL

RT: 13.659 min Scan# 1

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Instrument :

BNA_G

ClientSampleId :

SSTD020588





Tgt Ion:216 Resp: 137107

Ion Ratio Lower Upper

216 100

214 76.2 64.3 96.5

218 43.9 38.6 57.8



#76

Pentachlorobenzene

Concen: 23.251 ng/uL

RT: 15.157 min Scan# 2003

Delta R.T. -0.001 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18




Tgt Ion:250 Resp: 133776


Ion Ratio Lower Upper

250 100

248 61.4 50.4 75.6

252 72.2 55.8 83.6









#77
 Carbazole
 Concen: 24.401 ng/ul
 RT: 18.024 min Scan# 2492
 Delta R.T. -0.007 min
 Lab File: BG059327.D
 Acq: 6 Oct 2023 13:18




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020588


Tgt Ion:167 Resp: 498318
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.4 26.0
 139 11.8 9.0 13.4



#78
 Di-n-butylphthalate
 Concen: 23.539 ng/ul
 RT: 18.530 min Scan# 2577
 Delta R.T. -0.007 min
 Lab File: BG059327.D
 Acq: 6 Oct 2023 13:18

Tgt Ion:149 Resp: 569844
 Ion Ratio Lower Upper
 149 100
 150 8.5 7.2 10.8
 104 4.9 3.6 5.4









#80
Fluoranthene
Concen: 23.992 ng/ul
RT: 19.652 min Scan# 21
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:202 Resp: 678571
Ion Ratio Lower Upper
202 100
101 10.3 8.2 12.4
100 8.9 8.2 12.2



#82
Pyrene
Concen: 23.293 ng/ul
RT: 20.016 min Scan# 2830
Delta R.T. -0.007 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:202 Resp: 697813
Ion Ratio Lower Upper
202 100
101 12.9 10.2 15.4
100 12.3 9.6 14.4





#83

Butylbenzylphthalate

Concen: 25.367 ng/ul

RT: 20.868 min Scan# 2975

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Instrument :

BNA_G

ClientSampleId :

SSTD020588





Tgt Ion:149 Resp: 267987

Ion Ratio Lower Upper

149 100

91 73.6 64.5 96.7

206 22.2 17.0 25.6



#84

3,3'-Dichlorobenzidine

Concen: 25.966 ng/ul

RT: 21.808 min Scan# 3135

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18




Tgt Ion:252 Resp: 246527


Ion Ratio Lower Upper

252 100

254 66.6 50.2 75.2

126 10.0 7.9 11.9









#85
Benzo(a)anthracene
Concen: 24.106 ng/ul
RT: 21.891 min Scan# 3150
Delta R.T. -0.013 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:228 Resp: 695706
Ion Ratio Lower Upper
228 100
229 18.7 15.5 23.3
226 25.3 21.4 32.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 24.638 ng/ul
RT: 21.720 min Scan# 3120
Delta R.T. -0.019 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:149 Resp: 395247
Ion Ratio Lower Upper
149 100
167 25.1 19.1 28.7
279 3.3 2.9 4.3





#87

Chrysene

Concen: 23.176 ng/ul

RT: 21.967 min Scan# 31

Delta R.T. -0.007 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Instrument :

BNA_G

ClientSampleId :

SSTD020588





Tgt Ion:228 Resp: 631722

Ion Ratio Lower Upper

228 100

226 30.4 24.2 36.4

229 20.1 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 23.797 ng/ul




RT: 23.013 min Scan# 3340


Delta R.T. -0.019 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Tgt Ion:149 Resp: 668532









#90
Benzo(b)fluoranthene
Concen: 23.124 ng/ul
RT: 24.270 min Scan# 3554
Delta R.T. -0.013 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18




Instrument :
BNA_G
ClientSampleId :
SSTD020588


Tgt Ion:252 Resp: 684130
Ion Ratio Lower Upper
252 100
253 20.5 18.0 27.0
125 10.0 8.1 12.1



#91
Benzo(k)fluoranthene
Concen: 22.979 ng/ul
RT: 24.346 min Scan# 3567
Delta R.T. -0.013 min
Lab File: BG059327.D
Acq: 6 Oct 2023 13:18

Tgt Ion:252 Resp: 688835
Ion Ratio Lower Upper
252 100
253 21.5 17.6 26.4
125 9.6 8.1 12.1





#93

Benzo(a)pyrene

Concen: 23.111 ng/ul

RT: 25.228 min Scan# 31

Delta R.T. -0.013 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18

Instrument :

BNA_G

ClientSampleId :

SSTD020588





Tgt Ion:252 Resp: 648612

Ion Ratio Lower Upper

252 100

253 21.7 17.2 25.8

125 10.6 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 23.346 ng/ul

RT: 29.423 min Scan# 4431

Delta R.T. -0.013 min

Lab File: BG059327.D

Acq: 6 Oct 2023 13:18




Tgt Ion:276 Resp: 741712


Ion Ratio Lower Upper

276 100

138 19.1 17.0 25.6

277 24.1 18.6 28.0





#95
 Dibenzo(a,h)anthracene
 Concen: 23.074 ng/ul
 RT: 29.505 min Scan# 44
 Delta R.T. -0.007 min
 Lab File: BG059327.D
 Acq: 6 Oct 2023 13:18

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020588

Tgt Ion:	278	Resp:	596126
Ion Ratio	Lower	Upper	
278	100		
139	16.3	12.1	18.1
279	23.6	18.6	27.8

