

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG100618\
 Data File : BG037191.D
 Acq On : 5 Oct 2018 15:48
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 05 16:49:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 24 10:41:23 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	126	-0.02
2	1,4-Dioxane	0.477	0.599	-25.6#	148	0.00
3	Pyridine	1.378	1.575	-14.3	137	-0.01
4	n-Nitrosodimethylamine	0.612	0.702	-14.7	141	0.00
5 S	2-Fluorophenol	1.043	1.171	-12.3	137	-0.02
6	Aniline	2.094	2.083	0.5	123	-0.02
7 S	Phenol-d6	1.613	1.715	-6.3	129	-0.02
8	2-Chlorophenol	1.211	1.322	-9.2	133	-0.02
9	Benzaldehyde	1.066	1.057	0.8	122	-0.02
10 C	Phenol	1.665	1.807	-8.5	133	-0.02
11	bis(2-Chloroethyl)ether	1.540	1.524	1.0	129	-0.02
12	1,3-Dichlorobenzene	1.485	1.510	-1.7	125	-0.02
13 C	1,4-Dichlorobenzene	1.519	1.504	1.0	124	-0.02
14	1,2-Dichlorobenzene	1.472	1.467	0.3	127	-0.02
15	Benzyl Alcohol	1.309	1.298	0.8	122	-0.02
16	2,2'-oxybis(1-Chloropropane	2.957	3.012	-1.9	125	-0.02
17	2-Methylphenol	1.160	1.165	-0.4	125	-0.02
18	Hexachloroethane	0.559	0.611	-9.3	135	-0.02
19 P	n-Nitroso-di-n-propylamine	1.342	1.309	2.5	121	-0.02
20	3+4-Methylphenols	1.614	1.663	-3.0	126	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	117	-0.02
22	Acetophenone	0.470	0.477	-1.5	121	-0.02
23 S	Nitrobenzene-d5	0.373	0.400	-7.2	124	-0.02
24	Nitrobenzene	0.399	0.416	-4.3	121	-0.02
25	Isophorone	0.682	0.720	-5.6	123	-0.02
26 C	2-Nitrophenol	0.159	0.182	-14.5	129	-0.02
27	2,4-Dimethylphenol	0.248	0.258	-4.0	119	-0.02
28	bis(2-Chloroethoxy)methane	0.421	0.423	-0.5	116	-0.02
29 C	2,4-Dichlorophenol	0.287	0.308	-7.3	120	-0.02
30	1,2,4-Trichlorobenzene	0.359	0.355	1.1	114	-0.02
31	Naphthalene	0.952	0.946	0.6	117	-0.02
32	Benzoic acid	0.175	0.184	-5.1	122	0.00
33	4-Chloroaniline	0.433	0.423	2.3	114	-0.02
34 C	Hexachlorobutadiene	0.248	0.245	1.2	115	-0.02
35	Caprolactam	0.118	0.120	-1.7	115	0.00
36 C	4-Chloro-3-methylphenol	0.343	0.368	-7.3	119	-0.02
37	2-Methylnaphthalene	0.725	0.708	2.3	116	-0.02
38 I	Acenaphthene-d10	1.000	1.000	0.0	110	-0.02
39	1,2,4,5-Tetrachlorobenzene	0.698	0.692	0.9	109	-0.02
40 P	Hexachlorocyclopentadiene	0.359	0.313	12.8	92	-0.02
41 S	2,4,6-Tribromophenol	0.239	0.244	-2.1	110	-0.02
42 C	2,4,6-Trichlorophenol	0.387	0.423	-9.3	117	-0.02
43	2,4,5-Trichlorophenol	0.413	0.464	-12.3	118	-0.02
44 S	2-Fluorobiphenyl	1.343	1.347	-0.3	110	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG100618\
 Data File : BG037191.D
 Acq On : 5 Oct 2018 15:48
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 05 16:49:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 24 10:41:23 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.532	1.540	-0.5	112	-0.02
46	2-Chloronaphthalene	1.205	1.207	-0.2	111	-0.02
47	2-Nitroaniline	0.417	0.463	-11.0	118	-0.02
48	Acenaphthylene	1.888	1.926	-2.0	113	-0.02
49	Dimethylphthalate	1.610	1.581	1.8	108	-0.02
50	2,6-Dinitrotoluene	0.351	0.353	-0.6	110	-0.01
51 C	Acenaphthene	1.141	1.130	1.0	110	-0.02
52	3-Nitroaniline	0.370	0.382	-3.2	111	-0.01
53 P	2,4-Dinitrophenol	0.136	0.152	-11.8	119	-0.01
54	Dibenzofuran	1.930	1.915	0.8	109	-0.02
55 P	4-Nitrophenol	0.236	0.257	-8.9	115	-0.02
56	2,4-Dinitrotoluene	0.490	0.495	-1.0	109	-0.01
57	Fluorene	1.442	1.426	1.1	110	-0.02
58	2,3,4,6-Tetrachlorophenol	0.419	0.439	-4.8	110	-0.02
59	Diethylphthalate	1.624	1.605	1.2	109	-0.02
60	4-Chlorophenyl-phenylether	0.913	0.889	2.6	107	-0.02
61	4-Nitroaniline	0.389	0.372	4.4	103	-0.02
62	Azobenzene	1.560	1.659	-6.3	117	-0.02
63 I	Phenanthrene-d10	1.000	1.000	0.0	106	-0.02
64	4,6-Dinitro-2-methylphenol	0.105	0.116	-10.5	117	-0.01
65 c	n-Nitrosodiphenylamine	0.552	0.569	-3.1	108	-0.02
66	4-Bromophenyl-phenylether	0.227	0.227	0.0	104	-0.02
67	Hexachlorobenzene	0.234	0.227	3.0	101	-0.02
68	Atrazine	0.224	0.214	4.5	99	-0.02
69 C	Pentachlorophenol	0.148	0.141	4.7	98	-0.02
70	Phenanthrene	0.964	0.972	-0.8	107	-0.02
71	Anthracene	0.953	0.981	-2.9	109	-0.02
72	Carbazole	0.929	0.949	-2.2	107	-0.02
73	Di-n-butylphthalate	1.032	1.090	-5.6	111	-0.02
74 C	Fluoranthene	1.160	1.151	0.8	105	-0.02
75 I	Chrysene-d12	1.000	1.000	0.0	107	-0.02
76	Benzidine	0.605	0.694	-14.7	124	-0.02
77	Pyrene	1.200	1.181	1.6	104	-0.02
78 S	Terphenyl-d14	0.761	0.726	4.6	103	-0.02
79	Butylbenzylphthalate	0.478	0.506	-5.9	114	-0.02
80	Benzo(a)anthracene	1.167	1.147	1.7	107	-0.02
81	3,3'-Dichlorobenzidine	0.444	0.443	0.2	104	-0.02
82	Chrysene	1.128	1.106	2.0	107	-0.02
83	Bis(2-ethylhexyl)phthalate	0.668	0.706	-5.7	115	-0.02
84 c	Di-n-octyl phthalate	1.100	1.179	-7.2	114	-0.03
85	Indeno(1,2,3-cd)pyrene	1.211	1.237	-2.1	107	-0.05
86 I	Perylene-d12	1.000	1.000	0.0	110	-0.04
87	Benzo(b)fluoranthene	1.066	1.046	1.9	105	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG100618\
Data File : BG037191.D
Acq On : 5 Oct 2018 15:48
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 05 16:49:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 24 10:41:23 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.069	1.034	3.3	108	-0.03
89 C	Benzo(a)pyrene	1.030	1.015	1.5	107	-0.03
90	Dibenzo(a,h)anthracene	0.966	0.966	0.0	109	-0.05
91	Benzo(a,h,i)perylene	0.969	0.964	0.5	106	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0