



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100620\
 Data File : BG046788.D
 Acq On : 6 Oct 2020 8:04
 Operator : CG/JU
 Sample : DFTPP344
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP344

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM01-EPA-BG092820.M
 Title : SVOA CALIBRATION
 Last Update : Tue Oct 06 10:28:44 2020



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2473

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.4	48408	PASS
68	69	0.00	2	0.5	277	PASS
69	198	0.00	100	38.6	51226	PASS
70	69	0.00	2	0.5	252	PASS
127	198	10	80	43.1	57293	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	132819	PASS
199	198	5	9	7.2	9524	PASS
275	198	10	60	29.1	38624	PASS
365	198	1	100	3.4	4476	PASS
441	198	0.01	100	10.7	14172	PASS
442	198	50	100	70.5	93624	PASS
443	442	15	24	19.3	18104	PASS

m/z	Abundance
39.90	274.0
42.90	216.0
44.00	418.0
48.80	173.0
57.20	194.0
58.10	152.0
66.90	209.0
69.20	161.0
73.00	153.0
77.90	190.0
88.10	153.0
126.80	171.0
132.80	224.0
192.80	157.0
206.90	1295.0
207.80	472.0
210.60	154.0
267.00	154.0
280.80	193.0
354.90	199.0
356.20	221.0

Instrument :
BNA_G
ClientSampleId :
DFTPP344

m/z	Abundance
38.00	512.0
39.10	2982.0
40.00	593.0
41.10	215.0
44.00	476.0
45.20	154.0
49.00	352.0
50.00	10815.0
51.10	41768.0
52.10	2184.0
53.10	262.0
55.10	368.0
56.00	1097.0
57.10	2895.0
58.00	193.0
61.00	649.0
62.20	438.0
63.00	1572.0
65.00	700.0
67.90	830.0
69.00	42736.0
70.10	200.0
71.00	279.0
73.00	607.0
74.00	4207.0
75.00	6709.0
76.10	2306.0
77.10	43384.0
78.10	3072.0
79.00	2784.0
80.00	2531.0
81.00	3782.0
82.00	717.0
82.90	595.0
85.10	569.0
85.90	973.0
87.00	282.0
90.90	302.0
91.10	318.0
92.00	683.0
93.00	5995.0
93.90	531.0
95.90	246.0
98.00	4283.0
98.90	3060.0
101.00	1892.0
101.90	185.0
102.20	178.0
102.90	637.0
104.00	1181.0
104.90	1119.0
105.90	409.0
107.00	13273.0
107.90	1996.0
110.00	23552.0
111.00	3472.0
111.90	319.0
113.00	224.0
116.00	750.0
117.00	11615.0
118.00	898.0
118.80	390.0
121.80	472.0
122.90	1408.0
123.80	358.0
124.10	419.0
125.10	412.0
127.00	43088.0
128.00	3269.0
129.00	17608.0
130.00	1699.0
131.10	576.0
133.90	464.0
135.00	1834.0
136.10	503.0
136.80	616.0
138.30	219.0
141.00	1899.0
142.00	731.0
143.10	557.0
144.00	223.0
145.90	486.0
146.90	1476.0
147.90	2684.0
149.00	574.0
150.90	382.0
151.90	261.0
153.00	927.0
153.90	811.0
154.90	1377.0
155.90	2087.0
157.00	515.0

Instrument :
BNA_G
ClientSampleId :
DFTPP344

157.70	244.0
158.00	311.0
158.90	358.0
160.00	1091.0
160.90	1038.0
162.00	403.0
163.10	265.0
164.90	778.0
166.00	1077.0
167.00	3722.0
168.00	2689.0
168.90	855.0
170.00	259.0
171.00	217.0
171.70	708.0
172.90	710.0
173.90	1053.0
175.00	1880.0
175.90	497.0
176.90	853.0
177.70	453.0
178.90	3769.0
179.90	2768.0
180.80	690.0
182.00	309.0
182.90	154.0
183.90	428.0
185.00	2063.0
185.90	13947.0
186.90	3279.0
188.00	322.0
188.90	903.0
189.90	528.0
190.90	699.0
191.90	1228.0
192.90	1063.0
193.90	200.0
194.70	239.0
195.90	2979.0
197.90	90936.0
198.90	6799.0
199.80	400.0
201.70	342.0
203.00	760.0
203.90	3098.0
205.00	6329.0
205.90	26736.0
206.90	5566.0
207.90	1332.0
208.90	405.0
209.90	561.0
210.80	1306.0
211.80	310.0
213.10	157.0
214.90	351.0
215.80	427.0
216.90	6677.0
217.90	1020.0
219.90	160.0
220.90	4350.0
222.90	1362.0
223.90	17056.0
224.90	4288.0
226.10	372.0
226.90	6689.0
228.00	1003.0
228.90	1455.0
230.90	617.0
231.70	175.0
232.90	297.0
233.80	518.0
234.80	204.0
235.70	165.0
237.00	657.0
238.80	583.0
239.90	339.0
240.80	500.0
241.90	948.0
242.80	1481.0
243.90	11140.0
245.00	1399.0
245.80	2014.0
246.90	847.0
249.00	395.0
251.00	251.0
251.80	160.0
253.10	381.0
254.90	57816.0
255.90	9186.0
256.90	530.0
257.80	3117.0
258.90	667.0
259.80	230.0
260.90	184.0
262.70	274.0
263.80	247.0

Instrument :
BNA_G
ClientSampleId :
DFTPP344

264.90	1713.0
265.70	189.0
271.80	198.0
272.90	1528.0
273.90	4327.0
274.90	23064.0
275.90	2976.0
276.90	1560.0
277.70	331.0
278.90	188.0
280.90	179.0
283.10	218.0
284.90	235.0
285.80	167.0
289.90	323.0
291.80	243.0
292.90	373.0
293.70	198.0
295.90	6522.0
296.80	1345.0
297.70	246.0
301.00	265.0
301.80	248.0
302.80	463.0
303.80	418.0
310.10	276.0
313.90	360.0
314.80	798.0
315.90	310.0
320.90	295.0
322.90	2445.0
323.80	387.0
326.80	504.0
332.70	263.0
333.90	926.0
334.80	405.0
340.90	527.0
345.80	185.0
347.00	153.0
349.60	157.0
351.80	696.0
353.90	422.0
354.90	231.0
364.80	2780.0
365.80	270.0
370.90	259.0
371.80	998.0
372.90	238.0
382.70	462.0
390.80	255.0
401.80	383.0
402.70	411.0
403.90	151.0
420.80	506.0
422.10	266.0
422.80	2851.0
423.80	889.0
440.80	8657.0
441.80	51920.0
442.80	9857.0
443.70	906.0

Instrument :
BNA_G
ClientSampleId :
DFTPP344

m/z	Abundance
37.00	289.0
38.00	579.0
39.10	4005.0
40.00	447.0
40.90	157.0
44.10	530.0
49.00	442.0
50.10	14252.0
51.10	55728.0
52.10	2655.0
53.10	166.0
55.10	252.0
56.00	1779.0
57.00	4139.0
61.00	530.0
62.00	711.0
63.00	2623.0
64.00	326.0
65.10	1009.0
69.00	59688.0
70.00	228.0
73.00	560.0
74.00	6674.0
75.00	9408.0
76.10	2871.0
77.00	64072.0
78.00	5557.0
79.00	5260.0
80.00	4149.0
81.00	5518.0
82.10	1439.0
82.90	772.0
85.10	1291.0
86.00	1466.0
86.90	1019.0
87.90	276.0
91.10	1378.0
92.00	1247.0
93.00	8783.0
93.90	758.0
96.20	185.0
98.00	6377.0
98.90	4797.0
100.00	611.0
100.90	3047.0
101.80	229.0
102.20	193.0
102.90	1322.0
103.90	2079.0
105.00	2246.0
107.00	20352.0
108.00	3106.0
109.00	727.0
110.00	34168.0
111.00	5697.0
112.00	884.0
112.90	236.0
114.30	171.0
117.00	15449.0
118.10	754.0
122.00	1838.0
122.90	2736.0
123.80	1135.0
125.00	626.0
127.00	69232.0
128.00	5523.0
128.90	28800.0
129.90	2434.0
131.10	225.0
133.90	398.0
135.00	2280.0
135.90	987.0
137.00	1306.0
138.10	234.0
138.60	372.0
140.00	357.0
140.90	4131.0
142.00	902.0
143.00	668.0
144.00	380.0
145.00	382.0
146.00	997.0
146.90	2442.0
148.00	3941.0
148.90	714.0
149.80	547.0
151.10	754.0
152.90	1396.0
153.90	729.0
155.00	2443.0
155.90	3545.0
156.60	498.0

Instrument :
BNA_G
ClientSampleId :
DFTPP344

156.90	514.0
157.90	778.0
159.00	438.0
159.90	1582.0
161.00	1916.0
161.80	510.0
162.80	219.0
163.80	162.0
164.80	1981.0
166.00	1869.0
166.90	7848.0
168.00	2819.0
169.00	773.0
170.10	176.0
171.00	455.0
171.80	687.0
173.00	765.0
174.00	1651.0
174.90	3126.0
175.90	302.0
176.20	342.0
176.90	1668.0
177.90	583.0
178.90	7365.0
179.90	4284.0
180.90	2259.0
182.00	502.0
183.00	262.0
183.80	699.0
185.00	3423.0
185.90	23504.0
187.00	6660.0
187.90	562.0
188.90	2490.0
189.80	178.0
190.90	973.0
192.00	1890.0
192.90	2186.0
193.90	553.0
195.80	3673.0
197.90	156544.0
198.90	11285.0
199.90	931.0
201.60	810.0
202.90	1556.0
203.90	6272.0
204.90	11019.0
205.90	46416.0
206.90	6734.0
207.90	2364.0
208.80	800.0
210.00	626.0
211.00	1983.0
214.80	590.0
215.90	1299.0
216.90	14187.0
217.90	1538.0
219.00	240.0
220.80	5983.0
222.80	2431.0
223.90	25440.0
225.00	6508.0
225.90	810.0
226.90	12255.0
227.90	1336.0
229.00	1781.0
229.60	175.0
229.90	184.0
230.80	781.0
231.90	298.0
232.80	292.0
233.80	775.0
234.90	469.0
235.90	758.0
236.90	535.0
238.80	869.0
240.00	152.0
240.80	744.0
241.90	1364.0
243.00	1410.0
243.90	19440.0
245.00	2383.0
245.90	4185.0
246.80	524.0
248.00	193.0
248.90	1283.0
252.90	690.0
254.90	104792.0
255.90	14320.0
256.90	662.0
257.90	5793.0
258.90	1431.0
259.90	334.0
260.90	157.0
263.90	234.0
264.80	2184.0

Instrument :
BNA_G
ClientSampleId :
DFTPP344

265.70	670.0
269.80	192.0
271.00	301.0
271.70	242.0
272.80	2533.0
273.90	8987.0
274.90	44736.0
275.90	6405.0
276.90	3820.0
277.90	681.0
282.90	389.0
283.70	402.0
284.80	979.0
285.90	222.0
289.00	240.0
291.00	209.0
291.90	188.0
292.90	922.0
293.90	543.0
295.90	13531.0
296.70	1794.0
301.00	169.0
301.80	262.0
302.90	1385.0
303.70	491.0
308.00	355.0
308.80	220.0
309.90	156.0
311.90	215.0
313.80	576.0
314.90	1194.0
315.80	452.0
320.80	201.0
321.90	277.0
322.90	3751.0
324.00	714.0
326.80	722.0
327.80	288.0
331.90	296.0
332.90	295.0
333.90	3289.0
334.80	819.0
338.90	246.0
339.80	199.0
340.80	631.0
341.90	204.0
345.90	663.0
346.80	303.0
351.80	1232.0
352.80	610.0
353.90	1167.0
364.80	4736.0
365.80	635.0
369.70	202.0
370.80	360.0
371.80	1570.0
373.00	372.0
376.70	208.0
382.80	351.0
383.90	162.0
388.80	196.0
389.80	288.0
390.80	191.0
401.80	883.0
402.80	1017.0
403.90	239.0
420.70	821.0
421.70	583.0
422.80	5939.0
423.90	1277.0
440.80	15348.0
441.80	102216.0
442.80	20312.0
443.80	2983.0

Instrument :
BNA_G
ClientSampleId :
DFTPP344

m/z	Abundance
37.10	332.0
38.00	548.0
39.10	4027.0
40.10	545.0
40.90	268.0
42.90	248.0
44.00	857.0
50.10	13134.0
51.10	47728.0
52.00	1822.0
55.00	172.0
56.00	1381.0
57.00	3224.0
61.00	773.0
62.10	751.0
63.00	3127.0
63.90	217.0
65.00	675.0
69.00	51736.0
70.00	329.0
72.80	504.0
74.00	4934.0
75.00	8589.0
76.10	3059.0
77.00	56664.0
78.10	3606.0
79.00	3356.0
80.00	3985.0
81.00	4677.0
82.10	702.0
83.00	1690.0
84.10	208.0
85.00	619.0
86.00	1407.0
87.00	820.0
87.90	284.0
88.90	213.0
90.90	1054.0
92.00	1445.0
93.00	6611.0
94.00	628.0
96.00	362.0
98.00	5582.0
99.00	4147.0
100.00	390.0
100.90	2989.0
102.00	172.0
103.00	1090.0
103.90	1462.0
105.00	1534.0
106.10	732.0
107.00	18920.0
108.00	2829.0
110.00	30784.0
111.00	4149.0
111.90	852.0
113.00	198.0
114.10	203.0
115.10	191.0
117.00	14164.0
117.90	1213.0
118.90	284.0
120.20	208.0
122.00	1013.0
123.00	2123.0
124.00	790.0
125.00	540.0
127.00	60072.0
127.90	4594.0
129.00	25488.0
130.00	2041.0
131.00	551.0
134.00	884.0
135.00	1815.0
136.00	1039.0
137.10	927.0
139.00	160.0
139.90	482.0
140.90	3990.0
141.90	807.0
143.00	435.0
143.70	191.0
144.10	185.0
145.80	918.0
146.90	1806.0
147.90	4670.0
149.00	1105.0
150.90	802.0
152.90	1410.0
153.90	1024.0
155.00	1990.0
156.00	3238.0

Instrument :
BNA_G
ClientSampleId :
DFTPP344

156.80	285.0
158.00	978.0
159.00	717.0
159.80	1378.0
161.00	985.0
161.80	483.0
163.00	299.0
165.00	1400.0
165.90	1142.0
166.90	6528.0
168.00	3090.0
169.00	375.0
170.00	388.0
170.90	287.0
171.90	842.0
173.00	875.0
173.90	2153.0
175.00	3403.0
175.90	894.0
176.80	1143.0
178.10	1006.0
178.90	5532.0
179.90	3622.0
180.90	2146.0
182.60	193.0
183.30	289.0
183.90	399.0
185.00	3187.0
185.90	20936.0
186.90	5312.0
188.10	591.0
188.90	1211.0
189.90	589.0
190.90	881.0
192.00	1724.0
192.90	2136.0
193.90	325.0
195.90	4783.0
197.90	150976.0
198.90	10487.0
199.80	794.0
201.20	597.0
201.50	741.0
203.00	1579.0
204.00	7017.0
204.90	11028.0
205.90	46128.0
206.90	7252.0
207.90	1760.0
208.90	705.0
209.80	875.0
210.80	2329.0
211.50	438.0
214.70	480.0
216.90	14050.0
217.80	1770.0
218.80	209.0
219.60	195.0
220.90	6101.0
222.90	2427.0
223.90	25720.0
224.90	6346.0
226.90	11903.0
227.90	1644.0
228.90	2268.0
229.90	222.0
230.90	1162.0
231.70	164.0
232.90	378.0
233.90	778.0
234.90	1731.0
235.80	374.0
236.90	692.0
237.80	263.0
238.80	391.0
240.00	507.0
240.90	572.0
241.90	1378.0
242.90	1753.0
243.90	20704.0
245.00	2079.0
245.90	5004.0
246.80	1354.0
247.80	434.0
248.90	776.0
250.00	161.0
251.00	203.0
251.70	279.0
252.80	792.0
254.90	105088.0
255.90	16520.0
257.00	1708.0
257.90	7374.0
258.80	1149.0
259.80	224.0
263.10	393.0

Instrument :
BNA_G
ClientSampleId :
DFTPP344

263.90	197.0
264.80	2608.0
265.80	484.0
269.90	175.0
270.60	240.0
271.20	219.0
271.90	461.0
272.80	3257.0
273.90	8112.0
274.90	48072.0
275.90	6545.0
276.90	3385.0
277.90	550.0
281.90	195.0
282.80	518.0
283.80	308.0
285.00	899.0
285.70	208.0
288.80	416.0
291.90	492.0
292.80	1199.0
293.80	207.0
294.90	548.0
295.90	12157.0
296.90	1628.0
297.60	197.0
299.70	199.0
300.90	388.0
301.90	202.0
302.90	2045.0
303.90	389.0
307.80	256.0
310.00	284.0
313.90	525.0
314.80	1610.0
315.90	827.0
320.90	647.0
322.90	4693.0
323.90	524.0
326.80	1064.0
327.90	772.0
331.90	230.0
332.80	325.0
333.80	2909.0
335.00	921.0
338.70	154.0
341.00	439.0
345.80	1416.0
351.80	1611.0
352.90	990.0
353.80	1039.0
354.90	321.0
356.70	164.0
360.10	155.0
360.80	159.0
364.80	5912.0
365.70	891.0
370.80	406.0
371.90	1799.0
372.80	555.0
382.80	673.0
383.90	203.0
389.80	382.0
401.80	767.0
402.70	699.0
403.90	232.0
420.60	621.0
421.90	805.0
422.80	6294.0
423.70	1144.0
440.80	18512.0
441.80	126736.0
442.80	24144.0
443.80	2275.0
505.10	178.0

Instrument :
BNA_G
ClientSampleId :
DFTPP344