

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100621\
 Data File : BG050469.D
 Acq On : 6 Oct 2021 15:43
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG100621

Quant Time: Oct 06 16:27:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 06 15:51:01 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.693	0.652	5.9	90	0.00
3	Pyridine	1.892	1.869	1.2	95	0.00
4	n-Nitrosodimethylamine	0.892	0.863	3.3	95	0.00
5 S	2-Fluorophenol	1.337	1.311	1.9	96	0.00
6	Aniline	2.250	2.234	0.7	99	0.00
7 S	Phenol-d6	1.936	1.901	1.8	97	0.00
8	2-Chlorophenol	1.288	1.266	1.7	97	0.00
9	Benzaldehyde	1.429	1.299	9.1	96	0.00
10 C	Phenol	1.883	1.827	3.0	97	0.00
11	bis(2-Chloroethyl)ether	1.469	1.419	3.4	96	0.00
12	1,3-Dichlorobenzene	1.486	1.434	3.5	97	0.00
13 C	1,4-Dichlorobenzene	1.498	1.447	3.4	96	0.00
14	1,2-Dichlorobenzene	1.431	1.377	3.8	97	0.00
15	Benzyl Alcohol	1.530	1.550	-1.3	101	0.00
16	2,2'-oxybis(1-Chloropropane	2.396	2.308	3.7	97	0.00
17	2-Methylphenol	1.239	1.230	0.7	100	0.00
18	Hexachloroethane	0.567	0.558	1.6	97	0.00
19 P	n-Nitroso-di-n-propylamine	1.411	1.366	3.2	96	0.00
20	3+4-Methylphenols	1.744	1.725	1.1	97	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
22	Acetophenone	0.569	0.543	4.6	97	0.00
23 S	Nitrobenzene-d5	0.483	0.475	1.7	97	0.00
24	Nitrobenzene	0.480	0.465	3.1	97	0.00
25	Isophorone	0.857	0.838	2.2	98	0.00
26 C	2-Nitrophenol	0.176	0.178	-1.1	98	0.00
27	2,4-Dimethylphenol	0.305	0.309	-1.3	100	0.00
28	bis(2-Chloroethoxy)methane	0.490	0.484	1.2	99	0.00
29 C	2,4-Dichlorophenol	0.311	0.312	-0.3	99	0.00
30	1,2,4-Trichlorobenzene	0.353	0.345	2.3	98	0.00
31	Naphthalene	1.071	1.047	2.2	98	0.00
32	Benzoic acid	0.226	0.249	-10.2	104	0.00
33	4-Chloroaniline	0.449	0.447	0.4	100	0.00
34 C	Hexachlorobutadiene	0.222	0.217	2.3	99	0.00
35	Caprolactam	0.119	0.123	-3.4	102	0.00
36 C	4-Chloro-3-methylphenol	0.388	0.383	1.3	99	0.00
37	2-Methylnaphthalene	0.739	0.719	2.7	98	0.00
38	1-Methylnaphthalene	0.716	0.699	2.4	99	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	0.591	0.574	2.9	99	0.00
41 P	Hexachlorocyclopentadiene	0.341	0.348	-2.1	99	0.00
42 S	2,4,6-Tribromophenol	0.191	0.190	0.5	100	0.00
43 C	2,4,6-Trichlorophenol	0.416	0.414	0.5	99	0.00
44	2,4,5-Trichlorophenol	0.435	0.445	-2.3	102	0.00
45 S	2-Fluorobiphenyl	1.329	1.302	2.0	100	0.00
46	1,1'-Biphenyl	1.459	1.430	2.0	100	0.00
47	2-Chloronaphthalene	1.120	1.093	2.4	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100621\
 Data File : BG050469.D
 Acq On : 6 Oct 2021 15:43
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG100621

Quant Time: Oct 06 16:27:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 06 15:51:01 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.446	0.450	-0.9	99	0.00
49	Acenaphthylene	1.835	1.805	1.6	100	0.00
50	Dimethylphthalate	1.512	1.488	1.6	100	0.00
51	2,6-Dinitrotoluene	0.303	0.305	-0.7	100	0.00
52 C	Acenaphthene	1.179	1.188	-0.8	101	0.00
53	3-Nitroaniline	0.358	0.366	-2.2	102	0.00
54 P	2,4-Dinitrophenol	0.188	0.195	-3.7	101	0.00
55	Dibenzofuran	1.824	1.776	2.6	100	0.00
56 P	4-Nitrophenol	0.288	0.300	-4.2	104	0.00
57	2,4-Dinitrotoluene	0.427	0.430	-0.7	101	0.00
58	Fluorene	1.401	1.366	2.5	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.379	0.382	-0.8	100	0.00
60	Diethylphthalate	1.588	1.569	1.2	101	0.00
61	4-Chlorophenyl-phenylether	0.767	0.756	1.4	100	0.00
62	4-Nitroaniline	0.373	0.373	0.0	101	0.00
63	Azobenzene	1.839	1.801	2.1	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.120	0.119	0.8	99	0.00
66 c	n-Nitrosodiphenylamine	0.568	0.558	1.8	100	0.00
67	4-Bromophenyl-phenylether	0.202	0.197	2.5	99	0.00
68	Hexachlorobenzene	0.200	0.199	0.5	101	0.00
69	Atrazine	0.224	0.222	0.9	100	0.00
70 C	Pentachlorophenol	0.136	0.141	-3.7	102	0.00
71	Phenanthrene	1.054	1.026	2.7	100	0.00
72	Anthracene	1.047	1.023	2.3	100	0.00
73	Carbazole	1.044	1.022	2.1	101	0.00
74	Di-n-butylphthalate	1.225	1.214	0.9	102	0.00
75 C	Fluoranthene	1.296	1.244	4.0	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzydine	0.648	0.575	11.3	93	0.00
78	Pyrene	1.419	1.378	2.9	99	0.00
79 S	Terphenyl-d14	1.026	0.998	2.7	100	0.00
80	Butylbenzylphthalate	0.600	0.600	0.0	102	0.00
81	Benzo(a)anthracene	1.323	1.280	3.3	101	0.00
82	3,3'-Dichlorobenzidine	0.439	0.425	3.2	99	0.00
83	Chrysene	1.245	1.203	3.4	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.853	0.854	-0.1	102	0.00
85 c	Di-n-octyl phthalate	1.422	1.428	-0.4	101	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	1.393	1.358	2.5	102	0.01
88	Benzo(b)fluoranthene	1.225	1.178	3.8	100	0.00
89	Benzo(k)fluoranthene	1.205	1.178	2.2	102	0.00
90 C	Benzo(a)pyrene	1.041	1.011	2.9	101	0.00
91	Dibenzo(a,h)anthracene	1.152	1.118	3.0	102	0.00
92	Benzo(g,h,i)perylene	1.128	1.100	2.5	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100621\
Data File : BG050469.D
Acq On : 6 Oct 2021 15:43
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG100621

Quant Time: Oct 06 16:27:11 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 06 15:51:01 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0