

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100723\
 Data File : BG059344.D
 Acq On : 7 Oct 2023 16:04
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV406

Quant Time: Oct 07 22:48:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG100723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 07 22:45:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.230	152	41301	20.000	ng/u1	0.00
20) Naphthalene-d8	11.073	136	211822	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.863	164	148647	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.619	188	357319	20.000	ng/u1	0.00
79) Chrysene-d12	21.920	240	310829	20.000	ng/u1	0.00
88) Perylene-d12	25.404	264	372111	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.529	96	10041	8.668	ng/uL	0.00
4) Pyridine-d5	3.964	84	74077	22.218	ng/u1	0.00
7) Phenol-d5	7.354	99	98050	21.501	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.548	67	56003	21.727	ng/u1	0.00
11) 2-Chlorophenol-d4	7.748	132	69205	21.698	ng/u1	0.00
15) 4-Methylphenol-d8	8.923	113	75437	21.231	ng/u1	0.00
21) Nitrobenzene-d5	9.428	128	38061	21.108	ng/u1	0.00
24) 2-Nitrophenol-d4	10.145	143	44202	22.648	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.674	165	78196	22.198	ng/u1	0.00
31) 4-Chloroaniline-d4	11.214	131	123041	21.989	ng/u1	0.00
46) Dimethylphthalate-d6	14.258	166	271179	22.268	ng/u1	0.00
49) Acenaphthylene-d8	14.569	160	317538	21.911	ng/u1	0.00
54) 4-Nitrophenol-d4	15.057	143	47659	20.884	ng/u1	0.00
60) Fluorene-d10	15.850	176	253197	21.782	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.015	200	641	0.272	ng/u1	0.05
73) Anthracene-d10	17.713	188	377759	21.551	ng/u1	0.00
81) Pyrene-d10	19.987	212	422968	22.463	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.157	264	429285	21.923	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.559	88	9868	7.980	ng/uL#	82
5) Pyridine	3.982	79	72316	21.482	ng/u1	92
6) Benzaldehyde	7.372	77	47379	25.046	ng/u1	92
8) Phenol	7.384	94	96852	21.477	ng/u1#	94
10) Bis(2-Chloroethyl)ether	7.642	93	72868	20.982	ng/u1	97
12) 2-Chlorophenol	7.777	128	72549	21.587	ng/u1	93
13) 2-Methylphenol	8.659	108	72168	21.417	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.735	45	157156	21.628	ng/u1	98
16) Acetophenone	9.070	105	116038	21.491	ng/u1	94
17) N-Nitroso-di-n-propyla...	9.029	70	56692	21.072	ng/u1#	95
18) 4-Methylphenol	8.988	108	79958	21.578	ng/u1	95
19) Hexachloroethane	9.311	117	27646	22.615	ng/u1#	75
22) Nitrobenzene	9.469	77	84410	21.353	ng/u1#	87
23) Isophorone	9.975	82	184791	21.892	ng/u1	99
25) 2-Nitrophenol	10.180	139	45349	22.530	ng/u1	94
26) 2,4-Dimethylphenol	10.204	107	85106	21.434	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.457	93	113637	22.403	ng/u1	98
29) 2,4-Dichlorophenol	10.703	162	77655	21.892	ng/u1	93
30) Naphthalene	11.120	128	268084	21.883	ng/u1	97
32) 4-Chloroaniline	11.244	127	113654	21.471	ng/u1	100
33) Hexachlorobutadiene	11.350	225	45870	21.417	ng/u1	98
34) Caprolactam	12.113	113	55	0.043	ng/u1	86
35) 4-Chloro-3-methylphenol	12.325	107	80573	21.168	ng/u1	97
36) 2-Methylnaphthalene	12.707	142	181158	22.181	ng/u1	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100723\
 Data File : BG059344.D
 Acq On : 7 Oct 2023 16:04
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SICV406


Quant Time: Oct 07 22:48:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG100723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 07 22:45:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.924	142	188934	22.281	ng/ul	92
39) 1,2,4,5-Tetrachloroben...	13.054	216	97257	21.965	ng/ul	96
40) Hexachlorocyclopentadiene	13.007	237	56015	21.073	ng/ul	99
41) 2,4,6-Trichlorophenol	13.300	196	66593	22.165	ng/ul	95
42) 2,4,5-Trichlorophenol	13.371	196	71329	21.860	ng/ul	98
43) 1,1'-Biphenyl	13.700	154	269570	22.034	ng/ul	99
44) 2-Chloronaphthalene	13.753	162	199636	21.228	ng/ul	99
45) 2-Nitroaniline	13.976	65	66875	21.716	ng/ul	93
47) Dimethylphthalate	14.305	163	269185	21.994	ng/ul	97
48) 2,6-Dinitrotoluene	14.452	165	54538	22.087	ng/ul	98
50) Acenaphthylene	14.599	152	340562	21.862	ng/ul	98
51) 3-Nitroaniline	14.793	138	56367	23.313	ng/ul	100
52) Acenaphthene	14.928	153	227762	21.609	ng/ul	97
53) 2,4-Dinitrophenol	14.992	184	26631	15.015	ng/ul	92
55) 4-Nitrophenol	15.075	109	34441	22.368	ng/ul	87
56) Dibenzofuran	15.263	168	315707	21.708	ng/ul	96
57) 2,4-Dinitrotoluene	15.233	165	81482	22.532	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	15.474	232	68411	21.850	ng/ul#	96
59) Diethylphthalate	15.650	149	260951	21.437	ng/ul	97
61) Fluorene	15.909	166	271201	21.917	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.891	204	136015	22.245	ng/ul	95
63) 4-Nitroaniline	15.956	138	53174	24.428	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.985	198	52652	20.984	ng/ul#	95
67) N-Nitrosodiphenylamine	16.115	169	226569	22.407	ng/ul	99
68) 4-Bromophenyl-phenylether	16.790	248	81185	21.396	ng/ul	94
69) Hexachlorobenzene	16.884	284	95177	21.976	ng/ul	95
70) Atrazine	17.043	200	89976	22.443	ng/ul	96
71) Pentachlorophenol	17.243	266	51941	20.635	ng/ul	92
72) Phenanthrene	17.660	178	463400	21.898	ng/ul	99
74) Anthracene	17.748	178	478276	22.222	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.659	216	103877	22.377	ng/ul	90
76) Pentachlorobenzene	15.157	250	106301	22.125	ng/ul	93
77) Carbazole	18.024	167	399064	22.280	ng/ul	97
78) Di-n-butylphthalate	18.529	149	462644	21.832	ng/ul	99
80) Fluoranthene	19.652	202	524920	22.404	ng/ul	96
82) Pyrene	20.016	202	550711	22.616	ng/ul	100
83) Butylbenzylphthalate	20.868	149	200070	22.237	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.808	252	187203	24.142	ng/ul	92
85) Benzo(a)anthracene	21.896	228	526835	22.193	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.726	149	296767	22.054	ng/ul#	99
87) Chrysene	21.967	228	489825	21.967	ng/ul	99
89) Di-n-octyl phthalate	23.018	149	504623	21.492	ng/ul	100
90) Benzo(b)fluoranthene	24.270	252	535832	21.855	ng/ul	98
91) Benzo(k)fluoranthene	24.346	252	528754	21.240	ng/ul	98
93) Benzo(a)pyrene	25.233	252	503317	21.551	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	29.423	276	508150	18.889	ng/ul#	91
95) Dibenzo(a,h)anthracene	29.505	278	470450	21.506	ng/ul	97
96) Benzo(g,h,i)perylene	30.703	276	454802	21.198	ng/ul	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 07 22:48:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG100723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 07 22:45:22 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.980 ng/uL
RT: 3.559 min Scan# 29
Delta R.T. -0.008 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion: 88 Resp: 9868
Ion Ratio Lower Upper
88 100
43 51.9 25.4 38.0#
58 93.4 84.6 127.0



#5
Pyridine
Concen: 21.482 ng/uL
RT: 3.982 min Scan# 101
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion: 79 Resp: 72316
Ion Ratio Lower Upper
79 100
52 112.5 82.5 123.7
51 48.5 41.8 62.8









#6
Benzaldehyde
Concen: 25.046 ng/ul
RT: 7.372 min Scan# 61
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion: 77 Resp: 47379
Ion Ratio Lower Upper
77 100
105 88.6 79.5 119.3
106 98.6 74.8 112.2



#8
Phenol
Concen: 21.477 ng/ul
RT: 7.384 min Scan# 680
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion: 94 Resp: 96852
Ion Ratio Lower Upper
94 100
65 34.3 21.9 32.9#
66 35.5 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.982 ng/ul
RT: 7.642 min Scan# 71
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion: 93 Resp: 72868
Ion Ratio Lower Upper
93 100
63 91.9 71.0 106.6
95 36.4 27.8 41.6



#12
2-Chlorophenol
Concen: 21.587 ng/ul
RT: 7.777 min Scan# 747
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:128 Resp: 72549
Ion Ratio Lower Upper
128 100
64 53.4 38.6 57.8
130 30.3 21.8 32.8









#13
2-Methylphenol
Concen: 21.417 ng/ul
RT: 8.659 min Scan# 896
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:108 Resp: 72168
Ion Ratio Lower Upper
108 100
107 80.4 64.7 97.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.628 ng/ul
RT: 8.735 min Scan# 910
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion: 45 Resp: 157156
Ion Ratio Lower Upper
45 100
77 7.3 6.6 10.0
79 7.1 5.8 8.8









#16
Acetophenone
Concen: 21.491 ng/ul
RT: 9.070 min Scan# 90
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:105 Resp: 116038
Ion Ratio Lower Upper
105 100
77 74.2 55.2 82.8
51 37.9 33.0 49.4



#17
N-Nitroso-di-n-propylamine
Concen: 21.072 ng/ul
RT: 9.029 min Scan# 960
Delta R.T. -0.008 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion: 70 Resp: 56692
Ion Ratio Lower Upper
70 100
42 76.7 58.2 87.4
101 9.8 6.4 9.6#
130 21.5 15.4 23.0









#18
4-Methylphenol
Concen: 21.578 ng/ul
RT: 8.988 min Scan# 91
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:108 Resp: 79958
Ion Ratio Lower Upper
108 100
107 108.5 83.0 124.6



#19
Hexachloroethane
Concen: 22.615 ng/ul
RT: 9.311 min Scan# 1008
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:117 Resp: 27646
Ion Ratio Lower Upper
117 100
201 131.8 80.6 120.8#
199 75.3 50.6 76.0









#22
Nitrobenzene
Concen: 21.353 ng/ul
RT: 9.469 min Scan# 1035
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion: 77 Resp: 84410
Ion Ratio Lower Upper
77 100
123 42.7 42.2 63.4
65 15.2 9.8 14.6#



#23
Isophorone
Concen: 21.892 ng/ul
RT: 9.975 min Scan# 1121
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion: 82 Resp: 184791
Ion Ratio Lower Upper
82 100
95 6.2 4.7 7.1
138 17.4 13.5 20.3









#25
2-Nitrophenol
Concen: 22.530 ng/ul
RT: 10.180 min Scan# 1156
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:139 Resp: 45349
Ion Ratio Lower Upper
139 100
65 44.7 39.0 58.4
109 27.6 19.1 28.7



#26
2,4-Dimethylphenol
Concen: 21.434 ng/ul
RT: 10.204 min Scan# 1160
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:107 Resp: 85106
Ion Ratio Lower Upper
107 100
121 54.7 43.9 65.9
122 94.6 74.1 111.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 22.403 ng/ul
 RT: 10.457 min Scan# 11
 Delta R.T. -0.002 min
 Lab File: BG059344.D
 Acq: 7 Oct 2023 16:04

Instrument :
 BNA_G
 ClientSampleId :
 SICV406

Tgt Ion: 93 Resp: 113637




Ion	Ratio	Lower	Upper
93	100		
95	34.3	27.0	40.6
123	8.5	8.5	12.7




#29
 2,4-Dichlorophenol
 Concen: 21.892 ng/ul
 RT: 10.703 min Scan# 1245
 Delta R.T. 0.004 min
 Lab File: BG059344.D
 Acq: 7 Oct 2023 16:04

Tgt Ion:162 Resp: 77655

Ion	Ratio	Lower	Upper
162	100		
164	59.1	53.5	80.3
98	34.7	28.5	42.7









#30
Naphthalene
Concen: 21.883 ng/ul
RT: 11.120 min Scan# 11
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:128 Resp: 268084
Ion Ratio Lower Upper
128 100
129 10.3 9.5 14.3
127 13.6 10.4 15.6



#32
4-Chloroaniline
Concen: 21.471 ng/ul
RT: 11.244 min Scan# 1337
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:127 Resp: 113654
Ion Ratio Lower Upper
127 100
129 32.1 25.9 38.9









#33
Hexachlorobutadiene
Concen: 21.417 ng/ul
RT: 11.350 min Scan# 11
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:225 Resp: 45870
Ion Ratio Lower Upper
225 100
223 63.5 48.2 72.2
227 62.4 49.5 74.3



#34
Caprolactam
Concen: 0.043 ng/ul
RT: 12.113 min Scan# 1485
Delta R.T. 0.092 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:113 Resp: 55
Ion Ratio Lower Upper
113 100
55 239.1 169.1 253.7
56 189.1 141.7 212.5









#35
4-Chloro-3-methylphenol
Concen: 21.168 ng/ul
RT: 12.325 min Scan# 1520
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:107 Resp: 80573
Ion Ratio Lower Upper
107 100
144 26.6 22.0 33.0
142 80.2 61.6 92.4



#36
2-Methylnaphthalene
Concen: 22.181 ng/ul
RT: 12.707 min Scan# 1586
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:142 Resp: 181158
Ion Ratio Lower Upper
142 100
141 89.1 77.8 116.6





#37

1-Methylnaphthalene

Concen: 22.281 ng/ul

RT: 12.924 min Scan# 1

Delta R.T. -0.002 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:142 Resp: 188934

Ion Ratio Lower Upper

142 100

141 85.2 74.3 111.5

116 3.1 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.965 ng/ul

RT: 13.054 min Scan# 1645

Delta R.T. -0.002 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Tgt Ion:216 Resp: 97257




Ion Ratio Lower Upper


216 100

214 75.9 63.0 94.6

179 22.5 15.0 22.6

108 14.2 10.4 15.6









#40
Hexachlorocyclopentadiene
Concen: 21.073 ng/ul
RT: 13.007 min Scan# 1637
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:237 Resp: 56015
Ion Ratio Lower Upper
237 100
235 60.1 48.4 72.6
272 15.1 13.3 19.9



#41
2,4,6-Trichlorophenol
Concen: 22.165 ng/ul
RT: 13.300 min Scan# 1687
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:196 Resp: 66593
Ion Ratio Lower Upper
196 100
198 104.3 88.6 132.8
200 35.7 27.2 40.8









#42
2,4,5-Trichlorophenol
Concen: 21.860 ng/ul
RT: 13.371 min Scan# 1698
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:196 Resp: 71329
Ion Ratio Lower Upper
196 100
198 109.2 87.8 131.8
200 37.5 26.4 39.6



#43
1,1'-Biphenyl
Concen: 22.034 ng/ul
RT: 13.700 min Scan# 1755
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:154 Resp: 269570
Ion Ratio Lower Upper
154 100
153 37.8 30.1 45.1
76 11.5 10.3 15.5





#44

2-Chloronaphthalene

Concen: 21.228 ng/ul

RT: 13.753 min Scan# 11

Delta R.T. -0.002 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:162 Resp: 199636

Ion Ratio Lower Upper

162 100

164 32.4 25.7 38.5

127 37.2 29.4 44.2



#45

2-Nitroaniline

Concen: 21.716 ng/ul

RT: 13.976 min Scan# 1802

Delta R.T. 0.004 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04




Tgt Ion: 65 Resp: 66875


Ion Ratio Lower Upper

65 100

92 69.3 53.8 80.8

138 109.5 80.0 120.0









#47
Dimethylphthalate
Concen: 21.994 ng/ul
RT: 14.305 min Scan# 1858
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:163 Resp: 269185
Ion Ratio Lower Upper
163 100
194 3.7 2.6 4.0
164 9.2 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 22.087 ng/ul
RT: 14.452 min Scan# 1883
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:165 Resp: 54538
Ion Ratio Lower Upper
165 100
89 54.6 43.6 65.4
121 21.9 14.7 22.1









#50
Acenaphthylene
Concen: 21.862 ng/ul
RT: 14.599 min Scan# 1907
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:152 Resp: 340562
Ion Ratio Lower Upper
152 100
151 18.0 15.8 23.8
153 12.8 10.3 15.5



#51
3-Nitroaniline
Concen: 23.313 ng/ul
RT: 14.793 min Scan# 1941
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:138 Resp: 56367
Ion Ratio Lower Upper
138 100
108 8.4 7.4 11.0
92 111.1 89.1 133.7









#52
Acenaphthene
Concen: 21.609 ng/ul
RT: 14.928 min Scan# 1964
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:153 Resp: 227762
Ion Ratio Lower Upper
153 100
152 45.9 34.6 52.0
154 92.9 73.0 109.4



#53
2,4-Dinitrophenol
Concen: 15.015 ng/ul
RT: 14.992 min Scan# 1975
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:184 Resp: 26631
Ion Ratio Lower Upper
184 100
63 81.0 54.9 82.3
154 56.1 44.9 67.3









#55
4-Nitrophenol
Concen: 22.368 ng/ul
RT: 15.075 min Scan# 1988
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:109 Resp: 34441
Ion Ratio Lower Upper
109 100
139 157.3 107.9 161.9
65 148.6 126.6 189.8



#56
Dibenzofuran
Concen: 21.708 ng/ul
RT: 15.263 min Scan# 2021
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:168 Resp: 315707
Ion Ratio Lower Upper
168 100
139 32.0 27.4 41.2





#57

2,4-Dinitrotoluene

Concen: 22.532 ng/ul

RT: 15.233 min Scan# 2015

Delta R.T. 0.004 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:165 Resp: 81482

Ion Ratio Lower Upper

165 100

63 46.0 36.4 54.6

182 2.1 1.2 1.8#



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.850 ng/ul

RT: 15.474 min Scan# 2057

Delta R.T. 0.004 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Tgt Ion:232 Resp: 68411




Ion Ratio Lower Upper


232 100

131 38.2 31.0 46.6

130 2.6 1.8 2.6

166 28.4 18.9 28.3#









#59
Diethylphthalate
Concen: 21.437 ng/ul
RT: 15.650 min Scan# 2087
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:149 Resp: 260951
Ion Ratio Lower Upper
149 100
177 22.9 20.1 30.1
150 12.4 9.4 14.2



#61
Fluorene
Concen: 21.917 ng/ul
RT: 15.909 min Scan# 2131
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:166 Resp: 271201
Ion Ratio Lower Upper
166 100
165 99.1 81.0 121.6
167 13.3 11.7 17.5









#62
4-Chlorophenyl-phenylether
Concen: 22.245 ng/ul
RT: 15.891 min Scan# 2128
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:204 Resp: 136015
Ion Ratio Lower Upper
204 100
206 28.6 26.9 40.3
141 48.8 40.7 61.1



#63
4-Nitroaniline
Concen: 24.428 ng/ul
RT: 15.956 min Scan# 2139
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:138 Resp: 53174
Ion Ratio Lower Upper
138 100
92 49.5 44.1 66.1
108 85.5 63.3 94.9









#66
4,6-Dinitro-2-methylphenol
Concen: 20.984 ng/ul
RT: 15.985 min Scan# 2143
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:198 Resp: 52652
Ion Ratio Lower Upper
198 100
182 2.6 3.1 4.7#
77 19.7 13.8 20.6



#67
N-Nitrosodiphenylamine
Concen: 22.407 ng/ul
RT: 16.115 min Scan# 2166
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:169 Resp: 226569
Ion Ratio Lower Upper
169 100
168 67.6 54.6 82.0
167 38.3 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 21.396 ng/ul
RT: 16.790 min Scan# 21
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:248 Resp: 81185
Ion Ratio Lower Upper
248 100
250 100.2 75.0 112.4
141 57.6 49.0 73.6



#69
Hexachlorobenzene
Concen: 21.976 ng/ul
RT: 16.884 min Scan# 2297
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:284 Resp: 95177
Ion Ratio Lower Upper
284 100
142 27.8 23.8 35.8
249 32.6 23.0 34.6





#70

Atrazine

Concen: 22.443 ng/ul

RT: 17.043 min Scan# 21

Delta R.T. -0.002 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:200 Resp: 89976

Ion Ratio Lower Upper

200 100

173 23.2 19.4 29.0

215 45.7 34.2 51.4



#71

Pentachlorophenol

Concen: 20.635 ng/ul

RT: 17.243 min Scan# 2358

Delta R.T. 0.004 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04




Tgt Ion:266 Resp: 51941


Ion Ratio Lower Upper

266 100

264 65.4 55.0 82.6

268 70.1 48.5 72.7









#72
Phenanthrene
Concen: 21.898 ng/ul
RT: 17.660 min Scan# 2428
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:178 Resp: 463400
Ion Ratio Lower Upper
178 100
179 14.5 12.4 18.6
176 18.7 15.2 22.8



#74
Anthracene
Concen: 22.222 ng/ul
RT: 17.748 min Scan# 2444
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:178 Resp: 478276
Ion Ratio Lower Upper
178 100
179 14.2 12.5 18.7
176 19.2 14.6 21.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.377 ng/uL

RT: 13.659 min Scan# 1

Delta R.T. -0.002 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:216 Resp: 103877

Ion Ratio Lower Upper

216 100

214 71.0 68.5 102.7

218 47.7 37.8 56.8



#76

Pentachlorobenzene

Concen: 22.125 ng/uL

RT: 15.157 min Scan# 2003

Delta R.T. -0.002 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04




Tgt Ion:250 Resp: 106301


Ion Ratio Lower Upper

250 100

248 64.1 45.4 68.2

252 74.4 56.6 84.8





#77

Carbazole

Concen: 22.280 ng/ul

RT: 18.024 min Scan# 2491

Delta R.T. -0.002 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:167 Resp: 399064

Ion Ratio Lower Upper

167 100

166 20.9 18.2 27.4

139 12.1 9.1 13.7



#78

Di-n-butylphthalate

Concen: 21.832 ng/ul

RT: 18.529 min Scan# 2577

Delta R.T. 0.004 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04




Tgt Ion:149 Resp: 462644


Ion Ratio Lower Upper

149 100

150 9.1 7.5 11.3

104 4.9 3.4 5.2





#80

Fluoranthene

Concen: 22.404 ng/ul

RT: 19.652 min Scan# 21

Delta R.T. -0.002 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:202 Resp: 524920

Ion Ratio Lower Upper

202 100

101 11.9 8.0 12.0

100 9.8 7.2 10.8



#82

Pyrene

Concen: 22.616 ng/ul

RT: 20.016 min Scan# 2830

Delta R.T. -0.002 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04




Tgt Ion:202 Resp: 550711


Ion Ratio Lower Upper

202 100

101 12.8 10.5 15.7

100 12.0 9.6 14.4









#83
Butylbenzylphthalate
Concen: 22.237 ng/ul
RT: 20.868 min Scan# 2975
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:149 Resp: 200070
Ion Ratio Lower Upper
149 100
91 78.3 62.0 93.0
206 22.1 16.3 24.5



#84
3,3'-Dichlorobenzidine
Concen: 24.142 ng/ul
RT: 21.808 min Scan# 3135
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:252 Resp: 187203
Ion Ratio Lower Upper
252 100
254 62.6 55.5 83.3
126 9.1 9.0 13.4





#85

Benzo(a)anthracene

Concen: 22.193 ng/ul

RT: 21.896 min Scan# 3149

Delta R.T. 0.004 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:228 Resp: 526835

Ion Ratio Lower Upper

228 100

229 18.9 15.8 23.8

226 26.2 23.3 34.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.054 ng/ul

RT: 21.726 min Scan# 3121

Delta R.T. 0.004 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04




Tgt Ion:149 Resp: 296767


Ion Ratio Lower Upper

149 100

167 25.6 20.9 31.3

279 4.2 2.8 4.2#









#87
Chrysene
Concen: 21.967 ng/ul
RT: 21.967 min Scan# 3162
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:228 Resp: 489825
Ion Ratio Lower Upper
228 100
226 31.3 25.2 37.8
229 19.9 15.0 22.4



#89
Di-n-octyl phthalate
Concen: 21.492 ng/ul
RT: 23.018 min Scan# 3341
Delta R.T. 0.004 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:149 Resp: 504623









#90
Benzo(b)fluoranthene
Concen: 21.855 ng/ul
RT: 24.270 min Scan# 3554
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04




Instrument :
BNA_G
ClientSampleId :
SICV406


Tgt Ion:252 Resp: 535832
Ion Ratio Lower Upper
252 100
253 21.0 17.3 25.9
125 8.8 8.0 12.0



#91
Benzo(k)fluoranthene
Concen: 21.240 ng/ul
RT: 24.346 min Scan# 3567
Delta R.T. -0.002 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:252 Resp: 528754
Ion Ratio Lower Upper
252 100
253 22.7 19.0 28.4
125 10.1 9.0 13.4





#93

Benzo(a)pyrene

Concen: 21.551 ng/ul

RT: 25.233 min Scan# 3718

Delta R.T. 0.004 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04

Instrument :

BNA_G

ClientSampleId :

SICV406





Tgt Ion:252 Resp: 503317

Ion Ratio Lower Upper

252 100

253 21.3 18.5 27.7

125 9.7 8.7 13.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.889 ng/ul

RT: 29.423 min Scan# 4431

Delta R.T. -0.002 min

Lab File: BG059344.D

Acq: 7 Oct 2023 16:04




Tgt Ion:276 Resp: 508150


Ion Ratio Lower Upper

276 100

138 16.2 16.2 24.4#

277 26.0 17.1 25.7#









#95
Dibenzo(a,h)anthracene
Concen: 21.506 ng/ul
RT: 29.505 min Scan# 4444
Delta R.T. 0.010 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Instrument :
BNA_G
ClientSampleId :
SICV406

Tgt Ion:278 Resp: 470450
Ion Ratio Lower Upper
278 100
139 16.0 13.7 20.5
279 24.9 18.6 27.8



#96
Benzo(g,h,i)perylene
Concen: 21.198 ng/ul
RT: 30.703 min Scan# 4649
Delta R.T. -0.008 min
Lab File: BG059344.D
Acq: 7 Oct 2023 16:04

Tgt Ion:276 Resp: 454802
Ion Ratio Lower Upper
276 100
138 19.5 19.0 28.4
277 25.1 17.2 25.8

