

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100723\
 Data File : BG059367.D
 Acq On : 8 Oct 2023 8:29
 Operator : MA/JU
 Sample : 04607-16
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BBJS9

Quant Time: Oct 09 00:59:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG100723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 07 22:45:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.226	152	43671	20.000	ng/u1	0.00
20) Naphthalene-d8	11.070	136	215178	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.859	164	141410	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.615	188	340536	20.000	ng/u1	0.00
79) Chrysene-d12	21.916	240	304675	20.000	ng/u1	# 0.00
88) Perylene-d12	25.400	264	356588	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.526	96	5306	4.332	ng/uL	0.00
4) Pyridine-d5	3.960	84	41164	11.676	ng/u1	0.00
7) Phenol-d5	7.350	99	33148	6.874	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.550	67	96427	35.379	ng/u1	0.00
11) 2-Chlorophenol-d4	7.744	132	88983	26.385	ng/u1	0.00
15) 4-Methylphenol-d8	8.919	113	62869	16.734	ng/u1	0.00
21) Nitrobenzene-d5	9.430	128	63828	34.846	ng/u1	0.00
24) 2-Nitrophenol-d4	10.141	143	68636	34.620	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.670	165	118834	33.208	ng/u1	0.00
31) 4-Chloroaniline-d4	11.217	131	172323	30.316	ng/u1	0.00
46) Dimethylphthalate-d6	14.260	166	494894	42.718	ng/u1	0.00
49) Acenaphthylene-d8	14.566	160	553960	40.180	ng/u1	0.00
54) 4-Nitrophenol-d4	15.053	143	14464	6.662	ng/u1	0.00
60) Fluorene-d10	15.852	176	465175	42.065	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.964	200	94264	41.926	ng/u1	0.00
73) Anthracene-d10	17.715	188	766398	45.878	ng/u1	0.00
81) Pyrene-d10	19.989	212	842195	45.630	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.153	264	856578	45.649	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100723\
Data File : BG059367.D
Acq On : 8 Oct 2023 8:29
Operator : MA/JU
Sample : 04607-16
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BBJS9

Quant Time: Oct 09 00:59:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG100723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 07 22:45:22 2023
Response via : Initial Calibration

