

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100723\
 Data File : BG059365.D
 Acq On : 8 Oct 2023 7:08
 Operator : MA/JU
 Sample : 04624-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BBJQ0

Quant Time: Oct 09 00:59:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG100723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 07 22:45:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.230	152	39444	20.000	ng/u1	0.00
20) Naphthalene-d8	11.068	136	195718	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.863	164	137137	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.613	188	313274	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.914	240	284086	20.000	ng/u1	0.00
88) Perylene-d12	25.398	264	337056	20.000	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.535	96	4471	4.042	ng/uL	0.00
4) Pyridine-d5	3.958	84	46937	14.740	ng/u1	0.00
7) Phenol-d5	7.354	99	38409	8.819	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.548	67	97858	39.752	ng/u1	0.00
11) 2-Chlorophenol-d4	7.748	132	97827	32.116	ng/u1	0.00
15) 4-Methylphenol-d8	8.923	113	68319	20.133	ng/u1	0.00
21) Nitrobenzene-d5	9.428	128	67178	40.321	ng/u1	0.00
24) 2-Nitrophenol-d4	10.145	143	74420	41.269	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.674	165	126543	38.879	ng/u1	0.00
31) 4-Chloroaniline-d4	11.215	131	182833	35.363	ng/u1	0.00
46) Dimethylphthalate-d6	14.258	166	547496	48.731	ng/u1	0.00
49) Acenaphthylene-d8	14.564	160	609361	45.576	ng/u1	0.00
54) 4-Nitrophenol-d4	15.051	143	18430	8.754	ng/u1	0.00
60) Fluorene-d10	15.850	176	514850	48.008	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.962	200	98090	47.424	ng/u1	0.00
73) Anthracene-d10	17.713	188	831759	54.124	ng/u1	0.00
81) Pyrene-d10	19.987	212	920552	53.490	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.157	264	939919	52.993	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100723\
Data File : BG059365.D
Acq On : 8 Oct 2023 7:08
Operator : MA/JU
Sample : 04624-01
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BBJQ0

Quant Time: Oct 09 00:59:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG100723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 07 22:45:22 2023
Response via : Initial Calibration

