


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100923\
Data File : BG059391.D
Acq On : 9 Oct 2023 8:54
Operator : MA/JU
Sample : DFTPP544
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP544

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG100723.MA.M
Title : SVOA CALIBRATION
Last Update : Sat Oct 07 22:45:22 2023



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2473

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	53.2	71349	PASS
68	69	0.00	2	1.3	606	PASS
69	198	0.00	100	34.1	45776	PASS
70	69	0.00	2	0.3	154	PASS
127	198	10	80	44.9	60181	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	134080	PASS
199	198	5	9	7.2	9627	PASS
275	198	10	60	24.6	33035	PASS
365	198	1	100	3.5	4639	PASS
441	198	0.01	100	15.8	21176	PASS
442	198	50	100	92.1	123501	PASS
443	442	15	24	18.7	23143	PASS

m/z	Abundance
35.90	206.0
40.00	43240.0
44.00	286.0
65.80	209.0
149.90	156.0
191.00	156.0
403.30	152.0
418.70	215.0
533.90	163.0

Instrument :
BNA_G
ClientSampleId :
DFTPP544

m/z	Abundance
36.00	454.0
38.10	389.0
39.10	3945.0
40.00	42560.0
44.80	224.0
49.10	408.0
50.10	15922.0
51.10	63008.0
52.10	3285.0
53.00	327.0
53.90	215.0
55.10	491.0
56.00	1543.0
57.00	3844.0
58.00	321.0
61.10	436.0
62.10	1086.0
63.00	2114.0
63.90	191.0
65.00	1039.0
68.10	947.0
69.00	41288.0
69.90	187.0
73.10	402.0
74.10	3875.0
75.10	6267.0
76.10	2414.0
77.10	52328.0
78.20	3105.0
79.10	1735.0
80.10	2100.0
81.00	3446.0
82.00	761.0
83.10	790.0
84.10	275.0
85.10	1124.0
85.90	809.0
87.10	323.0
87.90	172.0
90.90	372.0
92.10	1320.0
93.00	5111.0
94.00	282.0
96.10	165.0
98.00	4793.0
99.10	3590.0
99.90	481.0
101.00	1995.0
102.10	159.0
102.90	375.0
104.00	987.0
105.10	1004.0
106.00	394.0
107.10	15089.0
108.00	1631.0
109.00	758.0
110.00	22776.0
110.90	4494.0
112.00	379.0
113.10	199.0
116.10	460.0
117.00	9265.0
117.90	1054.0
120.10	193.0
122.00	747.0
123.00	1296.0
123.90	582.0
125.00	294.0
127.10	51784.0
128.10	4083.0
129.00	19368.0
130.10	2468.0
131.10	503.0
131.80	233.0
134.00	463.0
135.00	1343.0
136.00	784.0
137.00	441.0
137.90	329.0
139.00	308.0
139.70	290.0
140.10	339.0
141.00	2138.0
142.00	609.0
142.90	529.0
146.00	556.0
147.00	1294.0
148.00	2295.0
149.10	592.0
150.10	176.0
151.10	249.0
151.40	259.0

Instrument :
BNA_G
ClientSampleId :
DFTPP544

153.00	630.0
154.00	256.0
155.10	1507.0
156.10	1998.0
157.20	320.0
157.80	437.0
158.10	465.0
158.70	184.0
160.10	993.0
160.90	1204.0
161.90	382.0
162.90	246.0
164.00	202.0
165.10	878.0
166.10	1152.0
167.10	4238.0
168.10	2251.0
169.00	445.0
170.30	254.0
172.00	461.0
172.90	927.0
174.00	1046.0
175.10	1846.0
175.80	575.0
176.40	271.0
177.10	714.0
179.00	4332.0
180.00	3083.0
181.00	1760.0
184.00	171.0
185.00	1728.0
186.00	13737.0
187.00	3896.0
188.10	432.0
188.90	708.0
189.60	169.0
190.10	176.0
190.70	461.0
192.00	1255.0
192.90	1366.0
193.90	248.0
195.90	2327.0
198.00	107136.0
199.00	8108.0
199.70	557.0
200.20	590.0
201.10	370.0
201.50	488.0
203.10	600.0
204.00	3445.0
205.00	5410.0
206.10	25136.0
207.00	4315.0
208.00	1173.0
208.90	157.0
210.40	832.0
210.90	949.0
213.00	291.0
214.80	212.0
216.00	843.0
217.00	6155.0
218.10	757.0
221.00	5356.0
221.90	860.0
223.00	1226.0
224.00	14236.0
225.10	4083.0
227.00	6034.0
227.80	382.0
229.00	1260.0
230.10	197.0
231.00	272.0
231.80	240.0
234.10	338.0
235.00	356.0
235.90	292.0
237.00	715.0
240.30	293.0
241.00	279.0
242.00	771.0
243.00	834.0
244.00	11998.0
244.90	1850.0
245.90	1781.0
247.20	579.0
248.90	542.0
250.90	221.0
252.00	283.0
252.80	343.0
255.00	64904.0
256.00	8534.0
257.00	881.0
257.90	2813.0
259.00	394.0
261.10	238.0
265.00	1303.0

Instrument :
BNA_G
ClientSampleId :
DFTPP544

266.10	332.0
271.00	155.0
272.10	419.0
272.90	1924.0
274.00	3832.0
275.00	22920.0
275.90	3860.0
277.00	1767.0
277.70	163.0
278.20	333.0
279.00	290.0
282.20	173.0
283.90	310.0
285.00	530.0
286.20	157.0
288.90	268.0
292.70	527.0
295.90	6043.0
296.90	1339.0
301.70	340.0
303.20	789.0
304.00	190.0
304.50	190.0
314.00	785.0
315.00	435.0
316.10	330.0
321.10	351.0
323.00	2940.0
323.90	446.0
325.20	211.0
326.70	550.0
327.90	547.0
331.60	153.0
332.20	400.0
332.80	592.0
333.90	1342.0
335.00	779.0
335.90	243.0
340.80	462.0
345.70	308.0
346.00	383.0
346.70	186.0
352.00	921.0
353.10	936.0
354.10	758.0
358.70	175.0
364.90	3648.0
365.90	603.0
370.00	199.0
371.80	1054.0
372.90	237.0
381.60	178.0
383.20	279.0
390.90	308.0
397.90	175.0
400.90	203.0
401.80	513.0
402.90	999.0
404.00	452.0
409.80	240.0
420.50	315.0
421.00	689.0
421.90	811.0
423.00	5141.0
423.90	975.0
424.70	528.0
432.50	221.0
441.00	12159.0
442.00	76872.0
442.90	13860.0
444.00	1089.0
479.20	170.0
504.90	152.0

Instrument :
BNA_G
ClientSampleId :
DFTPP544

m/z	Abundance
36.30	193.0
37.00	270.0
38.20	359.0
39.10	5316.0
40.00	43384.0
41.10	205.0
43.90	212.0
45.10	251.0
49.00	624.0
50.10	22424.0
51.10	82152.0
52.10	4586.0
54.40	155.0
55.10	499.0
56.00	2958.0
57.00	5512.0
58.00	196.0
61.10	1018.0
62.00	543.0
63.10	2962.0
64.20	437.0
65.00	1088.0
66.50	343.0
69.00	53680.0
70.00	275.0
73.00	650.0
74.00	5448.0
75.10	9147.0
76.10	3012.0
77.10	64320.0
78.10	5444.0
79.00	3097.0
80.10	2604.0
81.00	4344.0
82.10	994.0
83.00	1295.0
85.10	735.0
86.10	1283.0
87.00	374.0
87.90	265.0
88.90	177.0
90.10	162.0
91.20	660.0
92.10	901.0
93.00	6385.0
94.00	817.0
95.20	330.0
95.90	189.0
98.10	5427.0
98.90	4375.0
100.00	554.0
101.00	2149.0
103.10	1073.0
104.10	1252.0
105.00	1682.0
105.90	463.0
107.10	19704.0
108.00	2802.0
109.10	565.0
110.00	35304.0
111.00	5537.0
112.00	587.0
112.90	284.0
113.90	253.0
115.90	945.0
117.10	14928.0
118.00	1258.0
120.00	292.0
122.00	928.0
123.00	1968.0
124.10	941.0
125.00	590.0
127.00	69960.0
128.00	5625.0
129.00	24320.0
130.00	2275.0
131.10	606.0
132.00	307.0
133.10	225.0
134.00	939.0
135.10	1576.0
136.00	1226.0
137.10	858.0
138.10	528.0
141.00	2873.0
142.00	1746.0
143.00	895.0
144.10	368.0
146.00	670.0
147.10	1527.0
148.10	3650.0
149.10	874.0

Instrument :
BNA_G
ClientSampleId :
DFTPP544

150.10	393.0
151.40	376.0
152.10	244.0
153.00	1478.0
154.20	1039.0
155.00	1861.0
156.10	2612.0
157.10	614.0
157.90	1049.0
158.90	625.0
160.00	1328.0
161.00	1507.0
162.90	303.0
163.90	331.0
165.00	914.0
166.00	1118.0
167.10	7122.0
168.00	3479.0
169.00	1026.0
170.10	367.0
171.00	494.0
171.90	332.0
173.00	473.0
174.00	1649.0
175.10	2802.0
175.80	621.0
177.20	1201.0
178.10	671.0
179.00	4847.0
180.10	5134.0
181.00	2050.0
182.00	288.0
182.90	202.0
184.10	785.0
185.10	1729.0
186.10	19672.0
187.00	5625.0
187.80	598.0
189.10	1230.0
191.10	235.0
192.00	1665.0
193.00	1808.0
193.90	288.0
196.00	4261.0
198.00	154304.0
199.00	11316.0
200.00	1031.0
201.60	1056.0
203.00	974.0
204.00	4460.0
205.00	7915.0
206.00	41728.0
207.00	5749.0
207.90	1840.0
209.00	498.0
209.80	464.0
211.00	1192.0
211.60	698.0
215.10	417.0
216.10	1101.0
217.00	10606.0
217.80	1119.0
221.00	6005.0
221.60	2488.0
223.00	2114.0
224.10	22360.0
225.10	5648.0
226.00	573.0
227.00	10250.0
228.00	1429.0
229.00	1898.0
230.00	459.0
230.90	692.0
231.70	230.0
232.90	329.0
234.10	775.0
235.10	849.0
236.00	524.0
237.10	644.0
239.80	279.0
240.80	548.0
242.00	949.0
242.90	737.0
244.10	16792.0
244.90	1933.0
245.90	3826.0
247.10	1070.0
248.00	340.0
248.70	502.0
249.20	359.0
250.10	211.0
251.20	264.0
252.10	349.0
253.00	680.0
255.00	94840.0
256.00	12778.0

Instrument :
BNA_G
ClientSampleId :
DFTPP544

m/z	Abundance
37.20	280.0
38.10	540.0
39.00	4256.0
40.00	43128.0
41.00	292.0
43.00	210.0
44.00	320.0
45.00	295.0
49.00	438.0
50.10	16080.0
51.10	68888.0
52.10	2537.0
53.20	339.0
54.90	169.0
56.00	1613.0
57.00	4008.0
58.00	176.0
59.10	159.0
61.10	1202.0
62.10	681.0
63.00	1828.0
64.10	240.0
65.10	1169.0
68.10	870.0
69.00	42360.0
73.10	787.0
74.00	5148.0
75.00	7338.0
76.10	2673.0
77.10	49136.0
78.10	4224.0
79.10	3397.0
80.10	2292.0
81.10	2834.0
82.00	546.0
83.00	643.0
85.10	645.0
86.10	1307.0
86.90	273.0
88.00	248.0
91.00	736.0
92.10	773.0
93.00	6808.0
93.90	312.0
96.00	428.0
98.00	4120.0
99.00	4148.0
99.90	483.0
100.90	2249.0
102.00	150.0
103.00	426.0
104.00	1195.0
105.10	1045.0
106.10	968.0
107.00	15211.0
108.00	2736.0
110.00	30136.0
111.10	4787.0
111.90	439.0
113.00	199.0
115.00	190.0
116.00	975.0
117.00	12553.0
118.00	838.0
121.90	524.0
123.10	1945.0
124.00	604.0
124.90	854.0
127.10	58800.0
128.10	5401.0
129.00	20072.0
130.00	1955.0
131.00	584.0
132.10	206.0
134.00	865.0
135.00	1264.0
135.70	636.0
136.10	759.0
137.00	682.0
138.10	344.0
140.10	209.0
141.00	2490.0
142.10	806.0
143.00	905.0
143.70	234.0
144.90	191.0
146.40	644.0
147.00	1115.0
148.00	3057.0
149.10	229.0
150.00	307.0
151.30	579.0

Instrument :
BNA_G
ClientSampleId :
DFTPP544

152.10	346.0
153.10	1174.0
154.00	869.0
155.10	1182.0
156.10	2792.0
157.00	379.0
157.50	174.0
158.10	508.0
159.00	649.0
160.00	820.0
161.00	1159.0
162.10	606.0
163.80	201.0
165.00	1204.0
166.00	818.0
167.00	6800.0
168.00	2079.0
168.70	226.0
171.00	237.0
172.00	550.0
173.10	837.0
173.90	1624.0
175.00	2211.0
175.80	778.0
176.80	1167.0
177.80	246.0
179.00	4541.0
180.00	2745.0
181.00	1641.0
181.70	184.0
182.00	159.0
183.80	399.0
185.10	2317.0
186.00	16153.0
187.10	5458.0
188.20	763.0
189.10	634.0
189.80	409.0
190.80	400.0
192.00	1412.0
193.10	1751.0
194.00	500.0
196.00	4934.0
198.00	140800.0
199.00	9457.0
200.00	744.0
201.10	483.0
201.40	484.0
202.00	255.0
203.00	1042.0
204.10	5353.0
205.00	8710.0
206.10	35408.0
207.00	5048.0
208.10	1148.0
209.10	691.0
210.20	962.0
211.00	1338.0
211.60	317.0
214.90	480.0
216.10	1180.0
217.00	9557.0
218.00	1609.0
218.60	178.0
219.20	202.0
219.70	167.0
221.00	6356.0
221.80	1474.0
222.10	1419.0
222.90	2199.0
224.10	21624.0
225.10	4691.0
225.70	444.0
227.00	9202.0
228.00	2004.0
229.00	2180.0
230.10	568.0
230.50	461.0
231.10	377.0
233.90	278.0
234.90	864.0
235.90	575.0
236.90	779.0
238.80	175.0
239.70	178.0
241.00	482.0
241.90	721.0
243.00	1273.0
244.00	17616.0
245.10	2026.0
246.00	3403.0
247.10	636.0
248.00	176.0
249.00	531.0
250.20	321.0
250.80	162.0

Instrument :
BNA_G
ClientSampleId :
DFTPP544

252.20	411.0
253.20	623.0
255.00	95160.0
256.00	14478.0
257.10	636.0
258.00	5356.0
259.00	607.0
259.70	322.0
264.00	201.0
265.00	1606.0
266.10	158.0
266.80	348.0
270.90	195.0
271.90	341.0
272.90	2382.0
274.00	6409.0
275.00	37160.0
275.90	4862.0
277.00	2589.0
278.00	511.0
279.10	199.0
281.90	198.0
282.90	321.0
283.10	328.0
284.10	272.0
285.10	583.0
289.00	241.0
289.80	173.0
291.00	270.0
292.70	404.0
294.10	365.0
296.00	9822.0
297.00	1359.0
298.20	187.0
299.90	250.0
302.80	857.0
304.10	302.0
307.80	196.0
309.90	250.0
313.70	511.0
315.00	1870.0
315.90	784.0
317.20	159.0
320.70	344.0
321.90	242.0
323.00	4321.0
324.20	745.0
327.00	1117.0
327.90	572.0
328.90	189.0
332.80	406.0
334.00	3003.0
335.10	920.0
340.90	431.0
345.70	999.0
346.80	292.0
351.00	439.0
351.90	1464.0
353.00	798.0
354.00	1441.0
358.80	244.0
364.90	5097.0
366.00	928.0
372.00	2019.0
373.10	619.0
383.00	319.0
384.00	196.0
389.70	247.0
391.20	195.0
391.90	201.0
392.80	172.0
400.70	162.0
401.90	811.0
403.00	1729.0
403.90	347.0
413.10	151.0
420.90	1492.0
422.00	1856.0
422.90	7004.0
424.00	1674.0
425.00	162.0
426.20	161.0
441.00	27024.0
442.00	154240.0
442.90	29776.0
443.90	3246.0
445.00	188.0

Instrument :
BNA_G
ClientSampleId :
DFTPP544

257.10	858.0
257.90	4608.0
259.00	924.0
260.00	469.0
265.00	2180.0
265.80	558.0
269.90	187.0
271.00	228.0
273.00	2073.0
274.00	5674.0
275.00	39024.0
276.00	6274.0
277.00	3524.0
278.00	645.0
279.00	189.0
282.50	334.0
283.20	384.0
284.90	749.0
286.10	239.0
289.00	202.0
290.10	247.0
292.30	395.0
293.00	1166.0
293.90	370.0
295.90	11309.0
297.00	1774.0
300.90	269.0
302.00	230.0
302.90	1078.0
304.20	476.0
308.10	169.0
309.10	198.0
312.80	181.0
313.90	843.0
315.00	1201.0
315.90	918.0
320.90	538.0
321.90	419.0
323.00	4721.0
324.00	1082.0
326.80	623.0
327.80	447.0
331.60	226.0
332.10	351.0
332.70	187.0
334.10	1619.0
334.70	654.0
336.00	227.0
341.00	516.0
341.80	232.0
343.00	215.0
345.80	1595.0
347.10	276.0
347.90	188.0
351.00	237.0
351.90	1554.0
353.10	1113.0
354.10	1434.0
355.20	249.0
359.00	167.0
359.50	226.0
363.90	162.0
365.00	5172.0
366.00	593.0
370.80	279.0
372.10	1823.0
373.00	560.0
382.60	349.0
383.10	304.0
387.40	316.0
389.80	448.0
390.70	429.0
391.20	203.0
400.80	247.0
402.00	955.0
403.00	1161.0
403.90	553.0
414.80	167.0
420.90	1301.0
423.00	6701.0
424.00	1086.0
425.00	218.0
441.00	24344.0
442.00	139392.0
443.00	25792.0
444.10	1706.0
444.90	259.0
445.70	183.0
475.30	161.0

Instrument :
BNA_G
ClientSampleId :
DFTPP544