

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG101218\
 Data File : BG037292.D
 Acq On : 12 Oct 2018 15:43
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 12 16:26:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 11 13:59:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2	1,4-Dioxane	0.638	0.625	2.0	121	0.00
3	Pyridine	1.515	1.574	-3.9	119	0.00
4	n-Nitrosodimethylamine	0.588	0.570	3.1	116	0.00
5 S	2-Fluorophenol	1.136	1.168	-2.8	121	0.00
6	Aniline	2.258	2.208	2.2	115	0.00
7 S	Phenol-d6	1.725	1.771	-2.7	119	0.00
8	2-Chlorophenol	1.330	1.338	-0.6	115	0.00
9	Benzaldehyde	1.188	1.165	1.9	115	0.00
10 C	Phenol	1.817	1.851	-1.9	118	0.00
11	bis(2-Chloroethyl)ether	1.484	1.459	1.7	117	0.00
12	1,3-Dichlorobenzene	1.564	1.547	1.1	119	0.00
13 C	1,4-Dichlorobenzene	1.583	1.553	1.9	117	0.00
14	1,2-Dichlorobenzene	1.534	1.474	3.9	116	0.00
15	Benzyl Alcohol	1.349	1.360	-0.8	115	0.00
16	2,2'-oxybis(1-Chloropropane	2.934	2.748	6.3	111	0.00
17	2-Methylphenol	1.217	1.232	-1.2	117	0.00
18	Hexachloroethane	0.540	0.532	1.5	116	0.00
19 P	n-Nitroso-di-n-propylamine	1.306	1.270	2.8	114	0.00
20	3+4-Methylphenols	1.701	1.776	-4.4	119	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
22	Acetophenone	0.504	0.485	3.8	114	0.00
23 S	Nitrobenzene-d5	0.304	0.340	-11.8	124	0.00
24	Nitrobenzene	0.326	0.357	-9.5	122	0.00
25	Isophorone	0.694	0.677	2.4	113	0.00
26 C	2-Nitrophenol	0.116	0.138	-19.0	142	0.00
27	2,4-Dimethylphenol	0.290	0.289	0.3	117	0.00
28	bis(2-Chloroethoxy)methane	0.434	0.426	1.8	114	0.00
29 C	2,4-Dichlorophenol	0.295	0.316	-7.1	118	0.00
30	1,2,4-Trichlorobenzene	0.359	0.350	2.5	117	0.00
31	Naphthalene	0.972	0.954	1.9	116	0.00
32	Benzoic acid	0.160	0.209	-30.6#	151#	0.00
33	4-Chloroaniline	0.456	0.456	0.0	115	0.00
34 C	Hexachlorobutadiene	0.223	0.209	6.3	113	0.00
35	Caprolactam	0.120	0.135	-12.5	122	0.00
36 C	4-Chloro-3-methylphenol	0.348	0.362	-4.0	115	0.00
37	2-Methylnaphthalene	0.709	0.701	1.1	116	0.00
38	1-Methylnaphthalene	0.693	0.676	2.5	114	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
40	1,2,4,5-Tetrachlorobenzene	0.648	0.622	4.0	115	0.00
41 P	Hexachlorocyclopentadiene	0.267	0.268	-0.4	117	0.00
42 S	2,4,6-Tribromophenol	0.196	0.210	-7.1	119	0.00
43 C	2,4,6-Trichlorophenol	0.376	0.421	-12.0	124	0.00
44	2,4,5-Trichlorophenol	0.404	0.455	-12.6	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG101218\
 Data File : BG037292.D
 Acq On : 12 Oct 2018 15:43
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 12 16:26:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 11 13:59:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.332	1.272	4.5	114	0.00
46	1,1'-Biphenyl	1.645	1.598	2.9	115	0.00
47	2-Chloronaphthalene	1.243	1.199	3.5	115	0.00
48	2-Nitroaniline	0.300	0.358	-19.3	131	0.00
49	Acenaphthylene	1.901	1.855	2.4	115	0.00
50	Dimethylphthalate	1.572	1.561	0.7	114	0.00
51	2,6-Dinitrotoluene	0.284	0.323	-13.7	129	0.00
52 C	Acenaphthene	1.174	1.133	3.5	113	0.00
53	3-Nitroaniline	0.329	0.375	-14.0	129	0.00
54 P	2,4-Dinitrophenol	0.079	0.083	-5.1	127	0.00
55	Dibenzofuran	1.905	1.823	4.3	113	0.00
56 P	4-Nitrophenol	0.256	0.283	-10.5	124	0.00
57	2,4-Dinitrotoluene	0.356	0.427	-19.9	135	0.00
58	Fluorene	1.441	1.403	2.6	116	0.00
59	2,3,4,6-Tetrachlorophenol	0.358	0.397	-10.9	120	0.00
60	Diethylphthalate	1.629	1.612	1.0	113	0.00
61	4-Chlorophenyl-phenylether	0.842	0.818	2.9	114	0.00
62	4-Nitroaniline	0.345	0.378	-9.6	125	0.00
63	Azobenzene	1.562	1.484	5.0	112	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	0.056	0.069	-23.2	138	0.00
66 c	n-Nitrosodiphenylamine	0.587	0.572	2.6	113	0.00
67	4-Bromophenyl-phenylether	0.216	0.211	2.3	114	0.00
68	Hexachlorobenzene	0.224	0.215	4.0	113	0.00
69	Atrazine	0.218	0.221	-1.4	113	0.00
70 C	Pentachlorophenol	0.144	0.151	-4.9	118	0.00
71	Phenanthrene	1.018	0.965	5.2	112	0.00
72	Anthracene	0.996	0.963	3.3	114	0.00
73	Carbazole	1.027	0.984	4.2	112	0.00
74	Di-n-butylphthalate	1.100	1.120	-1.8	112	0.00
75 C	Fluoranthene	1.190	1.126	5.4	110	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
77	Benzidine	0.765	0.860	-12.4	115	0.00
78	Pyrene	1.303	1.269	2.6	110	0.00
79 S	Terphenyl-d14	0.952	0.908	4.6	108	0.00
80	Butylbenzylphthalate	0.507	0.537	-5.9	116	0.00
81	Benzo(a)anthracene	1.198	1.172	2.2	110	0.00
82	3,3'-Dichlorobenzidine	0.462	0.465	-0.6	109	0.00
83	Chrysene	1.143	1.107	3.1	111	0.00
84	Bis(2-ethylhexyl)phthalate	0.728	0.760	-4.4	114	0.00
85 c	Di-n-octyl phthalate	1.182	1.258	-6.4	116	-0.02
86	Indeno(1,2,3-cd)pyrene	1.266	1.266	0.0	111	-0.03
87 I	Perylene-d12	1.000	1.000	0.0	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG101218\
Data File : BG037292.D
Acq On : 12 Oct 2018 15:43
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 12 16:26:24 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 11 13:59:58 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.087	1.065	2.0	110	0.00
89	Benzo(k)fluoranthene	1.073	1.065	0.7	112	0.00
90 C	Benzo(a)pyrene	1.043	1.027	1.5	111	0.00
91	Dibenzo(a,h)anthracene	0.998	0.996	0.2	112	-0.03
92	Benzo(q,h,i)perylene	0.996	0.983	1.3	111	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0