


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG101618\
 Data File : BG037357.D
 Acq On : 16 Oct 2018 13:36
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG101118.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 11 13:59:58 2018



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2478

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.6	30090	PASS
68	69	0.00	2	0.9	251	PASS
69	198	0.00	100	33.4	26733	PASS
70	69	0.00	2	0.5	125	PASS
127	198	10	80	45.6	36464	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	80018	PASS
199	198	5	9	7.2	5731	PASS
275	198	10	60	24.9	19939	PASS
365	198	1	100	3.0	2405	PASS
441	198	0.01	100	8.6	6907	PASS
442	198	50	100	53.7	43000	PASS
443	442	15	24	18.5	7937	PASS