

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101723\
 Data File : BG059465.D
 Acq On : 17 Oct 2023 19:39
 Operator : MA/JU
 Sample : SSTD01050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD010414

Quant Time: Oct 18 04:00:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG101723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 18 03:51:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.190	152	23135	20.000	ng/ul	0.00
20) Naphthalene-d8	11.028	136	108745	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.829	164	66335	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.579	188	152131	20.000	ng/ul	0.00
79) Chrysene-d12	21.868	240	145528	20.000	ng/ul	0.00
88) Perylene-d12	25.311	264	170446	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.478	96	2854	4.028	ng/uL	0.00
4) Pyridine-d5	3.919	84	26587	12.410	ng/ul	0.00
7) Phenol-d5	7.315	99	23393	9.299	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.508	67	13950	9.811	ng/ul	0.00
11) 2-Chlorophenol-d4	7.708	132	16961	9.425	ng/ul	0.00
15) 4-Methylphenol-d8	8.889	113	16221	8.545	ng/ul	0.00
21) Nitrobenzene-d5	9.383	128	8967	10.023	ng/ul	0.00
24) 2-Nitrophenol-d4	10.105	143	8324	9.501	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.634	165	16857	9.171	ng/ul	0.00
31) 4-Chloroaniline-d4	11.181	131	26591	9.256	ng/ul	0.00
46) Dimethylphthalate-d6	14.224	166	55825	9.946	ng/ul	0.00
49) Acenaphthylene-d8	14.530	160	64039	9.382	ng/ul	0.00
54) 4-Nitrophenol-d4	15.017	143	8336	8.488	ng/ul	0.00
60) Fluorene-d10	15.816	176	53341	9.993	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.928	200	7014	7.593	ng/ul	0.00
73) Anthracene-d10	17.673	188	76414	9.738	ng/ul	0.00
81) Pyrene-d10	19.953	212	87634	9.774	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.070	264	91429	9.668	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.519	88	3551	3.985	ng/uL#	91
5) Pyridine	3.936	79	27697	13.241	ng/ul	97
6) Benzaldehyde	7.338	77	11698	10.628	ng/ul	92
8) Phenol	7.350	94	22496	9.039	ng/ul	94
10) Bis(2-Chloroethyl)ether	7.602	93	18597	9.387	ng/ul	97
12) 2-Chlorophenol	7.743	128	17759	9.497	ng/ul	93
13) 2-Methylphenol	8.619	108	17257	9.443	ng/ul	88
14) 2,2'-oxybis(1-Chloropr...	8.707	45	39205	9.633	ng/ul#	95
16) Acetophenone	9.030	105	27541	9.608	ng/ul	94
17) N-Nitroso-di-n-propyla...	8.995	70	12247	9.279	ng/ul#	90
18) 4-Methylphenol	8.948	108	17384	8.750	ng/ul	86
19) Hexachloroethane	9.271	117	7134	10.039	ng/ul#	83
22) Nitrobenzene	9.436	77	18514	9.242	ng/ul	98
23) Isophorone	9.929	82	39883	9.479	ng/ul#	95
25) 2-Nitrophenol	10.147	139	8498	8.801	ng/ul#	80
26) 2,4-Dimethylphenol	10.164	107	17728	8.962	ng/ul	92
27) Bis(2-Chloroethoxy)met...	10.417	93	26231	9.693	ng/ul#	92
29) 2,4-Dichlorophenol	10.664	162	17919	10.087	ng/ul#	89
30) Naphthalene	11.081	128	64674	9.798	ng/ul	98
32) 4-Chloroaniline	11.198	127	25714	9.404	ng/ul	95
33) Hexachlorobutadiene	11.310	225	11778	10.056	ng/ul	97
34) Caprolactam	11.968	113	5203	8.493	ng/ul	80
35) 4-Chloro-3-methylphenol	12.279	107	16200	9.115	ng/ul	96
36) 2-Methylnaphthalene	12.673	142	40126	9.526	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101723\
 Data File : BG059465.D
 Acq On : 17 Oct 2023 19:39
 Operator : MA/JU
 Sample : SSTD01050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD010414


Quant Time: Oct 18 04:00:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG101723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 18 03:51:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.890	142	43103	9.955	ng/ul#	89
39) 1,2,4,5-Tetrachloroben...	13.014	216	21509	9.885	ng/ul#	88
40) Hexachlorocyclopentadiene	12.973	237	11052	8.768	ng/ul#	89
41) 2,4,6-Trichlorophenol	13.261	196	12375	9.325	ng/ul	89
42) 2,4,5-Trichlorophenol	13.331	196	13617	9.435	ng/ul	89
43) 1,1'-Biphenyl	13.666	154	57467	9.580	ng/ul	96
44) 2-Chloronaphthalene	13.719	162	44438	9.641	ng/ul	94
45) 2-Nitroaniline	13.936	65	10826	8.903	ng/ul#	86
47) Dimethylphthalate	14.271	163	53421	9.514	ng/ul	99
48) 2,6-Dinitrotoluene	14.418	165	9087	8.835	ng/ul#	87
50) Acenaphthylene	14.559	152	68842	9.582	ng/ul	99
51) 3-Nitroaniline	14.759	138	9807	8.805	ng/ul#	51
52) Acenaphthene	14.894	153	47207	9.543	ng/ul	90
53) 2,4-Dinitrophenol	14.953	184	3084	5.522	ng/ul#	75
55) 4-Nitrophenol	15.029	109	5400	7.660	ng/ul#	85
56) Dibenzofuran	15.223	168	67271	10.003	ng/ul	99
57) 2,4-Dinitrotoluene	15.199	165	13332	8.963	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.434	232	12180	9.313	ng/ul#	92
59) Diethylphthalate	15.617	149	50658	9.303	ng/ul	96
61) Fluorene	15.869	166	56829	10.220	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.858	204	28428	10.040	ng/ul	89
63) 4-Nitroaniline	15.916	138	10264	9.377	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.940	198	7300	7.463	ng/ul	97
67) N-Nitrosodiphenylamine	16.075	169	47403	10.534	ng/ul#	88
68) 4-Bromophenyl-phenylether	16.751	248	16978	10.235	ng/ul	97
69) Hexachlorobenzene	16.845	284	18875	9.957	ng/ul	97
70) Atrazine	17.009	200	16742	9.556	ng/ul#	91
71) Pentachlorophenol	17.197	266	8640	8.011	ng/ul#	82
72) Phenanthrene	17.620	178	92929	9.799	ng/ul	98
74) Anthracene	17.708	178	95098	9.780	ng/ul	95
75) 1,2,3,4-Tetrachloroben...	13.625	216	20657	9.404	ng/ul	94
76) Pentachlorobenzene	15.117	250	21572	10.077	ng/ul	97
77) Carbazole	17.990	167	79039	9.399	ng/ul	98
78) Di-n-butylphthalate	18.496	149	88681	9.355	ng/ul	97
80) Fluoranthene	19.618	202	107781	10.086	ng/ul	98
82) Pyrene	19.982	202	116039	10.112	ng/ul	98
83) Butylbenzylphthalate	20.834	149	35574	9.200	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.768	252	38063	9.993	ng/ul	95
85) Benzo(a)anthracene	21.845	228	112203	9.751	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.680	149	54788	9.721	ng/ul	99
87) Chrysene	21.921	228	105494	9.718	ng/ul	99
89) Di-n-octyl phthalate	22.961	149	89114	8.925	ng/ul	100
90) Benzo(b)fluoranthene	24.201	252	112453	9.756	ng/ul#	97
91) Benzo(k)fluoranthene	24.271	252	117276	10.007	ng/ul	98
93) Benzo(a)pyrene	25.141	252	108427	9.699	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	29.277	276	61648	4.910	ng/ul	97
95) Dibenzo(a,h)anthracene	29.365	278	99631	9.536	ng/ul#	90
96) Benzo(g,h,i)perylene	30.534	276	50279	4.900	ng/ul#	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 18 04:00:11 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG101723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 18 03:51:17 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.985 ng/uL
RT: 3.519 min Scan# 21
Delta R.T. 0.005 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion: 88 Resp: 3551
Ion Ratio Lower Upper
88 100
43 30.3 30.8 46.2#
58 86.9 65.0 97.4



#5
Pyridine
Concen: 13.241 ng/uL
RT: 3.936 min Scan# 93
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion: 79 Resp: 27697
Ion Ratio Lower Upper
79 100
52 104.1 80.9 121.3
51 53.4 45.0 67.4









#6
Benzaldehyde
Concen: 10.628 ng/ul
RT: 7.338 min Scan# 671
Delta R.T. 0.005 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion: 77 Resp: 11698
Ion Ratio Lower Upper
77 100
105 99.3 83.0 124.6
106 90.7 82.8 124.2



#8
Phenol
Concen: 9.039 ng/ul
RT: 7.350 min Scan# 674
Delta R.T. 0.005 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion: 94 Resp: 22496
Ion Ratio Lower Upper
94 100
65 23.0 19.7 29.5
66 34.8 24.5 36.7









#10
Bis(2-Chloroethyl)ether
Concen: 9.387 ng/ul
RT: 7.602 min Scan# 717
Delta R.T. -0.007 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion: 93 Resp: 18597
Ion Ratio Lower Upper
93 100
63 77.4 60.0 90.0
95 33.4 27.6 41.4



#12
2-Chlorophenol
Concen: 9.497 ng/ul
RT: 7.743 min Scan# 741
Delta R.T. 0.005 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:128 Resp: 17759
Ion Ratio Lower Upper
128 100
64 40.2 36.8 55.2
130 35.6 26.6 40.0





#13

2-Methylphenol

Concen: 9.443 ng/ul

RT: 8.619 min Scan# 890

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :





SSTD010414

Tgt Ion:108 Resp: 17257

Ion Ratio Lower Upper

108 100

107 91.5 64.4 96.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 9.633 ng/ul

RT: 8.707 min Scan# 905

Delta R.T. 0.011 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion: 45 Resp: 39205


Ion Ratio Lower Upper

45 100

77 6.9 6.4 9.6

79 5.7 6.4 9.6#









#16
Acetophenone
Concen: 9.608 ng/ul
RT: 9.030 min Scan# 960
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:105 Resp: 27541
Ion Ratio Lower Upper
105 100
77 74.0 55.4 83.0
51 34.9 31.6 47.4



#17
N-Nitroso-di-n-propylamine
Concen: 9.279 ng/ul
RT: 8.995 min Scan# 954
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion: 70 Resp: 12247
Ion Ratio Lower Upper
70 100
42 86.6 62.1 93.1
101 5.2 7.3 10.9#
130 19.7 18.6 28.0









#18
4-Methylphenol
Concen: 8.750 ng/ul
RT: 8.948 min Scan# 946
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:108 Resp: 17384
Ion Ratio Lower Upper
108 100
107 112.5 78.7 118.1



#19
Hexachloroethane
Concen: 10.039 ng/ul
RT: 9.271 min Scan# 1001
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:117 Resp: 7134
Ion Ratio Lower Upper
117 100
201 107.9 76.1 114.1
199 62.6 65.0 97.4#





#22

Nitrobenzene

Concen: 9.242 ng/ul

RT: 9.436 min Scan# 1028

Delta R.T. 0.005 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion: 77 Resp: 18514

Ion Ratio Lower Upper

77 100

123 46.6 38.5 57.7

65 12.6 11.1 16.7



#23

Isophorone

Concen: 9.479 ng/ul

RT: 9.929 min Scan# 1113

Delta R.T. -0.007 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion: 82 Resp: 39883


Ion Ratio Lower Upper

82 100

95 4.7 5.6 8.4#

138 18.1 13.0 19.6









#25
2-Nitrophenol
Concen: 8.801 ng/ul
RT: 10.147 min Scan# 1149
Delta R.T. 0.005 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:139 Resp: 8498
Ion Ratio Lower Upper
139 100
65 39.4 40.5 60.7#
109 17.8 25.7 38.5#



#26
2,4-Dimethylphenol
Concen: 8.962 ng/ul
RT: 10.164 min Scan# 1153
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:107 Resp: 17728
Ion Ratio Lower Upper
107 100
121 57.0 44.9 67.3
122 98.9 69.6 104.4





#27

Bis(2-Chloroethoxy)methane

Concen: 9.693 ng/ul

RT: 10.417 min Scan# 1196

Delta R.T. -0.007 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion: 93 Resp: 26231

Ion Ratio Lower Upper

93 100

95 35.2 24.8 37.2

123 8.0 9.1 13.7#



#29

2,4-Dichlorophenol

Concen: 10.087 ng/ul

RT: 10.664 min Scan# 1238

Delta R.T. 0.005 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion:162 Resp: 17919


Ion Ratio Lower Upper

162 100

164 57.4 40.9 61.3

98 34.9 22.2 33.4#









#30
Naphthalene
Concen: 9.798 ng/ul
RT: 11.081 min Scan# 1309
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:128 Resp: 64674
Ion Ratio Lower Upper
128 100
129 9.2 8.0 12.0
127 12.5 9.5 14.3



#32
4-Chloroaniline
Concen: 9.404 ng/ul
RT: 11.198 min Scan# 1329
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:127 Resp: 25714
Ion Ratio Lower Upper
127 100
129 30.1 21.9 32.9









#33
Hexachlorobutadiene
Concen: 10.056 ng/ul
RT: 11.310 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:225 Resp: 11778
Ion Ratio Lower Upper
225 100
223 64.5 48.7 73.1
227 63.7 50.5 75.7



#34
Caprolactam
Concen: 8.493 ng/ul
RT: 11.968 min Scan# 1460
Delta R.T. -0.013 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:113 Resp: 5203
Ion Ratio Lower Upper
113 100
55 276.3 191.8 287.6
56 157.9 109.6 164.4









#35
4-Chloro-3-methylphenol
Concen: 9.115 ng/ul
RT: 12.279 min Scan# 11
Delta R.T. -0.007 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:107 Resp: 16200
Ion Ratio Lower Upper
107 100
144 26.5 22.5 33.7
142 71.5 60.6 90.8



#36
2-Methylnaphthalene
Concen: 9.526 ng/ul
RT: 12.673 min Scan# 1580
Delta R.T. 0.005 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:142 Resp: 40126
Ion Ratio Lower Upper
142 100
141 89.6 69.4 104.0





#37

1-Methylnaphthalene

Concen: 9.955 ng/ul

RT: 12.890 min Scan# 1

Delta R.T. 0.005 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion:142 Resp: 43103

Ion Ratio Lower Upper

142 100

141 81.2 73.4 110.2

116 2.8 3.8 5.6#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 9.885 ng/ul

RT: 13.014 min Scan# 1638

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Tgt Ion:216 Resp: 21509




Ion Ratio Lower Upper


216 100

214 65.4 62.8 94.2

179 20.8 16.4 24.6

108 13.0 15.1 22.7#





#40

Hexachlorocyclopentadiene

Concen: 8.768 ng/ul

RT: 12.973 min Scan# 1

Delta R.T. 0.005 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion:237 Resp: 11052

Ion Ratio Lower Upper

237 100

235 50.6 46.2 69.2

272 5.8 10.9 16.3#



#41

2,4,6-Trichlorophenol

Concen: 9.325 ng/ul

RT: 13.261 min Scan# 1680

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion:196 Resp: 12375


Ion Ratio Lower Upper

196 100

198 86.4 81.2 121.8

200 36.6 29.8 44.6









#42
 2,4,5-Trichlorophenol
 Concen: 9.435 ng/ul
 RT: 13.331 min Scan# 1
 Delta R.T. -0.001 min
 Lab File: BG059465.D
 Acq: 17 Oct 2023 19:39




Instrument :
 BNA_G
 ClientSampleId :
 SSTD010414


Tgt Ion:196 Resp: 13617
 Ion Ratio Lower Upper
 196 100
 198 113.9 81.0 121.6
 200 34.3 25.0 37.6



#43
 1,1'-Biphenyl
 Concen: 9.580 ng/ul
 RT: 13.666 min Scan# 1749
 Delta R.T. -0.001 min
 Lab File: BG059465.D
 Acq: 17 Oct 2023 19:39

Tgt Ion:154 Resp: 57467
 Ion Ratio Lower Upper
 154 100
 153 34.8 30.3 45.5
 76 11.4 9.4 14.0









#44
2-Chloronaphthalene
Concen: 9.641 ng/ul
RT: 13.719 min Scan# 1758
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:162 Resp: 44438
Ion Ratio Lower Upper
162 100
164 32.1 28.0 42.0
127 38.4 27.7 41.5



#45
2-Nitroaniline
Concen: 8.903 ng/ul
RT: 13.936 min Scan# 1795
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion: 65 Resp: 10826
Ion Ratio Lower Upper
65 100
92 71.1 47.0 70.4#
138 98.6 89.8 134.8









#47
Dimethylphthalate
Concen: 9.514 ng/ul
RT: 14.271 min Scan# 1852
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:163 Resp: 53421
Ion Ratio Lower Upper
163 100
194 4.2 3.1 4.7
164 9.4 7.3 10.9



#48
2,6-Dinitrotoluene
Concen: 8.835 ng/ul
RT: 14.418 min Scan# 1877
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:165 Resp: 9087
Ion Ratio Lower Upper
165 100
89 46.9 45.3 67.9
121 27.9 17.4 26.0#









#50
Acenaphthylene
Concen: 9.582 ng/ul
RT: 14.559 min Scan# 1901
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:152 Resp: 68842
Ion Ratio Lower Upper
152 100
151 17.5 14.5 21.7
153 13.3 10.4 15.6



#51
3-Nitroaniline
Concen: 8.805 ng/ul
RT: 14.759 min Scan# 1935
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:138 Resp: 9807
Ion Ratio Lower Upper
138 100
108 12.9 5.3 7.9#
92 142.5 74.4 111.6#





#52

Acenaphthene

Concen: 9.543 ng/ul

RT: 14.894 min Scan# 1958

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion:153 Resp: 47207

Ion Ratio Lower Upper

153 100

152 46.4 40.0 60.0

154 101.9 72.0 108.0



#53

2,4-Dinitrophenol

Concen: 5.522 ng/ul

RT: 14.953 min Scan# 1968

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion:184 Resp: 3084


Ion Ratio Lower Upper

184 100

63 89.3 43.0 64.6#

154 58.3 48.9 73.3









#55
4-Nitrophenol
Concen: 7.660 ng/ul
RT: 15.029 min Scan# 1982
Delta R.T. -0.007 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:109 Resp: 5400
Ion Ratio Lower Upper
109 100
139 173.2 109.6 164.4#
65 150.8 121.4 182.0



#56
Dibenzofuran
Concen: 10.003 ng/ul
RT: 15.223 min Scan# 2014
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:168 Resp: 67271
Ion Ratio Lower Upper
168 100
139 34.0 27.0 40.4









#57
2,4-Dinitrotoluene
Concen: 8.963 ng/ul
RT: 15.199 min Scan# 2009
Delta R.T. 0.005 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:165 Resp: 13332
Ion Ratio Lower Upper
165 100
63 54.3 41.1 61.7
182 4.0 3.0 4.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 9.313 ng/ul
RT: 15.434 min Scan# 2050
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:232 Resp: 12180
Ion Ratio Lower Upper
232 100
131 32.5 28.6 43.0
130 0.0 2.2 3.2#
166 28.3 18.2 27.2#





#59

Diethylphthalate

Concen: 9.303 ng/ul

RT: 15.617 min Scan# 2081

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion:149 Resp: 50658

Ion Ratio Lower Upper

149 100

177 25.0 18.2 27.2

150 11.7 9.9 14.9



#61

Fluorene

Concen: 10.220 ng/ul

RT: 15.869 min Scan# 2124

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion:166 Resp: 56829


Ion Ratio Lower Upper

166 100

165 107.5 85.6 128.4

167 12.4 9.8 14.8









#62
4-Chlorophenyl-phenylether
Concen: 10.040 ng/ul
RT: 15.858 min Scan# 2122
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:204 Resp: 28428
Ion Ratio Lower Upper
204 100
206 35.9 25.7 38.5
141 55.7 37.6 56.4



#63
4-Nitroaniline
Concen: 9.377 ng/ul
RT: 15.916 min Scan# 2132
Delta R.T. -0.007 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:138 Resp: 10264
Ion Ratio Lower Upper
138 100
92 41.5 33.3 49.9
108 68.7 55.4 83.2









#66
4,6-Dinitro-2-methylphenol
Concen: 7.463 ng/ul
RT: 15.940 min Scan# 2137
Delta R.T. -0.007 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:198 Resp: 7300
Ion Ratio Lower Upper
198 100
182 3.4 2.7 4.1
77 17.9 13.0 19.6



#67
N-Nitrosodiphenylamine
Concen: 10.534 ng/ul
RT: 16.075 min Scan# 2159
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:169 Resp: 47403
Ion Ratio Lower Upper
169 100
168 56.7 56.9 85.3#
167 36.5 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 10.235 ng/ul
RT: 16.751 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:248 Resp: 16978
Ion Ratio Lower Upper
248 100
250 91.4 71.8 107.8
141 72.2 54.8 82.2



#69
Hexachlorobenzene
Concen: 9.957 ng/ul
RT: 16.845 min Scan# 2290
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:284 Resp: 18875
Ion Ratio Lower Upper
284 100
142 24.4 21.4 32.0
249 33.8 28.1 42.1





#70

Atrazine

Concen: 9.556 ng/ul

RT: 17.009 min Scan# 21

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion:200 Resp: 16742

Ion Ratio Lower Upper

200 100

173 20.8 21.7 32.5#

215 48.3 34.8 52.2



#71

Pentachlorophenol

Concen: 8.011 ng/ul

RT: 17.197 min Scan# 2350

Delta R.T. -0.007 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion:266 Resp: 8640


Ion Ratio Lower Upper

266 100

264 51.0 56.9 85.3#

268 61.0 56.4 84.6





#72

Phenanthrene

Concen: 9.799 ng/ul

RT: 17.620 min Scan# 2422

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion:178 Resp: 92929

Ion Ratio Lower Upper

178 100

179 15.5 11.8 17.8

176 19.1 14.2 21.4



#74

Anthracene

Concen: 9.780 ng/ul

RT: 17.708 min Scan# 2437

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion:178 Resp: 95098


Ion Ratio Lower Upper

178 100

179 14.0 13.2 19.8

176 17.4 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 9.404 ng/uL
RT: 13.625 min Scan# 1742
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:216 Resp: 20657
Ion Ratio Lower Upper
216 100
214 84.3 64.5 96.7
218 49.8 34.7 52.1



#76
Pentachlorobenzene
Concen: 10.077 ng/uL
RT: 15.117 min Scan# 1996
Delta R.T. -0.007 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:250 Resp: 21572
Ion Ratio Lower Upper
250 100
248 59.6 46.5 69.7
252 69.6 53.3 79.9









#77
 Carbazole
 Concen: 9.399 ng/ul
 RT: 17.990 min Scan# 2485
 Delta R.T. -0.001 min
 Lab File: BG059465.D
 Acq: 17 Oct 2023 19:39




Instrument :
 BNA_G
 ClientSampleId :
 SST010414


Tgt Ion:167 Resp: 79039
 Ion Ratio Lower Upper
 167 100
 166 21.0 17.0 25.4
 139 12.2 8.6 12.8



#78
 Di-n-butylphthalate
 Concen: 9.355 ng/ul
 RT: 18.496 min Scan# 2571
 Delta R.T. -0.001 min
 Lab File: BG059465.D
 Acq: 17 Oct 2023 19:39

Tgt Ion:149 Resp: 88681
 Ion Ratio Lower Upper
 149 100
 150 9.6 6.9 10.3
 104 4.8 4.6 7.0





#80

Fluoranthene

Concen: 10.086 ng/ul

RT: 19.618 min Scan# 2761

Delta R.T. 0.005 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion:202 Resp: 107781

Ion Ratio Lower Upper

202 100

101 10.2 7.4 11.0

100 8.9 6.6 10.0



#82

Pyrene

Concen: 10.112 ng/ul

RT: 19.982 min Scan# 2824

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion:202 Resp: 116039


Ion Ratio Lower Upper

202 100

101 13.1 10.5 15.7

100 9.8 9.2 13.8





#83

Butylbenzylphthalate

Concen: 9.200 ng/ul

RT: 20.834 min Scan# 2969

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion:149 Resp: 35574

Ion Ratio Lower Upper

149 100

91 77.2 58.4 87.6

206 22.7 17.8 26.8



#84

3,3'-Dichlorobenzidine

Concen: 9.993 ng/ul

RT: 21.768 min Scan# 3128

Delta R.T. 0.005 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion:252 Resp: 38063


Ion Ratio Lower Upper

252 100

254 67.9 50.6 75.8

126 10.0 7.8 11.8









#85
Benzo(a)anthracene
Concen: 9.751 ng/ul
RT: 21.845 min Scan# 3142
Delta R.T. -0.007 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:228 Resp: 112203
Ion Ratio Lower Upper
228 100
229 19.5 14.6 21.8
226 27.5 20.2 30.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 9.721 ng/ul
RT: 21.680 min Scan# 3113
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:149 Resp: 54788
Ion Ratio Lower Upper
149 100
167 24.1 18.9 28.3
279 3.2 2.6 4.0









#87
Chrysene
Concen: 9.718 ng/ul
RT: 21.921 min Scan# 3153
Delta R.T. 0.005 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39




Instrument :
BNA_G
ClientSampleId :
SSTD010414


Tgt Ion:228 Resp: 105494
Ion Ratio Lower Upper
228 100
226 29.8 23.4 35.0
229 18.5 14.3 21.5



#89
Di-n-octyl phthalate
Concen: 8.925 ng/ul
RT: 22.961 min Scan# 3331
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:149 Resp: 89114





#90

Benzo(b)fluoranthene

Concen: 9.756 ng/ul

RT: 24.201 min Scan# 3542

Delta R.T. 0.005 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion:252 Resp: 112453

Ion Ratio Lower Upper

252 100

253 21.6 17.2 25.8

125 9.7 9.8 14.6#



#91

Benzo(k)fluoranthene

Concen: 10.007 ng/ul

RT: 24.271 min Scan# 3554

Delta R.T. -0.001 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion:252 Resp: 117276


Ion Ratio Lower Upper

252 100

253 20.8 17.6 26.4

125 9.5 7.2 10.8





#93

Benzo(a)pyrene

Concen: 9.699 ng/u1

RT: 25.141 min Scan# 31

Delta R.T. -0.007 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39

Instrument :

BNA_G

ClientSampleId :

SSTD010414





Tgt Ion:252 Resp: 108427

Ion Ratio Lower Upper

252 100

253 22.7 16.2 24.4

125 11.9 9.6 14.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.910 ng/u1

RT: 29.277 min Scan# 4406

Delta R.T. -0.007 min

Lab File: BG059465.D

Acq: 17 Oct 2023 19:39




Tgt Ion:276 Resp: 61648


Ion Ratio Lower Upper

276 100

138 19.6 16.1 24.1

277 19.9 17.4 26.2









#95
Dibenzo(a,h)anthracene
Concen: 9.536 ng/ul
RT: 29.365 min Scan# 4421
Delta R.T. -0.001 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Instrument :
BNA_G
ClientSampleId :
SSTD010414

Tgt Ion:278 Resp: 99631
Ion Ratio Lower Upper
278 100
139 12.9 12.6 18.8
279 17.5 18.6 27.8#



#96
Benzo(g,h,i)perylene
Concen: 4.900 ng/ul
RT: 30.534 min Scan# 4620
Delta R.T. -0.013 min
Lab File: BG059465.D
Acq: 17 Oct 2023 19:39

Tgt Ion:276 Resp: 50279
Ion Ratio Lower Upper
276 100
138 22.9 15.1 22.7#
277 20.7 17.4 26.2

