

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101723\
 Data File : BG059469.D
 Acq On : 17 Oct 2023 22:23
 Operator : MA/JU
 Sample : SSTD16054
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160418

Quant Time: Oct 18 03:58:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG101723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 18 03:51:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.188	152	25813	20.000	ng/ul	0.00
20) Naphthalene-d8	11.037	136	166096	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.833	164	72861	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.582	188	156148	20.000	ng/ul	0.00
79) Chrysene-d12	21.883	240	138764	20.000	ng/ul	0.01
88) Perylene-d12	25.326	264	208819	20.000	ng/ul	# 0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.552	96	644	0.815	ng/uL	0.07
4) Pyridine-d5	3.916	84	512553	214.418	ng/ul	0.00
7) Phenol-d5	7.330	99	461156	164.300	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	7.524	67	246524	155.394	ng/ul	0.01
11) 2-Chlorophenol-d4	7.712	132	351317	174.968	ng/ul	0.00
15) 4-Methylphenol-d8	8.899	113	347590	164.114	ng/ul	0.01
21) Nitrobenzene-d5	9.398	128	183322	134.152	ng/ul	0.00
24) 2-Nitrophenol-d4	10.115	143	179775	134.343	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.644	165	322237	114.777	ng/ul	0.01
31) 4-Chloroaniline-d4	11.190	131	695727	158.559	ng/ul	0.01
46) Dimethylphthalate-d6	14.233	166	907870	147.255	ng/ul	0.00
49) Acenaphthylene-d8	14.539	160	1120154	149.415	ng/ul	0.00
54) 4-Nitrophenol-d4	15.050	143	158316	146.766	ng/ul	0.03
60) Fluorene-d10	15.826	176	769940	131.322	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.955	200	153647	162.041	ng/ul	0.03
73) Anthracene-d10	17.688	188	1223708	151.931	ng/ul	0.00
81) Pyrene-d10	19.962	212	1333349	155.959	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.103	264	1437132	124.038	ng/ul	0.03

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.517	88	55384	55.712	ng/uL	88
5) Pyridine	3.940	79	531044	227.528	ng/ul	91
6) Benzaldehyde	7.336	77	134084	109.182	ng/ul	95
8) Phenol	7.359	94	448976	161.678	ng/ul#	89
10) Bis(2-Chloroethyl)ether	7.618	93	344115	155.673	ng/ul	92
12) 2-Chlorophenol	7.747	128	355230	170.266	ng/ul	98
13) 2-Methylphenol	8.628	108	335455	164.514	ng/ul	89
14) 2,2'-oxybis(1-Chloropr...	8.699	45	361867	79.686	ng/ul	98
16) Acetophenone	9.051	105	510339	159.565	ng/ul	97
17) N-Nitroso-di-n-propyla...	9.028	70	232510	157.885	ng/ul	91
18) 4-Methylphenol	8.975	108	369797	166.827	ng/ul	89
19) Hexachloroethane	9.275	117	130751	164.903	ng/ul	85
22) Nitrobenzene	9.445	77	366880	119.900	ng/ul	94
23) Isophorone	9.962	82	677618	105.442	ng/ul	95
25) 2-Nitrophenol	10.150	139	179142	121.471	ng/ul	95
26) 2,4-Dimethylphenol	10.179	107	332675	110.108	ng/ul	94
27) Bis(2-Chloroethoxy)met...	10.432	93	428170	103.593	ng/ul	99
29) 2,4-Dichlorophenol	10.673	162	315195	116.168	ng/ul#	83
30) Naphthalene	11.090	128	1611346	159.832	ng/ul	97
32) 4-Chloroaniline	11.213	127	676110	161.878	ng/ul	91
33) Hexachlorobutadiene	11.313	225	265522	148.425	ng/ul	93
34) Caprolactam	12.118	113	10870	11.616	ng/ul	89
35) 4-Chloro-3-methylphenol	12.306	107	297440	109.572	ng/ul	90
36) 2-Methylnaphthalene	12.676	142	701179	108.979	ng/ul	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101723\
 Data File : BG059469.D
 Acq On : 17 Oct 2023 22:23
 Operator : MA/JU
 Sample : SSTD16054
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD160418

Quant Time: Oct 18 03:58:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG101723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 18 03:51:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.894	142	703819	106.428	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	13.023	216	363115	151.934	ng/ul#	97
40) Hexachlorocyclopentadiene	12.970	237	248444	179.456	ng/ul	97
41) 2,4,6-Trichlorophenol	13.264	196	240425	164.940	ng/ul	96
42) 2,4,5-Trichlorophenol	13.346	196	261844	165.174	ng/ul	97
43) 1,1'-Biphenyl	13.675	154	964337	146.358	ng/ul	99
44) 2-Chloronaphthalene	13.728	162	744289	147.018	ng/ul	95
45) 2-Nitroaniline	13.951	65	219947	164.681	ng/ul	96
47) Dimethylphthalate	14.286	163	879632	142.632	ng/ul	98
48) 2,6-Dinitrotoluene	14.433	165	188913	167.232	ng/ul	98
50) Acenaphthylene	14.568	152	1163256	147.411	ng/ul	98
51) 3-Nitroaniline	14.774	138	180761	147.753	ng/ul#	80
52) Acenaphthene	14.903	153	787603	144.949	ng/ul	94
53) 2,4-Dinitrophenol	14.968	184	110764	180.566	ng/ul#	89
55) 4-Nitrophenol	15.062	109	113066	146.029	ng/ul	92
56) Dibenzofuran	15.232	168	952188	128.909	ng/ul	99
57) 2,4-Dinitrotoluene	15.215	165	240401	147.152	ng/ul#	98
58) 2,3,4,6-Tetrachlorophenol	15.444	232	211893	147.498	ng/ul#	90
59) Diethylphthalate	15.632	149	775152	129.607	ng/ul	97
61) Fluorene	15.884	166	793485	129.914	ng/ul	93
62) 4-Chlorophenyl-phenyle...	15.861	204	388627	124.953	ng/ul	92
63) 4-Nitroaniline	16.043	138	388	0.323	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.967	198	162532	161.876	ng/ul#	83
67) N-Nitrosodiphenylamine	16.084	169	653143	141.408	ng/ul	98
68) 4-Bromophenyl-phenylether	16.760	248	268504	157.702	ng/ul	99
69) Hexachlorobenzene	16.854	284	292872	150.528	ng/ul	91
70) Atrazine	17.024	200	274242	152.506	ng/ul	96
71) Pentachlorophenol	17.206	266	195466	176.579	ng/ul	91
72) Phenanthrene	17.629	178	1440564	147.990	ng/ul	98
74) Anthracene	17.723	178	1484843	148.769	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.628	216	360601	159.947	ng/ul	97
76) Pentachlorobenzene	15.127	250	306669	139.577	ng/ul	93
77) Carbazole	18.000	167	1279758	148.273	ng/ul	98
78) Di-n-butylphthalate	18.499	149	1475416	151.636	ng/ul	99
80) Fluoranthene	19.621	202	1642697	161.218	ng/ul	99
82) Pyrene	19.991	202	1649279	150.724	ng/ul	97
83) Butylbenzylphthalate	20.837	149	656589	178.089	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.772	252	552625	152.162	ng/ul	98
85) Benzo(a)anthracene	21.860	228	1711724	156.008	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.684	149	969964	180.486	ng/ul	96
87) Chrysene	21.936	228	1573196	151.987	ng/ul	99
89) Di-n-octyl phthalate	22.964	149	1680910	137.416	ng/ul	100
90) Benzo(b)fluoranthene	24.222	252	1795399	127.133	ng/ul#	96
91) Benzo(k)fluoranthene	24.298	252	1761359	122.678	ng/ul	100
93) Benzo(a)pyrene	25.185	252	1738448	126.933	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	29.333	276	1039397	67.565	ng/ul	99
95) Dibenzo(a,h)anthracene	29.433	278	1658623	129.585	ng/ul	99
96) Benzo(g,h,i)perylene	30.632	276	1642992	130.686	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 55.712 ng/uL
RT: 3.517 min Scan# 21
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion: 88 Resp: 55384
Ion Ratio Lower Upper
88 100
43 34.4 30.8 46.2
58 93.6 65.0 97.4



#5
Pyridine
Concen: 227.528 ng/uL
RT: 3.940 min Scan# 94
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion: 79 Resp: 531044
Ion Ratio Lower Upper
79 100
52 113.6 80.9 121.3
51 53.5 45.0 67.4









#6
Benzaldehyde
Concen: 109.182 ng/ul
RT: 7.336 min Scan# 671
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion: 77 Resp: 134084
Ion Ratio Lower Upper
77 100
105 97.8 83.0 124.6
106 99.1 82.8 124.2



#8
Phenol
Concen: 161.678 ng/ul
RT: 7.359 min Scan# 676
Delta R.T. 0.014 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion: 94 Resp: 448976
Ion Ratio Lower Upper
94 100
65 28.1 19.7 29.5
66 38.5 24.5 36.7#









#10
 Bis(2-Chloroethyl)ether
 Concen: 155.673 ng/ul
 RT: 7.618 min Scan# 718
 Delta R.T. 0.008 min
 Lab File: BG059469.D
 Acq: 17 Oct 2023 22:23




Instrument :
 BNA_G
 ClientSampleId :
 SSTD160418


Tgt Ion: 93 Resp: 344115
 Ion Ratio Lower Upper
 93 100
 63 82.9 60.0 90.0
 95 32.2 27.6 41.4



#12
 2-Chlorophenol
 Concen: 170.266 ng/ul
 RT: 7.747 min Scan# 742
 Delta R.T. 0.008 min
 Lab File: BG059469.D
 Acq: 17 Oct 2023 22:23

Tgt Ion:128 Resp: 355230
 Ion Ratio Lower Upper
 128 100
 64 47.5 36.8 55.2
 130 32.2 26.6 40.0





#13

2-Methylphenol

Concen: 164.514 ng/ul

RT: 8.628 min Scan# 890

Delta R.T. 0.008 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Instrument :

BNA_G

ClientSampleId :





SSTD160418

Tgt Ion:108 Resp: 335455

Ion Ratio Lower Upper

108 100

107 90.1 64.4 96.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 79.686 ng/ul

RT: 8.699 min Scan# 904

Delta R.T. 0.002 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23




Tgt Ion: 45 Resp: 361867


Ion Ratio Lower Upper

45 100

77 8.6 6.4 9.6

79 7.4 6.4 9.6









#16
Acetophenone
Concen: 159.565 ng/ul
RT: 9.051 min Scan# 90
Delta R.T. 0.020 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:105 Resp: 510339
Ion Ratio Lower Upper
105 100
77 71.9 55.4 83.0
51 37.3 31.6 47.4



#17
N-Nitroso-di-n-propylamine
Concen: 157.885 ng/ul
RT: 9.028 min Scan# 960
Delta R.T. 0.032 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion: 70 Resp: 232510
Ion Ratio Lower Upper
70 100
42 87.0 62.1 93.1
101 10.4 7.3 10.9
130 21.4 18.6 28.0









#18
4-Methylphenol
Concen: 166.827 ng/ul
RT: 8.975 min Scan# 91
Delta R.T. 0.026 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:108 Resp: 369797
Ion Ratio Lower Upper
108 100
107 109.0 78.7 118.1



#19
Hexachloroethane
Concen: 164.903 ng/ul
RT: 9.275 min Scan# 1002
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:117 Resp: 130751
Ion Ratio Lower Upper
117 100
201 113.5 76.1 114.1
199 71.8 65.0 97.4









#22
Nitrobenzene
Concen: 119.900 ng/ul
RT: 9.445 min Scan# 1028
Delta R.T. 0.014 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion: 77 Resp: 366880
Ion Ratio Lower Upper
77 100
123 52.2 38.5 57.7
65 15.3 11.1 16.7



#23
Isophorone
Concen: 105.442 ng/ul
RT: 9.962 min Scan# 1119
Delta R.T. 0.026 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion: 82 Resp: 677618
Ion Ratio Lower Upper
82 100
95 6.9 5.6 8.4
138 19.5 13.0 19.6









#25
2-Nitrophenol
Concen: 121.471 ng/ul
RT: 10.150 min Scan# 1151
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:139 Resp: 179142
Ion Ratio Lower Upper
139 100
65 54.4 40.5 60.7
109 29.4 25.7 38.5



#26
2,4-Dimethylphenol
Concen: 110.108 ng/ul
RT: 10.179 min Scan# 1156
Delta R.T. 0.014 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:107 Resp: 332675
Ion Ratio Lower Upper
107 100
121 54.3 44.9 67.3
122 94.1 69.6 104.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 103.593 ng/ul
 RT: 10.432 min Scan# 1199
 Delta R.T. 0.008 min
 Lab File: BG059469.D
 Acq: 17 Oct 2023 22:23




Instrument :
 BNA_G
 ClientSampleId :
 SSTD160418


Tgt Ion: 93 Resp: 428170
 Ion Ratio Lower Upper
 93 100
 95 31.2 24.8 37.2
 123 10.6 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 116.168 ng/ul
 RT: 10.673 min Scan# 1240
 Delta R.T. 0.014 min
 Lab File: BG059469.D
 Acq: 17 Oct 2023 22:23

Tgt Ion:162 Resp: 315195
 Ion Ratio Lower Upper
 162 100
 164 64.5 40.9 61.3#
 98 33.9 22.2 33.4#









#30
Naphthalene
Concen: 159.832 ng/ul
RT: 11.090 min Scan# 111
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:128 Resp: 1611346
Ion Ratio Lower Upper
128 100
129 11.0 8.0 12.0
127 13.4 9.5 14.3



#32
4-Chloroaniline
Concen: 161.878 ng/ul
RT: 11.213 min Scan# 1332
Delta R.T. 0.014 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:127 Resp: 676110
Ion Ratio Lower Upper
127 100
129 32.1 21.9 32.9





#33

Hexachlorobutadiene

Concen: 148.425 ng/ul

RT: 11.313 min Scan# 1

Delta R.T. 0.002 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Instrument :

BNA_G

ClientSampleId :

SSTD160418





Tgt Ion:225 Resp: 265522

Ion Ratio Lower Upper

225 100

223 67.3 48.7 73.1

227 67.1 50.5 75.7



#34

Caprolactam

Concen: 11.616 ng/ul

RT: 12.118 min Scan# 1486

Delta R.T. 0.138 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23




Tgt Ion:113 Resp: 10870


Ion Ratio Lower Upper

113 100

55 233.0 191.8 287.6

56 163.5 109.6 164.4









#35
4-Chloro-3-methylphenol
Concen: 109.572 ng/ul
RT: 12.306 min Scan# 11
Delta R.T. 0.020 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:107 Resp: 297440
Ion Ratio Lower Upper
107 100
144 27.4 22.5 33.7
142 86.6 60.6 90.8



#36
2-Methylnaphthalene
Concen: 108.979 ng/ul
RT: 12.676 min Scan# 1581
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:142 Resp: 701179
Ion Ratio Lower Upper
142 100
141 89.2 69.4 104.0





#37

1-Methylnaphthalene

Concen: 106.428 ng/ul

RT: 12.894 min Scan# 1

Delta R.T. 0.008 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Instrument :

BNA_G

ClientSampleId :

SSTD160418





Tgt Ion:142 Resp: 703819

Ion Ratio Lower Upper

142 100

141 92.1 73.4 110.2

116 4.0 3.8 5.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 151.934 ng/ul

RT: 13.023 min Scan# 1640

Delta R.T. 0.008 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Tgt Ion:216 Resp: 363115




Ion Ratio Lower Upper


216 100

214 79.8 62.8 94.2

179 21.0 16.4 24.6

108 14.1 15.1 22.7#









#40
Hexachlorocyclopentadiene
Concen: 179.456 ng/ul
RT: 12.970 min Scan# 1631
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:237 Resp: 248444
Ion Ratio Lower Upper
237 100
235 60.3 46.2 69.2
272 13.1 10.9 16.3



#41
2,4,6-Trichlorophenol
Concen: 164.940 ng/ul
RT: 13.264 min Scan# 1681
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:196 Resp: 240425
Ion Ratio Lower Upper
196 100
198 104.2 81.2 121.8
200 33.1 29.8 44.6









#42
2,4,5-Trichlorophenol
Concen: 165.174 ng/ul
RT: 13.346 min Scan# 1
Delta R.T. 0.014 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:196 Resp: 261844
Ion Ratio Lower Upper
196 100
198 104.0 81.0 121.6
200 33.5 25.0 37.6



#43
1,1'-Biphenyl
Concen: 146.358 ng/ul
RT: 13.675 min Scan# 1751
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:154 Resp: 964337
Ion Ratio Lower Upper
154 100
153 37.9 30.3 45.5
76 10.2 9.4 14.0









#44
2-Chloronaphthalene
Concen: 147.018 ng/ul
RT: 13.728 min Scan# 1758
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:162 Resp: 744289
Ion Ratio Lower Upper
162 100
164 31.0 28.0 42.0
127 36.1 27.7 41.5



#45
2-Nitroaniline
Concen: 164.681 ng/ul
RT: 13.951 min Scan# 1798
Delta R.T. 0.014 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion: 65 Resp: 219947
Ion Ratio Lower Upper
65 100
92 66.8 47.0 70.4
138 111.4 89.8 134.8









#47
Dimethylphthalate
Concen: 142.632 ng/ul
RT: 14.286 min Scan# 1855
Delta R.T. 0.014 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:163 Resp: 879632
Ion Ratio Lower Upper
163 100
194 3.3 3.1 4.7
164 9.7 7.3 10.9



#48
2,6-Dinitrotoluene
Concen: 167.232 ng/ul
RT: 14.433 min Scan# 1880
Delta R.T. 0.014 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:165 Resp: 188913
Ion Ratio Lower Upper
165 100
89 55.6 45.3 67.9
121 20.3 17.4 26.0





#50

Acenaphthylene

Concen: 147.411 ng/ul

RT: 14.568 min Scan# 1901

Delta R.T. 0.008 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Instrument :

BNA_G

ClientSampleId :

SSTD160418





Tgt Ion:152 Resp: 1163256

Ion Ratio Lower Upper

152 100

151 19.3 14.5 21.7

153 13.7 10.4 15.6



#51

3-Nitroaniline

Concen: 147.753 ng/ul

RT: 14.774 min Scan# 1938

Delta R.T. 0.014 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23




Tgt Ion:138 Resp: 180761


Ion Ratio Lower Upper

138 100

108 10.1 5.3 7.9#

92 112.6 74.4 111.6#









#52
Acenaphthene
Concen: 144.949 ng/ul
RT: 14.903 min Scan# 1960
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:153 Resp: 787603
Ion Ratio Lower Upper
153 100
152 43.1 40.0 60.0
154 93.2 72.0 108.0



#53
2,4-Dinitrophenol
Concen: 180.566 ng/ul
RT: 14.968 min Scan# 1971
Delta R.T. 0.014 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:184 Resp: 110764
Ion Ratio Lower Upper
184 100
63 67.3 43.0 64.6#
154 58.6 48.9 73.3









#55
4-Nitrophenol
Concen: 146.029 ng/ul
RT: 15.062 min Scan# 1982
Delta R.T. 0.026 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:109 Resp: 113066
Ion Ratio Lower Upper
109 100
139 132.4 109.6 164.4
65 136.0 121.4 182.0



#56
Dibenzofuran
Concen: 128.909 ng/ul
RT: 15.232 min Scan# 2016
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:168 Resp: 952188
Ion Ratio Lower Upper
168 100
139 34.2 27.0 40.4





#57

2,4-Dinitrotoluene

Concen: 147.152 ng/ul

RT: 15.215 min Scan# 2009

Delta R.T. 0.020 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Instrument :

BNA_G

ClientSampleId :

SSTD160418





Tgt Ion:165 Resp: 240401

Ion Ratio Lower Upper

165 100

63 50.6 41.1 61.7

182 1.9 3.0 4.4#



#58

2,3,4,6-Tetrachlorophenol

Concen: 147.498 ng/ul

RT: 15.444 min Scan# 2052

Delta R.T. 0.008 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Tgt Ion:232 Resp: 211893




Ion Ratio Lower Upper


232 100

131 40.2 28.6 43.0

130 2.5 2.2 3.2

166 29.3 18.2 27.2#





#59

Diethylphthalate

Concen: 129.607 ng/ul

RT: 15.632 min Scan# 2084

Delta R.T. 0.014 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Instrument :

BNA_G

ClientSampleId :

SSTD160418





Tgt Ion:149 Resp: 775152

Ion Ratio Lower Upper

149 100

177 24.6 18.2 27.2

150 11.9 9.9 14.9



#61

Fluorene

Concen: 129.914 ng/ul

RT: 15.884 min Scan# 2127

Delta R.T. 0.014 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23




Tgt Ion:166 Resp: 793485


Ion Ratio Lower Upper

166 100

165 98.8 85.6 128.4

167 13.0 9.8 14.8









#62
4-Chlorophenyl-phenylether
Concen: 124.953 ng/ul
RT: 15.861 min Scan# 2122
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:204 Resp: 388627
Ion Ratio Lower Upper
204 100
206 31.3 25.7 38.5
141 55.1 37.6 56.4



#63
4-Nitroaniline
Concen: 0.323 ng/ul
RT: 16.043 min Scan# 2154
Delta R.T. 0.120 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:138 Resp: 388
Ion Ratio Lower Upper
138 100
92 38.2 33.3 49.9
108 67.3 55.4 83.2









#66
4,6-Dinitro-2-methylphenol
Concen: 161.876 ng/ul
RT: 15.967 min Scan# 2137
Delta R.T. 0.020 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:198 Resp: 162532
Ion Ratio Lower Upper
198 100
182 2.4 2.7 4.1#
77 24.9 13.0 19.6#



#67
N-Nitrosodiphenylamine
Concen: 141.408 ng/ul
RT: 16.084 min Scan# 2161
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:169 Resp: 653143
Ion Ratio Lower Upper
169 100
168 68.8 56.9 85.3
167 37.9 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 157.702 ng/ul
RT: 16.760 min Scan# 21
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:248 Resp: 268504
Ion Ratio Lower Upper
248 100
250 89.6 71.8 107.8
141 70.4 54.8 82.2



#69
Hexachlorobenzene
Concen: 150.528 ng/ul
RT: 16.854 min Scan# 2292
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:284 Resp: 292872
Ion Ratio Lower Upper
284 100
142 31.2 21.4 32.0
249 29.8 28.1 42.1





#70

Atrazine

Concen: 152.506 ng/ul

RT: 17.024 min Scan# 21

Delta R.T. 0.014 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Instrument :

BNA_G

ClientSampleId :

SSTD160418





Tgt Ion:200 Resp: 274242

Ion Ratio Lower Upper

200 100

173 24.2 21.7 32.5

215 45.1 34.8 52.2



#71

Pentachlorophenol

Concen: 176.579 ng/ul

RT: 17.206 min Scan# 2352

Delta R.T. 0.002 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23




Tgt Ion:266 Resp: 195466


Ion Ratio Lower Upper

266 100

264 63.2 56.9 85.3

268 63.0 56.4 84.6









#72
Phenanthrene
Concen: 147.990 ng/ul
RT: 17.629 min Scan# 2422
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:178 Resp: 1440564
Ion Ratio Lower Upper
178 100
179 15.2 11.8 17.8
176 18.9 14.2 21.4



#74
Anthracene
Concen: 148.769 ng/ul
RT: 17.723 min Scan# 2440
Delta R.T. 0.014 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:178 Resp: 1484843
Ion Ratio Lower Upper
178 100
179 15.9 13.2 19.8
176 18.4 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 159.947 ng/uL
RT: 13.628 min Scan# 11
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:216 Resp: 360601
Ion Ratio Lower Upper
216 100
214 81.2 64.5 96.7
218 47.6 34.7 52.1



#76
Pentachlorobenzene
Concen: 139.577 ng/uL
RT: 15.127 min Scan# 1998
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:250 Resp: 306669
Ion Ratio Lower Upper
250 100
248 62.7 46.5 69.7
252 72.7 53.3 79.9









#77
Carbazole
Concen: 148.273 ng/ul
RT: 18.000 min Scan# 2487
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:167 Resp: 1279758
Ion Ratio Lower Upper
167 100
166 21.3 17.0 25.4
139 12.7 8.6 12.8



#78
Di-n-butylphthalate
Concen: 151.636 ng/ul
RT: 18.499 min Scan# 2572
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:149 Resp: 1475416
Ion Ratio Lower Upper
149 100
150 9.1 6.9 10.3
104 5.3 4.6 7.0









#80
Fluoranthene
Concen: 161.218 ng/ul
RT: 19.621 min Scan# 21
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:202 Resp: 1642697
Ion Ratio Lower Upper
202 100
101 9.9 7.4 11.0
100 8.3 6.6 10.0



#82
Pyrene
Concen: 150.724 ng/ul
RT: 19.991 min Scan# 2826
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:202 Resp: 1649279
Ion Ratio Lower Upper
202 100
101 11.5 10.5 15.7
100 10.6 9.2 13.8









#83
 Butylbenzylphthalate
 Concen: 178.089 ng/ul
 RT: 20.837 min Scan# 2969
 Delta R.T. 0.002 min
 Lab File: BG059469.D
 Acq: 17 Oct 2023 22:23




Instrument :
 BNA_G
 ClientSampleId :
 SSTD160418


Tgt Ion:149 Resp: 656589
 Ion Ratio Lower Upper
 149 100
 91 77.7 58.4 87.6
 206 20.5 17.8 26.8



#84
 3,3'-Dichlorobenzidine
 Concen: 152.162 ng/ul
 RT: 21.772 min Scan# 3129
 Delta R.T. 0.008 min
 Lab File: BG059469.D
 Acq: 17 Oct 2023 22:23

Tgt Ion:252 Resp: 552625
 Ion Ratio Lower Upper
 252 100
 254 65.3 50.6 75.8
 126 9.9 7.8 11.8









#85
Benzo(a)anthracene
Concen: 156.008 ng/ul
RT: 21.860 min Scan# 3142
Delta R.T. 0.008 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:228 Resp: 1711724
Ion Ratio Lower Upper
228 100
229 19.7 14.6 21.8
226 26.6 20.2 30.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 180.486 ng/ul
RT: 21.684 min Scan# 3114
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:149 Resp: 969964
Ion Ratio Lower Upper
149 100
167 25.8 18.9 28.3
279 3.2 2.6 4.0









#87
Chrysene
Concen: 151.987 ng/ul
RT: 21.936 min Scan# 3157
Delta R.T. 0.020 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:228 Resp: 1573196
Ion Ratio Lower Upper
228 100
226 29.4 23.4 35.0
229 19.0 14.3 21.5



#89
Di-n-octyl phthalate
Concen: 137.416 ng/ul
RT: 22.964 min Scan# 3332
Delta R.T. 0.002 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:149 Resp: 1680910









#90
Benzo(b)fluoranthene
Concen: 127.133 ng/ul
RT: 24.222 min Scan# 3546
Delta R.T. 0.026 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:252 Resp: 1795399
Ion Ratio Lower Upper
252 100
253 22.3 17.2 25.8
125 9.5 9.8 14.6#



#91
Benzo(k)fluoranthene
Concen: 122.678 ng/ul
RT: 24.298 min Scan# 3559
Delta R.T. 0.026 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:252 Resp: 1761359
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 9.5 7.2 10.8









#93
Benzo(a)pyrene
Concen: 126.933 ng/u1
RT: 25.185 min Scan# 31
Delta R.T. 0.038 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23




Instrument :
BNA_G
ClientSampleId :
SSTD160418


Tgt Ion:252 Resp: 1738448
Ion Ratio Lower Upper
252 100
253 20.9 16.2 24.4
125 10.6 9.6 14.4



#94
Indeno(1,2,3-cd)pyrene
Concen: 67.565 ng/u1
RT: 29.333 min Scan# 4416
Delta R.T. 0.049 min
Lab File: BG059469.D
Acq: 17 Oct 2023 22:23

Tgt Ion:276 Resp: 1039397
Ion Ratio Lower Upper
276 100
138 19.7 16.1 24.1
277 22.8 17.4 26.2





#95

Dibenzo(a,h)anthracene

Concen: 129.585 ng/ul

RT: 29.433 min Scan# 44

Delta R.T. 0.067 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Instrument :

BNA_G

ClientSampleId :

SSTD160418





Tgt Ion:278 Resp: 1658623

Ion Ratio Lower Upper

278 100

139 15.6 12.6 18.8

279 22.4 18.6 27.8



#96

Benzo(g,h,i)perylene

Concen: 130.686 ng/ul

RT: 30.632 min Scan# 4637

Delta R.T. 0.085 min

Lab File: BG059469.D

Acq: 17 Oct 2023 22:23

Tgt Ion:276 Resp: 1642992

Ion Ratio Lower Upper

276 100

138 21.3 15.1 22.7

277 21.8 17.4 26.2

