


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102220\
 Data File : BG047024.D
 Acq On : 22 Oct 2020 8:21
 Operator : CG/JU
 Sample : DFTPP45
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP45

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG102220MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Oct 22 05:48:14 2020



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.6	58053	PASS
68	69	0.00	2	1.3	803	PASS
69	198	0.00	100	36.4	61120	PASS
70	69	0.00	2	0.6	369	PASS
127	198	10	80	41.9	70309	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	167851	PASS
199	198	5	9	6.9	11501	PASS
275	198	10	60	28.4	47712	PASS
365	198	1	100	3.6	6119	PASS
441	198	0.01	100	11.3	18904	PASS
442	198	50	100	70.2	117757	PASS
443	442	15	24	20.1	23676	PASS

Scan 2466 (17.827 min): BG047024.D
DFTPP45

m/z	Abundance
44.00	323.0
57.20	228.0
167.00	241.0
172.80	167.0
195.20	152.0
198.20	263.0
207.10	535.0
208.00	156.0

Instrument :
BNA_G
ClientSampleId :
DFTPP45

m/z	Abundance
37.00	228.0
38.00	577.0
39.10	3158.0
40.10	344.0
44.00	577.0
50.10	11921.0
51.10	44416.0
52.10	2122.0
55.20	292.0
56.00	1529.0
57.10	2835.0
59.20	173.0
61.10	586.0
62.10	521.0
63.10	1799.0
63.90	269.0
65.10	776.0
69.00	46408.0
70.10	340.0
73.20	770.0
74.10	4594.0
75.10	7992.0
76.10	2897.0
77.10	50352.0
78.10	3126.0
79.10	3599.0
80.00	2727.0
81.00	3575.0
82.00	1146.0
83.10	1192.0
85.00	536.0
86.00	1455.0
87.00	463.0
88.00	190.0
91.00	834.0
92.00	881.0
93.10	6401.0
94.00	398.0
95.00	150.0
98.10	5382.0
99.00	3501.0
100.10	264.0
101.00	2173.0
103.00	471.0
104.00	1386.0
105.00	1414.0
105.90	486.0
107.00	16872.0
108.00	2631.0
110.00	26888.0
111.00	4391.0
111.90	299.0
113.00	377.0
114.80	297.0
116.10	1137.0
117.00	12296.0
118.00	1050.0
120.00	311.0
122.00	955.0
123.00	2151.0
124.00	696.0
124.90	937.0
127.00	51416.0
128.00	4709.0
129.00	23712.0
130.00	1704.0
130.90	333.0
133.20	217.0
134.00	741.0
135.00	1777.0
136.10	1058.0
137.00	1207.0
138.80	204.0
140.10	399.0
141.10	2709.0
142.00	998.0
142.90	563.0
144.00	511.0
146.00	690.0
147.00	1541.0
148.00	3676.0
149.00	720.0
151.10	418.0
152.10	241.0
153.00	1428.0
154.00	1163.0
155.10	1833.0
156.10	2874.0
157.00	497.0
158.00	573.0
158.80	507.0
160.10	1121.0

Instrument :
BNA_G
ClientSampleId :
DFTPP45

161.00	1548.0
161.90	573.0
162.90	229.0
163.50	171.0
164.30	237.0
164.90	1470.0
166.20	778.0
167.10	5285.0
168.10	3155.0
169.10	710.0
170.00	405.0
170.90	310.0
172.00	629.0
173.00	744.0
174.00	1212.0
175.00	2872.0
176.10	610.0
177.00	1236.0
178.00	562.0
179.00	4563.0
180.00	3291.0
181.00	1454.0
182.00	232.0
183.10	332.0
184.20	202.0
185.10	2454.0
186.00	17064.0
187.00	4927.0
188.10	675.0
189.10	1058.0
189.80	269.0
190.90	540.0
192.00	1249.0
193.10	1597.0
194.10	176.0
195.10	349.0
196.00	4137.0
198.00	113176.0
199.00	7658.0
200.00	624.0
201.40	689.0
203.10	768.0
204.10	4277.0
205.00	7986.0
206.00	33296.0
207.00	4786.0
208.00	1265.0
209.00	525.0
210.30	455.0
211.10	899.0
212.00	248.0
215.00	375.0
216.00	1056.0
217.00	9221.0
217.90	1186.0
218.90	273.0
221.00	5949.0
221.80	1341.0
223.00	1581.0
224.00	17976.0
225.00	4991.0
226.00	653.0
227.00	7433.0
228.10	1372.0
229.00	1487.0
229.90	329.0
231.00	538.0
234.00	421.0
235.00	796.0
235.80	205.0
236.20	156.0
237.10	438.0
237.70	212.0
238.90	553.0
239.80	175.0
241.10	387.0
242.00	569.0
243.00	989.0
244.10	13866.0
245.00	1278.0
246.00	2321.0
247.00	565.0
247.90	225.0
248.90	419.0
250.00	167.0
252.10	195.0
252.90	552.0
255.00	71112.0
256.00	9714.0
257.00	1048.0
258.00	4818.0
259.00	1021.0
259.90	196.0
264.00	310.0
265.00	1550.0
266.10	355.0

Instrument :
BNA_G
ClientSampleId :
DFTPP45

270.10	423.0
270.90	239.0
273.10	1624.0
274.00	5025.0
275.00	29632.0
276.10	4492.0
277.00	2889.0
278.00	794.0
281.80	160.0
283.00	350.0
284.00	221.0
285.00	311.0
293.00	402.0
293.90	273.0
296.00	8300.0
297.00	1529.0
297.80	156.0
300.80	198.0
303.10	793.0
304.10	280.0
309.10	177.0
312.60	233.0
314.00	376.0
315.10	727.0
315.90	547.0
321.10	337.0
322.00	204.0
323.00	2622.0
324.00	350.0
326.90	596.0
328.00	237.0
331.80	180.0
333.10	275.0
334.00	1455.0
334.90	447.0
341.00	314.0
346.00	667.0
352.10	734.0
353.10	701.0
354.00	662.0
360.90	166.0
365.00	3802.0
366.00	714.0
372.00	988.0
373.20	364.0
390.10	426.0
402.10	462.0
403.00	939.0
404.10	294.0
421.10	501.0
422.00	505.0
423.10	3059.0
423.90	472.0
441.00	10018.0
442.10	62888.0
443.00	12806.0
444.10	1131.0
445.10	174.0
536.30	210.0

Instrument :
BNA_G
ClientSampleId :
DFTPP45

m/z	Abundance
37.00	293.0
38.10	1049.0
39.10	4657.0
40.00	461.0
44.10	545.0
50.10	17592.0
51.10	71568.0
52.10	3097.0
53.00	313.0
55.10	300.0
56.10	2385.0
57.10	4339.0
58.00	265.0
61.00	1268.0
62.00	1127.0
63.10	3239.0
64.00	656.0
65.10	1485.0
67.10	242.0
68.00	1062.0
69.00	74272.0
70.10	267.0
73.00	892.0
74.10	7701.0
75.00	12040.0
76.10	4177.0
77.10	76808.0
78.10	4857.0
79.10	5379.0
80.00	4401.0
81.00	6757.0
82.00	2055.0
83.10	1304.0
85.10	1032.0
86.00	1507.0
87.00	687.0
87.80	183.0
89.10	224.0
91.00	1226.0
91.90	1505.0
93.00	10067.0
94.10	567.0
95.10	214.0
95.90	297.0
97.00	267.0
98.00	8128.0
99.00	5406.0
100.00	556.0
101.00	3839.0
102.00	206.0
103.00	998.0
104.00	2453.0
105.10	2980.0
107.00	24432.0
108.00	3125.0
110.00	45608.0
111.00	7783.0
111.90	863.0
113.20	328.0
115.00	203.0
117.00	22456.0
118.00	1728.0
118.90	163.0
120.10	504.0
121.10	196.0
121.90	1650.0
123.10	2492.0
124.10	1073.0
125.00	1080.0
127.00	84880.0
128.00	6524.0
129.00	35864.0
130.00	3379.0
131.00	667.0
132.10	245.0
134.10	1333.0
135.10	2708.0
136.10	1141.0
137.10	1443.0
138.40	230.0
139.00	313.0
140.00	375.0
141.10	5046.0
141.90	1436.0
143.10	1121.0
144.00	316.0
144.70	157.0
146.00	1351.0
147.00	2981.0
148.00	5754.0
149.10	1387.0
150.10	250.0

Instrument :
BNA_G
ClientSampleId :
DFTPP45

151.10	654.0
151.90	241.0
153.00	1342.0
154.00	1211.0
155.00	3239.0
156.00	4066.0
157.30	777.0
158.00	1097.0
159.00	1078.0
160.10	1459.0
161.10	2546.0
162.10	705.0
164.20	349.0
165.00	1324.0
166.10	1739.0
167.00	8786.0
168.00	4374.0
169.10	872.0
170.10	280.0
171.10	450.0
172.00	923.0
173.00	1098.0
174.10	2209.0
175.00	3844.0
176.00	1412.0
177.00	2004.0
178.00	809.0
179.00	7502.0
180.00	5898.0
181.00	2326.0
182.00	360.0
183.00	279.0
184.10	685.0
185.10	3329.0
186.10	26688.0
187.00	8390.0
188.10	1111.0
189.10	2068.0
189.90	386.0
191.10	837.0
192.00	2256.0
193.00	2732.0
194.10	524.0
195.10	464.0
196.00	6518.0
198.00	196800.0
199.00	14018.0
200.00	898.0
201.60	737.0
203.10	1561.0
204.00	7172.0
205.00	13901.0
206.10	57816.0
207.00	7955.0
208.00	2411.0
208.80	754.0
210.10	1274.0
210.90	2654.0
212.00	763.0
212.90	231.0
215.00	950.0
216.00	1347.0
217.00	17552.0
218.00	1697.0
218.80	174.0
221.00	8310.0
221.90	2292.0
223.00	3785.0
224.00	33760.0
225.00	7671.0
226.10	1044.0
227.00	14380.0
228.00	2099.0
229.00	2253.0
229.90	379.0
231.00	1164.0
232.00	262.0
233.20	358.0
234.00	910.0
235.00	1248.0
236.00	975.0
236.90	1184.0
237.90	242.0
239.00	560.0
239.90	353.0
241.00	1034.0
242.00	1608.0
243.10	1755.0
244.10	24240.0
245.00	3306.0
246.00	4855.0
247.10	1247.0
249.00	1164.0
250.10	272.0
251.10	169.0
252.00	674.0

Instrument :
BNA_G
ClientSampleId :
DFTPP45

253.10	579.0
255.00	130880.0
256.00	18120.0
257.10	1547.0
258.00	7918.0
259.00	1912.0
260.10	317.0
261.00	338.0
264.10	329.0
265.00	3191.0
266.10	429.0
270.00	194.0
270.40	172.0
271.00	434.0
271.90	339.0
273.00	5252.0
274.00	8935.0
275.00	54440.0
276.00	7748.0
277.00	4237.0
278.00	1019.0
279.00	164.0
283.00	661.0
283.90	425.0
285.00	842.0
285.90	170.0
290.00	235.0
291.10	270.0
293.00	1275.0
293.90	218.0
296.00	14948.0
297.10	2455.0
298.10	267.0
302.00	484.0
303.00	1860.0
304.10	798.0
308.10	152.0
310.10	244.0
311.00	179.0
313.10	252.0
314.00	1245.0
315.00	2252.0
316.10	1415.0
320.90	624.0
322.00	313.0
323.10	4999.0
324.10	1374.0
327.10	690.0
328.10	812.0
332.10	369.0
333.00	568.0
334.00	3120.0
335.00	697.0
341.10	433.0
341.80	153.0
345.90	1026.0
346.90	455.0
352.10	1845.0
353.00	894.0
354.00	1327.0
359.10	151.0
365.00	6762.0
366.00	1035.0
370.90	553.0
372.00	2565.0
373.10	432.0
383.00	748.0
390.00	278.0
391.00	213.0
392.00	229.0
395.00	154.0
402.00	1161.0
403.00	1334.0
404.10	342.0
405.00	201.0
421.00	1105.0
422.00	1194.0
423.00	7098.0
424.10	2031.0
441.00	20384.0
442.10	123408.0
443.10	25352.0
444.00	2673.0
445.00	212.0

Instrument :
BNA_G
ClientSampleId :
DFTPP45

m/z	Abundance
38.10	556.0
39.10	4708.0
40.30	242.0
41.10	261.0
43.90	318.0
45.00	268.0
49.10	328.0
50.10	15650.0
51.10	58176.0
52.10	3682.0
53.20	219.0
54.00	190.0
55.00	304.0
56.00	1447.0
57.10	4533.0
58.00	398.0
58.90	263.0
61.10	1034.0
62.10	846.0
63.10	3299.0
64.10	356.0
65.10	1251.0
68.00	1347.0
69.00	62680.0
70.00	500.0
72.90	831.0
74.10	6910.0
75.00	10436.0
76.10	3153.0
77.10	68928.0
78.10	4587.0
79.00	5335.0
80.00	4399.0
81.00	5493.0
82.10	1301.0
83.00	871.0
84.20	162.0
85.00	832.0
86.10	1523.0
87.10	492.0
89.10	265.0
90.10	253.0
90.90	1099.0
92.10	1341.0
93.00	8397.0
94.00	652.0
94.80	234.0
96.20	384.0
98.00	6932.0
99.00	4976.0
99.90	592.0
101.00	2657.0
102.10	223.0
103.10	1037.0
104.00	1572.0
105.10	1588.0
106.00	596.0
107.00	24072.0
108.10	3715.0
110.00	41112.0
111.00	5827.0
112.00	769.0
113.00	218.0
116.00	1609.0
117.00	19440.0
118.00	1417.0
120.10	703.0
120.90	286.0
122.00	2148.0
123.00	3009.0
124.00	1225.0
125.00	1190.0
127.00	74632.0
128.00	6312.0
129.00	34504.0
130.00	3341.0
131.10	595.0
131.90	355.0
133.10	279.0
133.90	779.0
135.00	2791.0
136.00	1167.0
137.00	1374.0
138.00	554.0
140.00	450.0
141.00	3869.0
142.00	1724.0
143.00	892.0
144.00	405.0
145.00	608.0
145.90	806.0
147.00	2518.0

Instrument :
BNA_G
ClientSampleId :
DFTPP45

148.00	4964.0
149.00	1466.0
149.90	169.0
151.10	1045.0
153.00	913.0
154.10	1235.0
155.00	2835.0
156.00	3451.0
157.00	553.0
157.90	859.0
159.00	725.0
160.10	1358.0
161.10	1928.0
162.00	651.0
163.00	165.0
164.10	334.0
165.00	1880.0
166.10	1410.0
167.00	8737.0
168.00	3992.0
169.10	1054.0
169.90	257.0
170.20	214.0
171.00	578.0
172.00	1326.0
173.00	1089.0
174.00	2372.0
175.10	4100.0
176.10	1151.0
177.00	2089.0
178.10	824.0
179.00	8190.0
180.00	5462.0
181.00	2698.0
182.00	573.0
183.10	413.0
184.10	767.0
185.00	4358.0
186.00	27256.0
187.10	7684.0
188.10	903.0
188.90	2240.0
191.10	1224.0
192.00	2852.0
193.10	2732.0
194.00	469.0
195.00	193.0
196.00	5149.0
198.00	194368.0
199.00	12827.0
200.10	969.0
201.60	1452.0
202.30	356.0
203.00	2043.0
204.10	8227.0
205.00	13240.0
206.00	61520.0
207.00	8144.0
208.00	2478.0
209.00	891.0
210.20	835.0
211.00	2457.0
213.00	218.0
214.00	226.0
214.30	190.0
215.00	756.0
216.10	1848.0
217.00	16374.0
217.90	1905.0
219.90	167.0
221.00	8829.0
223.00	3747.0
224.00	35560.0
225.10	9553.0
225.90	1174.0
227.00	15526.0
228.10	2389.0
229.00	3035.0
230.10	457.0
231.00	1358.0
232.20	374.0
232.90	254.0
234.00	1071.0
235.00	1237.0
235.90	813.0
237.00	1014.0
238.20	309.0
238.90	353.0
239.90	403.0
241.00	780.0
242.00	1962.0
243.20	1617.0
244.10	27368.0
245.10	2940.0
246.00	5121.0
247.00	1097.0

Instrument :
BNA_G
ClientSampleId :
DFTPP45

247.90	246.0
249.00	1464.0
250.10	170.0
251.10	381.0
252.30	428.0
253.10	751.0
255.00	139648.0
256.00	18576.0
257.00	1714.0
258.00	8833.0
259.10	1157.0
259.90	333.0
264.00	296.0
265.00	3489.0
266.00	431.0
269.90	249.0
270.80	362.0
271.20	295.0
273.00	5139.0
274.00	9771.0
275.00	59064.0
276.00	7920.0
277.00	5878.0
278.10	1010.0
280.90	312.0
282.90	583.0
284.00	451.0
285.10	1054.0
286.00	161.0
289.10	163.0
290.00	264.0
290.60	151.0
292.20	314.0
293.00	919.0
294.10	253.0
296.00	19168.0
297.00	2902.0
301.10	184.0
302.00	294.0
303.10	1924.0
304.10	503.0
307.80	310.0
310.00	392.0
314.00	723.0
315.00	2335.0
316.10	885.0
317.00	311.0
321.00	683.0
322.10	321.0
323.00	5479.0
324.10	1168.0
325.00	216.0
326.10	157.0
327.00	1142.0
328.00	525.0
329.10	170.0
332.10	670.0
333.10	579.0
334.00	3550.0
335.10	932.0
339.10	155.0
341.10	688.0
342.20	255.0
346.00	1181.0
347.00	314.0
351.00	154.0
352.00	1959.0
353.00	1068.0
354.00	1472.0
355.20	166.0
365.00	7793.0
366.00	1557.0
370.10	326.0
371.00	513.0
372.00	2727.0
373.00	513.0
383.00	1110.0
384.20	246.0
389.90	322.0
391.00	295.0
391.90	222.0
402.00	1093.0
403.00	1689.0
404.00	606.0
421.00	1372.0
422.00	1110.0
423.10	9354.0
424.10	2393.0
438.90	182.0
439.60	161.0
441.00	26312.0
442.10	166976.0
443.10	32872.0
444.00	2371.0
445.00	298.0

Instrument :
BNA_G
ClientSampleId :
DFTPP45