

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102318\
 Data File : BG037506.D
 Acq On : 24 Oct 2018 11:35
 Operator : JU/SJ
 Sample : PB114112BL
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB114112BL

Quant Time: Oct 24 13:42:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.11	152	57164	20.00	ng	0.00
21) Naphthalene-d8	10.93	136	249995	20.00	ng	0.00
39) Acenaphthene-d10	14.74	164	177177	20.00	ng	0.00
64) Phenanthrene-d10	17.48	188	435987	20.00	ng	0.00
76) Chrysene-d12	21.77	240	400689	20.00	ng	0.00
87) Perylene-d12	25.10	264	388305	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.65	112	415703	121.58	ng	0.00
7) Phenol-d6	7.25	99	616834	119.30	ng	0.00
23) Nitrobenzene-d5	9.28	82	347326	77.83	ng	0.00
42) 2,4,6-Tribromophenol	16.23	330	231770	119.94	ng	0.00
45) 2-Fluorobiphenyl	13.36	172	872863	74.29	ng	0.00
79) Terphenyl-d14	20.08	244	1341389	71.53	ng	0.00


Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102318\
Data File : BG037506.D
Acq On : 24 Oct 2018 11:35
Operator : JU/SJ
Sample : PB114112BL
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB114112BL

Quant Time: Oct 24 13:42:19 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 14:06:42 2018
Response via : Initial Calibration






#1



1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.11 min Scan# 820
Delta R.T. -0.00 min
Lab File: BG037506.D
Acq: 24 Oct 2018 11:35

Instrument :
BNA_G
ClientSampleId :
PB114112BL

Tot Ion:152 Resp: 57164
Ion Ratio Lower Upper
152 100
150 152.8 126.0 189.0
115 52.1 45.5 68.3




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5

2-Fluorophenol
Concen: 121.579 ng
RT: 5.65 min Scan# 401
Delta R.T. -0.00 min
Lab File: BG037506.D
Acq: 24 Oct 2018 11:35

Tgt Ion:112 Resp: 415703
Ion Ratio Lower Upper
112 100
64 66.1 53.1 79.7
63 34.2 27.3 40.9



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BG
Ion 63.00 (62.70 to 63.70): BG





#7

Phenol-d6

Concen: 119.304 ng

RT: 7.25 min Scan# 674

Delta R.T. -0.00 min

Lab File: BG037506.D

Acq: 24 Oct 2018 11:35

Instrument :

BNA_G

ClientSampleId :

PB114112BL


Tot Ion: 99 Resp: 616834

Ion Ratio Lower Upper



99 100

42 18.5 15.0 22.4

71 30.8 25.5 38.3



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.93 min Scan# 1300

Delta R.T. -0.00 min

Lab File: BG037506.D

Acq: 24 Oct 2018 11:35

Tgt Ion: 136 Resp: 249995


Ion Ratio Lower Upper

136 100



137 11.6 9.6 14.4

54 10.3 7.9 11.9

68 6.9 4.8 7.2



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
Nitrobenzene-d5
Concen: 77.829 ng
RT: 9.28 min Scan# 1020
Delta R.T. -0.00 min
Lab File: BG037506.D
Acq: 24 Oct 2018 11:35

Instrument :
BNA_G
ClientSampleId :
PB114112BL

Tot Ion: 82 Resp: 347326
Ion Ratio Lower Upper
82 100
128 43.8 34.6 51.8
54 53.5 45.3 67.9




Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 128.00 (127.70 to 128.70): BGC
Ion 54.00 (53.70 to 54.70): BGC





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.74 min Scan# 1949
Delta R.T. -0.00 min
Lab File: BG037506.D
Acq: 24 Oct 2018 11:35

Tgt Ion:164 Resp: 177177
Ion Ratio Lower Upper
164 100
162 102.1 81.3 121.9
160 45.4 36.3 54.5



Abundance Ion 164.00 (163.70 to 164.70): BGC
Ion 162.00 (161.70 to 162.70): BGC
Ion 160.00 (159.70 to 160.70): BGC






#42
 2,4,6-Tribromophenol
 Concen: 119.936 ng
 RT: 16.23 min Scan# 2202
 Delta R.T. -0.00 min
 Lab File: BG037506.D
 Acq: 24 Oct 2018 11:35

Instrument :
 BNA_G
 ClientSampleId :
 PB114112BL

Tot Ion	Ratio	Lower	Upper
330	100		
332	96.6	78.4	117.6
141	39.6	32.2	48.2





Abundance

Ion 329.80 (329.50 to 330.50): E


Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45
 2-Fluorobiphenyl
 Concen: 74.289 ng
 RT: 13.36 min Scan# 1714
 Delta R.T. -0.01 min
 Lab File: BG037506.D
 Acq: 24 Oct 2018 11:35

Tgt Ion	Ratio	Lower	Upper
172	100		
171	37.7	31.0	46.4
170	25.2	20.2	30.2





Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.48 min Scan# 2416
Delta R.T. -0.00 min
Lab File: BG037506.D
Acq: 24 Oct 2018 11:35

Instrument :
BNA_G
ClientSampleId :
PB114112BL

Tot Ion:188 Resp: 435987
Ion Ratio Lower Upper
188 100
94 9.3 7.2 10.8
80 10.0 7.7 11.5




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BGC
Ion 80.00 (79.70 to 80.70): BGC






#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.77 min Scan# 3145
Delta R.T. -0.01 min
Lab File: BG037506.D
Acq: 24 Oct 2018 11:35

Tgt Ion:240 Resp: 400689
Ion Ratio Lower Upper
240 100
120 10.4 8.1 12.1
236 26.4 21.4 32.0



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#79

Terphenyl-d14

Concen: 71.533 ng

RT: 20.08 min Scan# 2858

Delta R.T. -0.01 min

Lab File: BG037506.D

Acq: 24 Oct 2018 11:35

Instrument :


BNA_G

ClientSampleId :

PB114112BL

Tot Ion:244 Resp: 1341389

Ion	Ratio	Lower	Upper
244	100		
212	8.3	6.5	9.7
122	12.0	9.0	13.4






Abundance

Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#87

Perylene-d12

Concen: 20.000 ng

RT: 25.10 min Scan# 3712


Delta R.T. -0.01 min

Lab File: BG037506.D

Acq: 24 Oct 2018 11:35

Tgt Ion:264 Resp: 388305

Ion	Ratio	Lower	Upper
264	100		
260	23.0	18.2	27.4
265	22.0	17.9	26.9



Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

