

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
 Data File : BG064579.D
 Acq On : 24 Oct 2025 18:08
 Operator : RC/JU
 Sample : Q3405-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PAD-10-20-25-B

Quant Time: Oct 24 18:44:37 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 24 16:40:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.217	152	41405	20.000	ng	0.00
21) Naphthalene-d8	11.044	136	161271	20.000	ng	0.00
39) Acenaphthene-d10	14.839	164	132688	20.000	ng	0.00
64) Phenanthrene-d10	17.577	188	322915	20.000	ng	0.00
76) Chrysene-d12	21.854	240	335079	20.000	ng	0.00
86) Perylene-d12	25.198	264	363557	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.744	112	180380	73.123	ng	0.00
7) Phenol-d6	7.348	99	261537	77.265	ng	0.00
23) Nitrobenzene-d5	9.381	82	143596	53.480	ng	0.00
42) 2,4,6-Tribromophenol	16.320	330	134140	70.044	ng	0.00
45) 2-Fluorobiphenyl	13.470	172	448518	51.236	ng	0.00
79) Terphenyl-d14	20.168	244	958557	53.986	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064579.D
Acq On : 24 Oct 2025 18:08
Operator : RC/JU
Sample : Q3405-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PAD-10-20-25-B

Quant Time: Oct 24 18:44:37 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration

