

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG102518\
 Data File : BG037544.D
 Acq On : 25 Oct 2018 14:23
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 25 18:02:27 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	38.817	3.0	99	0.00
3	Pyridine	40.000	38.954	2.6	98	0.00
4	n-Nitrosodimethylamine	40.000	40.949	-2.4	104	0.00
5 S	2-Fluorophenol	80.000	79.426	0.7	99	0.00
6	Aniline	40.000	40.075	-0.2	100	0.00
7 S	Phenol-d6	80.000	80.416	-0.5	100	0.00
8	2-Chlorophenol	40.000	40.343	-0.9	101	0.00
9	Benzaldehyde	40.000	37.417	6.5	94	0.00
10 C	Phenol	40.000	40.257	-0.6	100	0.00
11	bis(2-Chloroethyl)ether	40.000	39.707	0.7	100	0.00
12	1,3-Dichlorobenzene	40.000	39.316	1.7	100	0.00
13 C	1,4-Dichlorobenzene	40.000	39.111	2.2	99	0.00
14	1,2-Dichlorobenzene	40.000	38.413	4.0	97	0.00
15	Benzyl Alcohol	40.000	41.405	-3.5	101	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	40.334	-0.8	102	0.00
17	2-Methylphenol	40.000	40.709	-1.8	100	0.00
18	Hexachloroethane	40.000	41.221	-3.1	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.432	-1.1	98	0.00
20	3+4-Methylphenols	40.000	41.034	-2.6	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	38.231	4.4	99	0.00
23 S	Nitrobenzene-d5	80.000	79.566	0.5	103	0.00
24	Nitrobenzene	40.000	39.715	0.7	102	0.00
25	Isophorone	40.000	39.950	0.1	101	0.00
26 C	2-Nitrophenol	40.000	44.270	-10.7	111	0.00
27	2,4-Dimethylphenol	40.000	38.894	2.8	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.463	3.8	98	0.00
29 C	2,4-Dichlorophenol	40.000	39.317	1.7	101	0.00
30	1,2,4-Trichlorobenzene	40.000	38.531	3.7	101	0.00
31	Naphthalene	40.000	37.855	5.4	99	0.00
32	Benzoic acid	40.000	40.569	-1.4	102	0.00
33	4-Chloroaniline	40.000	38.903	2.7	99	0.00
34 C	Hexachlorobutadiene	40.000	38.463	3.8	98	0.00
35	Caprolactam	40.000	39.677	0.8	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.042	2.4	100	0.00
37	2-Methylnaphthalene	40.000	38.179	4.6	98	0.00
38	1-Methylnaphthalene	40.000	37.935	5.2	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.010	5.0	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.354	6.6	95	0.00
42 S	2,4,6-Tribromophenol	80.000	79.334	0.8	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.474	-1.2	102	0.00
44	2,4,5-Trichlorophenol	40.000	38.853	2.9	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG102518\
 Data File : BG037544.D
 Acq On : 25 Oct 2018 14:23
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 25 18:02:27 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	80.000	75.017	6.2	99	0.00
46	1,1'-Biphenyl	40.000	37.563	6.1	99	0.00
47	2-Chloronaphthalene	40.000	37.930	5.2	100	0.00
48	2-Nitroaniline	40.000	41.377	-3.4	106	0.00
49	Acenaphthylene	40.000	38.187	4.5	99	0.00
50	Dimethylphthalate	40.000	38.135	4.7	99	0.00
51	2,6-Dinitrotoluene	40.000	39.892	0.3	101	0.00
52 C	Acenaphthene	40.000	37.408	6.5	99	0.00
53	3-Nitroaniline	40.000	39.449	1.4	99	0.00
54 P	2,4-Dinitrophenol	40.000	43.870	-9.7	121	0.00
55	Dibenzofuran	40.000	37.350	6.6	99	0.00
56 P	4-Nitrophenol	40.000	38.283	4.3	96	0.00
57	2,4-Dinitrotoluene	40.000	42.164	-5.4	104	0.00
58	Fluorene	40.000	37.400	6.5	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.338	4.2	98	0.00
60	Diethylphthalate	40.000	38.503	3.7	100	0.00
61	4-Chlorophenyl-phenylether	40.000	37.602	6.0	98	0.00
62	4-Nitroaniline	40.000	38.650	3.4	97	0.00
63	Azobenzene	40.000	37.619	6.0	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.429	-1.1	110	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.474	3.8	97	0.00
67	4-Bromophenyl-phenylether	40.000	39.321	1.7	100	0.00
68	Hexachlorobenzene	40.000	38.037	4.9	97	0.00
69	Atrazine	40.000	38.748	3.1	96	0.00
70 C	Pentachlorophenol	40.000	38.926	2.7	96	0.00
71	Phenanthrene	40.000	37.788	5.5	97	0.00
72	Anthracene	40.000	38.193	4.5	98	0.00
73	Carbazole	40.000	38.611	3.5	97	0.00
74	Di-n-butylphthalate	40.000	39.332	1.7	98	0.00
75 C	Fluoranthene	40.000	37.768	5.6	96	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzidine	40.000	35.321	11.7	86	0.00
78	Pyrene	40.000	38.459	3.9	96	0.00
79 S	Terphenyl-d14	80.000	77.131	3.6	97	0.00
80	Butylbenzylphthalate	40.000	41.140	-2.9	99	0.00
81	Benzo(a)anthracene	40.000	38.679	3.3	97	0.00
82	3,3'-Dichlorobenzidine	40.000	38.509	3.7	93	0.00
83	Chrysene	40.000	38.366	4.1	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.062	-2.7	98	0.00
85 c	Di-n-octyl phthalate	40.000	41.441	-3.6	99	-0.02
86	Indeno(1,2,3-cd)pyrene	40.000	35.310	11.7	86	-0.02
87 I	Perylene-d12	20.000	20.000	0.0	100	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG102518\
Data File : BG037544.D
Acq On : 25 Oct 2018 14:23
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 25 18:02:27 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 14:06:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	40.000	37.769	5.6	93	0.00
89	Benzo(k)fluoranthene	40.000	38.643	3.4	96	0.00
90 C	Benzo(a)pyrene	40.000	38.478	3.8	95	0.00
91	Dibenzo(a,h)anthracene	40.000	35.634	10.9	86	-0.02
92	Benzo(q,h,i)perylene	40.000	33.739	15.7	83	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0