

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102618\  
 Data File : BG037607.D  
 Acq On : 27 Oct 2018 8:13  
 Operator : JU/SJ  
 Sample : J5662-13  
 Misc :  
 ALS Vial : 27 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 102318-TIPC-F-D-B

Quant Time: Oct 29 02:23:38 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Oct 18 14:06:42 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min)  |
|-----------------------------|-------|------|----------|-------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.10  | 152  | 59477    | 20.00 | ng    | 0.00      |
| 21) Naphthalene-d8          | 10.93 | 136  | 250284   | 20.00 | ng    | 0.00      |
| 39) Acenaphthene-d10        | 14.74 | 164  | 172649   | 20.00 | ng    | 0.00      |
| 64) Phenanthrene-d10        | 17.48 | 188  | 422010   | 20.00 | ng    | 0.00      |
| 76) Chrysene-d12            | 21.77 | 240  | 381100   | 20.00 | ng    | 0.00      |
| 87) Perylene-d12            | 25.10 | 264  | 355999   | 20.00 | ng    | 0.00      |
| System Monitoring Compounds |       |      |          |       |       |           |
| 5) 2-Fluorophenol           | 5.64  | 112  | 147947   | 41.59 | ng    | 0.00      |
| 7) Phenol-d6                | 7.24  | 99   | 135640   | 25.21 | ng    | 0.00      |
| 23) Nitrobenzene-d5         | 9.28  | 82   | 430145   | 96.28 | ng    | 0.00      |
| 42) 2,4,6-Tribromophenol    | 16.22 | 330  | 168302   | 89.38 | ng    | 0.00      |
| 45) 2-Fluorobiphenyl        | 13.36 | 172  | 1038392  | 90.69 | ng    | 0.00      |
| 79) Terphenyl-d14           | 20.09 | 244  | 1573835  | 88.24 | ng    | 0.00      |
| Target Compounds            |       |      |          |       |       |           |
| 10) Phenol                  | 7.26  | 94   | 16659    | 2.935 | ng    | Qvalue 92 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102618\  
Data File : BG037607.D  
Acq On : 27 Oct 2018 8:13  
Operator : JU/SJ  
Sample : J5662-13  
Misc :  
ALS Vial : 27 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
102318-TIPC-F-D-B

Quant Time: Oct 29 02:23:38 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Oct 18 14:06:42 2018  
Response via : Initial Calibration

